

DMB

Reception Date	11/16/99
Classification Date	11/17/99
Control	unclassified

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. 99P-4064]

Medical Devices; Exemptions From Premarket Notification; Class II Devices

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA) is announcing that it has received a petition requesting an exemption from the premarket notification requirements for vascular tunnelers. FDA is publishing this notice in order to obtain comments on this petition in accordance with procedures established by the Food and Drug Administration Modernization Act of 1997 (FDAMA).

DATES: Written comments by (*insert date 30 days after date of publication in the Federal Register*).

ADDRESSES: Submit written comments on this notice to the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852.

FOR FURTHER INFORMATION CONTACT: Heather S. Rosecrans, Center for Devices and Radiological Health (HFZ-404), Food and Drug Administration, 9200 Corporate Blvd., Rockville, MD 20850, 301-594-1190.

SUPPLEMENTARY INFORMATION:

I. Statutory Background

Under section 513 of the Federal Food, Drug, and Cosmetic Act (the act) (21 U.S.C. 360c), FDA must classify devices into one of three regulatory classes: Class I, class II, or class III. FDA classification of a device is determined by the amount of regulation necessary to provide a reasonable assurance of safety and effectiveness. Under the Medical Device Amendments of 1976

(the 1976 amendments (Public Law 94–295)), as amended by the Safe Medical Devices Act of 1990 (Public Law 101–629), devices are to be classified into class I (general controls) if there is information showing that the general controls of the act are sufficient to assure safety and effectiveness; into class II (special controls), if general controls, by themselves, are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls to provide such assurance; and into class III (premarket approval), if there is insufficient information to support classifying a device into class I or class II and the device is a life-sustaining or life-supporting device or is for a use that is of substantial importance in preventing impairment of human health, or presents a potential unreasonable risk of illness or injury.

Most generic types of devices that were on the market before the date of the 1976 amendments (May 28, 1976) (generally referred to as preamendments devices) have been classified by FDA under the procedures set forth in section 513(c) and (d) of the act through the issuance of classification regulations into one of these three regulatory classes. Devices introduced into interstate commerce for the first time on or after May 28, 1976 (generally referred to as postamendments devices) are classified through the premarket notification process under section 510(k) of the act (21 U.S.C. 360(k)). Section 510(k) of the act and the implementing regulations (21 CFR part 807) require persons who intend to market a new device to submit a premarket notification report (510(k)) containing information that allows FDA to determine whether the new device is “substantially equivalent” within the meaning of section 513(i) of the act to a legally marketed device that does not require premarket approval.

On November 21, 1997, the President signed into law FDAMA (Public Law 105–115). Section 206 of FDAMA, in part, added a new section 510(m) to the act. Section 510(m)(1) of the act requires FDA, within 60 days after enactment of FDAMA, to publish in the **Federal Register** a list of each type of class II device that does not require a report under section 510(k) of the act to provide reasonable assurance of safety and effectiveness. Section 510(m) of the act further

provides that a 510(k) will no longer be required for these devices upon the date of publication of the list in the **Federal Register**. FDA published that list in the **Federal Register** of January 21, 1998 (63 FR 3142). In the **Federal Register** of November 3, 1998 (63 FR 59222), FDA published a final rule codifying those exemptions.

Section 510(m)(2) of the act provides that, 1 day after date of publication of the list under section 510(m)(1), FDA may exempt a device on its own initiative or upon petition of an interested person, if FDA determines that a 510(k) is not necessary to provide reasonable assurance of the safety and effectiveness of the device. This section requires FDA to publish in the **Federal Register** a notice of intent to exempt a device, or of the petition, and to provide a 30-day comment period. Within 120 days of publication of this document, FDA must publish in the **Federal Register** its final determination regarding the exemption of the device that was the subject of the notice. If FDA fails to respond to a petition under this section within 180 days of receiving it, the petition shall be deemed granted.

II. Criteria for Exemption


There are a number of factors FDA may consider to determine whether a 510(k) is necessary to provide reasonable assurance of the safety and effectiveness of a class II device. These factors are discussed in the guidance the agency issued on February 19, 1998, entitled “Procedures for Class II Device Exemptions from Premarket Notification, Guidance for Industry and CDRH Staff.” That guidance can be obtained through the World Wide Web (WWW) on the CDRH home page at <http://www.fda.gov/cdrh> or by facsimile through CDRH Facts-on-Demand at 1-800-899-0381 or 301-827-0111. Specify “159” when prompted for the document shelf number.

III. Petitions

On September 14, 1999, FDA received a petition from IMPRA, Inc., requesting an exemption from premarket notification for vascular tunnelers. Vascular tunnelers are currently classified under 21 CFR 870.3460 as an accessory.

IV. Comments

Interested persons may, on or before (*insert date 30 days after date of publication in the Federal Register*), submit to the Dockets Management Branch (address above) written comments regarding this notice. Two copies of any comments are to be submitted, except that individuals



may submit one copy. Comments are to be identified with the docket number found in brackets in the heading of this document. The petitions and received comments may be seen in the office above between 9 a.m. and 4 p.m., Monday through Friday.

Dated: 11. 4. 99
November 4, 1999



Linda S. Kahan
Deputy Director for Regulations Policy
Center for Devices and Radiological Health

CERTIFIED TO BE A TRUE COPY OF THE ORIGINAL



[FR Doc. 99-???? Filed ??-??-99; 8:45 am]

BILLING CODE 4160-01-F