

*Draft – Not for Implementation*

Guidance for Manufacturers  
and FDA Reviewers

**Guidance on Medical Device  
Patient Labeling**

***Draft Guidance – Not for Implementation***

**This guidance document is being distributed for comment purposes only.  
Draft released for comment on [release date as stated in FR Notice]**



**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Devices and Radiological Health  
Systems Analysis and Human Factors Branch I  
Division of Device User Programs and Systems Analysis  
Office of Health and Industry Programs**

## **Preface**

### **Public Comment:**

For 90 days following the date of publication in the Federal Register of the notice announcing the availability of this guidance, comments and suggestions regarding this document should be submitted to the Docket No. assigned to that notice, Dockets Management Branch, Division of Management Systems and Policy, Office of Human Resources and Management Services, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD 20852.

### **Additional Copies:**

World Wide Web/CDRH home page at <http://www.fda.gov/cdrh/HumanFactors.html> or CDRH Facts on Demand at 1-800-899-0381 or 301-827-0111, specify number 1128 when prompted for the document shelf number.

**Table of Contents**

**Introduction:** .....5

    What is the purpose of this guidance? .....5

    What is medical device patient labeling? .....5

    Why is medical device patient labeling important?.....6

    When should you use medical device patient labeling? .....6

    When is medical device patient labeling not usually necessary?.....7

    What are the general types of information that may be included in medical device patient labeling? .....8

    What should you consider when identifying a method to distribute the medical device patient labeling? .....9

**Suggested Content of Medical Device Patient Labeling:** .....10

**Sequence:** .....10

**Table of Contents:**.....10

**Descriptive Information:**.....10

        Purpose of the device (indications for use):.....10

        Description of the device: .....10

        When the device should not be used (contraindications):.....11

        Risks and benefits:.....11

        Expectations of the device and the procedure associated with the device .....13

        General warnings and precautions: .....14

        Explain the importance of the need to adhere to a care regimen:.....14

        Alternatives to the device and treatment: .....14

**Operating Information:** .....15

        Give clear setup instructions:.....15

        Give clear check-out procedures: .....16

        Give clear and easy-to-follow operating instructions: .....16

        Explain the need to monitor the activity of the device: .....16

        Give clear, complete cleaning instructions: .....17

        Describe maintenance and who should do it: .....17

        Give clear storage instructions:.....17

        Give the expected failure time and mode and its effect on the patient: .....17

        Give instructions on accessories: .....18

        Give instructions on related, additional devices: .....18

*Draft – Not for Implementation*

<b>Troubleshooting Information:</b> .....	18
Easy-to-find troubleshooting section:.....	18
<b>Additional Information:</b> .....	19
Clinical studies:.....	19
Disease and self-care information:.....	19
Adverse events:.....	20
Warranty:.....	20
Travel or international use:.....	20
<b>Index/Glossary:</b> .....	20
<b>Date of Printing:</b> .....	21
<b>User Assistance Information:</b> .....	21
<b>Appendix A</b> Readability:.....	22
<b>Appendix B</b> Writing for increased comprehension (writeability):.....	24
<b>Appendix C</b> Appearance of text:.....	33
<b>Appendix D</b> Appearance of graphics:.....	35
<b>Appendix E</b> Warnings and precautions:.....	37
<b>Appendix F</b> Pretesting: .....	41
<b>Checklist Summary:</b> .....	43
<b>References:</b> .....	45
Sequence:.....	45
Risks and benefits:.....	45
Additional Information: .....	46
User Assistance Information: .....	46
Readability:.....	46
Writing for increased comprehension (writeability):.....	47
Appearance of text:.....	48
Appearance of graphics:.....	48
Warnings and precautions:.....	49
Pretesting: .....	51
Usability testing: .....	51

# **Guidance<sup>1</sup> on Medical Device Patient Labeling**

## **Introduction:**

### **What is the purpose of this guidance?**

This guidance serves a dual purpose:

- (1) to assist manufacturers in their development, and
- (2) to assist Center reviewers in their review and evaluation

of medical device patient labeling to help make it understandable to and usable by patients (or family members or other lay persons caring for patients).

The writing style we have adopted in this guidance is targeted to manufacturers, since they will be developing the medical device patient labeling.

### **What is medical device patient labeling?**

Medical device patient labeling is any information associated with a device targeted to the patient or lay caregiver. It is intended to help assure that the device will be used safely and effectively. This labeling may pertain to therapeutic, restorative, or diagnostic devices. Medical device patient labeling is usually supplied as patient brochures, patient leaflets, or videotapes. It is intended to be supplied, or given to and used by patients or their lay caregivers with or without accompanying professional counseling. It may accompany devices intended solely for physicians to operate, devices for both physicians and patients or lay caregivers to operate, and devices operated solely by patients or their lay caregivers.

---

<sup>1</sup> This document is intended to provide guidance. It represents the Agency's current thinking on this topic. It does not create or confer any rights for or on any person and does not operate to bind FDA or the public. An alternative approach may be used if such approach satisfies the requirements of the applicable statute, regulations, or both.

## **Why is medical device patient labeling important?**

Medical device patient labeling is essential to assure safe and effective use of many devices. It informs patients or their lay caregivers about proper use, risks, and benefits of the device in language they can understand.

Medical device patient labeling assists patients or their lay caregivers in understanding the device; its operation, care, and maintenance; the way it interacts with the body to accomplish its purpose; and its place and purpose in the patient care regimen.

Patient labeling is important for all devices operated by lay users. Adequate directions for operating the devices are needed to make devices safe and effective. For example, as more patients use prescription devices at home, medical device patient labeling becomes necessary to better communicate to the lay person how to operate the device.

## **When should you use medical device patient labeling?**

Medical device patient labeling should be supplied whenever it can benefit patients or lay caregivers. Patients or lay caregivers benefit from this labeling by increasing their knowledge about the device. Knowledgeable users are more likely to use the device as the manufacturer intends.

Medical device patient labeling is needed when patients or lay caregivers need to:

- operate, interpret, and manipulate the device;  
(e.g., programmable implants, home pregnancy test kits, ostomy supplies)
- understand the effect of the device on the patient or others;  
(e.g., HIV testing; rods, screws, and fixation devices)
- give personal health information to aid their health care practitioner in deciding to use or not use devices in prevention, treatment, or diagnosis;  
(e.g., MRI, endoscopes)
- cooperate with the prevention, treatment, or diagnosis;  
(e.g., ultrasound devices, in vitro diagnostics)

## *Draft – Not for Implementation*

- monitor and report on the operation or output of a device;  
(e.g., pacemakers, glucose monitors)
- maintain the device;
- explain the operation of the device to others, such as a practitioner caring for another condition of the patient;
- explain the patient’s medical situation to others, such as when lay caregivers or others need to understand the requirements of care or the alterations of lifestyle associated with care;
- select among similar devices or device procedures;
- choose (buy, rent) among similar devices;
- know how to alter his or her lifestyle or care regimen to properly integrate the use of the device.
- be involved in deciding whether to have a procedure involving the device; or
- know how to be careful in using the device, such as understanding the basis for warnings, precautions, and contraindications.

### **When is medical device patient labeling not usually necessary?**

Medical device patient labeling is not usually necessary when:

- a patient will have no opportunity to benefit from the labeling. The following are typical situations.
  - The device is a tool of the health care practitioner and the patient is not involved in the choice of the device.
  - The device is a tool of the health care practitioner and the patient has no control over or access to the device.
  - The patient can contribute no useful information, such as sensitivities or aversions to materials or abilities to cope with sequelae of device use, that would help the health care practitioner choose or use the device.  
(e.g., blood pumps, scalpels, or other surgical instruments)
- a patient’s opportunity to benefit from patient labeling is outweighed by the risk of allowing him the opportunity. The following is a typical situation.

## *Draft – Not for Implementation*

- Time is of the essence so that involving the patient, family member, or lay caregiver to help decide on whether to use the device or select among several devices or procedures using devices would be more risky to the patient's health than not involving him. In this case, the health care practitioner subordinates the patient's right to know, choose, and decide to his obligation to give prudent care.

(e.g., devices and device procedures whose need arises intraoperatively and the patient had no opportunity to benefit from patient labeling before the procedure, such as some catheters or anesthesia gas machines)

### **What are the general types of information that may be included in medical device patient labeling?**

There are two general categories of information that may be included in medical device patient labeling:

- (1) risk/benefit information and
- (2) instructions for use.

Risk/benefit information is information a person needs to decide to use a device or have it used on them. This information also allows the user to become aware of potential problems with the device. It might include, as appropriate to the device:

- sufficient descriptive information to tell what the device is and what it is used for,
- types of people and situations for whom the device would not be a good choice,
- risks and benefits associated with the uses of the device,
- information about maintaining the device or identifying potential problems with the device,
- alternative therapeutic and diagnostic choices available,
- other information to enable the person to make an informed decision about the device.

Devices that would have patient labeling that would consist primarily or completely of risk/benefit information might be: implants that have no external patient interface once they are implanted, or prescription diagnostic or therapeutic devices that the patient is actively involved in choosing (e.g., laser eye surgery, lithotripsy, intraocular lenses).

Instructions for Use are the procedural steps to follow in setting up, using, cleaning, troubleshooting, and storing a device. This information constitutes the "how to" for the device.

Devices that might have patient labeling that would include instructions for use would be those the patient or lay care giver has to set up, operate, clean, etc. They might include such devices

## *Draft – Not for Implementation*

as suction equipment, intravenous infusion pumps, physical therapy equipment, or transdermal electrical nerve stimulation (TENS) devices.

There are many types of devices for which the medical device patient labeling would have both risk/benefit and instructions for use information.

### **What should you consider when identifying a method to distribute the medical device patient labeling?**

To increase the likelihood that the patient labeling will reach the patient and be read by the patient or their caregiver, consider:

- Must an intermediary, such as a health professional or a supplier, pass along the patient labeling or can it be provided directly to the patient?
- If an intermediary is involved, will there be any logistical difficulties in passing the patient labeling along and can these be overcome?
- When should the patient labeling be provided?
- What are the obstacles to the patient labeling reaching the patient, and what are the solutions to these obstacles?

During pretesting with your target audience, you can ask what method of distribution the audience would find best and assess experience they have had in the past with the distribution of patient labeling. (See Appendix F Pretesting.)

## **Suggested Content of Medical Device Patient Labeling:**

### **Sequence:**

The sequence of the following subheadings is being suggested, but is not a requirement for patient labeling. Medical device patient labeling should flow logically in brief, simple, clear, language with highlighting to guide the user to the desired information to meet the need at the moment. The content should be sequenced in a way that is logical for the intended user, with the most important information presented first. Notice that similar information in this suggested sequence is grouped (chunked) together, an organizational technique that participants in FDA's medical device patient labeling focus groups said they liked. (See References – Sequence.)

Not all of the following subheadings apply to all medical device labeling. For example, medical device patient labeling presenting only risk/benefit information might not contain subheadings related to instructions for use.

### **Table of Contents:**

Include a Table of Contents if the medical device patient labeling is lengthy and/or complex. This helps readers quickly and easily find the information they need.

### **Descriptive Information:**

#### **Purpose of the device (indications for use):**

Briefly describe the FDA cleared or approved indications for use.

#### **Description of the device:**

Give a brief physical description of the device, its parts and accessories. A graphic may be the simplest and clearest way to describe a device. All parts of the device shown in the graphic, such as switches, dials and meters, should be labeled with numbers, letters or words. The function or purpose of each labeled item should be briefly described in the text of this section.

## **When the device should not be used (contraindications):**

Tell when a device should **not** be used (contraindications). Contraindications are conditions under which the device should not be used because the risk of use clearly outweighs any possible benefit. There may be persons in whom the device should not be used because of their health status. For example, the device may be contraindicated for pregnant women.

List known and reasonably foreseeable hazards, not theoretical possibilities. For example, if hypersensitivity to a material in the device has not been demonstrated, it should not be a contraindication.

Contraindications to the use of a device could include:

- demonstrated hypersensitivity to a material where there is patient contact with the device,
- substantial risk of being harmed because of patient characteristics (e.g., age, gender, accompanying therapy, disease state, health status), or
- continued use in the face of an unacceptably hazardous adverse event.

If no contraindications are known, this section of the medical device patient labeling should state “None known.”

## **Risks and benefits:**

To make an informed choice about a medical device, a patient must have a thorough understanding of the effects and expectations of that device. We characterize this as “risk/benefit” information because the decision making process typically involves the weighing of the positive and negative effects and expectations. It is a global concept that may include a number of specific types of information, depending on the device and the target audience.

The goal of risk/benefit communication as applied to medical device patient labeling is to provide the patients with information about the risks and benefits associated with a device or procedure in a manner that is meaningful to the user. Risks and benefits conveyed in an effective and meaningful way to the user should aid the user in deciding whether to use a device, or undergo a procedure that uses a device, and to motivate the user to use the device as labeled.

## **General guidance for risk/benefit development:**

- Anticipate and respond to people’s concerns about their personal risk.
- Carefully word messages that may arouse fear and anxiety.
- Tell patients what they can do specifically to avoid risks.

## *Draft – Not for Implementation*

- Describe risks and benefits clearly and specifically.
- State both benefit and risk information in the same way (e.g., qualitative or quantitative), if possible.
- Be positive in the risk messages. For example, state what the patient can do to overcome risk.  
(e.g., “In rare cases a sweat rash can develop in the contact area. If a sweat rash develops, take the silicone sheet off for a few days, then start again...”)
- Balance risk and benefit information. Present factual risk and benefit information without any attempt to influence the patient.
- Pretest risk messages to indicate how effective they are. (See Appendix F Pretesting.)
- Use graphs and other visual materials to communicate risk information. Visuals should be colorful enough to attract attention and simple enough to be understandable at a glance. Combining visuals with brief text that contains the “take-home” message can help to ensure more accurate interpretation of risk information.
- Inform and educate people about risks by using risk comparisons.

### **The content of risk messages:**

Risk messages should closely reflect the perspective, technical capacity, and concerns of the target audience. The message should:

- include specific actions that people can take, even if it is only to tell them where to go to get further information or assistance.
- avoid using vague or unfamiliar terms to characterize the risk (e.g., “some women,” “fourfold,” “lifetime risk”). Word choice greatly influences how people attend to risk information (e.g., “risk” raises alarm, “chance” minimizes alarm).
- respect the audience by addressing people’s values, preferences, and concerns (e.g., Tell patients about a change in their appearance, such as scarring.),
- seek strictly to inform, not influence.
- assume the audience has little technical knowledge when in doubt about the audience.
- show empathy – with statements that illustrate caring.
- provide facts - two to three bits of supporting data.
- use vivid, concrete images, examples, and anecdotes that communicate on a personal level and make technical risk data come alive.
- limit the comparisons to risks that are similar or closely related when making comparisons.
- caution against unwarranted conclusions.
- provide information on consequences of decisions.

## *Draft – Not for Implementation*

Risk messages:

- should provide a source for more information.
- should acknowledge uncertainties.
- should include analogies.
- may discuss the nature of the risk.
- may include alternatives.
- may discuss benefits.

### **The content of benefit messages:**

- Present balanced benefit information.
  - Identify outcomes.
  - Estimate the magnitude of outcomes.
  - Evaluate the benefit for individuals.

### **Evaluation of the effectiveness of the risk/benefit communication:**

It is important to evaluate the effectiveness of the risk/benefit communication message by testing it. (See Appendix F Pretesting.)

- Is there adequate *awareness* of the risks and benefits, and their potential consequences?
- Is there *knowledge* of the risks and benefits, and what steps can be taken to lessen the risks?
- What are the *attitudes* toward the risks and benefits?
- What is the *behavior* toward the risks and benefits?

### **Expectations of the device and the procedure associated with the device:**

Tell the patient what to expect before, during, and after a surgical procedure and/or the use of the device. If appropriate, give instructions on post-operative or post-procedural care. If the device is one that the patient operates, and the medical device patient labeling has an “Operating Information” section, the information on what to expect could be included in that section.

## **General warnings and precautions:**

Embedded in the concept of risk/benefit information is the type of information known as “**general warnings and precautions**”. This is the specific hazard alert information that a patient needs to know before using the device. In presenting this specific information, it might be generally assumed that the decision had already been made to use the device. At this point, the users know the global risks and benefits, but need specific information to avoid or reduce a particular hazard associated with the use of the device. It may be appropriate to include general warnings and precautions in a presentation of risk/benefit information for a device where that information might have bearing on the decision to use the device as well as on the actual safe use/operation of the device.

Warnings and precautions tell the reader about hazards, other than those that are contraindications to device use. Warnings and precautions provide information on how to avoid these hazards, i.e., sources of harm in the use of the device. (See Appendix E for development and use of effective warnings and precautions.)

## **Explain the importance of the need to adhere to a care regimen:**

State why it is important to follow the care regimen explained in the medical device patient labeling. This will help to motivate the user to follow the instructions.

## **Alternatives to the device and treatment:**

When the patient has a choice among viable, effective, appropriate alternatives, the patient should be informed of the alternatives. If alternative devices or treatments are available that involve significant differences in such factors as risks, discomfort, or accuracy, information about the alternatives should be provided.

There are several approaches to discussing alternatives to the device and treatment. The medical device patient labeling can mention the existence of alternatives and refer the patient to the physician to get this information. Mentioning the existence of alternatives and referring the patient to the physician may be appropriate when there are too many alternatives to state in the text, or there are special circumstances surrounding the alternatives (e.g., investigational device). Other approaches include listing alternatives or listing and describing alternatives.

## **Operating Information:**

The user needs to know what to do, how to do it, and when to do it. The operating instructions should:

- focus on **how to** operate the device. It is usually not necessary to provide a detailed explanation for lay users of the mechanism of action of the device or why it does what it does. That approach can lead to information overload.
- assume that the user does not have device or medical knowledge or ability.
- provide logically ordered steps for the task and make the user aware of the importance of doing the steps in order.
- state the purpose and the expected outcome of each task.
- tell the user what steps are essential and which ones are optional.
- be written at an eighth-grade reading level or below to reach most of the population.
- be clear the first time they are read. Many people do not reread something they do not understand.

## **Give clear setup instructions:**

If a user is not responsible for the setup of the device, tell the user this and omit the setup instructions.

Include in setup instructions for the lay user:

- a parts list, if appropriate.
- list of materials and tools needed for setup.
- unpacking instructions, if appropriate.
- directions for where the device should be placed, such as a table top or floor. Also state if the device should remain in one place after setup.
- any warnings or safety instructions specifically related to setup, placed right before the corresponding task or instruction.
- results of incorrect setup.
- numbered setup instruction steps in logical order.
- any special preparation required before first use of the device, such as cleaning or disinfecting.
- space to write in user-specific instructions.
- who to call if there is a problem. Users may be referred to the assistance section in the medical device patient labeling.

## **Give clear check-out procedures:**

If the device requires any type of check-out procedure for safety and effectiveness, this process should be completely explained. This task may be as simple as a visual inspection of the device. Other examples of check-out procedures are calibration and quality control checks.

Include:

- when the check-out should be done, such as at the time of setup and/or before each use.
- step-by-step procedures for checking proper function of necessary parts of the device.
- what to do if the check-out shows that the device is not working properly.
- who to call if there is a problem. Users may be referred to the assistance section in the medical device patient labeling.

A clock or calendar graphic may be useful to show the user correct times or days to check the device.

## **Give clear and easy-to-follow operating instructions:**

Operating instructions should include:

- special preparation the user needs before operation, such as handwashing or device warm-up procedures.
- any warnings or safety instructions specifically related to operation, placed immediately before the corresponding task or instruction.
- results of incorrect operation.
- operating steps in logical order, with the expected results.
- space for user-specific instructions.
- who to call if there is a problem. Users may be referred to the assistance section in the medical device patient labeling.

## **Explain the need to monitor the activity of the device:**

State the importance of monitoring the activity of the device. This section explains that the user needs to make sure the device is acting the way it should. Give examples of what to check to make sure the device still works properly. For example, check to see that leads are connected, wires are not damaged, and the power is supplied to the device.

### **Give clear, complete cleaning instructions:**

- List the supplies needed and give step-by-step procedures.
- State how often the device should be cleaned.
- Tell the user what cleaning accomplishes.
- Tell the user what the results of failure to clean will be.
- Include appropriate warnings and cautions for cleaning agents.
- Describe the results of using improper cleaning solutions or methods.

### **Describe maintenance and who should do it:**

- Clearly describe what maintenance actions are the responsibility of the user.
- If a particular kind of user, such as the lay user at home, is not responsible for maintenance, proper maintenance actions should be briefly outlined, who is responsible and how often the action should be done. The lay user will then know what to expect and can take action if proper maintenance is not provided.
- If the device has some maintenance procedures to be done by the user and some done by others, such as the biomedical engineer, you may wish to write this section in two parts. The two parts will help make clear to users what they should and should not be doing to maintain the device.

### **Give clear storage instructions:**

Clearly describe proper preparation for storage and storage conditions. State the results of improper storage conditions. If extended storage may affect the device, inform the user. It may be necessary to include information that addresses extended storage in the sections on setup, check-out, operation, and maintenance.

### **Give the expected failure time and mode and its effect on the patient:**

State how long the device will last. State what to expect when the device fails. Let the user know if it fails safely or dangerously and the effect on the patient. Include this type of information in the Risks and Benefits section of the medical device patient labeling if appropriate.

### **Give instructions on accessories:**

If the device comes with or is used with accessories, discuss all appropriate content areas for each accessory. You could have a separate accessories section or include information on the accessories in the content areas that apply.

You may need to include a general warning at the beginning of the medical device patient labeling advising users of problems that may occur if accessories are used other than those recommended by the manufacturer.

### **Give instructions on related, additional devices:**

If the patient will receive an additional device after an operation or procedure, information about the additional device should be included in the medical device patient labeling. Describe why this additional device would be needed, what symptoms might be expected post-treatment, and any other information about the additional device that impacts the safe and effective use of the primary device. A typical situation is when a patient who has had a home infusion pump with a peripheral line also receives a central line.

### **Troubleshooting Information:**

#### **Easy-to-find troubleshooting section:**

When a problem occurs, troubleshooting helps determine if the problem is with the device or with the patient's condition. Anticipate any problems the user may have with setup, operation, or maintenance. Provide solutions for these problems in the troubleshooting section. Group similar problems, such as problems with alarms, and highlight each group heading. Highlighting makes it easy to find each group of problems and the specific problems in it. Put the most life-threatening problems first in each section.

Format this section so that the user can locate specific problems quickly. The troubleshooting section could be a table with a column for signs of trouble and a column for actions.

Clearly describe the symptom of each problem in as few words as possible so that the user can easily match the description to the problem observed. If your device displays error messages, list them and what they mean. Explain the steps necessary to correct the problem. Do not confuse the reader with technical reasons for problems unless the reasons are important to the corrective action.

## *Draft – Not for Implementation*

If there are problems that users cannot or should not try to solve themselves, include warnings or cautions and tell them how to get help.

Instruct users to call their health care professionals for emergency assistance if troubleshooting reveals a patient health problem rather than a device problem.

**Note:** Putting assistance information in this section does not eliminate the need for a separate assistance section. (User assistance information is discussed in a later section.)

Tell the user how to report undesirable outcomes (adverse events). For example, the user would report malfunctioning of the device, mistakes in using the device, or injury from the use of the device.

### **Additional Information:**

There are certain categories of information that not all patients want in medical device patient labeling. We refer to these categories as “additional information.” Additional information includes such categories as clinical studies and adverse events. Whether or not to include a particular category depends on the needs of the target audience of the particular medical device. Include additional information in appendices in the medical device patient labeling, or advise the reader that it is available on demand. For example, tell the reader to call the user assistance toll-free telephone number, contact the physician, or go to an Internet site. FDA’s medical device patient labeling focus testing has shown that some patients like all the additional information, but most prefer to see it at the end of the document or have it available on demand so they can choose to read it or not. (See References – Additional Information.)

### **Clinical studies:**

Provide the clinical study information in simple, plain language.

### **Disease and self-care information:**

Members of some target audiences may benefit from disease and self-care information. For example, a patient undergoing bone densitometry benefits from information about osteoporosis.

## **Adverse events:**

Provide information about any adverse events. Devices whose applications are supported by clinical trials will have data about adverse events that occurred during these trials and that may be of value to the device user. Other devices may have adverse event data from other sources, e.g., published literature or experience with similar devices. The detail in, and the need to include, an Adverse Events section depend on the benefit of these data to the device user. For a device cleared under Premarket Notification, which was not supported by clinical studies, the adverse events section might include only potential adverse events and a statement of the source of the information.

Potentially fatal adverse events should be included in the Warnings and Precautions section, or the Contraindications section.

The listing of adverse events should be followed, if appropriate, by statements directing the reader to other sections of the labeling for additional information regarding these adverse events and any steps that should be taken to avoid them.

## **Warranty:**

Provide any warranty information.

## **Travel or international use:**

If a patient could travel with the device, provide any appropriate travel instructions. For example, tell the patient if the medical device is not compatible with foreign power systems. The patient will need an adapter, and may need a converter to convert to the proper voltage. Also, tell the patient to check with the carrier to confirm that the device can be carried and/or used on the airplane.

## **Index/Glossary:**

Provide an index if the medical device patient labeling is lengthy and/or, complex. This helps readers quickly and easily find the information they need. If a glossary is used, it should be placed after the Table of Contents to alert readers that it is there to help them. Whether or not a glossary is used, definitions should appear in the text.

**Date of Printing:**

Put the date that the medical device patient labeling is issued or revised where it can be easily found (e.g., cover or last page). It is required (21CFR801.109(e)) to date labeling for prescription devices and it is recommended for all other devices.

**User Assistance Information:**

Design a clearly marked section that advises users on how to get help for problems with the device. FDA’s medical device patient labeling focus groups indicated that users look at the end of the medical device patient labeling for the user assistance information, although it can be included in other places in addition to the end. (See References – User Assistance Information.) This section should be very easy for the user to find. It may be as simple as putting the customer assistance number near the company name, device name, and model number. The medical device patient labeling can include a toll-free number or the number for customer assistance, as well as an Internet address.

Provide space near the user assistance information, with appropriate identifying terms, for phone numbers of the medical equipment supplier, the home health care agency, the doctor, and/or any other appropriate points of contact for the typical user.

*Note:* Ideally, the manufacturer’s toll-free number should be on the device.

## **Appendix A**

### **Readability:**

#### **Readability defined:**

Readability is defined as the ease of understanding or comprehension achieved by the style of writing. Reading involves both decoding and comprehension. The reader must be able to recognize (decode) the words in the medical device patient labeling as well as comprehend the meaning of the text. You are encouraged to follow the following guidance to assess and enhance readability.

#### **Assessing readability:**

To assess readability, analyze qualitative factors (e.g., explanation of jargon, careful organization) in combination with quantitative factors (results of readability formulas). You should test the medical device patient labeling with a sample of likely, representative users of the device.

#### **A qualitative analysis:**

Use these qualities in the text to enhance the reader's comprehension of the medical device patient labeling:

- Define complex medical terminology and jargon when it first appears in the text.
- Carefully organize medical device patient labeling from a user's perspective (organized in the way the user will use the information). Repeat important points and summarize important information, to increase the reader's recall and reading comprehension. The reader will remember the message when key points are reinforced.
- Organize sections with headings and questions.

#### **A quantitative analysis (readability formulas):**

The reading level of the medical device patient labeling should be no higher than the eighth-grade level, the average reading level among adults. To predict the reading level of the medical device patient labeling, use the readability formulas (manual or software programs) available. (See References: Readability.)

Readability formulas use semantic (vocabulary difficulty) and syntactic (sentence length) factors to predict the readability of the medical device patient labeling. If the reading level is predicted to be above the eighth-grade level, the patient labeling should be rewritten by applying

## *Draft – Not for Implementation*

writeability principles (see Appendix B,) and qualitative factors that can enhance reading comprehension. These formulas should be used to predict (not measure) the reading level of the text.

**The formulas should not be used to write, rewrite and revise medical device patient labeling to specific readability levels. While readability formulas can predict readability, the reader must actually read the text to determine if it is readable.** To

mechanically substitute easier for harder words does not necessarily make the text more readable. Such text could become harder to understand through loss of organization and cohesion and loss of clarity of concepts. As reading expert George Klare explains, merely shortening words and sentences to improve readability is like holding a lighted match under a thermometer when you want to make your house warmer. The thermometer certainly goes up, but the room does not get any warmer.

### **Testing:**

Test the medical device patient labeling with a sample of appropriate users of the device. This is the only way to know if the medical device patient labeling is understandable and useful. User-oriented testing helps to find places where the medical device patient labeling may be inaccurate, incomprehensible or poorly organized. Since reading is an interactive process between the reader and the text, testing the medical device patient labeling with the target audience is the best way to determine how well the audience understands the labeling.

### **A summary of readability:**

Use qualitative factors in concert with readability formulas to assess the readability of medical device patient labeling. Reader background knowledge and interest, context clues (words that surround a particular word or passage and can throw light on its meaning), text organization and opportunity for reinforcement are just a few factors other than formula-derived readability that should be used to predict the readability of medical device patient labeling. The best solution is to pretest the medical device patient labeling with the target audience to see if they comprehend it.

## **Writing for increased comprehension (writeability):**

Writeability refers to readable writing. The following principles increase the writeability of medical device patient labeling. You are encouraged to follow these principles in order to write for increased comprehension.

### **General principles:**

Write with a specific type of person in mind.

Stress the “need to know” information.

Use concrete examples to clarify abstract ideas.

Consider who will be using the device:

- Are they elderly, disabled or children?
- Would their vision or hearing likely be impaired?

Consider where the devices will likely be used.

Group (chunk) similar information together.

Use well-mapped, carefully organized writing. Well-organized material provides occasional repetition and thoughtful summary.

Repeat the most important points to increase patient recall and comprehension.

Emphasize and summarize main points.

Use headings and summaries to aid organization and provide message repetition.

Organize medical device patient labeling to meet varied skill and knowledge needs.

- One approach is to provide one well-segmented, highlighted document with a table of contents and the most-desired, basic information up front. Put additional information in appendices to the primary document, or make it available on demand.
- Another approach is to distribute a quick reference card of important reminder information for the experienced user, in addition to the full patient labeling. Also, the experienced user

## *Draft – Not for Implementation*

can benefit from an intermediate form (a mini manual) with an expansion of the important information in the quick reference card, especially key use information, in addition to the full patient labeling.

### **Locating information in lengthy and/or, complex medical device patient labeling:**

Include a Summary Page with critical information.

Include a Table of Contents.

Include an Index.

Include page numbers to make it easier to locate information.

Include chapter names and numbers, if chapters are used.

### **Principles for clear, concise writing:**

#### **Eliminate unnecessary words:**

- Avoid “aware of the fact that.”

#### Example

Poor: Be aware of the fact that dampness may affect the device and cause rust.

Better: Dampness may affect the device and cause rust.

- Avoid the unnecessary use of make, made and making.

#### Example

Poor: Make an attempt to clean your braces twice a day.

Better: Clean your braces twice a day.

- Substitute a single word for a phrase.

#### Example

## *Draft – Not for Implementation*

Poor: It may take a good deal of practice to operate the device.

Better: It may take much practice to operate the device.

- Use clear and simple phrases whenever possible.

### Example

Poor: Endeavor to ascertain the hospital closest to your home.

Better: Try to find the hospital closest to your home.

- Avoid overuse of “it is.”

### Example

Poor: It is possible that you may need to use more cleaner.

Better: You may need to use more cleaner.

- Reduce long, complicated phrases.

### Example

Poor: You are not able to use another manufacturer’s cable.

Better: You cannot use another manufacturer’s cable.

- Simplify prepositional phrases.

### Example

Poor: Store the device in a dry area at all times.

Better: Always store the device in a dry area.

## *Draft – Not for Implementation*

### **Technical vocabulary:**

- Define technical words or use them in context to help increase comprehension. Use lay language first with the technical word in parentheses. In addition to parentheses, use italics or other highlighting techniques to give the reader a signal for the technical vocabulary.

#### Example

Poor: 65 mm is the tolerance level.

Better: Do not set this gauge above 65 mm (*tolerance level*).

- Define terms the first time they occur in the text. Keep the definitions simple and concise. If you need to define many words on one page, define them in a set off section of the page on which the words first appear. For example, use a sidebar.
- Provide examples to explain technical words.
- Provide a glossary of technical words. If a glossary is used, it should be placed after the Table of Contents to alert readers that it is there to help them. Whether or not a glossary is used, definitions should appear in the text.

### **Word choice:**

- Personalize the material by using the second person “you” instead of the third person “he,” “she,” or “they.” Using “you” focuses the information directly to the patient, which makes it more important and personal.

#### Example

Poor: The user should not operate this device near water.

Better: You should not operate this device near water.

- Use terms consistently throughout the text. Use the same term to identify the device and its parts throughout the medical device patient labeling. Avoid synonyms or alternate phrases.

## *Draft – Not for Implementation*

### Example

If you start with “dial,” do not call it a “knob” later.

- Put adjectives and adverbs close to the words they modify.

### Example

Poor: Use the wire that is covered with green plastic.

Better: Use the green wire.

- Avoid adverbs that are difficult to define or interpret.

### Example

Poor: Respond quickly.

Better: Respond within one minute.

- Use active rather than passive verbs.

### Example

Poor: The dial should be turned clockwise.

Better: Turn the dial clockwise.

- Use action verbs, not nouns created from verbs.

### Example

Poor: Avoidance of cellular phones is necessary when operating the device.

Better: Avoid cellular phones when operating the device.

- Use specific terms. Vague terms may be misinterpreted.

## *Draft – Not for Implementation*

### Example

Poor: Device operates poorly in a cool room.

Better: Device will not operate below 60 degrees F.

- Avoid abbreviations or acronyms. If abbreviations or acronyms are necessary, define them the first time. Use them consistently.

### Example

Abbreviation: oxygen instead of O<sub>2</sub>

Acronym: home medical equipment  
instead of HME

### **Sentences:**

- The burden for short-term memory is greater for longer sentences. Use as few words as possible to present an idea or describe an action.

### Example

Poor: Find the opaque plastic container that has a blue line on the upper half of it and fill it with any type of water until you reach the blue line.

Better: Fill the plastic container to the blue line with tap water.

- Use no more than one clause in a sentence.

### Example

Poor: Check the power cord and do not use it if you find cuts or frays or it is loosely connected to the device.

## *Draft – Not for Implementation*

- Better:
1. Look at the power cord for cuts or frays. If it is cut or frayed, do not use the device.
  2. Tug lightly on the power cord. If it slips out of the device, do not use the device.
  3. Call 1-800-xxx-xxxx if you need help.

- Place phrases that describe or explain at the end of the sentence. Phrases at the beginning or in the middle of a sentence may be confusing.

### Example

Poor: Before using this device, you should read the instruction manual.

Better: Read the instruction manual before using this device.

- Write the way you talk. Avoid formal language.

### Example

Poor: Insert the blue cable into the blue socket on the anterior section of the machine to form a completed circuit of the electrical system.

Better: Plug the blue cord into the blue hole on the front of the machine.

- Express ideas of similar content in similar form.

### Example

Poor: Twist the large dial clockwise until it stops. Turning the small dial, move it 3 notches counterclockwise.

Better:

1. Turn the large dial marked X clockwise until it stops.
2. Turn the small dial marked Y counterclockwise 3 notches from the “Off” position.

- Users should be able to read instructions aloud. Do not use parentheses for information that should be read. Parentheses cause the reader to hesitate, making it hard to read. Use parentheses only for extra information such as technical terms.

## *Draft – Not for Implementation*

- Don't promote the product in the medical device patient labeling. Ads or promotions in the text will interfere with the patient's ability to follow instructions.
- Use bullets, lists or more than one sentence instead of a long sentence that requires a lot of punctuation.

### **Paragraphs:**

- Begin paragraphs with a simple topic sentence that states the main idea.
- Paragraphs should be cohesive about a single thought.

### **Motivation:**

- Focus on what the target audience should **do** as well as **know**.
- Use questions throughout the text as headings and summary points to encourage active learning.

### **Writing procedures:**

- Write procedures in short, identifiable steps. Put the steps in the order they should be performed.
- Before each set of steps, tell the reader how many steps are in the procedure. This helps the reader avoid missing steps.
- Number each step in Arabic numbers such as 1, 2, 3. Do not use Roman numerals such as I, II, III; letters such as A, B, C; or words such as one, two, three. People most readily identify Arabic numbers with steps in a sequence.
- Limit each step to no more than three logically connected actions. If actions are not related, they should be in separate steps.
- Make the instructions for each action clear and definite to prevent misunderstandings. This approach is especially critical for steps that require more than one action.

### Example

Poor: Turn the machine on.

## *Draft – Not for Implementation*

Better: To turn the machine on:

1. Plug the power cord into an AC outlet.
  2. Face the front of the machine. Find the black power switch on the right side.
  3. Turn the power switch to the “ON” position.
- Tell the user what to expect from an action.

### Example

Poor: Flip the switch to the “ON” position.

Better: Face the front of the machine. Flip the black switch on the left side, marked “ON/OFF,” to the “ON” position.

The green light will go on.

- Discuss common errors at the point in the procedures where they are likely to occur. Provide information to prevent and correct use errors.
- Each step should be contained on one page. If the entire step will not fit on a single page, break the step into smaller substeps, each fitting on a page or less. Put more than one step on a page only if each step and accompanying graphics are complete on that page.
- Avoid referring the user to another place in the manual for other information (cross referencing). It is confusing to the reader and interrupts the flow of the procedures. If cross referencing is absolutely necessary, make sure the reader knows when to return to the original place.

### Example

Poor: If the alarm sounds, go back to the beginning of chapter X.

Better: If the alarm sounds:

1. Turn to page X.
2. Repeat steps 1 and 2 on page X.
3. Return to step 1 on page Y.

**Appearance of text:**

**What principles should be applied to the physical features of the text?**

The speed by which letters and words can be recognized (legibility), can be enhanced by applying the following principles.

**Sections:**

Divide the text appropriately into short sections. Use headings or other highlighting/separation devices. Major headings, including captions and subtitles should capture the main points. Within the text, bullet text is helpful.

When creating headings, it is acceptable to use uppercase letters and bold print.

**Type:**

Base the type size and font on the needs of the target audience.

Use at least a 12 point type size whenever possible. For elderly or visually impaired users, use at least 14 points. For headings, use ranges from 18-36 points.

Use a serif font for the text. It allows more variation among letters than sans serif. This makes the letters easier to recognize.

Use both upper and lower case letters in the body of the text. It is more difficult to read all capital letters.

Use black type on a white background for best contrast. It is the easiest to read.

**Paper:**

Avoid glossy paper. It may cause a glare.

Use paper heavy enough to prevent text and graphics from showing through.

## *Draft – Not for Implementation*

### **Highlighting:**

Use highlighting techniques to emphasize important words, thoughts or phrases. Highlighting techniques include bolding, underlining, italics, capital letters, color, background patterns and white space.

Do not overdo highlighting. It will decrease the impact of the message.

Be consistent in the highlighting methods used.

### **White space:**

Use white space carefully to keep the medical device patient labeling from looking too cramped or too spread out. White space between blocks of text aids the ease of reading.

Justify the left-hand margin. Use ragged right margins.

Use at least 1/16<sup>th</sup> of an inch of white space between lines of text.

Increase the amount of white space around important individual words, text and graphics for emphasis.

### **Formatting and organizing instructions:**

Organize instructions with text, flowchart, or list formats as appropriate.

Use tables to present information graphically.

- It should complement and supplement the text.
- It should condense statistical information.

Use lists when the use of the device requires checking off steps that are completed.

Number steps that must be completed in order.

Bullet lists that have no specific order of importance.

## **Appearance of graphics:**

### **What principles should be applied to the graphics of the medical device patient labeling?**

Photographs and line art attract and keep a reader's interest and are often remembered longer than words. Properly chosen and placed illustrations make the text more meaningful and reduce the burden of details in the text. Use the following principles in the graphics of the medical device patient labeling.

Graphics should:

- attract attention,
- re-emphasize the text of the medical device patient labeling,
- be simple and clearly drawn, without clutter, unneeded background or extraneous detail,
- demonstrate one concept, single idea, or point of information at a time,
- be placed next to corresponding text,
- use cues such as circles or arrows to point out key information,
- be clearly labeled,
- be easy to understand,
- improve understanding of essential information,
- be recognizable to the audience, and
- fit the target audience.

Set off text and graphic that go together by the use of lines, white space or titles. If a graphic is referred to in the text, it should have a title, for example, Figure 1.

Use accurate and precise graphics.

- When comparing two illustrations, show the difference. If the difference is not distinct, the reader may get confused.
- Represent only simple concepts in your graphics, either of actions or of the device and its surroundings.
- Confine action graphics to a single action whenever possible.
- Use a separate graphic for each distinct idea.

Graphics should be large enough to see the focal point and important words clearly. Use as few words as possible. Captions should tell readers what to look for in the illustration. Eliminate detail that is not necessary. The clearest graphics have dark, sharp lines for good contrast.

## *Draft – Not for Implementation*

Line drawings and illustrations are clearer than photographs. Photos may have distracting extra images and poor contrast. Simple exploded views or cut-away views may be helpful. Use exploded views only for devices that the user should put together or take apart.

In instruction manuals, tables and graphs are normally not appropriate and their use should be minimized. If a table or graph is necessary, include instructions on its use. Label each table or graph clearly.

Apply the following to symbols and icons.

- A symbol is a sign or picture that has been developed to represent an idea. A symbol should be defined or explained because it doesn't mean anything by itself.
- An icon is a drawing that looks like the idea it is meant to represent. Use icons only with text to explain them.
- Use standardized symbols and icons, or those already understood by the general population. Make sure that your population understands the symbols that you use.

Do not use icons with commonly understood usages to illustrate examples inconsistent with those usages (e.g., a red octagon to indicate “go”).

## **Warnings and precautions:**

The purpose of this section is to:

- define and explain the terms *warnings* and *precautions*,
- discuss their use in medical device labeling,
- recommend approaches to effective presentation based on literature and research findings, and
- present some of the common issues associated with warnings and precautions.

*NOTE: Labeling a device with warnings and precautions is the least preferable method of controlling accidents and injuries. You should make every effort to design the device so that the hazard is eliminated. Only when this is clearly impossible should you resort to a warning or precaution in the labeling.*

## **What are warnings and precautions?**

*Warnings* and *precautions* are written, pictorial, and/or audible alerts to a hazard. The term used to identify the particular hazard presents the reader with a cue to the seriousness of the hazard.

A warning alerts the reader about a situation which, if not avoided, could result in death or serious injury. [ANSI Z535.4-1991] It may also describe potential serious adverse reactions and safety hazards. The term *warning* is generally used as the signal word for this type of hazard alert. If a problem may lead to death or serious injury, FDA may expect you to highlight the warning by placing it in a box.

The term *precaution* is used for the statement of a hazard alert that warns the reader of a potentially hazardous situation which, if not avoided, may result in minor or moderate injury to the user or patient or damage to the equipment or other property. [ANSI Z535.4-1998] It may also be used to alert against unsafe practices. This includes the special care necessary for the safe and effective use of the device and the care necessary to avoid damage to a device that may occur as a result of use or misuse. The word CAUTION! is generally used as the signal word for a precaution statement.

The distinction between warnings and precautions is a matter of degree of likelihood and seriousness of the hazard. The target audience for medical device labeling (health care practitioners and lay users of home use devices) generally recognize a hierarchy of hazard alerts, with warnings being those of a more serious nature and precautions being of a less serious, but important, nature.

## **What is the purpose of warnings and precautions in medical device labeling?**

The basic purpose of a warning or precaution is twofold:

- to inform users of potential hazards, and
- to persuade them to modify their behavior to avoid injury or device damage.

For a warning or precaution to be effective, readers must:

- perceive the threat to be both severe and relevant to them,
- believe that they can perform the recommended response, and
- believe that response will be effective in avoiding the hazard.

Effective warnings and precautions capture the reader's attention, are understood, are consistent enough with the reader's beliefs and attitudes to be accepted, and are persuasive enough to motivate the reader to comply. They invoke an appropriate level of fear arousal, conveying the nature and extent of the hazard, without being so strong that they backfire, causing the reader to select an alternative action or no action.

## **What is appropriate content of an effective warning or precaution?**

There are four elements generally recognized by the courts and research (See References – Warnings and Precautions) as necessary for an effective warning or precaution:

- **a signal word** (WARNING, CAUTION) to alert the reader that what follows is important hazard information. A symbol or icon may emphasize the effect of the signal word. Additional enhancement, such as bolding, larger type, underlining, italics, or color may help the information stand out from the rest of the text. However, studies have demonstrated that a large difference in font size between the signal word and the text may de-emphasize the importance of the text and therefore reduce the likelihood that the text will be read.
- **a hazard avoidance directive** in the form: **Do Not, Never, Avoid...** (or **Do**, if more appropriate) followed by the action to avoid (or perform). The objective of this directive is to give clear instructions to the user on how to avoid the hazard.
- **a clear statement of the nature of the hazard** associated with the warning (e.g., allergic reaction to material, strong magnetic field) or precaution (e.g., environmental effect, damage from reesterilization) that characterizes the severity and the likelihood.
- **the consequences**, specifying the serious adverse events, potential safety hazards and limitations in device use that result if users do not follow instructions. The purpose is to give them a clear idea of the risk, which is likely to increase compliance. Hazard alert research has shown that this element has a significant effect on readers. If the consequences are not included, the alert is likely to be less effective.

## *Draft – Not for Implementation*

The signal word should appear first. The order provided here for the other three elements (hazard avoidance, hazard identification, and consequences) will be appropriate for most instances, but may be altered as necessary to best communicate the information to the reader.

### **How is an effective warning or precaution formatted?**

No magic formula works in every instance. The writer must apply the principles discussed here for writing effective warnings and precautions as appropriate to the user and the device.

We suggest that warnings and precautions be presented in “clusters” of issues that provide the reader with the information in the most logical and usable order for the particular device. For instance, a surgical device may have the warnings and precautions grouped, or clustered, into Preoperative Information, Intraoperative Information, and Postoperative Information, with appropriately highlighted headers. A home use device may have warnings and precautions grouped according to Setup, Calibration, Use and Storage. There are many possible groupings of this important hazard avoidance information, including the use of a “Before You Start” subsection for devices deemed to need a select and very limited group of warnings and precautions to be read before the device is handled at all. Within each cluster, the information should be prioritized by clinical significance.

Warnings and precautions should be as concise and focused as possible while providing sufficient information. We recommend using bullets, rather than full sentences. Each bullet should contain a single item. We recommend against grouping a number of warnings or precautions in paragraph form. Users are more likely to read and comply with warnings or precautions presented in outline form using plenty of white space and consistent indentations, rather than paragraphs.

Additional approaches that can assist readers to notice, understand and comply with warnings and precautions include:

- Integrating warnings and precautions into the task/hazard-related context. This increases the chance that the reader will encounter the information when he is most receptive to it.
- Using concrete rather than abstract terms and jargon. Frank language makes the writing more interesting and therefore more likely to be remembered. For example, say “Contact with this product will produce severe burns” rather than “Contact with this product will result in serious injury”.
- Using the simplest possible construction.
- Making each warning and precaution conspicuous. The use of white space and simple highlighting techniques can call attention to the warning or precaution. Do not bury this important information in extensive text.
- Standardizing terms and formatting across the labeling and, if possible, across products to increase consumer recognition.

## **Where should warnings and precautions be placed for maximum effect?**

Research has determined that hazard alerts are most effective when integrated into the task information at the most relevant location. Medical device labeling has traditionally contained separate sections in which the warnings and precautions are presented. This approach predates much of the research into hazard alert effectiveness. However, there is a reluctance to abandon the placement of this important information in separate sections in medical device labeling because of user familiarity with this format. It is interesting to note that some research has shown that short, well highlighted warning messages can actually get users to read a set of longer, more detailed warnings in another section of the labeling. We recommend that warnings and precautions included in the separate sections be limited to those that can be taken out of procedural context and still be effective.

Procedurally related warnings and precautions that cannot be taken out of context should be located with the associated procedural step. Research differs on the most effective location relative to relevant text, but placement immediately before the associated procedural task has been shown to be effective. We suggest considering this approach for warnings and precautions that are included in procedural instructions.

## **What other issues should be considered in designing warnings and precautions?**

Because it is difficult to get readers to notice and read warnings and precautions, it is important to know what works with the target audience of the particular device. This is most effectively done by testing draft warnings and precautions on a sample of the target audience. Recommended testing points are when the labeling is being designed and again after it has been implemented. The latter testing can provide information to be used at the next labeling change.

Over-warning, that is including too many warnings and precautions, dilutes the strength of all of the hazard alerts. Researchers suggest the elimination of warnings and precautions about generally known hazards as a possible solution. Also, we recommend that writers use care in what is designated as a warning or precaution. Careless designation can have the same diluting effect as over-warning (e.g. “WARNING! Batteries not included.”). Repeated exposure to unnecessary hazard alerts (not relevant or already known) reduces the effectiveness of the important warnings and precautions.

We know that readability indexes can predict, but do not measure the reader’s actual ability to comprehend labeling. Because of the complexity of the process by which individuals interact with hazard alerts, the readability indexes should be considered particularly unreliable in predicting warning and precaution comprehension.

## **Appendix F**

### **Pretesting:**

#### **What is pretesting?**

Pretesting of medical device patient labeling is the systematic and formal gathering of target audience reactions to medical device patient labeling’s content and format, before the medical device patient labeling is issued in final form. It is typically one of the evaluations conducted in the early stages of message development. Pretesting may assist the manufacturer in determining which of a number of labeling presentations is most effective for the intended audience and in identifying strengths and weaknesses in the presentation. It can help to identify sensitive and controversial elements. Findings can be used to revise and improve materials before distribution to users. Also, pretesting can help the manufacturer identify the best method to distribute the labeling to the target audience.

#### **What methods of pretesting the medical device patient labeling should be used?**

Pretesting methods include methods such as individual in-depth interviews, focus group interviews, self-administered questionnaires, usability testing, and readability testing (See Appendix A Readability).

Pretesting can check the potential users’ comprehension of the medical device patient labeling and their ability to follow instructions in the medical device patient labeling in order to operate the device. User-oriented testing helps to find places where the patient labeling may be inaccurate, incomprehensible, or poorly organized.

#### **Individual in-depth interviews:**

A potential user provides ideas and impressions of possible ways that the medical device patient labeling could be most effectively written.

#### **Focus group interviews:**

A small group of potential users, usually 8 to 10 people, discusses their perceptions, opinions, beliefs, and attitudes (POBAs) toward the medical device patient labeling. The discussion is guided by a skilled moderator.

### **Self-administered questionnaires:**

Potential users are asked to review the medical device patient labeling, complete the written questions in the questionnaire about the medical device patient labeling, and return it within a specified time.

### **Usability testing:**

The concept of *usability* refers to the extent to which people who use a product can use it quickly and easily to accomplish specific tasks. The usability of a product is composed of the combined usability of the product's sub-components which can include hardware, software, menus, icons, messages, labels, manuals, reference materials, and software-based help. Consideration of the usability of a device may focus on all or some of the characteristics of the device that pertain to its use. For many kinds of devices, including medical devices, an important aspect of usability is the medical device patient labeling materials.

*Usability testing* is a technique designed to determine how usable a product is. This technique involves systematic observation of actual users trying out a product (or sub-component) and the collection of information from the users about aspects of the product that are easy and those that are difficult for them. Performing usability testing on medical device patient labeling materials involves the use of the device and its labeling materials by a group of people that correspond to the intended users of the device. Data are then collected on how well the labeling materials supports the users, how effectively they are able to use the device, how many and what kind of errors they make, and any difficulties they encounter.

From the perspective of medical device patient labeling reviews, it is desirable for manufacturers to demonstrate that labeling materials can be used safely and effectively through the application of *usability testing*. When usability test results are submitted, they can be used to assist the process of reviewing patient labeling. To the extent that safety concerns are adequately reflected in the test, these results can be considered an indication of the adequacy of the medical device patient labeling.

## **Checklist Summary:**

Use this checklist to make sure that you have considered all the recommendations in this guidance.

The medical device patient labeling contains the following content:

- \_\_\_ **Table of Contents**
- \_\_\_ **Descriptive Information:**
- \_\_\_ Purpose of the device (indications for use)
- \_\_\_ Description of the device
- \_\_\_ When the device should not be used (contraindications)
- \_\_\_ Risks and benefits
- \_\_\_ Expectations of the device and the procedure associated with the device
- \_\_\_ General warnings and precautions
- \_\_\_ Explanation of the importance of the need to adhere to a care regimen
- \_\_\_ Alternatives to the device and treatment
- \_\_\_ **Operating Information:**
- \_\_\_ Clear setup instructions
- \_\_\_ Clear check-out procedures
- \_\_\_ Clear and easy-to-follow operating instructions
- \_\_\_ Explanation of the need to monitor the activity of the device
- \_\_\_ Clear, complete cleaning instructions
- \_\_\_ Description of maintenance and who should do it
- \_\_\_ Clear storage instructions
- \_\_\_ Expected failure time and mode and its effect on the patient
- \_\_\_ Instructions on accessories
- \_\_\_ Instructions on related, additional devices
- \_\_\_ **Troubleshooting Information:**
- \_\_\_ Easy-to-find troubleshooting section
- \_\_\_ **Additional Information:**
- \_\_\_ Clinical studies
- \_\_\_ Disease and self-care information
- \_\_\_ Adverse events
- \_\_\_ Warranty
- \_\_\_ Travel or international use
- \_\_\_ **Index/Glossary**
- \_\_\_ **Date of Printing**
- \_\_\_ **User Assistance Information**

*Draft – Not for Implementation*

The medical device patient labeling adheres to the guidance recommended for:

- \_\_\_ Readability
- \_\_\_ Writing for increased comprehension (writeability)
- \_\_\_ Appearance of text
- \_\_\_ Appearance of graphics
- \_\_\_ Warnings and precautions
- \_\_\_ Pretesting

**References:**

**Sequence:**

Food and Drug Administration, Center for Devices and Radiological Health. Draft Report on Medical Device Labeling: Patients' and Lay Caregivers' Medical Device Information and Labeling Needs, Results of Qualitative Research. Rockville, MD; 1999.

**Risks and benefits:**

Covello, Vincent T.; McCallum, David B.; and Pavlova, Maria T. Effective Risk Communication. New York: Plenum Press; 1989. Chapter 1, Principles and Guidelines for Improving Risk Communication.

Health and Environmental Risk Communication - Advanced Training Workshop, November 13-15, 1996; Message Development by Erin Donovan.

Kasperson, Roger E. and Stallen, Pieter Jan M. Communicating Risks to the Public. The Netherlands: Kluwer Academic Publishers; 1991. Chapter 5, Risk comparisons and risk communication: Issues and problems in comparing health and environmental risks. by Vincent T. Covello.

National Cancer Institute, Office of Cancer Communications. How the Public Perceives, Processes, and Interprets Risk Information: Findings from Focus Group Research with the General Public. Bethesda, MD: June, 1998.

National Research Council, Improving Risk Communication. Washington, D. C.: National Academy Press: 1989.

Risk Communication Workshop given by the U.S. Environmental Protection Agency. September 21, 1993.

Tinker, T.; Pavlova, M.; Gotsch, A.; Arkin, EB. Communicating Risks in a Changing World. Maryland/Massachusetts: The Ramazzini Institute/OEM Press; 1998. Section IV, Conclusions and Recommendations.

U.S. Environmental Protection Agency. Communicating Environmental Risks - A Guide to Practical Evaluations 1990.

## *Draft – Not for Implementation*

### **Additional Information:**

Food and Drug Administration, Center for Devices and Radiological Health. Draft Report on Medical Device Labeling: Patients' and Lay Caregivers' Medical Device Information and Labeling Needs, Results of Qualitative Research. Rockville, MD; 1999.

### **User Assistance Information:**

Food and Drug Administration, Center for Devices and Radiological Health. Draft Report on Medical Device Labeling: Patients' and Lay Caregivers' Medical Device Information and Labeling Needs, Results of Qualitative Research. Rockville, MD; 1999.

### **Readability:**

Bormuth, JR. The cloze readability procedure. *Elementary English* 1968;45:429-436.

Boyd, MD, Citro, K. Cardiac patient education literature: can patients read what we give them? *Health Educator* 1983 July;3(7):513-516.

Chall, Jeanne S. Readability and Prose Comprehension: Continuities and Discontinuities. J. Flood editor, *Understanding Reading Comprehension*. Newark, DE: International Reading Association; 1984.

Chall, Jeanne S.; Dale, Edgar. *Readability Revisited: The New Dale-Chall Readability Formula*. Cambridge, Massachusetts: Brookline Books; 1995.

Coe, L. Readability of printed educational materials. *Journal of Clinical Nursing* 1996 5(6):359-366.

Dixon E, Park R. Do patients understand written health information? *Nursing Outlook* 1990 Nov/Dec;38(6):278-281.

Doak, Cecilia C.; Doak, Leonard G.; Root, Jane R. *Teaching patients with low literacy skills*. 2<sup>nd</sup> ed. Philadelphia: J.B. Lippincott; 1996.

Kintsch, Walter; Miller, James R. *Readability: A View from Cognitive Psychology*. In *Teaching: Research Reviews*. Newark, DE: International Reading Association 1981.

Klare, George R. *The Measurement of Readability*. Ames, Iowa: Iowa State University Press; 1996.

## *Draft – Not for Implementation*

Klare, George R. Readability. New York: Longman; 1984 (p. 681-744) In P.D. Pearson (Ed.) Handbook of reading research.

McLaughlin, HG. SMOG grading - a new readability formula. Journal of Reading 1969 May;12:639-646.

National Institutes of Health, National Cancer Institute. Clear and Simple: Developing Effective Print Materials for Low-Literate Readers. Rockville, MD: NIH Publication No. 95-3594; 1994.

Nelson GD, Nelson B. Are your patient education materials readable? Health Educator 1985 Nov/Dec;3(6):10-11.

Redish JC, Selzer J. The place of readability formulas in technical communication. Technical Communication 1985;32(4):46-52.

Sansgiry SS, Cady PS, Patil S. Readability of over-the-counter medication labels. Journal of the American Pharmaceutical Association 1997 Sept/Oct;NS37(5):522-528.

Selzer, J. Readability is a four-letter word. Journal of Business Communication, 1981, 18(4), 23-34.

Taylor, WL. "Cloze Procedure": a new tool for measuring readability. Journalism Quarterly 1953;30:415-433.

Weiss BD, Coyne, C. Communicating with patients who cannot read. The New England Journal of Medicine 1997 July;337(4):272-273.

### **Writing for increased comprehension (writeability):**

Bernier, MJ. Developing and evaluating printed education materials: a prescriptive model for quality. Orthopaedic Nursing 1993; 12(6):39-46.

Chall, Jeanne S.; Dale, Edgar. Readability Revisited: The New Dale-Chall Readability Formula. Cambridge, Massachusetts: Brookline Books; 1995.

Farrell-Miller, P, Gentry, P. How effective are your patient education materials? Guidelines for developing and evaluating written educational materials. Diabetes Educator 1989;15(5):418-422.

Food and Drug Administration, Center for Devices and Radiological Health. Write It Right. Rockville, MD: HHS Publication FDA 93-4258; 1993.

## *Draft – Not for Implementation*

Food and Drug Administration, Center for Devices and Radiological Health. Draft Report on Medical Device Labeling: Patients' and Lay Caregivers' Medical Device Information and Labeling Needs, Results of Qualitative Research. Rockville, MD; 1999.

Securities and Exchange Commission, Office of Investor Education and Assistance. A Plain English Handbook: How to create clear SEC disclosure documents. Washington, DC; August, 1998.

### **Appearance of text:**

Allensworth, DD, Luther, CR. Evaluating printed materials. Nurse Educator March/April 1986; 11(2):18-22.

Bernier, MJ. Developing and evaluating printed education materials: a prescriptive model for quality. Orthopaedic Nursing 1993; 12(6):39-46.

Coey, L. Readability of printed educational materials. Journal of Clinical Nursing 1996 5(6):359-366.

Farrell-Miller, P, Gentry, P. How effective are your patient education materials? Guidelines for developing and evaluating written educational materials. Diabetes Educator 1989;15(5):418-422.

Food and Drug Administration, Center for Devices and Radiological Health. Write It Right. Rockville, MD: HHS Publication FDA 93-4258; 1993.

Sidebottom, Charles B International Labeling Requirements For Medical Devices, Medical Equipment, and Diagnostic Products. Buffalo Grove, IL: Interpharm Press, Inc. 1995.

White, Jan V. Graphic Design for the Electronic Age. New York: A Xerox Press Book Watson- Guphill Publications; 1988.

### **Appearance of graphics:**

Bernier, MJ. Developing and evaluating printed education materials: a prescriptive model for quality. Orthopaedic Nursing 1993; 12(6):39-46.

Coey, L. Readability of printed educational materials. Journal of Clinical Nursing 1996 5(6):359-366.

## *Draft – Not for Implementation*

Daniels, P., Gatson, N. Guidelines: Writing For Adults With Limited Reading Skills. U.S. Department of Agriculture, Food and Nutrition Service 1988.

Farrell-Miller, P, Gentry, P. How effective are your patient education materials? Guidelines for developing and evaluating written educational materials. *Diabetes Educator* 1989;15(5):418-422.

Food and Drug Administration, Center for Devices and Radiological Health. Write It Right. Rockville, MD: HHS Publication FDA 93-4258; 1993.

### **Warnings and precautions:**

American National Standard for Product Safety Signs and Labels. ANSI Z535.4-1991. Washington, DC: National Electrical Manufacturers Association, 1991.

Ayres TJ, Gross MM, Wood CT, Horst DP, Beyer RR, Robensen JN. What is a warning and when will it be read? In Laughery Kenneth R; Wogalter Michael S; Young Stephen L., editors. *Human Factors Perspectives on Warnings*. [place unknown] Human Factors and Ergonomics Society; 1994; p. 476-480.

Barnett RL. The principle of uniform safety. *Safety Brief Triodyne, Inc.* 1994 Apr;10(3).

Braun CC, Silver NC, Stock BR. Likelihood of reading warnings: the effect of fonts and font size. *Proceedings of the Human Factors and Ergonomic Society Annual Meeting.* 1992:926-30.

CDRH General Program Memorandum #G91-1. Device Labeling Guidance. Mar 1991.

Desaulniers DR. Layout, organization and effectiveness of consumer product warnings. In Laughery Kenneth R; Wogalter Michael S; Young Stephen L., editors. *Human Factors Perspectives on Warnings*. [place unknown] Human Factors and Ergonomics Society; 1994: p. 26-30.

Discenza R, Ferguson JM. The instrumental role of product information: a study of warning labels for non-prescription drugs. *Health Marketing Quarterly* 1992;10(1/2): 155-168.

Dunne KJ, Patterson KD, Baroumand CR. Drafting warnings for medical products: practical considerations for minimizing litigation. *Healthspan* 1993 Sep;10(8):11-20.

Edworthy Judy, Adams Austin. *Warning design: a research perspective*. London: Taylor & Francis, Ltd; 1996. 219 p.

*Draft – Not for Implementation*

Lehto MR, Papostavrou JD. Models of the warning process: important implications toward effectiveness. *Safety Science* 1993;16:569-95.

Leonard SD, Matthews d, Karnes EW. How does the population interpret warning signals? In Laughery Kenneth R; Wogalter Michael S; Young Stephen L., editors. *Human Factors Perspectives on Warnings*. [place unknown]Human Factors and Ergonomics Society; 1994; p. 140.

Magurno AB, Wogalter MS. Behavioral compliance with warnings: effect of stress and placement. *Proceedings of the Human Factors and Ergonomic Society Annual Meeting*. 1994:826-30.

Miller JA. *Labeling research-the state of the art*. Cambridge, MA: Marketing Science Institute; 1978. Report #78-115.

Ross K. Evaluating the adequacy of warnings and instructions: to test or not to test, that is the question. 1993(?) Manuscript of presentation by member of Bowman and Brooke, Attorneys at Law, Minneapolis, MN.

Silver NC, Wogalter MS. Broadening the range of signal words. In Laughery Kenneth R; Wogalter Michael S; Young Stephen L., editors. *Human Factors Perspectives on Warnings*. [place unknown]Human Factors and Ergonomics Society; 1994; p. 186-90.

Taber MD. Developing litigation-resistant warnings. *Product Safety and Liability Reporter* 1992 Oct;20(39):1081-1085.

Wogalter MS. Factors influencing the effectiveness of warnings. *Proceedings of Public Graphics*. 1994 Sep; Lunteren, The Netherlands.

Wogalter MS, Barlow T, Murphy SA. Compliance to owners manual warnings: the influence of familiarity and the placement of a supplemental directive. Draft manuscript submitted to *Ergonomics* 1994 July.

Wogalter MS, Brelsford JW, Desaulniers DR, Laughery KR. Consumer product warnings: the role of hazard perception. *Journal of Safety Research* 1991;22:71-82.

Wogalter MS, Jarrard SW, Simpson SN. Influence of warning label signal words on perceived hazard level. *Human Factors* 1994;36(3):547-556.

Wogalter MS, Kalsher MJ, Racicat BM. Behavioral compliance with warnings: effect of voice, context, location. *Safety Science* 1993;16:637-654.

## *Draft – Not for Implementation*

Wogalter MS, Kalsher MJ, Rashid R. Effect of warning signal word and source on perceived credibility and compliance likelihood. Proceedings of 13<sup>th</sup> Triennial Congress of the International Ergonomics Association 1997;3:478-480.

Wogalter MS, Silver NC. Warning signal words: connoted strength and understandability by children, elders, and non-native English speakers. Ergonomics 1995;38(11):2188-2206.

Wright P, Creighton P, Thulfall SM. Some factors determining when instructions will be read. Ergonomics 1982;25(3):225-237.

Zeitlin LR. Failure to follow safety instructions: faulty communication or risky decisions? Human Factors 1994;36(1):172-81.

### **Pretesting:**

Department of Health and Human Services, Public Health Service, National Institutes of Health. Making health communication programs work. NIH Publication No. 89-1493; April 1989 Stage, Developing material and Pre-testing.

### **Usability testing:**

Dumas, Joseph S., Redish, Janice C. A practical guide to usability testing. New Jersey: Ablex Publishing Corporation; 1993. 412 p.

O'Brien, Thomas G., Charlton, Samuel G. Handbook of human factors testing and evaluation. New Jersey: Lawrence Erlbaum Associates; 1996. 359 p.

Salvendy, Gavriel, editor. Handbook of human factors and ergonomics. New York: John Wiley and Sons, Inc. 1997. 2137 p.