

# Blood Grouping Reagent

## Anti-Jk<sup>a</sup> (JK1)

Seraclone<sup>®</sup> Human Monoclonal  
(MS15)

## Anti-Jk<sup>b</sup> (JK2)

Seraclone<sup>®</sup> Human Monoclonal  
(MS8)

FOR IN-VITRO DIAGNOSTIC USE

For Tube Testing

MEETS FDA POTENCY REQUIREMENTS

U.S. License Number: 1798

### Package size

**REF** 808179100 **VOL** 2 mL Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> (JK1)

**REF** 808184100 **VOL** 2 mL Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> (JK2)

### Intended Use

For the determination of the Kidd antigens Jk<sup>a</sup> (JK1) and Jk<sup>b</sup> (JK2) of red blood cells using the tube test.

### Summary

The Kidd antigen was first identified in 1951 when the corresponding antibody was found to cause hemolytic disease of the fetus and newborn (HDFN). Although Kidd antibodies have been shown to cause generally mild HDFN, they have been implicated in severe hemolytic transfusion reactions (HTR). The HTR are often delayed due to an anamnestic response to the Kidd antigen<sup>1</sup>.

The frequencies of the common phenotypes are shown in the table.

Phenotypes and Frequencies in the Kidd System <sup>1</sup>		
Phenotype	Whites	Blacks
Jk (a+b-)	28	57
Jk (a+b+)	49	34
Jk (a-b+)	23	9
Jk (a-b-)	Exceedingly rare	

Biotest Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> and Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> Blood Group Reagents are used to test for the presence or absence of the Jk<sup>a</sup> and Jk<sup>b</sup> antigens. Biotest Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> and Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> are used principally in the resolution of antibody problems or in family studies.

### Principle of the Test

The test principle is hemagglutination. The antibodies in Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> (JK1) and Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> (JK2) bind to the corresponding antigen on red blood cells and cause an antigen-antibody reaction visible as red blood cell agglutination.

### Reagent

As the reactive components Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> (JK1) and Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> (JK2) contain human monoclonal antibodies of the immunoglobulin class IgM. They are derived from cell culture supernatant and demonstrate the consistent specificity and reproducibility characteristic for monoclonal antibodies.

Antibodies are diluted in a isotonic saline solution containing bovine albumin.

The following antibodies are produced using intermediate products produced for Biotest Medical Diagnostics GmbH in a shared manufacturing agreement with Millipore (UK) Ltd., 9 Fleming Road, Kirkton Campus, EH547BN, Livingston, UK; License Number 1721.

Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> (JK1)      clone MS15 (IgM)  
Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> (JK2)      clone MS8 (IgM)

Preservative: 0.1% sodium azide.

### Precautions

- For In-vitro diagnostic use.
- Store at 2 to 8°C.
- Do not use beyond the expiration date.
- Do not use if turbid.
- Handle and dispose of reagents as potentially infectious
- Caution: Do not pipette by mouth. The absence of all viruses has not been determined.
- Caution: This product contains Natural Rubber Latex Which May Cause Allergic Reactions.
- Warning: Contains sodium azide (NaN<sub>3</sub>), which may react with lead or copper plumbing to form explosive azides. If discarded in the sink, flush with large amounts of water to prevent the build-up of explosive metal azides.
- The bovine albumin used for the production of this reagent is purchased from BSE-free US sources, Boval Company L.P. in Cleburne, Tx, USA and Millipore in Kankakee, IL, USA.

### Specimen Collection

Fresh samples of clotted, EDTA or citrate anticoagulated whole blood collected following general blood sampling guidelines are acceptable. The specimen should be tested as soon as possible after collection. If testing is delayed, EDTA and clotted specimens should be stored at 2 to 8°C, citrated specimens (donor segments) at 1 to 6°C.

Blood specimens exhibiting gross hemolysis or contamination should not be used.

Clotted samples or those collected in EDTA may be tested within ten days from collection. Donor blood stored in citrate anticoagulant may be tested until the expiration date of the donor unit.

### Materials

#### Materials provided

- Seraclone<sup>®</sup> Anti-Jk<sup>a</sup> (JK1) and/or Seraclone<sup>®</sup> Anti-Jk<sup>b</sup> (JK2)

#### Materials required but not provided

- Pipettes (drop volume 40 to 50 µl)
- Isotonic saline solution
- Glass tubes 10 x 75mm or 12 x 75 mm
- Serological Centrifuge
- Interval Timer
- Markers
- Optical aid (optional). The use of an optical aid for agglutination reading must be validated by the user.

### Test Procedure

#### Tube test

1. Prepare a 3 to 5% suspension of red blood cells to be tested in isotonic saline.
2. Place one drop reagent into an appropriately labeled tube.
3. Add one drop of red blood cell suspension into the tube and mix.
4. Incubate at room temperature (20 to 24°C) for 15 minutes.
5. Centrifuge 60 seconds at 800 - 1000 x g.
6. Gently dislodge red blood cell button and observe for agglutination.
7. Record results

#### Stability of the Reaction

Following centrifugation, all tube tests should be read immediately and results interpreted without delay. Time delays may cause a dissociation of the antigen-antibody complexes resulting to false negative or more often weak positive reactions.

#### Quality Control

The reactivity of all blood typing reagents should be confirmed by testing with known positive and negative red blood cells on each day of use.

To confirm the reactivity or specificity of Biotest Monoclonal Anti-Jk<sup>a</sup>, Anti-Jk<sup>b</sup> Blood Grouping Reagents, each should be tested with antigen-positive (preferably from heterozygous individuals) and antigen-negative red blood cells, respectively. Each reagent is satisfactory for use if it reacts only with antigen-positive red blood cells.

It is recommended that a positive and a negative control be performed in parallel with testing.

#### Interpretation of results

Agglutination of the red blood cells is a positive result and indicates the presence of the corresponding antigen. No agglutination is a negative result and indicates the absence of the corresponding antigen. An agglutination viewer may facilitate the reading of tube tests (as recommended by the AABB Technical Manual, 15th edition). Frequencies in the population are listed in the "Summary" section.

**Limitations**

- Samples with a positive direct antiglobulin test, cold agglutinins, or rouleaux formation may show false positive results in testing with monoclonal antibodies. Results on these samples must be interpreted with caution. False positive results or reaction suspected to be due to cold agglutinins should be resolved according to in-house procedures. It is recommended that an appropriate control be tested in parallel.
- Some conditions that may cause false positive results are:
  - Contamination of sample or reagents
  - Autoantibodies
  - Improper storage or preparation of red blood cells
  - Antibodies to antibiotics or other reagents i
  - Cold Antibodies

**Specific Performance Characteristics**

Testing is performed in accordance with FDA recommended methods. The final release testing is performed according to the product specific SOPs. Each lot of Biotest Blood Group Reagent is tested in the Quality control by package insert method against a panel of antigen positive red blood cells (heterozygous antigen expression and if possible weakened antigen expression) to insure suitable reactivity. The products meet FDA potency requirements. The specificity testing for the presence of contaminating antibodies is performed according to the product specific SOPs.

For the product performance it is necessary to adhere to the recommended method in the instructions for use.

The performance of the Biotest Anti-Jk<sup>a</sup> and Anti-Jk<sup>b</sup> was confirmed against FDA approved reference reagents in a Multi Center Field Trial.

For Technical Support or further product information, contact Biotest Diagnostics Corporation at 800-522-0090.

**Note**

Each facility should verify the optimum spin time for the specific centrifuge in use.

Manual techniques are to be performed according to the manufacturer's instructions. Each deviation from these instructions is the sole responsibility of the user.

Used tests must be discarded as hazardous material. Manage waste according to local, state and national regulations.

**Glossary of Symbols**

Symbol	Definition	Symbol	Definition
	Batch Code		<i>In vitro</i> diagnostic medical device
	Caution, consult accompanying documents		Consult instructions for use.
	Manufacturer		Use by YYYY-MM-DD
	Contains sufficient quantity for <n> tests.		Catalog number
	Temperature limitation		Volume

**Bibliography**

1. Mark E. Brecher, MD et al. Technical Manual 15th Edition, Bethesda, MA: AABB, 2005.