

Page Two

Scott R. Schultz, M.D.
October 2, 2001

having taught or completed at least 15 Category I medical education units in 36 months. Credits documented = zero.

Individuals failing to meet either the "Initial" and/or "Continuing" MQSA requirements must immediately cease performing mammography independently. For physicians the "Continuing" requirements include either the lack of appropriate CME/24 months, or Number of Interpretations/36 months. Requirements for re-qualification are listed in the Final Regulation that became effective on April 28, 1999.

The specific problems noted above appeared on your MQSA Facility Inspection Report which was issued to your facility following the close of the inspection.

Because these conditions may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, they represent a serious violation of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction, charging your facility for the cost of on-site monitoring, assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, the Standards, suspension or revocation of your facility's FDA certificate, or obtaining a court injunction against further mammography.

It is necessary for you to act on this matter immediately. Please explain to this office in writing within 15 working days from the date you received this letter:

- the specific steps you have taken to correct all of the violations noted in this letter;
- each step your facility is taking to prevent the recurrence of similar violations;
- equipment settings (including technique factors), raw test data, and calculated final results, where appropriate; and
- sample records that demonstrate proper record keeping procedures if the findings relate to quality control or other records.

Please submit your response to Thomas W. Garvin, Radiological Health Specialist, Food and Drug Administration, 2675 No. Mayfair Road, Suite 200, Milwaukee, WI, 53226-1305.

Finally, you should understand that there are many FDA requirements pertaining to mammography. This letter pertains only to findings of your inspection and does not necessarily address other obligations you have under the law. You may obtain general information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug

Page Three

Scott R. Schultz
October 2, 2001

Administration, P.O. Box 6057, Columbia, MD 21045-6057 (1-800-838-7715) or through the Internet at <http://www.fda.gov/cdrh/mammography/index/html>.

If you have specific questions about mammography facility requirements or about the content of this letter please feel free to phone Mr. Garvin at (414) 771-7167 ext. 12.

Sincerely,


James A. Rahto
Director
Minneapolis District

TGP/ccl


xc:


Lead Interpreting Radiologist
Minneapolis Radiology Associates, Ltd.
3366 Oakdale Ave. No.
Minneapolis, MN 55422

Sue McClanahan
Supervisor, Radiation Unit
Minnesota Department of Health
1645 Energy Park Drive, Suite 300
St. Paul, MN 55108-2970

Priscilla F. Butler
Director, Breast Imaging Accreditation Programs
American College of Radiology
1891 Preston White Drive
Reston, VA 20191