



**DEPARTMENT OF HEALTH AND HUMAN SERVICES**

Food and Drug Administration  
Nashville District Office

*m24/6n*

297 Plus Park Blvd.  
Nashville, TN 37217

March 9, 1999

**CERTIFIED MAIL - RETURN RECEIPT REQUESTED**

**WARNING LETTER-99-NSV-7**

*Purged  
3/10/99  
JEP*

**FACILITY ID# 169847**

Mildred E. Howell, M.D.  
Chief Executive Officer  
Childersburg Medical - Surgical Center  
P.O. Drawer 369  
Childersburg, AL 35044

Dear Dr. Howell:

Your facility was inspected on February 24, 1999 by a representative of the State of Alabama, on contract to the Food and Drug Administration. This inspection revealed that your facility failed to comply with the Quality Standards for Mammography (Standards) as specified in Title 21, Code of Federal Regulations (CFR), Part 900.12, as follows:

**Level 1**

1. The interpreting physician did not meet the requirement of being certified by an FDA recognized board or having the alternative of 2 months training in the interpretation of mammograms: **[REDACTED]**
2. The interpreting physician did not meet the requirement of being certified by an FDA recognized board or having the alternative of 2 months training in the interpretation of mammograms: **[REDACTED]**
3. The interpreting physician did not meet the requirement of being certified by an FDA recognized board or having the alternative of 2 months training in the interpretation of mammograms: **[REDACTED]**
4. The interpreting physician did not meet the requirement of being certified by an FDA recognized board or having the alternative of 2 months training in the interpretation of mammograms: **[REDACTED]**

Level 2

1. The interpreting physician did not meet the requirement of having a minimum of 40 CME credit hours of initial training in mammography: [REDACTED]
2. The interpreting physician did not meet the requirement of having a minimum of 40 CME credit hours of initial training in mammography: [REDACTED]
3. The interpreting physician did not meet the requirement of having a minimum of 40 CME credit hours of initial training in mammography: [REDACTED]
4. The interpreting physician did not meet the requirement of having a minimum of 40 CME credit hours of initial training in mammography: [REDACTED]

These specific deficiencies appear on the Post Inspection Report which was sent to your facility on March 1, 1999. These deficiencies are symptomatic of serious underlying problems that could compromise the quality of mammography at your facility.

It is your responsibility to ensure adherence to each requirement of the Mammography Quality Standards Act of 1992 (MQSA) and FDA's regulations. You are responsible for investigating and determining the cause of these deficiencies that the inspection identifies and to promptly initiate permanent corrective actions.

If you fail to properly address these deficiencies, FDA may, without further notice, initiate regulatory action. Under MQSA, FDA may:

- impose civil money penalties on a facility of up to \$10,000 for each failure to substantially comply with, or each day of failure to comply with the Standards.
- suspend or revoke a facility's FDA certificate for failure to comply with the Standards.
- seek an injunction in federal court to prohibit any mammography activity that constitutes a serious risk to human health.

Within 15 working days after receiving this letter, you should notify FDA in writing of each step your facility is taking to prevent the recurrence of similar violations.

If your facility is unable to complete the corrective actions within 15 working days, you should state the reason for the delay and the time within which the correction will be completed.

Your reply should be directed to Joseph E. Hayes, Compliance Officer, U.S. Food and Drug Administration, 297 Plus Park Boulevard, Nashville, Tennessee 37217, telephone 615/781-5389, extension 125. If you should have any questions in regard to this letter or about how to ensure you are meeting MQSA standards, please call Karen Smallwood, Radiation Specialist, at 615/781-5380, extension 144.

Sincerely,



Raymond K. Hedblad  
Director, Nashville District

RKH/ks

cc: State of Alabama