



DEPARTMENT OF HEALTH & HUMAN SERVICES  
Public Health Service  
Food and Drug Administration  
SOUTHWEST REGION

m 24781

Office of the Regional  
Food and Drug Director  
7920 Elmbrook Drive, Suite 102  
Dallas, TX 75247-4982  
TELEPHONE: 214-655-8100  
FACSIMILE: 214-655-8130

March 24, 1999

**WARNING LETTER**

**CERTIFIED MAIL  
RETURN RECEIPT REQUESTED**

99-SWR-WL-0777  
CFN# 1650537  
Facility ID# 213009

Angela Crenshaw  
President  
Millennium Diagnostic Imaging  
12840 Hillcrest Road  
Dallas, TX 75230

Dear Ms. Crenshaw:

Your facility was inspected on March 9, 1999 by a representative of the State of Texas, acting on behalf of the Food and Drug Administration. This inspection revealed that your facility failed to comply with certain parts of the Quality Standards for Mammography (Standards) as specified in Title 21, *Code of Federal Regulations (CFR)*, Part 900.12, as follows:

21 CFR900.12(a)(1)(ii)(A)(B): The interpreting physician did not meet the requirement of being certified by a FDA-recognized board or having the alternative of 2 months training in the interpretation of mammograms: [REDACTED]

The specific deficiency noted above appeared under the level 1 heading on your MQSA Facility Inspection Report, which was issued at the close of the inspection. This deficiency may be symptomatic of serious underlying problems that could compromise the quality mammography at your facility.

In addition, level 2 noncompliances were listed on the inspection report provided to you at the close of the inspection. These noncompliances are:

21 CFR900.12(a)(1)(ii)(C): The interpreting physicians did not meet the requirement of having a minimum of 40 CME credit hours of initial training in mammography:  
Marc Berger, [REDACTED]

21 CFR900.12(a)(1)(iii)(A): The interpreting physicians did not meet the requirement of having initial experience in mammography (read or interpreted 240 patient examinations in a 6 month period): [REDACTED]

21 CFR900.12(a)(1)(iv)(A): The interpreting physician did not meet the continuing experience requirement of having read or interpreted 960 patient examinations in a 24 month period: [REDACTED]

**It is your responsibility to ensure adherence to each requirement of the Mammography Quality Standards Act of 1992 (MQSA) and FDA's regulations. You are responsible for investigating and determining the causes of the deficiencies that the inspection identifies and promptly initiating permanent corrective actions.**

If you fail to promptly correct these deficiencies, FDA may, without further notice, initiate regulatory action. Under MQSA, FDA may:

- ▶ **impose civil money penalties** on a facility of up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, the Standards.
- ▶ **suspend or revoke a facility's FDA certificate** for failure to comply with the Standards.
- ▶ **seek an injunction in federal court** to prohibit any mammography activity that constitutes a serious risk to human health.

Please note that FDA regulations do not preclude a State from enforcing its own State mammography laws and regulations. In some cases, these requirements may be more stringent than FDA's. When you plan your corrective actions, therefore, you should consider the more stringent State requirements, if any.

Within 15 working days after receiving this letter, you should notify FDA in writing of:

- ▶ the specific steps you have taken to correct the violations noted in this letter;
- ▶ each step your facility is taking to prevent the recurrence of similar violations;

If your facility is unable to complete the corrective action within 15 working days, you should state the reason for the delay and the time within which the corrections will be completed.

Please send the original copy of your response to Deborah M. McGee, Radiation Specialist, Food and Drug Administration, 7920 Elmbrook Drive, Suite 102, Dallas, Texas 75247-4982. Also, send a copy to the State radiation control office that conducted the inspection referenced in this letter. You may choose to address both FDA and State requirements in your response.

Page 3  
March 24, 1999

If you have any questions regarding this letter or how to ensure you are meeting MQSA standards, please call Ms. McGee at 214-655-8100, extension 138.

Sincerely yours,



Edward R. Esparza  
Regional Food and Drug Director

cc: Thomas C. Cardwell  
X-Ray Branch Administrator  
Bureau of Radiation Control  
Texas Department of Health  
1100 West 49th Street  
Austin, TX 78756-3189

Director, Government Relations  
American College of Radiology  
1891 Preston White Drive  
Reston, Virginia 22091