



DEPARTMENT OF HEALTH & HUMAN SERVICES

New York District

M3302 n

Food & Drug Administration
850 Third Avenue
Brooklyn, NY 11232

WARNING LETTER

January 4, 2000

REF: NYK-2000-23

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

Azad K. Anand, M.D.
Lake Success Imaging
444 Lakeville Road
Lake Success, New York 11042

Facility ID: 119917

Dear Dr. Anand:

Your facility was inspected on November 30, 1999 by a representative of the Nassau County Dept. of Health, acting in behalf of the Food and Drug Administration. This inspection revealed a serious regulatory problem involving the mammography at your facility. Under a United States Federal law, the Mammography Quality Standards Act of 1992, your facility must meet specific requirements for mammography. These requirements help protect the health of women by assuring that a facility can perform quality mammography. The inspection revealed the following Level 1 finding at your facility:

Processor QC records were missing 2 out of 4 days of operation, or 50% of the time, in June 1999, for the [REDACTED] processor.

The specific problem noted above appeared on your MQSA Facility Inspection Report which was issued to your facility at the close of the inspection. This problem is identified as Level 1 because it identifies a failure to meet significant MQSA requirements.

Because this condition may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, it represents a violation of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction, charging your facility for the cost of on-site monitoring, assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, MQSA standards, suspension or revocation of your facility's FDA certificate, or obtaining a court injunction against further mammography.

Lake Success Imaging
Warning Letter NYK-2000-23

In addition, your response should address the Level 2 findings that were listed on the inspection report provided at the close of the inspection. The Level 2 findings are:

- 1. Processor QC records were missing 2 consecutive days (related to above); and**
- 2. The radiologic technologist, [REDACTED], did not meet the continuing education requirement of having completed a minimum of 15 CEUs in a 36 month period (0 CEU in 36 months).**

With respect to processor QC records, the inspector also observed other occasions when patients were x-rayed but processor QC records were missing. Examples are November 1998 (27% of operating days had no QC recorded) and January 1999 (30% of operating days had no QC recorded).

It is necessary for you to act on these matters immediately. Please explain to this office in writing within fifteen (15) working days from the date you received this letter:

- The specific steps you have taken to correct the violation noted in this letter; and
- Each step your facility is taking to prevent the recurrence of similar violations.


Please submit your response to the attention of Lillian C. Aveta, Compliance Officer, U.S. Food and Drug Administration, 850 Third Avenue, Brooklyn, New York 11232, Tel. (718) 340-7000, ext. 5142.

Finally, you should understand there are many FDA requirements pertaining to mammography. This letter pertains only to findings of our inspection and does not necessarily address other obligations you have under the law. You may obtain general information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug Administration, P.O. Box 6057, Columbia, Maryland 21045-6057 (1-800-838-7715), or through the Internet at <http://www.fda.gov>.

Lake Success Imaging
Warning Letter NYK-2000-23

If you have any questions about mammography facility requirements in general, please feel free to contact Murray L. Kurzman, Radiation Programs Manager, at (516) 921-2035.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Brenda J. Holman". The signature is fluid and cursive, with a large initial "B" and "H".

Brenda J. Holman
District Director