



DEPARTMENT OF HEALTH & HUMAN SERVICES

PURGED
F.O.I
MATERIAL

Public Health Service
Food and Drug Administration

1990 MacArthur Blvd., Ste 300
Irvine, California 92612-2445
Telephone (949) 798-7600

WARNING LETTER

Certified Mail
Return Receipt Requested

November 2, 2000

Mr. Xavier C. Hernandez
Owner
X - Diagnostic Imaging
1635 3rd Avenue, unit #F
Chula Vista, CA 91911

W/L: #03-01
Inspection ID: 2202360002
CFN: 20-31,864
FEI: 3002819164

Dear Mr. Hernandez:

We are writing to you because on September 21, 2000, your facility was inspected by a representative of the State of California, acting in behalf of the Food and Drug Administration (FDA). This inspection revealed a serious regulatory problem involving the mammography at your facility.

Under a United States Federal law, the Mammography Quality Standards Act of 1992, your facility must meet specific requirements for mammography. These requirements help protect the health of women by assuring that a facility can perform quality mammography. The inspection revealed the following Level 1 findings at your facility:

- Level 1: Phantom QC records were missing for 4 weeks for unit #1, [REDACTED], model [REDACTED], in the Mammo Room
- Level 1: Processor QC records were missing 4 out of 9 days of operation (44%) in the month December 1999 for the processor dark room having the [REDACTED] equipment.

The specific problems noted above appeared on your MQSA Facility Inspection Report, which was issued to your facility at the close of the inspection. These problems are identified as Level 1, because they identify a failure to meet a significant MQSA requirement.

Page Two of Three
November 2, 2000

re: X - Diagnostic Imaging
W/L #03-01

Because these conditions may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, they represent a serious violation of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction, charging your facility for the cost of on-site monitoring, assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with, MQSA Standards, suspension or revocation of your facility's FDA certificate, or obtaining a court injunction against further mammography.

It is necessary for you to act on this matter immediately. Please explain to this office in writing within fifteen (15) working days from the date you received this letter:

- the specific steps you have taken to correct all of the violations noted in this letter;
- each step your facility is taking to prevent the recurrence of similar violations; and
- please provide sample records that demonstrate proper record keeping procedures, if the findings relate to quality control or other records (Note: Patient names or identification should be deleted from any copies submitted).


Please submit your response to:

Mr. Thomas L. Sawyer
Director, Compliance Branch
U.S. Food & Drug Administration
19900 MacArthur Blvd., suite #300
Irvine, CA 92612-2445
Phone: (949) 798-7600

Finally, you should understand that there are many FDA requirements pertaining to mammography. This letter pertains only to findings of your inspection and does not necessarily address other obligations you have under the law. You may obtain general information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug Administration, P.O. Box 6057, Columbia, MD 21045-6057 (1-800-838-7715) or through the Internet at <http://www.fda.gov>.

If you have more specific questions about mammography facility requirements, or about the content of this letter, please feel free to contact [REDACTED] (MQSA Auditor) at telephone number [REDACTED].

Sincerely,


Alonza E. Cruse
District Director

**PURGED
F.O.I
MATERIAL**

Page Three of Three
November 2, 2000

re: X - Diagnostic Imaging
W/L #03-01

cc:

State of California
Dept. of Health Services
Radiological Health Unit
550 South Vermont Avenue; suite #601
Los Angeles, CA 90020

Ms Priscilla F. Butler
Director, Breast Imaging Accreditation Programs
Standards & Accreditation Dept.
American College of Radiology
1891 Preston White Drive
Reston, VA 20191