

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 862

[Docket No. 2004N-0482]

Medical Devices; Clinical Chemistry and Clinical Toxicology Devices; Classification of Newborn Screening Test Systems for Amino Acids, Free Carnitine, and Acylcarnitines Using Tandem Mass Spectrometry

AGENCY: Food and Drug Administration, HHS.

ACTION: Final rule.

SUMMARY: The Food and Drug Administration (FDA) is classifying newborn screening test systems for amino acids, free carnitine, and acylcarnitines using tandem mass spectrometry into class II (special controls). The special control that will apply to the device is the guidance document entitled “Class II Special Controls Guidance Document: Newborn Screening Test Systems for Amino Acids, Free Carnitine, and Acylcarnitines Using Tandem Mass Spectrometry.” The agency is taking this action in response to a petition submitted under the Federal Food, Drug, and Cosmetic Act (the act) as amended by the Medical Device Amendments of 1976, the Safe Medical Devices Act of 1990, the Food and Drug Administration Modernization Act of 1997, and the Medical Device User Fee and Modernization Act of 2002. The agency is classifying the device into class II (special controls) in order to provide a reasonable assurance of safety and effectiveness of the device. Elsewhere in this issue of the **Federal Register**, FDA is publishing a notice

of availability of a guidance document that is the special control for this device.

DATES: This rule becomes effective [*insert date 30 days after date of publication in the Federal Register*]. The classification was effective August 24, 2004.

FOR FURTHER INFORMATION CONTACT: Carol Benson, Center for Devices and Radiological Health (HFZ-440), Food and Drug Administration, 2098 Gaither Rd., Rockville, MD 20850, 301-594-1243, ext. 144.

SUPPLEMENTARY INFORMATION:

I. Background

In accordance with section 513(f)(1) of the act (21 U.S.C. 360c(f)(1)), devices that were not in commercial distribution before May 28, 1976, the date of enactment of the Medical Device Amendments of 1976 (the amendments), generally referred to as postamendments devices, are classified automatically by statute into class III without any FDA rulemaking process. These devices remain in class III and require premarket approval, unless and until the device is classified or reclassified into class I or II or FDA issues an order finding the device to be substantially equivalent, in accordance with section 513(i) of the act, to a predicate device that does not require premarket approval. The agency determines whether new devices are substantially equivalent to previously marketed devices by means of premarket notification procedures in section 510(k) of the act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807) of FDA's regulations.

Section 513(f)(2) of the act provides that any person who submits a premarket notification under section 510(k) of the act for a device that has not previously been classified may, within 30 days after receiving an order classifying the device in class III under section 513(f)(1) of the act, request

FDA to classify the device under the criteria set forth in section 513(a)(1) of the act. FDA shall, within 60 days of receiving such a request, classify the device by written order. This classification shall be the initial classification of the device. Within 30 days after the issuance of an order classifying the device, FDA must publish a notice in the **Federal Register** announcing such classification (section 513(f)(2) of the act).

In accordance with section 513(f)(1) of the act, FDA issued a notice on June 9, 2004, classifying the Perkin Elmer Life and Analytical Sciences' NeoGram Amino Acids and Acylcarnitines Tandem Mass Spectrometry Kit in class III because it was not substantially equivalent to a device that was introduced or delivered for introduction into interstate commerce for commercial distribution before May 28, 1976, or a device which was subsequently reclassified into class I or class II. On July 2, 2004, Perkin Elmer Life and Analytical Sciences submitted a petition requesting classification of the NeoGram Amino Acids and Acylcarnitines Tandem Mass Spectrometry Kit under section 513(f)(2) of the act. The manufacturer recommended that the device be classified into class II.

In accordance with section 513(f)(2) of the act, FDA reviewed the petition in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the act. Devices are to be classified into class II if general controls, by themselves, are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls to provide reasonable assurance of the safety and effectiveness of the device for its intended use. After review of the information submitted in the petition, FDA determined that the Perkin Elmer Life and Analytical Sciences' NeoGram Amino Acids and Acylcarnitines Tandem Mass Spectrometry Kit can

be classified in class II with the establishment of special controls. FDA believes these special controls, in addition to general controls, will provide reasonable assurance of safety and effectiveness of the device.

The device is assigned the generic name “newborn screening test system for amino acids, free carnitine, and acylcarnitines using tandem mass spectrometry.” It is identified as a device intended for the measurement and evaluation of amino acids, free carnitine, and acylcarnitine concentrations from newborn whole blood filter paper samples. The quantitative analysis of amino acids, free carnitine, and acylcarnitines and their relationship with each other provides analyte concentration profiles that may aid in screening newborns for one or more inborn errors of amino acid, free carnitine, and acyl-carnitine metabolism.

FDA has identified no direct risks to health related to use of newborn screening test systems for amino acids, free carnitine, and acylcarnitines using tandem mass spectrometry. However, failure of the test to perform as indicated or error in interpretation of results may lead to the indirect risk of improper medical management of patients with inborn errors of metabolism. For example, a false negative (false normal) measurement could contribute to failure to detect a possible inborn error of metabolism, which could lead to functional impairment or death. A false positive (false abnormal) measurement could contribute to unnecessary additional patient testing and added concern and apprehension of parents and physicians.

The class II special controls guidance document provides information on how to meet premarket (510(k)) submission requirements for the device, including recommendations on validation of performance characteristics and labeling. FDA believes that following the class II special controls guidance

document generally addresses the risk to health identified in the previous paragraph. Therefore, on August 24, 2004, FDA issued an order to the petitioner classifying the device into class II. FDA is codifying this classification by adding 21 CFR 862.3840.

Following the effective date of this final classification rule, any firm submitting a 510(k) premarket notification for a newborn screening test system for amino acids, free carnitine, and acylcarnitines using tandem mass spectrometry will need to address the issues covered in the special controls guidance. However, the firm need only show that its device meets the recommendations of the guidance or in some other way provides equivalent assurance of safety and effectiveness.

Section 510(m) of the act provides that FDA may exempt a class II device from the premarket notification requirements under section 510(k) of the act, if FDA determines that premarket notification is not necessary to provide reasonable assurance of the safety and effectiveness of the device. For this type of device, however, FDA has determined that premarket notification is necessary because FDA review of performance characteristics, test methodology, and labeling to satisfy requirements of § 807.87(e), will provide reasonable assurance that acceptable levels of performance for both safety and effectiveness will be addressed before marketing clearance. Thus, persons who intend to market this type of device must submit to FDA a premarket notification containing information on the newborn test system before marketing the device.

II. Environmental Impact

The agency has determined under 21 CFR 25.34(b) that this action is of a type that does not individually or cumulatively have a significant effect on

the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

III. Analysis of Impacts

FDA has examined the impacts of the final rule under Executive Order 12866, the Regulatory Flexibility Act (5 U.S.C. 601–612), and the Unfunded Mandates Reform Act of 1995 (Public Law 104–4). Executive Order 12866 directs agencies to assess all costs and benefits of available regulatory alternatives and, when regulation is necessary, to select regulatory approaches that maximize net benefits (including potential economic, environmental, public health and safety, and other advantages; distributive impacts; and equity). The agency believes that this final rule is not a significant regulatory action under the Executive order.

The Regulatory Flexibility Act requires agencies to analyze regulatory options that would minimize any significant impact of a rule on small entities. Because classification of these devices into class II will relieve manufacturers of the device of the cost of complying with the premarket approval requirements of section 515 of the act (21 U.S.C. 360e), and may permit small potential competitors to enter the marketplace by lowering their costs, the agency certifies that the final rule will not have a significant impact on a substantial number of small entities.

Section 202(a) of the Unfunded Mandates Reform Act of 1995 requires that agencies prepare a written statement, which includes an assessment of anticipated costs and benefits, before proposing “any rule that includes any Federal mandate that may result in the expenditure by State, local, and tribal governments, in the aggregate, or by the private sector, of \$100,000,000 or more (adjusted annually for inflation) in any one year.” The current threshold after

adjustment for inflation is \$110 million. FDA does not expect this final rule to result in any 1-year expenditure that would meet or exceed this amount.

IV. Federalism

FDA has analyzed this final rule in accordance with the principles set forth in Executive Order 13132. FDA has determined that the rule does not contain policies that have substantial direct effects on the States, on the relationship between the National Government and the States, or on the distribution of power and responsibilities among the various levels of government. Accordingly, the agency has concluded that the rule does not contain policies that have federalism implications as defined in the Executive order and, consequently, a federalism summary impact statement is not required.

V. Paperwork Reduction Act of 1995

This final rule contains no collections of information. Therefore, clearance by the Office of Management and Budget under the Paperwork Reduction Act of 1995 is not required.

VI. Reference

The following reference has been placed on display in the Division of Dockets Management (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852, and may be seen by interested persons between 9 a.m. and 4 p.m., Monday through Friday.

1. Petition from Perkin Elmer Life and Analytical Sciences, dated July 2, 2004.

List of Subjects in 21 CFR Part 862

Medical devices.

■ Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 862 is amended as follows:

PART 862—CLINICAL CHEMISTRY AND CLINICAL TOXICOLOGY DEVICES

■ 1. The authority citation for 21 CFR part 862 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 371.

■ 2. Section 862.1055 is added to subpart B to read as follows:

§ 862.1055 Newborn screening test system for amino acids, free carnitine, and acylcarnitines using tandem mass spectrometry.

(a) *Identification.* A newborn screening test system for amino acids, free carnitine, and acylcarnitines using tandem mass spectrometry is a device that consists of stable isotope internal standards, control materials, extraction solutions, flow solvents, instrumentation, software packages, and other reagents and materials. The device is intended for the measurement and evaluation of amino acids, free carnitine, and acylcarnitine concentrations from newborn whole blood filter paper samples. The quantitative analysis of amino acids, free carnitine, and acylcarnitines and their relationship with each other provides analyte concentration profiles that may aid in screening newborns for one or more inborn errors of amino acid, free carnitine, and acyl-carnitine metabolism.

(b) *Classification.* Class II (special controls). The special control is FDA's guidance document entitled "Class II Special Controls Guidance Document: Newborn Screening Test Systems for Amino Acids, Free Carnitine, and Acylcarnitines Using Tandem Mass Spectrometry." See § 862.1(d) for the availability of this guidance document.

Dated: November 15, 2004.

Linda S. Kahan,

Deputy Director, Center for Devices and Radiological Health.

[FR Doc. 04-????? Filed ??-??-04; 8:45 am]

BILLING CODE 4160-01-S