Salmonella Plate Incorporation Mutagenicity Assay (Ames Test)
Using Test Article Extracts

CONCLUSION

All criteria for a valid study were met as described in the protocol. The results of the *Salmonella* Plate Incorporation Mutagenicity Assay (Ames Test) Using Test Article Extracts indicate that under the conditions of this study, extractions of **BIOCELL® Saline-filled Mammary Implant, Style 168** did not cause a positive response with any of the tester strains in the presence and absence of metabolic activation.

REFERENCES

Ames, B.N., J. McCann and E. Yamasaki (1975) Methods for Detecting Carcinogens and Mutagens with the *Salmonella*/Mammalian Microsome Mutagenicity Test, Mutation Research, 31:347-364.

Maron, D.M. and B.N. Ames (1983) Revised Methods for the *Salmonella* Mutagenicity Test, Mutation Research, 113:173-215.

McCann, J. and B.N. Ames (1976) Detection of Carcinogens as Mutagens in the *Salmonella*/Microsome Test: Assay of 300 Chemicals: Discussion, Proc. Natl. Acad. Sci. USA, 73:950-954.

McCann, J., E. Choi, E. Yamasaki and B.N. Ames (1975) Detection of Carcinogens as Mutagens in the *Salmonella*/Microsome Test: Assay of 300 Chemicals, Proc. Natl. Acad. Sci. USA, 72:5135-5139.

Vogel, H.J. and D.M. Bonner (1956) Acetylornithinase of *E. coli*: Partial Purification and Some Properties, J. Biol. Chem., 218:97-106.