

**Report on Assessment via ESCA and SIMS of the presence or Absence
of Silica at the Surface of Saline-Fill RTV Mammaries**

V. CONCLUSIONS:

Based on the results of this study, it is concluded that silica is not present at the surface of either smooth, BIOCELL® or abraded surfaces of saline-filled RTV mammaries.