



Regulatory & Clinical Research Institute, Inc.  
5353 Wayzata Boulevard, Suite 505  
Minneapolis, MN 55416-1334

tel: 952.746.8080  
fax: 952.884.6518  
www.rcri-inc.com

July 18, 2006

Food and Drug Administration  
Center for Devices and Radiological Health  
Regulations Staff (HFZ-215)  
1350 Piccard Drive  
Rockville, MD 20857

**Re: Amendment to 513(e) Petition For Reclassification  
Tissue Adhesive For Soft Tissue Approximation  
Docket Number: 2006P-0071/CCP1**

To Whom It May Concern,

This submission is provided to amend 513(e) Petition for Reclassification for Tissue Adhesives for Soft Tissue Approximation; Docket Number 2006P-0071/CCP1. In accordance with 21 CFR 860.123(b)(4), an original and two copies of this Amendment are enclosed.

Section 2.2 Intended Use of the Petition shall be amended to read:

"Topical cyanoacrylate tissue adhesives are intended for topical closure of surgical incisions, including laparoscopic incisions, and simple traumatic lacerations that have easily approximated skin edges. Topical cyanoacrylate tissue adhesives may be used in conjunction with, but not in place of, deep dermal stitches."

Please feel free to contact me at the numbers below should you have any comments or questions concerning this Petition for Reclassification.

Sincerely,

REGULATORY & CLINICAL RESEARCH INSTITUTE, INC.

Tierney Norsted, Ph.D., M.P.H.  
Executive Vice President & Principal Advisor  
Tel: 952-746-8080, ext 222  
Fax: 952-884-6518  
Email: [tnorsted@rcri-inc.com](mailto:tnorsted@rcri-inc.com)