

Largo Medical Center

201 14th Street SW
Largo, Florida 33770
Mail P.O. Box 2905
Largo, Florida 33779-2905
Phone (727) 588-5200
www.largomedical.com

2799 '00 MAR -6 P1:35

March 2, 2000

Docket No. 99D-4910, Dockets management Branch
Division of Management Systems & Policy
Office of Human Resources & Management Services
Food and Drug Administration
5630 Fishers Lane, Room 1061 (HFA-305)
Rockville, MD. 20852

RE: CFR 900.12(b)(8)(I)

To Whom It May Concern:

I would like to speak against the requirement for the fine adjustment compression controls operable from both sides of the patient.


Many mammography machines that otherwise comply with FDA regulations would have to be removed from service if this requirement goes into effect. I believe that this is an onerous requirement for two reasons:

First, I do not believe that the fine adjustment control provides any significant improvement in the quality of the mammographic image. After reading about this proposed requirement, I discussed it with several technologists (who are experienced with the GE DMR machine). As experienced mammographers, they all agree the amount of change in compression is minute as well as the minimal difference in images obtained. I do not believe that this specific requirement will help to diagnose any more breast cancers.

Second, the cost of this requirement is prohibitive. There is not some simple piece of add on equipment that can make otherwise good machines compliant. Entire units would have to be replaced at a cost of approximately \$80,000 each. I truly believe that this cost is not justified for such questionable benefit. This requirement could drive some privately owned breast-imaging centers out of business, decreasing access to high quality mammograms. It seems to me that this result is contrary to the goal of the MQSA.

Please consider my comments in drafting the final regulation. I would like to be personally contacted with your decision.

Most Sincerely,




A. Martin Landry III, M.D.
Medical Director

99D-4910

27

Largo Medical Center

P.O. Box 2905
Largo, Florida 33779-2905



** 05-05-00 FOR PRESORT BLUE TAPER #1 336

Docket No. 99D-4910,
Dockets management Branch
Division of Management Systems & Policy
Office of HR & Management Services
Food and Drug Administration
5630 Fishers Lane, Room 1061 (HFA-305)
Rockville, MD. 20852

WUNP 20857

