

# MAXXIM MEDICAL

2322 '99 OCT 28 A9:16

October 26, 1999

Dockets Management Branch (HFA-305)  
Food and Drug Administration  
5630 Fishers Lane, Rm. 1061  
Rockville, MD 20852

Subject: Comments: [Docket No. 98N-0313]

Dear Agency Member:

I wish to provide comments on the Food and Drug Administration's draft guidance entitled Guidance for Medical Gloves: A Workshop Manual. These comments are provided on behalf of Maxxim Medical, Inc., Clearwater, FL, a major medical glove manufacturer of both surgical and examination gloves.

Please refer to the attachment for applicable comments. Two copies are provided as requested.

Sincerely,



Joyce T. Moody  
Vice President,  
Regulatory Affairs/Quality Assurance

JTM:rnr

99D-2335

C15

CASE MANAGEMENT DIVISION

477 COMMERCE BLVD. • OLDSMAR, FL 34677 • (800) 727-7951 • FAX: (813) 855-2083

Docket No. 98N-0313  
Comments

Expiration Dating

FDA has historically rejected expiration dating based upon accelerated aging studies. Should the requirement for expiration dating on all medical gloves (sterile and non-sterile) be imposed, FDA must allow manufacturers to use data collected from accelerated aging studies to substantiate the claim. Failure to do so would force glove manufacturers to delay the introduction of new products to the market and possibly affect the level of care available to consumers. The final rule, when published, must include the agency's recommended accelerated aging stability protocol or allow manufacturers the option to utilize their own validated method(s).

Labeling for Medical Gloves

- The proposed caution statements are lengthy. FDA should be sensitive to the problems created by requirements for the translation of the caution statements into as many as fourteen different languages, all appearing on each device label and product packaging. Provisions must be made to allow alternate methods for conveyance of information other than the principle display panel, which will achieve the ultimate goal to alert the user concerning NRL, powder and protein content of the medical gloves.
- There is no threshold amount of extractable protein that can be identified as the level that will not cause an allergic response in some individuals. The same is true for residual powder content levels. For this reason, the statement "FDA recommends that this product contain no more than 120 mg of powder and no more than 1200 µg extractable protein per glove" should be deleted from the proposed caution statements for powdered gloves. The similar statement in the proposed caution statements for powder-free gloves should also be deleted. FDA can accomplish the very same goal by publishing this information in the Medical Glove Guidance Manual. The publishing of recommended levels may erroneously imply that the user of the products so labeled is assured that the risk of an allergic reaction due to powder or protein content would be minimal. FDA should only require that manufacturers state the powder and protein content on primary packaging with no recommended limits stated.
- The requirement to state the protein content per glove instead of per gram of glove will cause concern by medical glove users because it will appear that the levels have suddenly increased. Glove weights per size have never been regulated by FDA; however, stating the extractable protein limit per glove automatically mandates the implementation of additional process controls for this variable throughout the production run. Re-education of the medical community, multiple packaging changes and additional process controls for this requirement are costly and add no value to the product and no additional benefit to the user.

Glove in Kits

The proposed rule requires convenience kit manufacturers to comply with the medical glove labeling requirements. This requirement places needless restrictions on the kit manufacturer in facilitating product substitutions where there is no stipulated user preference. To comply with the rule, the kit manufacturer would have to make labeling changes based upon the protein and powder content levels of each and every medical glove supplier for the kit. It is also feasible to have more than one type of glove within one kit, each with different levels of protein and/or powder. Additionally, kit manufacturers routinely purchase non-sterile latex examination gloves in bulk for placement in kits. The proposed guidance would eliminate this option and even place additional requirements on kit manufacturers to individually package latex exam gloves prior to placement in kits. This seems impractical since other latex-containing devices may be contained within the kits (catheters, tubing, etc.) that are not enclosed in the manner stipulated for latex gloves.

Docket No. 98N-0313  
Comments

Proposed Effective Date

A two-year effective date should be allowed in order for industry to make necessary changes to processes, technology and labeling. The new requirements should not apply to the sale and distribution of products that were manufactured prior to the effective date of the guidance. Manufacturers must be given time to exhaust inventory of such stock without extreme negative impact on the company's ability to handle the cost of changes brought about by the new guidance requirements.

Reclassification of Medical Gloves

The rationale for reclassification of medical gloves into Class II is sound. We believe this to be a prudent action given the intended use of medical gloves in infection and contamination control.

Restrictions on Powder Glove Sales/Distribution/Use

The decision to ban the sale and use of powdered glove products does not appear to be warranted, based upon current facts and statistics.

- Although some reports suggest that powder may aid in the exposure of glove users to NRL proteins, similar concerns have not been established for powdered gloves made of synthetic materials. Sufficient studies must be conducted to compare and correlate user or patient reactions attributed to powdered synthetic gloves vs. powder free gloves of the same material.
- Sufficient time should be allowed to assess the impact of the proposed powder content limits on glove safety and efficacy. Perhaps the requirements to lower powder content on medical gloves will produce the desired level of patient/user safety without necessitating a complete ban on powdered products.
- Powder-free technology for medical gloves is fairly new compared to that of powdered glove processes. While manufacturers conduct all the required, and often additional, biocompatibility studies to ensure user safety, the technology and various powder-free coatings and formulations cannot purport to a long history of documented user safety. Given current knowledge of both the low incidence of powder-related user problems, and the short user history for powder-free technology, FDA's actions of issuing limits, for lower powder content on powdered gloves seems the most appropriate action to take at this time, to reduce adverse health effects from allergic reactions and foreign body reactions.

Limits: Recommended or Required?

FDA should state powder content and protein content limits as "required", rather than "recommended". The required limits should be stated in the Medical Glove guidance, not on the product labeling. This action would eliminate the proposed requirement to state the specific protein and powder contents on the primary labeling and reduce some of the costs associated with implementation of the guidance.



Shipments May Be Subject To Duties And Taxes At Destination

400-7187 9964

PACKAGE TRACKING NUMBER



Date of Shipment 10/29/97	Origin Station I.D. (FedEx use only)	Dest. Station I.D. (FedEx use only)	URSA Routing (FedEx use only)
------------------------------	---	--	----------------------------------

1 顧客番号 Sender's Federal Express Account Number 1581-2504-8	電話番号 (重要) Phone Number (Very Important) 0782653057
--	--

3 荷受人名 (活字体で記入) TC (Recipient's Name) Please Print: 会社/Company: FOOD AND DRUG ADMINISTRATION	電話番号 (確実な配達のために重要) Phone Number (Very Important To Ensure Delivery):
---	---

部署および階 /Dept./Floor No.: THE DOCKETS MANAGEMENT BRANCH (HFA-305)
--

荷送人名 (活字体で記入) /FROM (Your Name) Please Print: KEN MASAOKA
--

住所 /Exact Street Address: 5630 FISHERS LANE, RM. 1061
--

州名 /State/Province: MD
---------------------------

会社名 /Company: SUMITOMO RUBBER INDUSTRIES LTD	郵便番号 /ZIP/Postal Code: 20852
---	---------------------------------

住所 /Exact Street Address: ROCKVILLE,
---

市および区 /City: MD
--------------------

住所 /Exact Street Address: 4F/SANGYOHIN KIKAKU BU
---

市および区 /City: CHUD-KU KOBE
------------------------------

国名 /Country: U.S.A.
------------------------

住所 /Exact Street Address: 3-5-9 WAKINOHAMA-CHO
---

郵便番号 /ZIP/Postal Code: 6510072
-----------------------------------

国名 /Country: JP
--------------------

市および区 /City: CHUD-KU KOBE
------------------------------

国名 /Country: U.S.A.
------------------------

郵便番号 /ZIP/Postal Code: 20852
---------------------------------

国名 /Country: JP	郵便番号 /ZIP/Postal Code: 6510072
--------------------	-----------------------------------

B 通関業者名 /BROKER NAME International Priority Broker Selectionをご希望の場合のみご記入ください。	通関業者の電話番号 /Broker Phone Number (確実な配達のため非常に重要です。)
--	--

指定の通関ポイント (国および市名) /City/Country: 通関業者の郵便番号 /Broker ZIP/Postal Code:
---

2 発送人の参考欄 (オプション) (最初の24文字が請求書に記載されます。) Sender's Internal Billing Reference Information (Optional) (First 24 Characters Will Appear On Invoice)
--

4 サービス /SERVICES 当てはまる項目を1つだけチェックしてください Check Only One Box	1 <input checked="" type="checkbox"/> FedEx International Priority	2 <input type="checkbox"/> FedEx Intl. Economy	3 <input type="checkbox"/> FedEx International Priority Freight
--	--	--	---

9 貨物明細 /SHIPMENT INFORMATION FedExは関税額の見積は行いません			
パッケージ数 (荷送人種法及び個数確認 - SLAC) No. of packages (Shippers Load and Count - SLAC)	総重量 /Total Weight: 該当する欄に印を付け、重量を記入	貨物の製造国名 Country of Manufacture	申告価値の通貨 Specify Currency of Declared Value:
		JAPAN	USD

5 パッケージ /PACKAGING 当てはまる項目を1つだけチェックしてください /Must check one box.	1 <input checked="" type="checkbox"/> FEDEX Letter/Envelope	2 <input type="checkbox"/> FEDEX Pak	3 <input type="checkbox"/> そのほかのパッケージ /Other Packaging
---	---	--------------------------------------	--

内容及び税番を正確にご記入ください。 Complete Description of Contents/Harmonized Code: DOCUMENTS
--

6 配達 /取り扱い方法 DELIVERY/HANDLING INSTRUCTIONS	1 <input type="checkbox"/> 営業所および業務センター止め /Hold for Pick-up	2 <input checked="" type="checkbox"/> 平日配達 /Deliver Weekday	3 <input type="checkbox"/> 土曜配達 /Deliver Saturday
--	---	---	---

税関用の受取人識別番号 (I.N., V.A.T., E.I.N.等もしくはその他の現地で必要とされる番号) Recipient's ID Number for Customs Purposes (e.g., I.N./V.A.T./E.I.N., or as locally required):
--

7 送料支払方法 /TRANSPORTATION CHARGES 当てはまる項目を1つだけチェックしてください /Must check one box.	1 <input checked="" type="checkbox"/> 元払い /Bill Sender	2 <input type="checkbox"/> 着払い /Bill Recipient	3 <input type="checkbox"/> 第三者払い /Bill 3rd Party
---	--	--	--

10 荷送人の署名 /SENDER'S SIGNATURE 貨物を当社に引き渡すことにより、荷送人は、譲渡不能の航空貨物運送の責任を負担するものとします。契約条件に規定するように、この貨物の運送はワルソー条約の他の国際条約の適用があります。これらの条約上の貨物の損失、損害、到着の遅延に関する当社の責任が限定されることがあります。By giving us your shipment, you agree to the conditions on the back of this Non-Negotiable Air Waybill. Certain international treaties, including the Warsaw Convention, may apply to this shipment and limit our liability for damage, loss or delay as described in the Conditions of Contract. 貨物の配達の際には荷受人の署名を要するものとします。(This is not authorization to deliver this shipment without a recipient signature.)
---

8 税金の支払方法 /DUTIES AND TAXES 当てはまる項目を1つだけチェックしてください /Must check one box.	1 <input type="checkbox"/> 元払い /Bill Sender	2 <input type="checkbox"/> 着払い /Bill Recipient	3 <input type="checkbox"/> 第三者払い /Bill 3rd Party
--	---	--	--

荷送人の署名 SENDER'S SIGNATURE Required X 野田 孝子	重量 Weight	体積 Volume
---	--------------	--------------

11 送料支払方法 /TRANSPORTATION CHARGES 当てはまる項目を1つだけチェックしてください /Must check one box.	4 <input type="checkbox"/> クレジットカード払い /Credit Card	5 <input type="checkbox"/> 現金 /小切手払い /Cash/Check
--	--	--

Handling Units	CI Attached	SED Attached	CO Attached
Received At	Base Charges	Dec. Val. Chrg.	Other
Regular Stop	FedEx Emp. #	Audit Emp. #	Date

有効期限 /Exp. Date	承認コード /Approval Code
-----------------	----------------------

Recipient Name Printed	Del. Courier Emp. #	Date M D Y	Time
------------------------	---------------------	------------	------

RECIPIENT'S SIGNATURE X
-------------------------

300
-----

EXTRA/CUSTOMS COPY

RECEIVED ABOVE SHIPMENT IN GOOD ORDER AND CONDITION. WE AGREE TO PAY ALL CHARGES INCLUDING CUSTOMS DUTIES AND TAXES AS APPLICABLE AND TO THE CONDITIONS OF CONTRACT AS STATED ON THE REVERSE SIDE OF THE CONSIGNEE COPY.

RECEIVED BY THE ADDRESSEE AT THE ABOVE ADDRESS. WE AGREE TO PAY ALL CHARGES INCLUDING CUSTOMS DUTIES AND TAXES AS APPLICABLE AND TO THE CONDITIONS OF CONTRACT AS STATED ON THE REVERSE SIDE OF THE CONSIGNEE COPY.

REC-08