

**Section N**



ETYMOTIC RESEARCH, INC.

513(f) Reclassification Petition – TV-TIP Sound Amplifier

## SECTION N - APPLICABLE STANDARDS

The device is factory tested prior to shipment with a standard hearing aid analyzer per American National Standards Institute, *Specification of Hearing Aid Characteristics*, (ANSI S3.22-1996), American National Standards Institute, New York, 1996.