OCT - 5 2000

K002804

510(k) Summary
For N Protein Standard SL

1. **Manufacture's Name, Address, Telephone, and Contact Person, Date of Preparation:**

Manufacturer:	Dade Behring Marburg GmbH Emil-von-Behring Str. 76 Marburg/Germany
Contact Information:	Dade Behring Inc. Glasgow Site P.O. Box 6101 Newark, Delaware 19714 Attn: Rebecca S. Ayash Tel: 302-631-6276
Preparation date:	September 6, 2000

2. **Device Name/ Classification:**

N Protein Standard SL:	Calibrator, Multi-Analyte
Classification Number:	Class II (862.1150)

3. **Identification of the Legally Marketed Device:**

N Protein Standard SL (K991181)

4. **Device Description:**

N Protein Standard SL is a liquid standard prepared from human serum with stabilizers and preservative. It is intended to establish reference curves for the quantitative determination of human serum proteins by immunonephelometry with the Behring Nephelometer Systems (particle-enhanced nephelometry).

5. **Device Intended Use:**

Establishment of reference curves for the determination of IgG, IgG_{1-4}, IgA, IgM, C3c, C4, Transferrin, Albumin, α_1-antitrypsin, α_2-macroglobulin, Haptoglobin, α_1-acid glycoprotein, Prealbumin, Hemopexin, Ceruloplasmin, RbP, Ig/L-chain lambda & kappa, β_2-microglobulin, Soluble transferrin receptor (sTfR), IgE, and Ferritin by immunonephelometry with the Behring Nephelometer Systems.

6. **Medical device to which equivalence is claimed and comparison information:**

The N Protein Standard SL (modified to include hemopexin) is substantially equivalent in intended use to the Dade Behring N Protein Standard SL (K991181) currently marketed. The N Protein Standard SL (modified), like the current N Protein Standard SL is intended to be used for the calibration of human serum protein assays on the Behring Nephelometer Systems.

K002804

7. Device Performance Characteristics:

Stability:
Stability was evaluated according to Dade Behring protocols and the standard was found to be stable for at least 24 months at +2° to +8° C, as originally packaged and for at least 14 days at +2° to +8° C, once opened.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

OCT - 5 2000

Ms. Rebecca S. Ayash
Manager, Regulatory Affairs, Biology
Dade Behring Inc.
Glasgow Site
PO Box 6101
Newark, Delaware 19714

Re: K002804
Trade Name: N Protein Standard SL
Regulatory Class: II
Product Code: JIX
Dated: September 6, 2000
Received: September 8, 2000

Dear Ms. Ayash:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4588. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "http://www.fda.gov/cdrh/dsma/dsmamain.html".

Sincerely yours,

Steven Gutman

Steven I. Gutman, M.D., M.B.A.
Director
Division of Clinical Laboratory Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use Statement

K 002804

Device Name: N Protein Standard SL

Indications for Use:

Establishment of reference curves for the determination of IgG, IgG_{1-4}, IgA, IgM, C3c, C4, Transferrin, Albumin, α_1-antitrypsin, α_2-macroglobulin, Haptoglobin, α_1-acid glycoprotein, Prealbumin, Hemopexin, Ceruloplasmin, RbP, Ig/L-chain lambda & kappa, β_2-microglobulin, Soluble transferrin receptor (sTfR), IgE, and Ferritin by immunonephelometry with the Behring Nephelometer Systems.

Sean Cooper
(Division Sign-Off)
Division of Clinical Laboratory Devices
510(k) Number K002804

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓
(Per 21 CFR 801.109)

Over-The-Counter-Use _____
(Optional Format 1-2-96)