

Radiography and Fluoroscopy X-ray Systems

Yuan Fang, PhD, PEng

Division of Radiological Health

Office of In Vitro Diagnostics & Radiological Health

Center for Devices and Radiological Health

U.S. Food and Drug Administration

Yuan.Fang@fda.hhs.gov

Diagnostic X-ray: Product Overview

- **Radiography/Dental**
- Fluoroscopy/Interventional
- Hand-held radiographic units
- Third party components



Diagnostic X-ray: Product Overview


- Radiography/Dental
- **Fluoroscopy/Interventional**
- Hand-held radiographic units
- Third party components



Courtesy of Normaler Schluckakt and Anka Friedrich under the Creative Commons Attribution-Share Alike 3.0 Unported license.

Diagnostic X-ray: Product Overview

- Radiography/Dental
- Fluoroscopy/Interventional
- **Hand-held radiographic units**
- Third party components



Radiation Safety Concerns

- Tissue reactions
 - Acute radiation injuries from interventional procedures
- Stochastic
 - Patient dose
 - Operator dose



Shope TB. "Radiation-induced skin injuries from fluoroscopy," *RadioGraphics*. 16: 1195-1199, 1996.

Performance Standards

- 21 CFR 1020.30 Diagnostic x-ray system and their major components
- 21 CFR 1020.31 Radiographic equipment
- 21 CFR 1020.32 Fluoroscopic equipment
- 2005 Amendment
 - Mode of operation description and user instructions
 - Display of fluoroscopy time
 - Air Kerma (dose) meters
 - Schedule of maintenance

What are FDA's Current Concerns

- Inadequate quality control information for end user
- Availability of protocols and descriptions for imaging different-sized patients
- Insufficient shielding for hand-held x-ray systems
- Inadequate integration information for third party certified components
- Reporting of radiation dose

International Consensus Standards

- IEC 60601-2-54 Radiography and Radioscopy
- IEC 60601-2-43 Interventional X-ray Equipment
- IEC 60601-2-63 Extra-Oral Dental Equipment
- IEC 60601-2-65 Intra-Oral Dental Equipment

What would FDA like to do?

- Require certain features (already in existing or proposed IEC standards)
 - Manufacturer defined quality control procedures in the user manual
 - Requirement for Physics Mode
 - Easily removable anti-scatter grid
 - Size-specific presets (unless AEC present)
 - Radiation Dose Structured Report

What would FDA like to do?

- Require additional features (not in IEC standards)
 - Standardized quality control tests
 - User access control
 - Descriptions of parameters affecting dose/image
 - Skin dose mapping for interventional fluoroscopy equipment

What would FDA like to do?

- Consider a performance standard for hand-held x-ray radiography units
- Ensure hand-held devices include safety features
 - Shield the unit housing and identify necessary safety precautions
 - Provide shielding or a means to increase distance between the operator and unit

What would FDA like to do?

- Address integration issues with third-party certified components
 - Third-party certified components (x-ray detector with software to control the generator) may affect quantity, quality, or direction of radiation
 - Clarify how connections are made
 - Provide a list of compatible systems in labeling

Questions for Committee

- What is your opinion on the value of requiring manufacturers to include a QC phantom with radiographic and/or fluoroscopic x-ray systems free of charge similar to CT?

Questions for Committee

- What is the committee's opinion on including the proposed features above for radiography and fluoroscopic systems in a performance standard?

Questions for Committee

- Are there additional safety improvements that should be pursued for radiography and fluoroscopy?

Questions for Committee

- What information is necessary to ensure adequate integration of third-party certified components?
- Should third-party component integration issues be addressed in a performance standard?

Questions for Committee

- What is the committee's opinion on the importance of regulating hand-held x-ray systems through a specific performance standard?
- Does the committee have any additional concerns with the use of hand-held devices?

Questions for Committee

- Should FDA include requirements for Radiation Dose Structured Reports (RDSRs) for all imaging equipment that generates ionizing radiation (radiography, fluoroscopy, CT, and dental cone-beam CT (CBCT) in the performance standards?

