October 5, 2017



One Lambda, Inc. Attention: Ms. Angela Estany 21001 Kittridge Street Canoga Park, CA 91303

Re:BK170053Device Name:One Lambda LABType™ XR and CWD DNA Typing TestsRegulation Name:Qualitative Test for HLA, Non-diagnosticRegulatory Class:UnclassifiedProduct Code:MZIDated:October 3, 2017Received:October 3, 2017

Dear Ms. Estany:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to a legally marketed predicate device marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <u>http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm</u>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <u>http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm</u> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely,

Orieji Illoh, MD Director Division of Blood Components and Devices Office of Blood Research and Review Center for Biologics Evaluation and Research

Enclosure Indications For Use Page 3 – BK170053 – Ms. Estany

Indications for Use

510(k) Number: BK170053

Device Name: LABType[™] XR and CWD DNA Typing Test

Indications for Use:

For use to determine HLA A,B,C and DRB1 locus typing to aid in transfusion and transplantation donor recipient matching.

Prescription Use <u>X</u> (Part 21 CFR 801 Subpart D) Over-The-Counter Use ______ (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW LINE-CONTINUE ON ANOTHER PAGE IF NEEDED) Concurrence of CBER, Office of Blood Research and Review (OBRR)

AND/OR

Division Sign-Off, Office of Blood Research and Review