



December 13, 2017

Ortho-Clinical Diagnostics, Inc.
Attention: Ms. Alexandra Chamberlain, RAC
100 Indigo Creek Drive
Rochester, NY 14626-5101

Re: **BP160122/o**
Device: VITROS Immunodiagnostic Products HIV Combo Reagent Pack
VITROS Immunodiagnostic Products HIV Combo Calibrator
Filed: December 19, 2016
Amended: December 28, 2016; January 24, 2017; January 31, 2017;
February 9, 2017; March 9, 2017; March 27, 2017; July 26, 2017;
September 11, 2017; October 20, 2017; November 17, 2017;
November 20, 2017; November 30, 2017; December 4, 2017; December 5,
2017; December 12, 2017
Procode: MZF

Dear Ms. Chamberlain:

The Center for Biologics Evaluation and Research (CBER) of the Food and Drug Administration (FDA) has completed its evaluation of your premarket approval application (PMA) for the VITROS Immunodiagnostic Products HIV Combo Reagent Pack and the VITROS Immunodiagnostic Products HIV Combo Calibrator (VITROS HIV Combo test). The VITROS Immunodiagnostic Products HIV Combo Reagent Pack is for the simultaneous qualitative detection of antibodies to human immunodeficiency virus (HIV) types 1, including group M and O, and/or 2 (anti-HIV-1 and anti-HIV-2) and HIV p24 antigen in human serum and plasma (heparin and EDTA) in adults, pregnant women, adolescents and children (as young as 2 years of age), using the VITROS 3600 Immunodiagnostic System.

A reactive result with the VITROS HIV Combo test does not distinguish between the detection of HIV-1 p24 antigen, antibodies to HIV-1, and antibodies to HIV-2.

The results of the VITROS HIV Combo Test, in conjunction with other serological evidence and clinical information, may be used as an aid in the diagnosis of infection with HIV-1 and/or HIV-2. The VITROS HIV Combo Test is not intended for use in screening blood or plasma donors. The effectiveness of the VITROS HIV Combo assay for use in screening blood or plasma donors has not been established. However, this assay can be used as a blood donor screening assay in urgent situations where traditional licensed blood donor screening tests are unavailable or their use is impractical.

It is not intended for newborn screening or for use with cord blood specimens or specimens from individuals less than 2 years of age.

The VITROS HIV Combo Calibrator is for use in the calibration of the VITROS 3600 Immunodiagnostic System.

We are pleased to inform you that the PMA is approved. You may begin commercial distribution of the device upon receipt of this letter.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the Act). The device is further restricted under section 515(d)(1)(B)(ii) of the Act insofar as the labeling must specify the specific training or experience practitioners need in order to use the device. FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the Act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Expiration dating for this device has been established and approved at 26 weeks when stored at 2 °C to 8 °C. This is to advise you that the protocol you used to establish this expiration dating is considered an approved protocol for the purpose of extending the expiration dating as provided by 21 CFR 814.39(a)(7).

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as “Annual Report” (please use this title even if the specified interval is more frequent than one year) and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84. This is a reminder that as of September 24, 2014, class III devices are subject to certain provisions of the final UDI rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR 830 Subpart E). Additionally, 21 CFR 814.84 (b)(4) requires PMA annual reports submitted after September 24, 2014, to identify each device identifier currently in use for the subject device, and the device identifiers for devices that have been discontinued since the previous periodic report. It is not necessary to identify any device identifier discontinued prior to December 23, 2013. For more information on these requirements, please see the UDI website, <http://www.fda.gov/udi>.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. For more information, please refer to the FDA guidance document entitled, “Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process” (www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm.

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm.

CBER does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading. CBER will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found on the FDA website located at <https://www.fda.gov/BiologicsBloodVaccines/BloodBloodProducts/ApprovedProducts/default.htm>. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an

independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in triplicate, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing.

Food and Drug Administration
Center for Biologics Evaluation and Research
Document Control Center
10903 New Hampshire Ave.
WO71-G112
Silver Spring, MD 20993-0002

If you have questions concerning this approval order, please contact Dr. Vasantha Kumar at (240) 402-8413 or Vasantha.Kumar@fda.hhs.gov.

Sincerely,

Jay S. Epstein, MD
Director
Office of Blood Research and Review
Center for Biologics Evaluation and Research