



March 21, 2019

Bio-Rad Medical Diagnostics GmbH
Attention: Mr. David Bhend
Bio-Rad Laboratories, Inc.
6565 185th Avenue, NE
Redmond, WA 98052

Re: BK180275
Device Name: IH-Com V5.0
Regulation Number: 21 CFR 864.9175
Regulation Name: Automated Blood Grouping and Antibody Test System
Regulatory Class: II
Product Code: KSZ
Dated: March 8, 2019
Received: March 11, 2019

Dear Mr. Bhend:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for more information.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Orieji Illoh, MD
Director
Division of Blood Components and Devices
Office of Blood Research and Review
Center for Biologics Evaluation and Research

Enclosure
Indications For Use

Indications for Use

510(k) Number: BK180275

Device Name: IH-Com V5.0

Indications for Use:

IH-Com V5.0 is data management and result interpretation software used in immunohematology testing to analyze reaction images for blood samples that are tested on the IH-1000, IH-500 and IH-Reader 24 with IH-System reagents or the TANGO infinity[®] with reagents specifically formulated and manufactured by Bio-Rad Medical Diagnostics GmbH. It manages sample, patient, donor, QC and reagent data. IH-Com V5.0 may be used with:

- a standalone or multiple IH-1000 systems,
- a standalone or multiple IH-500 systems,
- a standalone or multiple IH-Reader 24 systems,
- a standalone TANGO infinity[®] or multiple TANGO infinity[®] systems;
or connected to a Laboratory Information System

IH-Com V5.0 should only be operated by trained personnel and is not intended for use in a direct patient environment. In the USA IH-Com is “Rx only”.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CBER, Office of Blood Research and Review

Division Sign-Off, OBRR