

1
2
3
4
5

Safer Technologies Program for Medical Devices

6
7
8

Draft Guidance for Industry and Food and Drug Administration Staff

9

DRAFT GUIDANCE

10
11
12

This draft guidance document is being distributed for comment purposes only.

13
14
15

Document issued on September 19, 2019.

16
17
18
19
20
21
22

You should submit comments and suggestions regarding this draft document within 60 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

23
24
25
26
27
28
29
30

For questions about this document regarding CDRH-regulated devices, contact OCEA: Office of Clinical Evidence and Analysis/DCEA1: Division of Clinical Science and Quality at 301-796-5550 or BreakthroughDevicesProgram@fda.hhs.gov. For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-8010.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Center for Biologics Evaluation and Research

37

38

Preface

39

40

Additional Copies

42

CDRH

44 Additional copies are available from the Internet. You may also send an e-mail request to
45 CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please include the document
46 number 19001 and complete title of the guidance in the request.

47

CBER

49 Additional copies are available from the Center for Biologics Evaluation and Research (CBER),
50 Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave.,
51 Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-
52 8010, by email, ocod@fda.hhs.gov or from the Internet at [https://www.fda.gov/vaccines-blood-
53 biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances](https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances).

54

55
56
57

58
59
60
61
62
63
64
65
66
67
68
69
70
71
72
73
74
75
76
77
78
79
80
81
82
83
84
85
86
87

Table of Contents

I. Introduction.....	4
II. Background.....	5
III. Program Principles.....	6
A. Interactive and Timely Communication.....	7
B. Review Team Support.....	7
C. Review of Regulatory Submissions	8
D. Benefit-Risk Assessments and Pre/Post-Market Balance of Data Collection.....	8
E. Efficient and Flexible Clinical Study Design.....	9
F. Manufacturing Considerations for PMA Submissions	9
IV. Factors for STeP Acceptance and Review Process.....	10
A. General Eligibility Factor.....	10
B. Specific Eligibility Factors for Inclusion in STeP	10
C. Considerations for Evaluating Specific STeP Eligibility Factors	10
(1) First Factor.....	10
(2) Second Factor	11
D. Additional Considerations for STeP Acceptance Review.....	13
(1) Regulatory Path	13
(2) Timeframe for STeP Acceptance	15
(3) Multiple Devices for the Same Expected Safety Benefit	15
E. Submitting a Request for Inclusion in STeP and FDA Review	15
F. Withdrawal and Disqualification from STeP After Program Acceptance.....	16
V. Mechanisms for Feedback on Development of Devices in STeP.....	16
A. Sprint Discussion.....	18
B. Data Development Plan (DDP).....	18
C. Other Pre-Submissions for STeP Devices.....	19
D. Regular Status Updates	19
Appendix 1: Illustrative Example: Recommended Contents of a Q-Submission Request for Inclusion in STeP.....	20

88
89 **Safer Technologies Program for**
90 **Medical Devices**

91
92
93 **Draft Guidance for Industry and**
94 **Food and Drug Administration Staff**
95

96 *This draft guidance, when finalized, will represent the current thinking of the Food and Drug*
97 *Administration (FDA or Agency) on this topic. It does not establish any rights for any person*
98 *and is not binding on FDA or the public. You can use an alternative approach if it satisfies*
99 *the requirements of the applicable statutes and regulations. To discuss an alternative*
100 *approach, contact the FDA staff or Office responsible for this guidance as listed on the title*
101 *page.*

102 **I. Introduction¹**

103 The FDA is introducing a new, voluntary program for certain medical devices and device-led
104 combination products² that are reasonably expected to significantly improve the safety of
105 currently available treatments or diagnostics that target an underlying disease or condition
106 associated with morbidities and mortalities less serious than those eligible for the Breakthrough
107 Devices Program; for example, this may include devices treating or diagnosing non-life-
108 threatening or reasonably reversible conditions. Devices and device-led combination products
109 are eligible for this program if they are subject to review under a premarket approval application
110 (PMA), De Novo classification request (“De Novo request”), or premarket notification (510(k)).
111 Consistent with the Agency’s statutory mission³ to protect and promote public health, FDA
112 believes that this “Safer Technologies Program” or “STeP” will help patients have more timely
113 access to these medical devices and device-led combination products by expediting their
114 development, assessment, and review, while preserving the statutory standards for premarket
115 approval, De Novo marketing authorization, and 510(k) clearance. FDA has modeled STeP on
116 the principles and features of FDA’s Breakthrough Devices Program as mandated in section
117 515B of the Federal Food, Drug and Cosmetic Act (FD&C Act) (21 U.S.C. 360e-3) and further

¹ The Office of Combination Products (OCP) and the Center for Drug Evaluation and Research (CDER) were consulted in the preparation of this guidance.

² A combination product is defined in 21 CFR 3.2(e). For purposes of this guidance, device-led combination products refer to combination products subject to review under a premarket approval application (PMA), premarket notification (510(k)), or De Novo classification request.

³ See section 1003(b) of the Federal Food, Drug, and Cosmetic Act (FD&C Act).

Contains Nonbinding Recommendations

Draft – Not for Implementation

118 described in the FDA guidance document entitled “[Breakthrough Devices Program](#)”⁴ (hereinafter
119 referred to as the “Breakthrough Devices Program guidance document”). As resources permit,
120 FDA intends for STeP to incorporate similar features offered under the Breakthrough Devices
121 Program, such as interactive and timely communications, early engagement on Data
122 Development Plans (DDP), prioritized review, and senior management engagement.

123
124 For the current edition of the FDA-recognized standard(s) referenced in this document, see the
125 [FDA Recognized Consensus Standards Database](#).⁵ For more information regarding use of
126 standards in regulatory submissions, please refer to the FDA guidance titled “[Appropriate Use of](#)
127 [Voluntary Consensus Standards in Premarket Submissions for Medical Devices](#)”⁶ and “[Standards](#)
128 [Development and the Use of Standards in Regulatory Submissions Reviewed in CBER](#).”⁷

129
130 FDA recognizes and anticipates that the Agency may need up to 60 days to perform activities to
131 operationalize this Safer Technologies Program following issuance of the final guidance. FDA
132 does not intend to accept requests for inclusion in STeP within this time period.

133
134 FDA’s guidance documents, including this draft guidance, do not establish legally enforceable
135 responsibilities. Instead, guidance documents describe the Agency’s current thinking on a topic
136 and should be viewed only as recommendations, unless specific regulatory or statutory
137 requirements are cited. The use of the word *should* in Agency guidance documents means that
138 something is suggested or recommended, but not required.

139

140 II. Background

141 FDA is responsible for protecting and promoting public health by ensuring the safety,
142 effectiveness, and security of medical products.⁸ Additionally, FDA is responsible for advancing
143 public health by helping to provide timely access to innovations that make medical products and
144 their use safer and more effective.⁹ In recent years, FDA has developed policies and
145 implemented new programs designed to promote patient access to innovative and safe new
146 therapies and diagnostics. An important example of this approach is the Breakthrough Devices
147 Program, which superseded the Expedited Access Pathway and Priority Review Program. The
148 Breakthrough Devices Program is intended to expedite the development and review of certain
149 devices that meet the designation criteria for the program and treat life-threatening or irreversibly
150 debilitating diseases or conditions.¹⁰

⁴ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/breakthrough-devices-program>.

⁵ <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>.

⁶ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices>.

⁷ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/standards-development-and-use-standards-regulatory-submissions-reviewed-center-biologics-evaluation>.

⁸ For more information about CDRH's vision for medical device safety, see “Medical Device Safety Action Plan: Protecting Patients, Promoting Public Health,” available at <https://www.fda.gov/about-fda/cdrh-reports/medical-device-safety-action-plan-protecting-patients-promoting-public-health>.

⁹ See FDA’s “Mission,” available at <https://www.fda.gov/about-fda/what-we-do>.

¹⁰ The designation criteria are defined in section 515B of the FD&C Act and described in the Breakthrough Devices Program guidance document.

Contains Nonbinding Recommendations

Draft – Not for Implementation

151
152 FDA recognizes that, although medical products provide great benefits to patients, they also
153 present risks. FDA strives to permit marketing only for products with a favorable benefit-risk
154 profile. However, patients may experience a wide range of adverse events attributed to use of the
155 medical product including those that are considered serious, resulting in death or serious injury.¹¹
156 These types of events may significantly impact patients and their quality of life. Safety and
157 innovation are both important priorities for the Agency, and improvements in each of these areas
158 are expected to result in increased quality of life and health benefits for patients, while providing
159 a reasonable assurance of both safety and effectiveness.

160
161 As a complement to the Breakthrough Devices Program, FDA believes that advancements in
162 medical devices that are ineligible for the Breakthrough Devices Program but offer a significant
163 safety advantage in treating and/or diagnosing less serious diseases or conditions can also
164 provide an important public health benefit. Therefore, FDA is developing STeP to help spur
165 safety innovation for medical devices and to provide patients timely access to devices that are
166 not eligible for the Breakthrough Devices Program and may offer significant improvements to
167 the safety profile of available medical treatments.¹² FDA believes that efforts to improve safety
168 are directly related to improving overall clinical benefits and may also help patients experience
169 fewer serious adverse events.

170

III. Program Principles

171
172 Similar to the Breakthrough Devices Program, STeP is comprised of two phases. In the first
173 phase, interested sponsors formally request inclusion in STeP through a Q-submission (Section
174 IV and Appendix 1). The second phase encompasses actions to expedite the development of the
175 device and the prioritized review of subsequent regulatory submissions (e.g., pre-submissions,
176 marketing submissions) (Section V).

177
178 The principles below describe the philosophy of STeP and the approach FDA intends to take for
179 review of devices accepted into the program. As resources permit, FDA intends to leverage many
180 of the principles of the Breakthrough Devices Program for STeP in order to expedite the
181 development and review of devices that have the potential to significantly improve safety. As
182 part of the program, FDA and the sponsor should work collaboratively to define an efficient
183 device development path towards obtaining an FDA marketing authorization. To benefit from
184 the policies outlined for STeP, the commitment on behalf of the sponsor to resolve all scientific
185 and regulatory issues in a timely manner should match that of FDA. FDA believes that effective
186 communication (e.g., interactive review), collaboration, and the sponsor's commitment to
187 fulfilling all regulatory and scientific requirements are necessary to expedite the availability of
188 safe and effective medical devices.

189

¹¹ The types of events that must be reported to FDA pursuant to 21 CFR part 803 are described in the FDA guidance document entitled "[Medical Device Reporting for Manufacturers](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-device-reporting-manufacturers)," available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-device-reporting-manufacturers>.

¹² See "Medical Device Safety Action Plan: Protecting Patients, Promoting Public Health," available at <https://www.fda.gov/about-fda/cdrh-reports/medical-device-safety-action-plan-protecting-patients-promoting-public-health>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

190 FDA intends to evaluate resources throughout the device development, assessment, and review
191 processes to make the best use of FDA’s resources and maximize the impact of STeP. However,
192 when necessary, FDA plans to prioritize resources for the Breakthrough Devices Program over
193 STeP because the Breakthrough Devices Program is statutorily mandated.

194 **A. Interactive and Timely Communication**

195 For devices accepted into STeP, FDA intends to provide interactive and timely communication
196 with the sponsor during device development and throughout the review process for Q-
197 submissions, Investigational Device Exemption (IDE) applications, PMAs, certain PMA
198 supplements (i.e., Panel Track Supplements, 180 Day PMA Supplements), 510(k)s, and/or De
199 Novo requests. To promote collaborative dialogue and interaction between FDA and the sponsor,
200 both parties should, as applicable:

- 201 • agree on the goals of the interaction and feasibility of response timeframes prior to
202 submission of, or early in the review of, one of the relevant regulatory submissions listed
203 above;
- 204 • utilize redlined versions of documents being reviewed and/or revised interactively for
205 transparent communication concerning proposed changes; and
- 206 • utilize summary tables, documents, and/or FDA correspondence (e.g., written feedback,
207 meeting minutes) to communicate points of agreement, disagreement, or unresolved
208 issues at the conclusion of a review period.

209
210 Given that there may be novel scientific aspects of products in STeP, FDA may need to interact
211 with external experts or an Advisory Committee to reach various regulatory decisions.¹³ In the
212 event that such consultation is undertaken, FDA intends to follow the approach outlined in
213 Section II.A, of the Breakthrough Devices Program guidance document.

214 **B. Review Team Support**

215 Regulatory submissions (i.e., Q-submissions, IDEs, and marketing applications as listed above in
216 Section III.A.) for devices accepted into STeP come with review team support and senior
217 management (e.g., Office director or designee representing Office director) engagement, as
218 resources permit. Specifically, senior management intend to be involved in regulatory
219 submissions for devices accepted into STeP to ensure adherence to programmatic principles and
220 to support efficient and timely dispute resolution when points of disagreement cannot be
221 resolved quickly.

222
223 FDA intends for review teams of devices in STeP to be trained on programmatic principles and
224 features so that they are prepared to apply novel approaches to regulatory and device
225 development challenges.

226
227
228

¹³ For more information, see FDA’s guidance, “[Procedures for Meetings of the Medical Devices Advisory Committee](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/procedures-meetings-medical-devices-advisory-committee),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/procedures-meetings-medical-devices-advisory-committee>.

229

C. Review of Regulatory Submissions

230 FDA intends that the reviews of regulatory submissions for devices in STeP are prioritized in the
231 appropriate review queue as resources permit and receive additional review resources, as
232 available. Although prioritizing the review of devices in STeP is intended to help expedite
233 patient access, FDA’s past experience with the Priority Review Program¹⁴ indicates that review
234 times of the marketing submission may take longer for devices accepted into STeP than for other
235 devices because their anticipated technological or design innovations may raise novel scientific
236 issues. Similar to the Breakthrough Devices Program, we believe that STeP may enable patients
237 to have more timely access to these devices than they would have otherwise had because of the
238 earlier interaction between FDA and sponsors during the device development process to address
239 issues related to meeting the statutory standard for marketing authorization.

240

241 Given that the purpose of STeP is earlier access to devices that address important safety issues,
242 sponsors of devices under this program are expected to work interactively with FDA and respond
243 to FDA requests, collect premarket and postmarket data, and market their devices, if authorized,
244 in a timely manner. Sponsors of these devices should commit to resolving all scientific and
245 regulatory issues during the review process.

246

D. Benefit-Risk Assessments and Pre/Post-Market Balance of Data Collection

247

248 As with all devices subject to either a PMA or De Novo request, devices accepted into STeP
249 must meet the statutory standard of reasonable assurance of safety and effectiveness at the time
250 of PMA approval or granting of a De Novo request. When deciding whether to approve a PMA
251 or grant a De Novo request, FDA conducts a benefit-risk determination as described in the FDA
252 guidance document “[Factors to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications](#).”¹⁵ As part of the benefit-risk
253 determination, FDA considers the totality of evidence regarding the extent of probable benefits
254 and extent of probable risks of a device, including the extent of uncertainty in the benefit-risk
255 information. For devices in STeP, FDA intends to use timely postmarket data collection, when
256 appropriate for certain submission types, to facilitate expedited and efficient development and
257 review as described in the FDA guidance documents, “[Balancing Premarket and Postmarket Data Collection for Devices Subject to Premarket Approval](#)”¹⁶ and “[Consideration of](#)

259

¹⁴ FDA’s guidance, “Priority Review of Premarket Submissions for Devices,” issued on May 17, 2013, implemented former section 515(d)(5) of the FD&C Act (as in effect prior to the date of enactment of the 21st Century Cures Act), which applied only to PMAs. Because of the potential public health importance of devices warranting priority review status, FDA also applied the priority review criteria to other types of premarket submissions for devices. FDA withdrew this guidance on August 3, 2017. See <https://www.fda.gov/medical-devices/guidance-documents-medical-devices-and-radiation-emitting-products/withdrawn-guidance>.

¹⁵ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/factors-consider-when-making-benefit-risk-determinations-medical-device-premarket-approval-and-de>.

¹⁶ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/balancing-premarket-and-postmarket-data-collection-devices-subject-premarket-approval>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

260 [Uncertainty in Making Benefit-Risk Determinations in Medical Device Premarket Approvals, De](#)
261 [Novo Classifications, and Humanitarian Device Exemptions.](#)¹⁷

262
263 When making substantial equivalence determinations for devices in STeP subject to 510(k)
264 premarket review, FDA intends to follow the principles described in the FDA guidance
265 document, “[The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications](#)
266 [\[510\(k\)\]](#)”¹⁸ and, when appropriate, to apply benefit-risk policies in accordance with those
267 described in the FDA guidance document “[Benefit-Risk Factors to Consider When Determining](#)
268 [Substantial Equivalence in Premarket Notifications \(510\(k\)\) with Different Technological](#)
269 [Characteristics.](#)”¹⁹ As with devices reviewed under the PMA and De Novo pathways, devices
270 accepted into STeP must meet the applicable statutory standard at the time of 510(k) clearance.

271 **E. Efficient and Flexible Clinical Study Design**

272 Specific indications or labeling statements regarding clinical benefit for devices in STeP should
273 be supported by valid scientific evidence, including clinical data, in a manner consistent with
274 least burdensome approaches as described in the FDA guidance document “[The Least](#)
275 [Burdensome Provisions: Concept and Principles.](#)”²⁰ For devices in STeP, FDA intends to
276 consider proposals for efficient and flexible clinical study designs, including those incorporating
277 real world data sources, that may be used to support the proposed indication and/or labeling.²¹

278 **F. Manufacturing Considerations for PMA Submissions**

279 A device must be in conformance with the Quality System regulation (“QS Reg”; 21 CFR part
280 820), and the sponsor must submit adequate information in a PMA to meet the requirements
281 under section 515(c)(1)(C) of the FD&C Act (21 U.S.C. 360e(c)(1)(C)) and 21 CFR
282 814.20(b)(4)(v). As with other PMAs, sponsors of a device accepted for inclusion into STeP
283 should submit PMA information as described in the FDA guidance, “[Quality System Information](#)
284 [for Certain Premarket Application Reviews.](#)”²²

285
286 For application types that typically require a preapproval inspection (i.e., PMA), FDA intends to
287 expedite the review of manufacturing and quality systems compliance, as applicable and as
288 resources permit, for devices in STeP using approaches consistent with those established in
289 Section II.G of the Breakthrough Devices Program guidance document.

290

¹⁷ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/consideration-uncertainty-making-benefit-risk-determinations-medical-device-premarket-approvals-de>.

¹⁸ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k>.

¹⁹ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/benefit-risk-factors-consider-when-determining-substantial-equivalence-premarket-notifications-510k>.

²⁰ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/least-burdensome-provisions-concept-and-principles>.

²¹ See, for example, the guidance document “[Use of Real-World Evidence to Support Regulatory Decision-Making for Medical Devices](#)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-real-world-evidence-support-regulatory-decision-making-medical-devices>.

²² <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/quality-system-information-certain-premarket-application-reviews>.

291 **IV. Factors for STeP Acceptance and Review Process**

292 Inclusion in STeP is only at the request of the sponsor and with FDA’s agreement. To request
293 acceptance into STeP, interested sponsors should first evaluate whether they believe their device
294 meets the general eligibility factor (Section IV.A) and the specific program factors (Section
295 IV.B).

296 **A. General Eligibility Factor**

297 To be eligible for STeP, the device should be subject to marketing authorization via the PMA,
298 De Novo request, or 510(k) pathways.

299
300 As specified in Section I, FDA intends to consider device-led combination products for inclusion
301 in STeP. However, as part of the review processes for acceptance into STeP, FDA intends to
302 evaluate which constituent part of the product (i.e., device or drug/biologic) is providing the
303 proposed safety improvement, and only consider products in which safety improvements are
304 made to the device constituent part.

305
306 FDA intends to accept devices into STeP if FDA determines that, as described by the sponsor,
307 the device meets the general eligibility factor specified here and both of the specific eligibility
308 factors identified below.

309 **B. Specific Eligibility Factors for Inclusion in STeP**

310 For inclusion in STeP, devices:

- 311
- 312 1. should not be eligible for the Breakthrough Devices Program due to the less serious
313 nature of the disease or condition treated, diagnosed, or prevented by the device; and
314
 - 315 2. should be reasonably expected to significantly improve the benefit-risk profile of a
316 treatment or diagnostic through substantial safety innovations that provide for one or
317 more of the following:
 - 318 a. a reduction in the occurrence of a known serious adverse event,
 - 319 b. a reduction in the occurrence of a known device failure mode,
 - 320 c. a reduction in the occurrence of a known use-related hazard or use error, or
 - 321 d. an improvement in the safety of another device or intervention.

322 **C. Considerations for Evaluating Specific STeP Eligibility**
323 **Factors**

324 **(1) First Factor**

325 The first specific eligibility factor in Section IV.B describes the severity of the disease or
326 condition that a device in STeP is intended to address. A central tenet of the Breakthrough
327 Devices Program is that only devices that treat or diagnose life-threatening or irreversibly
328 debilitating diseases or conditions may be considered for designation based on certain statutory

Contains Nonbinding Recommendations

Draft – Not for Implementation

329 criteria.²³ FDA recognizes, however, that medical products are used to treat a wide variety of
330 health conditions not all of which have such serious morbidities, but which may still impact
331 patients' health and quality of life. Timely access to safer, medical devices for less serious
332 conditions is expected to be important in improving health outcomes, and it is these devices that
333 are the focus of STeP. Specifically, FDA intends to include in STeP devices that have the
334 potential to significantly improve the safety of existing treatments or diagnostics intended for use
335 in diseases or conditions that would be considered non-life-threatening or reasonably reversible.
336 These diseases or conditions could affect patient quality of life or be debilitating for short
337 timeframes, their health consequences might not significantly impact daily function, and/or they
338 might not progress to a more serious disease or condition.

339 (2) **Second Factor**

340 While the first specific eligibility factor considers the severity of the disease or condition the
341 device is intended to address, the second factor in Section IV.B considers how a device in STeP
342 is expected to improve the benefit-risk profile of a treatment or diagnostic compared to
343 alternatives for the identified disease or condition, as well as the significance of the expected
344 improvement. This second eligibility factor encompasses several elements including the
345 anticipated significant improvement to the benefit-risk profile, the type of the safety innovation
346 proposed, and whether the device addresses one of four specific categories of safety
347 improvement. Below is a discussion of how FDA intends to consider each of these elements for
348 the purpose of evaluating the second eligibility factor.

349 First, FDA intends to consider whether the device is reasonably expected to have a significant
350 improvement in the benefit-risk profile relative to other available treatment or diagnostic
351 alternatives for the disease or condition where there are known serious adverse events and/or
352 safety concerns (e.g., as identified in an FDA Safety Communication or medical device recall,²⁴
353 or otherwise identified as a significant safety issue of public health importance). FDA expects
354 that safety improvements generally should not compromise the device's effectiveness.
355 Additionally, as part of this evaluation, FDA will consider whether the safety profile of the new
356 device introduces the potential for new serious adverse events or use-related hazards due to the
357 proposed innovation. For example, a modification to a device made for the purpose of realizing a
358 safety improvement should not also be reasonably expected to increase the rate of a different
359 type of serious adverse event associated with the device or its use. FDA anticipates that requests
360 for STeP inclusion will primarily focus on medical devices offering a potential significant safety
361 improvement over other medical devices that are legally marketed in the United States. FDA also
362 intends to consider devices for inclusion in STeP that have the potential for significant safety
363 improvements over the current standard of care, which may include FDA-approved drugs or
364 biologics, or other technologies.

365 Second, FDA intends to consider the significance of the anticipated safety benefit and if the
366 anticipated improvement in the benefit-risk profile is through substantial safety innovations. For
367 the purposes of this evaluation, a substantial safety innovation is one that incorporates an

²³ The designation criteria are defined in Section 515B of the Federal Food, Drug and Cosmetic Act (FD&C Act) and described in the Breakthrough Devices Program guidance document.

²⁴ A collection of medical device safety information can be found at the following link:
<https://www.fda.gov/medical-devices/medical-device-safety>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

368 innovative technological feature or represents an innovative use of a technology to accomplish
369 the safety improvement. Illustrative examples of innovative technological features may include
370 changes to surface physicochemical properties, software, or material manufacturing method. As
371 part of this evaluation, FDA intends to consider the principles of operation of the device and
372 preliminary data from non-clinical or clinical sources and/or literature analyses. A complete
373 dataset of clinical evidence is generally not expected in a request for inclusion in STeP. FDA
374 intends to evaluate if there is a reasonable expectation for technical and clinical success of the
375 device based on information submitted by the sponsor.

376 Finally, FDA intends to consider how the device is reasonably expected to achieve the
377 significantly improved benefit-risk profile by considering whether the device meets one of the
378 following four sub-parts.

379 **a. A reduction in the occurrence of a known serious adverse event**

380 For this sub-part, the device should be reasonably expected to result in a significant reduction in
381 the occurrence of a known serious adverse event. FDA recognizes that, while some medical
382 products do not directly treat or diagnose life-threatening diseases or conditions, their use may be
383 associated with very serious adverse events including patient death or serious injury or illness.
384 These would include serious injuries or illnesses that lead to development of life-threatening
385 conditions, disability or permanent damage, or subsequent treatment or intervention to prevent
386 permanent impairment or damage.²⁵ Modifications to an existing medical device that address
387 these serious adverse events or a proposed new device that would reduce the occurrence of these
388 serious adverse events, based on the principles of operation of the device, would likely be
389 considered to meet this sub-part. For the purposes of STeP, FDA intends to consider serious
390 adverse events that are attributable or reasonably attributed to use of the device that occur in
391 acute timeframes following treatment or diagnosis (days to months) as well as those that are
392 associated with long term adverse outcomes (occur months to years following treatment or
393 diagnosis). Illustrative examples of devices that meet this sub-part may be expected to improve
394 safety by:

- 395 • significantly reducing or eliminating infections associated with death, life-threatening
396 conditions, or permanent disability, or
- 397 • significantly reducing or eliminating debilitating symptoms that manifest after device
398 implantation.

399 **b. A reduction in the occurrence of a known device failure mode**

400 For this sub-part of Factor 2, FDA intends to consider whether the device reduces the occurrence
401 of a known failure mode²⁶ that results in serious adverse health consequences,²⁷ including those
402 that result in death, are life-threatening, or involve permanent or long-term injuries to patients.
403 Devices may be considered to meet this sub-part if the failure is known to occur and not solely if

²⁵ See “serious injury” as defined in 21 CFR 803.3(w).

²⁶ For the purpose of this guidance, “failure mode” means “the manner in which failure occurs” and is intended to be used within the context of a risk management framework which may include formal failure mode and effects analysis (FMEA). See, for example, IEC 60812: *Analysis techniques for system reliability – Procedure for failure mode and effects analysis (FMEA)* and ISO 14971: *Medical devices – Applications of risk management to medical devices*, for additional information.

²⁷ See “serious, adverse health consequence” as defined in 21 CFR 810.2(j).

Contains Nonbinding Recommendations

Draft – Not for Implementation

404 there is a risk for the failure. Additionally, the failure mode should be associated with serious
405 adverse health consequences for the patient.

406 **c. A reduction in the occurrence of a known use-related hazard or**
407 **use error**

408 Medical devices should be safe and effective for their intended use(s) and condition(s) of use
409 including, for example, intended users and use environments, and manufacturers should design
410 their devices such that they incorporate features that mitigate use-related hazards or use errors.
411 Generally, use-related hazards and use errors result from user operation of, or interaction with,
412 the device and do not represent hazards that are consequences of either device or component
413 failure or are inherent to device design or material features. FDA’s guidance document
414 “[Applying Human Factors and Usability Engineering to Medical Devices](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/applying-human-factors-and-usability-engineering-to-medical-devices),”²⁸ defines both use-
415 related hazards and use errors. Use-related hazards and use errors that result in serious safety
416 issues can and do occur and may affect not only the patient but also the user of the device. In
417 addition to the patient, the user of the device may include a clinician or other person directly
418 involved in the administration and use of the device. For the purposes of inclusion in STeP, FDA
419 intends to consider medical devices with substantial safety innovations that improve upon use-
420 related hazards or use errors associated with the device design or operational features rather than
421 those associated with inadequate or unclear labeling (e.g., instructions for use).

422 **d. An improvement in the safety of another device or intervention**

423 When evaluating this sub-part, FDA intends to consider if the medical device is reasonably
424 expected to offer a specific type of improved safety benefit for another medical device or
425 intervention. In some cases, this improved safety benefit might come from the device being
426 evaluated for inclusion in STeP acting as an accessory²⁹ to other medical devices. This subpart
427 may, however, also apply to finished devices that are not accessories.

428 **D. Additional Considerations for STeP Acceptance Review**

429 **(1) Regulatory Path**

430 As described in Section III.A, as part of the acceptance process for STeP, FDA considers
431 whether the planned marketing pathway is a PMA, De Novo request, or 510(k). However,
432 accepting or denying a proposed device in STeP does not constitute a formal decision regarding
433 the applicable regulatory pathway or device classification. Instead, accepting a device into STeP
434 indicates that, based on the information provided in the request and other information known at
435 the time, the Agency expects that submission of a PMA (or PMA supplement), 510(k), or De
436 Novo request will be necessary for marketing authorization. When communicating acceptance or
437 denial of a device into STeP, FDA does not intend to specify which type of marketing
438 submission the sponsor will need to submit for the device.

²⁸ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/applying-human-factors-and-usability-engineering-to-medical-devices>.

²⁹ Medical device accessories are defined in the FDA guidance document entitled “[Medical Device Accessories – Describing Accessories and Classification Pathways](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-device-accessories-describing-accessories-and-classification-pathways)” (available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-device-accessories-describing-accessories-and-classification-pathways>) as a finished device that is intended to support, supplement, and/or augment the performance of one or more parent devices.

Contains Nonbinding Recommendations

Draft – Not for Implementation

439
440 Additionally, some, but not all, devices accepted into this program that are intended to improve
441 the safe use of another device under specific eligibility Factor 2d may be accessories to other
442 devices. Acceptance of a device into STeP does not constitute a decision on whether the device
443 is an accessory or on its risk classification. Please refer to FDA’s guidance document “[Medical
444 Device Accessories – Describing Accessories and Classification Pathways](#)”³⁰ for a discussion of
445 how FDA evaluates whether a medical device is an accessory as well as the classification
446 processes for devices that are considered accessories.

447
448 STeP is predicated upon expediting the development and review of devices that are reasonably
449 expected to address significant safety issues associated with available treatments. Therefore, the
450 safety improvement planned for the STeP device is relative to available technologies for treating
451 or diagnosing the same disease or condition. FDA recognizes that many of the devices ultimately
452 accepted into STeP are expected to offer safety improvements as compared to the use of other
453 medical devices. Changes to a medical device that are intended to affect its safety profile and/or
454 mitigate known safety issues are likely to require approval of a new PMA or PMA supplement,³¹
455 granting of a new De Novo request, or clearance of a new 510(k).³² As described in the FDA
456 guidance document “[Deciding When to Submit a 510\(k\) for a Change to an Existing Device](#),”³³
457 FDA anticipates that changes made to 510(k) devices intended to significantly affect safety, as is
458 the intent of STeP, will require a new premarket submission to the FDA. The substantial
459 equivalence evaluation in a 510(k) will not be impacted by acceptance of the device into STeP.
460 However, proposed modification(s) may raise different questions of safety or effectiveness as
461 compared to the unmodified version or other predicates. If the device cannot be found
462 substantially equivalent through the 510(k) process,³⁴ the sponsor may choose to pursue
463 marketing through either a PMA or De Novo request.³⁵
464

³⁰ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-device-accessories-describing-accessories-and-classification-pathways>.

³¹ The criteria for determining what type of application mechanism is needed when making device design or manufacturing changes to lawfully marketed PMA devices are described in 21 CFR 814.39 and elaborated on in the FDA guidance document entitled “[Modifications to Devices Subject to Premarket Approval \(PMA\) - The PMA Supplement Decision-Making Process](#)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/modifications-devices-subject-premarket-approval-pma-pma-supplement-decision-making-process>.

³² See FDA’s guidance “[Deciding When to Submit a 510\(k\) for a Change to an Existing Device](#)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/deciding-when-submit-510k-change-existing-device>.

³³ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/deciding-when-submit-510k-change-existing-device>.

³⁴ The framework for FDA’s evaluation of substantial equivalence in 510(k) submissions is described in detail in the guidance document entitled “[The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \[510\(k\)\]](#)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k>.

³⁵ See also <https://www.fda.gov/medical-devices/premarket-submissions/premarket-approval-pma> and “[De Novo Classification Process \(Evaluation of Automatic Class III Designation\)](#)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/de-novo-classification-process-evaluation-automatic-class-iii-designation>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

465 Finally, acceptance of a modified device into this program will not impact the obligations or
466 responsibilities of a manufacturer with respect to any recall or correction (e.g., reporting
467 requirements under 21 CFR 806). Please refer to the FDA guidance document “[Distinguishing](#)
468 [Medical Device Recalls from Medical Device Enhancements](#),” which provides additional
469 considerations for these types of device changes.³⁶

470 (2) Timeframe for STeP Acceptance

471 Ideally, sponsors should submit a request for inclusion in STeP prior to FDA receipt of the
472 marketing submission for that device. Additionally, FDA may consider requests for inclusion in
473 STeP in parallel with a marketing submission or after a marketing submission has been
474 submitted. It should be noted, however, that devices included in STeP during review of the
475 marketing submission may not benefit from programmatic features to the same extent as those
476 devices for which requests for inclusion in the program occur earlier in their development
477 process.

478 (3) Multiple Devices for the Same Expected Safety Benefit

479 FDA might accept multiple devices into STeP that are intending to address the same safety issue
480 or improvement. As a consequence, multiple regulatory submissions for devices intending to
481 address the same safety issue may be pending simultaneously.

482
483 FDA recommends that each request for inclusion in STeP be limited to one device intending to
484 address a significant safety concern(s).

485 E. Submitting a Request for Inclusion in STeP and FDA 486 Review

487 Requests for inclusion in STeP should be submitted using the Q-submission process as described
488 in the FDA guidance document “[Requests for Feedback on Medical Device Submissions: The Q-](#)
489 [Submission Program](#)” (hereinafter, “the Q-Submission Guidance”).³⁷ Requests for program
490 inclusion should be sent to the Document Control Center at CDRH or CBER as applicable for
491 the regulation of the device.³⁸ A sponsor intending to request inclusion in STeP should submit a
492 Q-submission package containing the recommended information as described in **Appendix 1:**
493 **Illustrative Example of a Request for Inclusion in STeP.** The inclusion request should be the
494 only request contained in the Q-Submission. Requests for feedback on the device outside of the
495 request for acceptance into STeP should be submitted separately. Furthermore, if sponsors are
496 requesting inclusion in STeP and at the same time have other requests for feedback pending, they
497 may wish to consider submitting those additional questions after FDA accepts or denies a device
498 in STeP as FDA’s feedback may incorporate STeP’s additional programmatic features. In
499 addition, note that requests for inclusion in STeP should be submitted separately from the
500 submission of a marketing submission or IDE application.

³⁶ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/distinguishing-medical-device-recalls-medical-device-enhancements>.

³⁷ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>.

³⁸ See FDA’s guidance document “[eCopy Program for Medical Device Submissions](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

501
502 FDA intends to accept or deny each request for inclusion in STeP within 60 calendar days of
503 receipt. In general, FDA intends to interact with a sponsor by Day 30 regarding any requests for
504 additional information needed to evaluate the request. It is helpful when a sponsor is available
505 and responsive to FDA requests throughout FDA’s review. If FDA does not receive additional
506 information needed to evaluate a STeP request in a timely manner, it may result in denial of the
507 request for inclusion in STeP.

508 **F. Withdrawal and Disqualification from STeP After**
509 **Program Acceptance**

510 A sponsor may request to withdraw from STeP at any time. Such a request should be submitted
511 in writing to FDA as a withdrawal amendment to the Q-submission number under which
512 inclusion into STeP was requested.

513 FDA does not intend to disqualify a device from further participation in STeP on the basis of
514 another STeP device that was intended to address the same safety issue receiving PMA approval,
515 having a De Novo request granted, or receiving clearance of a 510(k).³⁹ However, FDA may
516 disqualify a device from further participation in STeP at any time upon written notice to the
517 sponsor if FDA determines that:

- 518 • for other reasons, the device is no longer eligible for STeP based on available
519 information; or
- 520 • the information submitted in support of a request for inclusion in STeP, including,
521 without limitation, the Q-submission requesting inclusion in STeP or any related
522 premarket submission, contained an untrue statement of material fact or omitted material
523 information, including false statements relating to data collection.
524

525 **V. Mechanisms for Feedback on Development of Devices**
526 **in STeP**

527 To facilitate an interactive and expedited approach to device development and similar to features
528 outlined for devices granted Breakthrough designation, as resources permit, FDA intends to offer
529 sponsors of devices accepted into STeP several voluntary options for early and regular
530 interaction with FDA as device development progresses. A sponsor who wishes to request
531 feedback on a device that has been accepted into STeP may select one or more of the options
532 described below in Sections A-C; use of these options is not mandatory.

533 The options available for STeP include (1) a sprint discussion (See Section V.A), and (2) review
534 of a Data Development Plan (DDP) (See Section V.B). We consider these options to be subsets
535 of pre-submissions. When submitting a request for feedback on a device accepted into STeP,
536 sponsors should specify if they are requesting one of these special program features available to
537 facilitate the expedited review. Additionally, sponsors of devices accepted into STeP also have
538 the option to request feedback from FDA through mechanisms that are available for devices,

³⁹ This policy is consistent with program requirements for the Breakthrough Devices Program as required by section 515B(d)(3) of the FD&C Act (21 U.S.C. 360e-3(d)(3)).

Contains Nonbinding Recommendations

Draft – Not for Implementation

539 generally. For example, they may submit traditional pre-submissions whose scope is more
540 consistent with typical requests for feedback received through the pre-submission program
541 (Section V.C). FDA intends to follow the approaches and review procedures for these optional
542 mechanisms for feedback as outlined in Section IV. of the Breakthrough Devices Program
543 guidance document.

544 The regulatory mechanisms described in this section for obtaining FDA feedback on devices in
545 STeP may also be used for device-led combination products accepted into the program.
546 However, it is important to note these products may raise additional scientific challenges which
547 could influence the feedback that FDA provides. Interactive review of complex scientific issues
548 requiring expertise from a different Center may require additional time to resolve. When CDRH
549 or CBER receives a Q-submission, IDE, or marketing submission for a device-led combination
550 product that has been accepted into STeP, CDRH or CBER intends to notify the consulting
551 Center(s) of its receipt. Furthermore, the appropriate review staff from the consulting Center(s)
552 should be included in relevant meetings to ensure that the entire combination product review
553 team is aware of the issues discussed and that they are engaged, as needed, in the review.⁴⁰

554 Sponsors should recognize that, even though the FDA may have already reviewed the sponsor's
555 protocols/plans in a sprint discussion, DDP, or pre-submission, this does not guarantee approval,
556 clearance, or granting of future marketing submissions. Additional questions may be raised
557 during the review of the future submission when all information is available and considered as a
558 whole. Although sprint discussions, DDP reviews, and pre-submissions are not decisional or
559 binding on the Agency or the sponsor, it is FDA's intent to provide the best advice possible
560 based on the information provided by the sponsor and other information known at that point in
561 time.

562 FDA intends that the feedback the Agency provides in response to DDP requests, as part of a
563 sprint discussion, or through the pre-submission process will not change, provided that the
564 information submitted in a future IDE or marketing submission is consistent with that provided
565 in the feedback request and that the data in the future submission do not raise any important new
566 issues materially affecting safety or effectiveness. FDA intends that modifications to its feedback
567 be limited to situations in which FDA concludes that the feedback given previously does not
568 adequately address important new issues materially relevant to a determination of safety or
569 effectiveness that have emerged since the time the feedback was provided. For example, FDA
570 might modify its previous feedback if new scientific findings emerge that indicate there is a new
571 risk or an increased frequency of a known risk that affects FDA's prior advice, or if there is a
572 new public health concern that affects FDA's prior advice. In such cases, FDA intends to
573 acknowledge a change in the advice and clearly document the rationale for the change and the
574 appropriate managerial concurrence.

575

⁴⁰ While the lead Center is the primary contact point for combination product sponsors, OCP is available to participate in meetings or otherwise engage on regulatory matters for these products upon request (see section 503(g)(8) of the FD&C Act, 21 U.S.C. 353(g)(8)). For further information on combination products and OCP, see the OCP webpage at <https://www.fda.gov/CombinationProducts/default.htm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

576 **A. Sprint Discussion**

577 To support sponsors needing timely resolution of non-clinical or clinical evaluation issues, FDA
578 offers “sprint” discussions with the goal of reaching mutual agreement on a specific topic within
579 a set time period (e.g., 45 days) which FDA intends to expedite from the review timelines for
580 traditional pre-submissions⁴¹ as resources permit. The number, format, and duration of
581 interactions within a sprint discussion may vary based on project needs and should be defined *a*
582 *priori* by the sponsor and FDA. FDA recommends that sponsors limit the content of the sprint
583 request to one general topic (e.g., animal study protocol) and specific goals thereof.

584 During an open sprint review period, FDA recommends that the sponsor email draft meeting
585 minutes to FDA for comment and inclusion in the administrative record for the sprint
586 submission. Following closure of a sprint discussion, sponsors may additionally submit a formal
587 meeting minutes Q-submission amendment to FDA for review that documents all of the
588 teleconferences and/or face-to-face meetings held throughout the sprint review. To submit the
589 formal meeting minutes Q-submission amendment, sponsors should use the process described in
590 the Q-submission Guidance. FDA intends to follow the timeline and procedures established for
591 other meetings under the Q-Submission Program when reviewing formal meeting minutes Q-
592 submission amendments.

593 Additional information regarding the general conduct of a sprint discussion and example formats
594 are included in Section IV.A of the Breakthrough Devices Program guidance document.

595 **B. Data Development Plan (DDP)**

596 Sponsors of devices in STeP may request coordination with FDA regarding review of a DDP.
597 The DDP is an optional, high-level document intended to help ensure predictable, efficient,
598 transparent, and timely device assessment and review by outlining data collection expectations
599 for the entire product lifecycle. The DDP may include either clinical evaluation strategies, non-
600 clinical testing approaches, or both, as well as the anticipated timeframe for submitting results of
601 these evaluations to FDA for review (e.g., in an IDE application for a pivotal study). While the
602 optimal timeframe for submission of a DDP will vary depending on the device, it may be most
603 beneficial to initiate DDP discussions with FDA soon after acceptance into STeP. For additional
604 information on DDPs, please refer to Section IV.B of the Breakthrough Devices Program
605 guidance document. FDA encourages sponsors to consider the non-clinical testing that will be
606 needed to support the regulatory review of their device early in development and to discuss the
607 planned approach with FDA. Additionally, sponsors of devices accepted into STeP may outline
608 in their DDP any proposals to evaluate the clinical impact of safety improvements that balance
609 the amount of data collected pre- and post-market for PMAs.

610 FDA review of a DDP may follow a similar model as the sprint discussion described above and
611 is not subject to an acceptance review. In general and as resources permit, FDA anticipates that
612 feedback on a DDP will be provided in less time than would be expected for a traditional pre-

⁴¹ Review timelines for Pre-Submissions are described in the FDA guidance document entitled “[Requests for Feedback on Medical Device Submissions: The Q-Submission Program](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program)” available at: <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>.

613 submission.

614 **C. Other Pre-Submissions for STeP Devices**

615 FDA recognizes that some sponsors of devices accepted for inclusion in STeP may wish to
616 engage with FDA on a broader scope of topics in a single pre-submission than may be discussed
617 in one of the options presented above in Sections V.A and B. For these requests, the sponsor may
618 submit a pre-submission as described in the Q-submission Guidance and specify that it is for a
619 device accepted into STeP. As resources permit, review teams will prioritize these submissions
620 and develop an appropriate timeline for feedback with the sponsor that, when possible, does not
621 exceed review timelines for traditional pre-submissions.⁴²

622 **D. Regular Status Updates**

623 FDA and the sponsor of devices accepted into STeP may agree to have regular (e.g., bimonthly)
624 status updates outside of a formal regulatory submission to the Agency. Through these
625 interactions, FDA and the sponsor may discuss general progress of the project (e.g., timeframe
626 for a planned marketing submission) and next steps or plans for future discussions. Importantly,
627 FDA does not plan to provide feedback on device development progress or data during status
628 updates. For more detail regarding status updates, please refer to Section IV.E of the
629 Breakthrough Devices Program guidance document.

630

⁴² Review timelines for Pre-Submissions are described in the FDA guidance document entitled “[Requests for Feedback on Medical Device Submissions: The Q-Submission Program](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>.

631 **Appendix 1: Illustrative Example: Recommended Contents**
632 **of a Q-Submission Request for Inclusion in STeP**

633 This appendix outlines the recommended information that should be included in a Q-submission
634 requesting inclusion in STeP.

635 **Background Information**

636 *Device Description:* This section should provide an overview of the device or device-led
637 combination product (including device, drug, and/or biologic constituent parts), including
638 principles of operation and properties relevant to clinical function, if known. Images or
639 engineering schematics are also encouraged for inclusion, as appropriate.

640 *Expected Safety Improvement:* This section should provide a clear description of the safety issue
641 that the device is intending to address and rationale for the seriousness of the adverse events
642 associated with the safety issue. Additionally, the sponsor should describe any technological
643 advances or features of the device intended to improve safety.

644 *Indications for Use:* This section should present the sponsor’s proposed indications for use. If the
645 sponsor plans on including specific claims related to safety improvement, those claims should be
646 included as well.

647 *Regulatory History:* This section should detail the history of previous FDA interactions and
648 submissions, including feedback received and resolution of that feedback, as applicable. All
649 relevant IDE, 513(g)⁴³, and Q-submission numbers should be included.

650 **Justification for Meeting General Eligibility Factor**

651 *What is the planned marketing submission?*

- 652
- 653 • PMA;
 - 654 • De Novo request; or
 - 510(k).

655 This section should provide a discussion of which marketing submission the sponsor plans to
656 submit for the device, including a rationale for such selection. Only one submission type should
657 be selected.

658 **Justification for Meeting STeP Eligibility Factors**

660 *Eligibility Factor 1: The device seeking inclusion in STeP is “not eligible for the Breakthrough*

⁴³ See FDA’s guidance “[FDA and Industry Procedures for Section 513\(g\) Requests for Information under the Federal Food, Drug, and Cosmetic Act](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/fda-and-industry-procedures-section-513g-requests-information-under-federal-food-drug-and-cosmetic)” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/fda-and-industry-procedures-section-513g-requests-information-under-federal-food-drug-and-cosmetic>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

661 *Devices Program due to the less serious nature of the disease or condition treated or*
662 *diagnosed.”*

663 This section should provide a discussion regarding how the first STeP eligibility factor is met by
664 the proposed device and indications for use.

665 *Eligibility Factor 2: The device seeking inclusion in STeP is reasonably expected to significantly*
666 *improve the benefit-risk profile of a treatment or diagnostic through substantial safety*
667 *innovations that provide for one or more of the following:*

668 a. *A reduction in the occurrence of a known serious adverse event,*

669 b. *A reduction in the occurrence of a known device failure mode,*

670 c. *A reduction in the occurrence of a known use-related hazard or use error, or*

671 d. *An improvement in the safety of another device or intervention.*

672 This section should provide a discussion of which sub-part(s) of Eligibility Factor 2 is/are met by
673 the proposed device and indications for use. Please note that multiple sub-parts of Factor 2 may
674 apply; however, meeting only one of these sub-parts would still support inclusion in STeP if the
675 other eligibility factor is otherwise met. For each sub-part of Eligibility Factor 2 identified as
676 being met, a discussion regarding how that sub-part is met should be included in the request for
677 inclusion in STeP.

678