FACT SHEET FOR HEALTHCARE PROVIDERS

Clinical Research Sequencing Platform (CRSP), LLC at the Broad Institute of MIT and Harvard CRSP SARS-CoV-2 Real-time Reverse Transcriptase (RT)-PCR Diagnostic Assay

Updated: June 10, 2020

Coronavirus
Disease 2019
(COVID-19)

This Fact Sheet informs you of the significant known and potential risks and benefits of the emergency use of the CRSP SARS-CoV-2 Real-time Reverse Transcriptase (RT)-PCR Diagnostic Assay.

The CRSP SARS-CoV-2 Real-time RT-PCR Diagnostic Assay is authorized for use with respiratory specimens collected from individuals suspected of COVID-19 by their healthcare provider.

All patients whose specimens are tested with this assay will receive the Fact Sheet for Patients: Clinical Research Sequencing Platform (CRSP), LLC at the Broad Institute of MIT and Harvard - CRSP SARS-CoV-2 Real-time Reverse Transcriptase (RT)-PCR Diagnostic Assay.

What are the symptoms of COVID-19?

Many patients with COVID-19 have developed fever and/or symptoms of acute respiratory illness (e.g., cough, dyspnea), although some individuals experience only mild symptoms or no symptoms at all. The durrent information available to characterize the spectrum of clinical illness associated with COVID-19 suggests that, when present, symptoms include cough, shortness of breath or dyspnea, fever, chills, myalgias, headache, sore throat, new loss of taste or smell, nausea or vomiting or diarrhea. Signs and symptoms may appear any time from 2 to 14 days after exposure to the virus, and the median time to symptom onset is approximately 5 days. For further information on the symptoms of COVID-19 please see the link provided in "Where can I go for updates and more information?" section.

Public health officials have identified cases of COVID-19 infection throughout the world, including the United States. Please check the CDC COVID-19 webpage (see link provided in "Where can I go for updates and more information?" section at the end of this document) or your local jurisdictions website for the most up to date information.

This test is to be performed only using respiratory specimens collected from individuals suspected of COVID-19 by their healthcare provider.

What do I need to know about COVID-19 testing? Current information on COVID-19 for healthcare providers is available at CDC's webpage, Information for Healthcare Professionals (see links provided in "Where can I go for updates and more information?" section).

- The CRSP SARS-CoV2 Real-time RT-PCR
 Diagnostic Assay (Version 1) can be used to test
 nasopharyngeal, propharyngeal, anterior nasal
 (nasal), and mid-turbinate swabs, nasopharyngeal
 wash/aspirate and nasal aspirate specimens, and
 bronchoalveolar lavage specimens
 - The CRSP SARS-CoV-2 Real-time RT-PCR Diagnostic Assay should be ordered for the detection of COVID-19 in individuals suspected of COVID-19 by their healthcare provider.
- The CRSP SARS-CoV-2 Real-time RT-PCR Diagnostic Assay is only authorized for use at the Clinical Research Sequencing Platform (CRSP), LLC at the Broad Institute of MIT and Harvard located at 320 Charles Street, Cambridge, MA 02141 which is certified under Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, and meets requirements to perform high complexity tests.

Specimens should be collected with appropriate infection control precautions. Current guidance is available at the CDC's website (see links provided in "Where can I go for updates and more information?" section).

When collecting and handling specimens from individuals suspected of being infected with COVID-19, appropriate personal protective equipment should be used as outlined in the CDC *Interim Laboratory Biosafety Guidelines for Handling and Processing Specimens Associated with Coronavirus Disease 2019 (COVID-19).* For additional information, refer to CDC

FACT SHEET FOR HEALTHCARE PROVIDERS

Clinical Research Sequencing Platform (CRSP), LLC at the Broad Institute of MIT and Harvard CRSP SARS-CoV-2 Real-time Reverse Transcriptase (RT)-PCR Diagnostic Assay

Updated: June 10, 2020

Coronavirus
Disease 2019
(COVID-19)

Interim Guidelines for Collecting, Handling, and Testing Clinical Specimens from Persons Under Investigation (PUIs) for Coronavirus Disease 2019 (COVID-19) (see links provided in "Where can I go for updates and more information?" section).

What does it mean if the specimen tests positive for the virus that causes COVID-19?

A positive test result for COVID-19 indicates that RNA from SARS-CoV-2 was detected, and therefore the patient is infected with the virus and presumed to be contagious. Laboratory test results should always be considered in the context of clinical observations and epidemiological data (such as local prevalence rates and current outbreak/epicenter locations) in making a final diagnosis and patient management decisions. Patient management should be made by a healthcare provider and follow current CDC guidelines.

The CRSP SARS-CoV-2 Real-time RT-PCR Diagnostic Assay has been designed to minimize the likelihood false positive test results. However, it is still possible hat this test can give a false positive result, even when ed in locations where the prevalence is below 5%. In the event of a false positive result, risks to patients could include the following: a recommendation for isolation of the patient, monitoring of household or other clos contacts for symptoms, patient isolation that might limit contact with family or friends and may increase contact with other potentially COVID-19 patients, limits in ability to work, delayed diagnosis and treatment for the true infection causing the symptoms, unnecessary prescription of a treatment or the apy, or other unintended adverse effects.

All laboratories using this test must follow the standard testing and reporting guidelines according to their appropriate public health authorities.

What does it mean if the specimen tests negative for the virus that causes COVID-19?

A negative test result to this test means that SARS-CoV-2 RNA was not present in the specimen above the limit of detection. However, a negative result does not rule out COVID-19 and should not be used as the sole basis for treatment or patient management decisions. It

is possible to test a person too early or too late during COVID-19 infection to make an accurate diagnosis via CRSP SARS-CoV-2 Real-time RT-PCR Diagnostic Assay.

When diagnostic testing is negative, the possibility of a false negative result should be considered in the context of a patient's recent exposures and the presence of clinical signs and symptoms consistent with COVID-19. The possibility of a false negative result should especially be considered if the patient's recent exposures or clinical presentation indicate that COVID-19 is likely, and diagnostic tests for other causes of illness (e.g., other respiratory illness) are negative.

If COVID-19 is still suspected based on exposure history together with other clinical findings, re-testing with an alternative method should be considered by healthcare providers in consultation with public health authorities. Additional testing may be helpful to ensure testing was not conducted too early.

Risks to a patient of a false negative test result include: delayed or lack of supportive treatment, lack of prohitoring of infected individuals and their household or other close contacts for symptoms resulting in increased risk of spread of COVID-19 within the community, or other unintended adverse events.

The performance of this test was established based on the evaluation of a limited number of clinical specimens. The clinical performance has not been established in all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation. Performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARSCoV-2 and their prevalence, which change over time.

What is an EUA?

The United States FDA has made this test available under an emergency access mechanism called an Emergency Use Authorization (EUA). The EUA is supported by the Secretary of Health and Human Service's (HHS's) declaration that circumstances exist to

FACT SHEET FOR HEALTHCARE PROVIDERS

Clinical Research Sequencing Platform (CRSP), LLC at the Broad Institute of MIT and Harvard CRSP SARS-CoV-2 Real-time Reverse Transcriptase (RT)-PCR Diagnostic Assay

Updated: June 10, 2020

Coronavirus
Disease 2019
(COVID-19)

justify the emergency use of in vitro diagnostics (IVDs) for the detection and/or diagnosis of the virus that causes COVID-19.

An IVD made available under an EUA has not undergone the same type of review as an FDA-approved or cleared IVD. FDA may issue an EUA when certain criteria are met, which includes that there are no adequate, approved, available alternatives, and based on the totality of scientific evidence available, it is reasonable to believe that this IVD may be effective in diagnosing COVID-19.

The EUA for this test is in effect for the duration of the COVID-19 declaration justifying emergency use of IVDs, unless terminated or revoked (after which the test may no longer be used).

What are the approved available alternatives?

Any tests that have received full marketing status (e.g. cleared, approved), as opposed to an EUA, by FDA can be found by searching the medical device databases here: https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases. A cleared or approved test should be used instead of a test made available under an EUA, when appropriate and available. FDA has issued EUAs for other tests that can be found at:

https://www.fda.gov/emergency_pleparedness_andresponse/mcm-legal-regulatory_and-policyframework/emergency-use_authorization.

Where can I go for updates and more information?

CDC webpages:

General: https://www.cdc.gov/coronavinus/2019-ncov/index.html

Symptoms:

https://www.cdc.gov/coronavirus/2019-ncov/symptoms-

testing/symptoms.html

Healthcare Professionals:

https://www.cde.gov/corenavirus/2019-nCoV/hcp/index.html

Information for Laboratories:

https://www.cdc.gov/coronavirus/2019-nCoV/lab/index.html
Laboratory Biosafety: https://www.cdc.gov/coronavirus/2019-

nCoV/lab-biosafety-ruidelines.html

Isolation Precautions in Healthcare Settings:

https://www.cdc.gov/infectioncontrol/guidelines/isolation/index.htm

Specimen Collection: https://www.cdc.gov/coronavirus/2019-

nCoV/quidelines-clinical-specimens.html

Infection Control: https://www.cdc.gov/coronavirus/2019-

ncov/php/infection-control.html

FDA webpages:

General: www.fda.gov/novelcoronavirus

EUAs:(includes links to fact sheet for individuals and manufacturer's instructions) https://www.fda.gov/medical-devices/in-vitro-diagnostics-euas

<u>Clinical Research Sequencing Platform (CRSP),</u> <u>LLC at the Broad Institute of MIT and Harvard:</u>

320 Charles Street, Cambridge, MA 02141

Customer Support:

+1 617 714-7250

covidsupport@broadinstitute.org

Technical Support:

+1 617 714-7250

covidsupport@broadinstitute.org