

August 21, 2020

Aviva Jacobs, Ph.D.
Vice President, Product Development
DxTerity Diagnostics, Inc.
19500 S. Rancho Way, Suite 116
Rancho Dominguez, CA 90220

Device: DxTerity SARS-CoV-2 RT-PCR Test

Laboratory: DxTerity Diagnostics, Inc.

Indication: Qualitative detection of nucleic acid from SARS-CoV-2 in saliva specimens from individuals suspected of COVID-19 by their healthcare provider.

This test is for use with saliva specimens from individuals with symptoms of COVID-19 collected in a healthcare setting under the supervision of a healthcare provider using the Spectrum Solutions LLC SDNA-1000 Saliva Collection Device.

Testing is limited to the DxTerity Diagnostics, Inc. laboratory located at 19500 S. Rancho Way, Suite 116, Rancho Dominguez, CA 90220 which is certified under Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, and meets requirements to perform high complexity tests.

Dear Dr. Jacobs:

This letter is in response to your¹ request that the Food and Drug Administration (FDA) issue an Emergency Use Authorization (EUA) for emergency use of your product,² pursuant to Section 564 of the Federal Food, Drug, and Cosmetic Act (the Act) (21 U.S.C. §360bbb-3).

On February 4, 2020, pursuant to Section 564(b)(1)(C) of the Act, the Secretary of the Department of Health and Human Services (HHS) determined that there is a public health emergency that has a significant potential to affect national security or the health and security of United States citizens living abroad, and that involves the virus that causes COVID-19. Pursuant to Section 564 of the Act, and on the basis of such determination, the Secretary of HHS then declared that circumstances exist justifying the authorization of emergency use of in

¹ For ease of reference, this letter will use the term “you” and related terms to refer to DxTerity Diagnostics, Inc.

² For ease of reference, this letter will use the term “your product” to refer to the DxTerity SARS-CoV-2 RT-PCR Test used for the indication identified above.

vitro diagnostics for detection and/or diagnosis of the virus that causes COVID-19 subject to the terms of any authorization issued under Section 564(a) of the Act.³

Having concluded that the criteria for issuance of this authorization under Section 564(c) of the Act are met, I am authorizing the emergency use of your product, described in the Scope of Authorization of this letter (Section II), subject to the terms of this authorization.

I. Criteria for Issuance of Authorization

I have concluded that the emergency use of your product meets the criteria for issuance of an authorization under Section 564(c) of the Act, because I have concluded that:

1. The SARS-CoV-2 can cause a serious or life-threatening disease or condition, including severe respiratory illness, to humans infected by this virus;
2. Based on the totality of scientific evidence available to FDA, it is reasonable to believe that your product may be effective in diagnosing COVID-19, and that the known and potential benefits of your product when used for diagnosing COVID-19, outweigh the known and potential risks of your product; and
3. There is no adequate, approved, and available alternative to the emergency use of your product.⁴

II. Scope of Authorization

I have concluded, pursuant to Section 564(d)(1) of the Act, that the scope of this authorization is limited to the indication above.

Authorized Product Details

Your product is a qualitative test for the detection of nucleic acid from SARS-CoV-2 in saliva specimens from individuals suspected of COVID-19 by their healthcare provider.

Saliva specimens are collected for use with your product using the Spectrum Solutions LLC SDNA-1000 Saliva Collection Device in a healthcare setting under the supervision of a trained healthcare provider.

The SARS-CoV-2 nucleic acid is generally detectable in saliva specimens during the acute phase of infection. Positive results are indicative of the presence of SARS-CoV-2 nucleic acid; clinical correlation with patient history and other diagnostic information is necessary to determine patient infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. Negative results do not preclude SARS-CoV-2 infection and should not be used as the

³ U.S. Department of Health and Human Services, *Determination of a Public Health Emergency and Declaration that Circumstances Exist Justifying Authorizations Pursuant to Section 564(b) of the Federal Food, Drug, and Cosmetic Act*, 21 U.S.C. § 360bbb-3. 85 FR 7316 (February 7, 2020).

⁴ No other criteria of issuance have been prescribed by regulation under Section 564(c)(4) of the Act.

sole basis for patient management decisions. Negative results must be combined with clinical observations, patient history, and epidemiological information. Negative results from saliva should be confirmed by testing of an alternative specimen type if clinically indicated.

To use your product, SARS-CoV-2 nucleic acid is first extracted, isolated and purified from saliva specimens collected using the Spectrum Solutions LLC SDNA-1000 Saliva Collection Device. The purified nucleic acid is then reverse transcribed into cDNA followed by PCR amplification and detection using an authorized real-time (RT) PCR instrument described in the authorized labeling (described below).

The product uses all commercially sourced materials or other authorized materials and authorized ancillary reagents commonly used in clinical laboratories as described in the authorized labeling.

Your product requires the following control materials, or other authorized control materials (as specified under Condition P below), that are processed in the same way as the patient specimens and are required to be included with each batch of specimens tested with your product. All controls listed below must generate expected results in order for a test to be considered valid, as outlined in the authorized labeling:

- Negative Control – consists of nuclease-free water to monitor for cross-contamination during RNA extraction and RT-PCR
- Positive Control – RNA control that contains the genomic regions targeted for amplification by the assay to monitor the integrity of the RT-PCR reagents and process
- Internal Control (MS2 Phage) – monitors the integrity of nucleic acid extraction and RT-PCR for each specimen

The above described product is authorized to be accompanied with the DxTerity SARS-CoV-2 RT-PCR Test Procedure,” the “Automated 30 min RNA Extraction from Stabilized Saliva Samples” protocol, and the EUA Summary (available at <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/vitro-diagnostics-euas>), and the following information pertaining to the emergency use, which is required to be made available to healthcare providers and patients:

- Fact Sheet for Healthcare Providers: DxTerity Diagnostics, Inc. – DxTerity SARS-CoV-2 RT-PCR Test
- Fact Sheet for Patients: DxTerity Diagnostics, Inc. – DxTerity SARS-CoV-2 RT-PCR Test

The above described product, when accompanied by the “DxTerity SARS-CoV-2 RT-PCR Test Procedure,” the “Automated 30 min RNA Extraction from Stabilized Saliva Samples” protocol, the EUA Summary (identified above) and the two Fact Sheets (collectively referenced as “authorized labeling”) is authorized to be used under this EUA, despite the fact that it does not meet certain requirements otherwise required by applicable federal law.

I have concluded, pursuant to Section 564(d)(2) of the Act, that it is reasonable to believe that

the known and potential benefits of your product, when used consistent with the Scope of Authorization of this letter (Section II), outweigh the known and potential risks of your product.

I have concluded, pursuant to Section 564(d)(3) of the Act, based on the totality of scientific evidence available to FDA, that it is reasonable to believe that your product may be effective in diagnosing COVID-19, when used consistent with the Scope of Authorization of this letter (Section II), pursuant to Section 564(c)(2)(A) of the Act.

FDA has reviewed the scientific information available to FDA, including the information supporting the conclusions described in Section I above, and concludes that your product (as described in the Scope of Authorization of this letter (Section II)) meets the criteria set forth in Section 564(c) of the Act concerning safety and potential effectiveness.

The emergency use of your product under this EUA must be consistent with, and may not exceed, the terms of this letter, including the Scope of Authorization (Section II) and the Conditions of Authorization (Section IV). Subject to the terms of this EUA and under the circumstances set forth in the Secretary of HHS's determination under Section 564(b)(1)(C) of the Act described above and the Secretary of HHS's corresponding declaration under Section 564(b)(1) of the Act, your product is authorized for the indication above.

III. Waiver of Certain Requirements

I am waiving the following requirements for your product during the duration of this EUA:

- Current good manufacturing practice requirements, including the quality system requirements under 21 CFR Part 820 with respect to the design, manufacture, packaging, labeling, and storage of your product.

IV. Conditions of Authorization

Pursuant to Section 564(e) of the Act, I am establishing the following conditions on this authorization:

DxTerity Diagnostics, Inc. (You)

- A. Your product must comply with the following labeling requirements under FDA regulations: the intended use statement (21 CFR 809.10(a)(2), (b)(2)); adequate directions for use (21 U.S.C. 352(f)), (21 CFR 809.10(b)(5), (7), and (8)); appropriate limitations on the use of the device including information required under 21 CFR 809.10(a)(4); and any available information regarding performance of the device, including requirements under 21 CFR 809.10(b)(12).
- B. You will inform relevant public health authorities of this EUA, including the terms and conditions herein, and any updates made to your product and authorized labeling.
- C. You will notify the relevant public health authorities of your intent to run your product.

- D. You will have a process in place for reporting test results to healthcare providers and relevant public health authorities, as appropriate.
- E. You will include with test result reports, all authorized Fact Sheets. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- F. You will make available on your website(s), if applicable, the Fact Sheet for Healthcare Providers and the Fact Sheet for Patients.
- G. You are authorized to make available additional information relating to the emergency use of your product that is consistent with, and does not exceed, the terms of this letter of authorization.
- H. You will use your product as outlined in the authorized labeling. Deviations from the authorized laboratory procedures, including the authorized instruments, authorized extraction methods, authorized clinical specimen types, authorized control materials, authorized other ancillary reagents and/or authorized materials required to use your product are not permitted.
- I. You will collect information on the performance of your product. You will report to Division of Microbiology (DMD)/Office of Health Technology 7 (OHT7)-Office of In Vitro Diagnostics and Radiological Health (OIR)/Office of Product Evaluation and Quality (OPEQ)/Center for Devices and Radiological Health (CDRH) (via email: CDRH-EUA-Reporting@fda.hhs.gov) any suspected occurrence of false positive or false negative results and significant deviations from the established performance characteristics of your product of which you become aware.
- J. You may request changes to the Scope of Authorization (Section II in this letter) of your product. Such requests will be made in consultation with DMD/OHT7-OIR/OPEQ/CDRH, and require concurrence of, Office of Counterterrorism and Emerging Threats (OCET)/Office of the Chief Scientist (OCS)/Office of the Commissioner (OC) and DMD/OHT7-OIR/OPEQ/CDRH.
- K. You may request changes to the authorized labeling. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- L. You may request the addition of other instruments and associated software for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- M. You may request the addition of other extraction methods for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.

- N. You may request the addition of other specimen types for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- O. You may request the addition and/or substitution of primers or probes for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- P. You may request the addition and/or substitution of control materials for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- Q. You may request the addition and/or substitution of other ancillary reagents and materials for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- R. You may request the addition and/or substitution of saliva collection kits or kit components for use with your product. Such requests will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- S. You will evaluate the analytical limit of detection and assess traceability of your product with any FDA-recommended reference material(s), if requested by FDA.⁵ After submission to FDA and DMD/OHT7-OIR/OPEQ/CDRH's review of and concurrence with the data, FDA will update the EUA summary to reflect the additional testing. Such updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- T. You will further perform a sample stability study using the authorized collection device in an FDA agreed upon post authorization study within 3 months of the date of this letter (unless otherwise agreed to with DMD/OHT7-OIR/OPEQ/CDRH). After submission to FDA and DMD/OHT7-OIR/OPEQ/CDRH's review of and concurrence with the data, you will update authorized labeling to reflect the additional testing. Such labeling updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.”
- U. You will further evaluate the negative percent agreement of your product in an FDA agreed upon post authorization clinical evaluation study within 3 months of the date of this letter (unless otherwise agreed to with DMD/OHT7-OIR/OPEQ/CDRH). After submission to FDA and DMD/OHT7-OIR/OPEQ/CDRH's review of and concurrence with the data, you will update authorized labeling to reflect the additional testing. Such labeling updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.

⁵ Traceability refers to tracing analytical sensitivity/reactivity back to an FDA-recommended reference material. FDA may request, for example, that you perform this study in the event that we receive reports of adverse events concerning your authorized test.

- V. You will track adverse events, including any occurrence of false results with your product and report any such events to FDA pursuant to 21 CFR Part 803.
- W. All laboratory personnel using your product must be appropriately trained in molecular techniques and use appropriate laboratory and personal protective equipment when handling this product, and use your product in accordance with the authorized laboratory procedure.
- X. You will ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records will be made available to FDA for inspection upon request.
- Y. You will track adverse events associated with the Spectrum Solutions LLC SDNA-1000 Saliva Collection Device, or any other saliva collection kit authorized for use with your product, including occurrences of false results and report to FDA pursuant to 21 CFR Part 803. Serious adverse events, especially unexpected biosafety concerns, should immediately be reported to DMD/OHT7-OIR/OPEQ/CDRH (via email: CDRH-EUA-Reporting@fda.hhs.gov).
- Z. You will submit to FDA a summary report within 30 calendar days of authorization summarizing the results of any testing performed using saliva specimens collect with the Spectrum Solutions LLC SDNA-1000 Saliva Collection Device or any new saliva collection kit authorized for use with your product during that timeframe, including the positivity rate for specimens collected with the authorized collection kit.
- AA. Upon request you will conduct post-authorization studies and/or data analysis concerning the performance of saliva specimens with your authorized test. Such studies and/or data analysis will be agreed upon between you and FDA. After submission to FDA and DMD/OHT7-OIR/OPEQ/CDRH's review of the data, FDA will consider whether additional action is appropriate, such as revision or revocation of the EUA.

Conditions Related to Printed Materials, Advertising and Promotion

- BB. All descriptive printed matter, including advertising and promotional materials, relating to the use of your product shall be consistent with the authorized labeling, as well as the terms set forth in this EUA and the applicable requirements set forth in the Act and FDA regulations.
- CC. No descriptive printed matter, including advertising or promotional materials, relating to the use of your product may represent or suggest that this test is safe or effective for the detection of SARS-CoV-2.
- DD. All descriptive printed matter, including advertising and promotional materials, relating to the use of your product shall clearly and conspicuously state that:

- This test has not been FDA cleared or approved;
- This test has been authorized by FDA under an EUA for use by DxTerity Diagnostics, Inc. located at 19500 S. Rancho Way, Suite 116, Rancho Dominguez, CA 90220;
- This test has been authorized only for the detection of nucleic acid from SARS-CoV-2, not for any other viruses or pathogens; and
- This test is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the authorization is terminated or revoked sooner.

The emergency use of your product as described in this letter of authorization must comply with the conditions and all other terms of this authorization.

V. Duration of Authorization

This EUA will be effective until the declaration that circumstances exist justifying the authorization of the emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 is terminated under Section 564(b)(2) of the Act or the EUA is revoked under Section 564(g) of the Act.

Sincerely,

RADM Denise M. Hinton
Chief Scientist
Food and Drug Administration

Enclosure