


Collection precautions:

Ensure the sponge tip does NOT come into contact with any surface prior to collection.

Donor should NOT eat, drink, smoke or chew gum for 30 minutes before collecting saliva sample.

Contents: Contains 1 collection kit.

Warning and precautions: Choking hazard. Caution should be used when inserting sponge into the mouth.

Wash with water if liquid comes in contact with eyes or skin. Do NOT ingest. See MSDS at www.dnagenotek.com


Storage: 15-25°C

Summary and explanation of the kit:


ORACollect[®]•RNA is a collection kit that provides the materials and instructions for collecting, stabilizing and maintaining during transport saliva specimens suspected of containing SARS-CoV-2 ribonucleic acid (RNA).

Label legend:

REF	Catalog number
MD	In vitro diagnostic medical device
M	Manufacturer
15°C-25°C	Storage instructions
U	Collect sample by (Use by)
LOT	Lot number
Rx	For Prescription Use Only




Open package and remove collector without touching sponge tip. Place sponge as far back in the mouth as comfortable and rub along the lower gums (see close up image) in a back and forth motion. Gently rub the gums 10 times. If possible, avoid rubbing the teeth.




Gently repeat rubbing motion on the opposite side of the mouth along the lower gums for an additional 10 times.



Hold the tube upright to prevent the liquid inside the tube from spilling. Unscrew the cap from the collection tube without touching the sponge.



Turn the cap upside down, insert the sponge into the tube and close cap tightly.



Invert the capped tube and shake vigorously 15 times.

For IVD, Rx and For Use Under Emergency Use Authorization.

This sample collection device has not been FDA cleared or approved.

This sample collection device has been authorized by FDA under an EUA.

This sample collection device has been authorized only to collect, stabilize, and maintain during transport, saliva specimens suspected of containing SARS-CoV-2 ribonucleic acid (RNA), not for any other viruses or pathogens.

This sample collection device is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of medical devices during the COVID-19 outbreak under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the authorization is terminated or revoked sooner.

DNAGENOTEK™

■ DNA Genotek Inc.
3000 - 500 Palladium Drive
Ottawa, ON, Canada K2V 1C2

Superior samples • Proven performance

Toll-free (North America): 1.866.813.6354
Tel: +1.613.723.5757 • Fax: +1.613.723.5057
info@dnagenotek.com • www.dnagenotek.com

*ORACollect is a registered trademark and DNA Genotek is a trademark of DNA Genotek Inc. Some DNA Genotek products may not be available in all geographic regions, contact your sales representative for details. All DNA Genotek protocols, white papers and application notes, are available in the support section of our website at www.dnagenotek.com.

Patent (www.dnagenotek.com/legal/notice)
© 2020 DNA Genotek Inc., a subsidiary of OraSure Technologies, Inc., all rights reserved.
PD-PR-01063 Issue 1/2020-10

Instructions for sample collection:



Collection precautions:

Ensure the sponge tip does NOT come into contact with any surface prior to collection.
Donor should NOT eat, drink, smoke or chew gum for 30 minutes before collecting saliva sample.

Contents: Contains 1 collection kit.

Warning and precautions: Choking hazard. Caution should be used when inserting sponge into the mouth.

Wash with water if liquid comes in contact with eyes or skin. Do NOT ingest. See MSDS at www.dnagenotek.com

Storage: 15-25°C

Summary and explanation of the kit:

ORACollect®-RNA is a collection kit that provides the materials and instructions for collecting, stabilizing and maintaining during transport saliva specimens suspected of containing SARS-CoV-2 ribonucleic acid (RNA).

Label legend:

- Catalog number
- CE Marking
- In vitro diagnostic medical device
- Manufacturer
- Authorized Representative
- 15°C-25°C Storage instructions
- Collect sample by (Use by)
- Lot number
- For Prescription Use Only

	<p>Open package and remove collector without touching sponge tip. Place sponge as far back in the mouth as comfortable and rub along the lower gums (see close up image) in a back and forth motion. Gently rub the gums 10 times. If possible, avoid rubbing the teeth.</p>
	<p>Gently repeat rubbing motion on the opposite side of the mouth along the lower gums for an additional 10 times.</p>
	<p>Hold the tube upright to prevent the liquid inside the tube from spilling. Unscrew the cap from the collection tube without touching the sponge.</p>
	<p>Turn the cap upside down, insert the sponge into the tube and close cap tightly.</p>
	<p>Invert the capped tube and shake vigorously 15 times.</p>



For IVD, Rx and For Use Under Emergency Use Authorization.

This sample collection device has not been FDA cleared or approved.

This sample collection device has been authorized by FDA under an EUA.

This sample collection device has been authorized only to collect, stabilize, and maintain during transport, saliva specimens suspected of containing SARS-CoV-2 ribonucleic acid (RNA), not for any other viruses or pathogens.

This sample collection device is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of medical devices during the COVID-19 outbreak under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the authorization is terminated or revoked sooner.



Superior samples • Proven performance

DNA Genotek Inc.
 3000 - 500 Palladium Drive
 Ottawa, ON, Canada K2V 1C2

Toll-free (North America): 1.866.813.6354
 Tel: +1.613.723.5757 • Fax: +1.613.723.5057
info@dnagenotek.com • www.dnagenotek.com



Emergo Europe, Prinsessegracht 20, 2514 AP The Hague, The Netherlands

Australian Sponsor: Emergo Australia, Level 20, Tower II, Darling Park, 201 Sussex Street, Sydney, NSW 2000 Australia

*ORACollect is a registered trademark and DNA Genotek is a trademark of DNA Genotek Inc. Some DNA Genotek products may not be available in all geographic regions, contact your sales representative for details. All DNA Genotek protocols, white papers and application notes, are available in the support section of our website at www.dnagenotek.com.

Patent (www.dnagenotek.com/legalnotices)
 © 2020 DNA Genotek Inc., a subsidiary of OraSure Technologies, Inc., all rights reserved.
 PD-PR-01062 Issue 1/2020-10