



October 12, 2021

Stacy Drakousis
Regulatory Affairs Manager
Thermo Fisher Scientific, Inc.
5781 Van Allen Way
Carlsbad, CA 92008

Device: Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit

EUA Number: EUA210257

Company: Thermo Fisher Scientific, Inc.

Indications: This test is authorized for the following indications for use:
Qualitative detection of nucleic acid from SARS-CoV-2 in nasopharyngeal and anterior nasal swab specimens from individuals suspected of COVID-19 by their healthcare provider.
Qualitative detection of nucleic acid from SARS-CoV-2 in saliva specimens collected using Spectrum Solutions SDNA 1000 Saliva Collection Device from individuals suspected of COVID-19 by their healthcare provider.
Emergency use of this test is limited to authorized laboratories.

Authorized Laboratories: Laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform high complexity tests.

REVIEWED

Dear Ms. Drakousis:

On April 9, 2021, based on your¹ request, the Food and Drug Administration (FDA) issued a letter authorizing the emergency use of the Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit pursuant to Section 564 of the Federal Food, Drug, and Cosmetic Act (the Act) (21 U.S.C. §360bbb-3) for the qualitative detection of nucleic acid from SARS-CoV-2 in nasopharyngeal and anterior nasal swab specimens from individuals suspected of COVID-19 by their healthcare provider. Testing was limited to laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet

¹ For ease of reference, this letter will use the term “you” and related terms to refer to Thermo Fisher Scientific, Inc.

requirements to perform high complexity tests. Based on your request, FDA revised and reissued the letter on September 7, 2021.²

On August 27, 2021, you requested to further revise your Emergency Use Authorization (EUA). Based on this request, and having concluded that revising the September 7, 2021 EUA is appropriate to protect the public health or safety under section 564(g)(2)(C) of the Act (21 U.S.C. § 360bbb 3(g)(2)(C)), FDA is reissuing the September 7, 2021 letter in its entirety with the revisions incorporated.³ Pursuant to section 564 of the Act and the Scope of Authorization (Section II) and Conditions of Authorization (Section IV) of this reissued letter, your product⁴ is now authorized for use consistent with the indication described above.

On February 4, 2020, pursuant to Section 564(b)(1)(C) of the Act, the Secretary of the Department of Health and Human Services (HHS) determined that there is a public health emergency that has a significant potential to affect national security or the health and security of United States citizens living abroad, and that involves the virus that causes COVID-19. Pursuant to Section 564 of the Act, and on the basis of such determination, the Secretary of HHS then declared that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of the virus that causes COVID-19 subject to the terms of any authorization issued under Section 564(a) of the Act.⁵

FDA considered the totality of scientific information available in authorizing the emergency use of your product for the indication above. A summary of the performance information FDA relied upon is contained in the Instructions for Use (identified below).

Having concluded that the criteria for issuance of this authorization under Section 564(c) of the Act are met, I am authorizing the emergency use of your product, described in the Scope of Authorization of this letter (Section II), subject to the terms of this authorization.

I. Criteria for Issuance of Authorization

I have concluded that the emergency use of your product meets the criteria for issuance of an authorization under Section 564(c) of the Act, because I have concluded that:

² On September 7, 2021, the revisions to the April 9, 2021 letter and authorized labeling included: (1) updating the Instructions for Use (IFU) to include additional information on instrument cleaning and decontamination and safety and electromagnetic compatibility standards, and additional edits to reflect language used in more recent authorizations, (2) updating the Healthcare Provider and Patient Fact Sheets to include language used in more recent authorizations, and (3) revisions to the letter and Conditions of Authorization (Section IV) for the addition of new Conditions related to circulating variants (Conditions Q and R in the September 7, 2021 letter) and additional edits to reflect language used in more recent authorizations.

³ The revisions to the September 7, 2021, letter and authorized labeling include: (1) addition of saliva specimens collected using Spectrum Solutions SDNA 1000 Saliva Collection Device from individuals suspected of COVID-19 by their healthcare provider as an indicated specimen type and (2) revision to add information on performance with viral mutations.

⁴ For ease of reference, this letter will use the term “your product” to refer to the Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit used for the indication identified above.

⁵ U.S. Department of Health and Human Services, *Determination of a Public Health Emergency and Declaration that Circumstances Exist Justifying Authorizations Pursuant to Section 564(b) of the Federal Food, Drug, and Cosmetic Act*, 21 U.S.C. § 360bbb-3. 85 FR 7316 (February 7, 2020).

1. The SARS-CoV-2 can cause a serious or life-threatening disease or condition, including severe respiratory illness, to humans infected by this virus;
2. Based on the totality of scientific evidence available to FDA, it is reasonable to believe that your product may be effective in diagnosing COVID-19, and that the known and potential benefits of your product when used for diagnosing COVID-19, outweigh the known and potential risks of your product; and
3. There is no adequate, approved, and available alternative to the emergency use of your product.⁶

II. Scope of Authorization

I have concluded, pursuant to Section 564(d)(1) of the Act, that the scope of this authorization is limited to the indication above.

Authorized Product Details

Your product is a qualitative test for the detection of nucleic acid from SARS-CoV-2 in nasopharyngeal and anterior nasal swab specimens collected from individuals suspected of COVID-19 by their healthcare provider. Your product is also for use with saliva specimens collected using Spectrum Solutions SDNA 1000 Saliva Collection Device from individuals suspected of COVID-19 by their healthcare provider. The SARS-CoV-2 nucleic acid is generally detectable in nasopharyngeal swab, anterior nasal swab and saliva specimens during the acute phase of infection. Positive results are indicative of the presence of SARS-CoV-2 nucleic acid; clinical correlation with patient history and other diagnostic information is necessary to determine patient infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. Negative results do not preclude SARS-CoV-2 infection and should not be used as the sole basis for patient management decisions. Negative results must be combined with clinical observations, patient history, and epidemiological information. Negative results for SARS CoV-2 RNA from saliva should be confirmed by testing of an alternative specimen type if clinically indicated.

To use your product, SARS-CoV-2 nucleic acid is first extracted, isolated and purified from upper respiratory and saliva specimens. The purified nucleic acid is then reverse transcribed into cDNA followed by PCR amplification and detection using an authorized real-time PCR instrument and interpretive software using various workflows outlined in the Instructions for Use. The Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit includes the following materials or other authorized materials: TaqPath COVID-19 Module 2 Assay Kit, TaqPath COVID-19 Module 1 MS2 Phage Control, and TaqPath COVID-19 Control.

Your product requires the following control materials, or other authorized control materials (as may be requested under Condition K below), that are processed in the same way as the patient specimens and are required to be included with each batch of specimens tested with your

⁶ No other criteria of issuance have been prescribed by regulation under Section 564(c)(4) of the Act.

product. All controls listed below must generate expected results in order for a test to be considered valid, as outlined in the Instructions for Use:

- Internal Positive Control – MS2 phage control which is required as an extraction, reverse transcription and PCR amplification positive control.
- External Positive control - TaqPath COVID-19 Control contains the SARS-CoV-2 RNA genomic regions targeted by the kit. The positive control is used to monitor for failures of rRT-PCR reagents and reaction conditions.
- Negative Control - molecular-grade, nuclease-free, non-DEPC-treated water used to monitor non-specific amplification, cross-contamination during experimental setup, and nucleic acid contamination of reagents.

Your product also requires the use of additional authorized materials and authorized ancillary reagents that are not included with your product and are described in the Instructions for Use.

The labeling entitled “Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit Instructions for Use”, “TaqPath COVID-19 High-Throughput Combo Kit Product Information Sheet” (available at <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorized-medical-devices/in-vitro-diagnostics-euas>), and the following fact sheets pertaining to the emergency use, are required to be made available as set forth in the Conditions of Authorization (Section IV) and are collectively referred to as “authorized labeling”:

- Fact Sheet for Healthcare Providers: Thermo Fisher Scientific, Inc. - Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit
- Fact Sheet for Patients: Thermo Fisher Scientific, Inc. - Amplitude Solution with the TaqPath COVID-19 High-Throughput Combo Kit

The above described product, when accompanied by the authorized labeling provided as set forth in the Conditions of Authorization (Section IV), is authorized to be distributed to and used by authorized laboratories under this EUA, despite the fact that it does not meet certain requirements otherwise required by applicable federal law.

I have concluded, pursuant to Section 564(d)(2) of the Act, that it is reasonable to believe that the known and potential benefits of your product, when used consistent with the Scope of Authorization of this letter (Section II), outweigh the known and potential risks of your product.

I have concluded, pursuant to Section 564(d)(3) of the Act, based on the totality of scientific evidence available to FDA, that it is reasonable to believe that your product may be effective in diagnosing COVID-19, when used consistent with the Scope of Authorization of this letter (Section II), pursuant to Section 564(c)(2)(A) of the Act.

FDA has reviewed the scientific information available to FDA, including the information supporting the conclusions described in Section I above, and concludes that your product (as

described in the Scope of Authorization of this letter (Section II)) meets the criteria set forth in Section 564(c) of the Act concerning safety and potential effectiveness.

The emergency use of your product under this EUA must be consistent with, and may not exceed, the terms of this letter, including the Scope of Authorization (Section II) and the Conditions of Authorization (Section IV). Subject to the terms of this EUA and under the circumstances set forth in the Secretary of HHS's determination under Section 564(b)(1)(C) of the Act described above and the Secretary of HHS's corresponding declaration under Section 564(b)(1) of the Act, your product is authorized for the indication above.

III. Waiver of Certain Requirements

I am waiving the following requirements for your product during the duration of this EUA:

- Current good manufacturing practice requirements, including the quality system requirements under 21 CFR Part 820 with respect to the design, manufacture, packaging, labeling, storage, and distribution of your product, but excluding Subpart H (Acceptance Activities, 21 CFR 820.80 and 21 CFR 820.85), Subpart I (Nonconforming Product, 21 CFR 820.90), and Subpart Q (Statistical Techniques, 21 CFR 820.250).

IV. Conditions of Authorization

Pursuant to Section 564(e) of the Act, I am establishing the following conditions on this authorization:

Thermo Fisher Scientific, Inc. (You) and Authorized Distributor(s)⁷

- A. Your product must comply with the following labeling requirements pursuant to FDA regulations: the intended use statement (21 CFR 809.10(a)(2), (b)(2)); adequate directions for use (21 U.S.C. 352(f)), (21 CFR 809.10(b)(5), (7), and (8)); appropriate limitations on the use of the device including information required under 21 CFR 809.10(a)(4); and any available information regarding performance of the device, including requirements under 21 CFR 809.10(b)(12).
- B. You and authorized distributor(s) must make your product available with the authorized labeling to authorized laboratories.
- C. You and authorized distributor(s) must make available on your website(s) the authorized labeling.
- D. You and authorized distributor(s) will include a physical copy of the “TaqPath COVID-19 High-Throughput Combo Kit Product Information Sheet” with each shipped product to authorized laboratories, and will make the “Amplitude Solution

⁷ “Authorized Distributor(s)” are identified by you, Thermo Fisher Scientific, Inc., in your EUA submission as an entity allowed to distribute your product.

with the TaqPath COVID-19 High-Throughput Combo Kit Instructions for Use” electronically available with the opportunity to request a copy in paper form, and after such request, you must promptly provide the requested information without additional cost.

- E. You and authorized distributor(s) must inform authorized laboratories and relevant public health authorities of this EUA, including the terms and conditions herein, and any updates made to your product and authorized labeling.
- F. Through a process of inventory control, you and authorized distributor(s) must maintain records of the authorized laboratories to which they distribute your product and number they distribute.
- G. You and authorized distributor(s) must collect information on the performance of your product. You will report to FDA any suspected occurrence of false positive or false negative results and significant deviations from the established performance characteristics of the product of which you become aware.
- H. You and authorized distributor(s) are authorized to make available additional information relating to the emergency use of your product that is consistent with, and does not exceed, the terms of this letter of authorization.

Thermo Fisher Scientific, Inc. (You)

- I. You must notify FDA of any authorized distributor(s) of your product, including the name, address, and phone number of any authorized distributor(s).
- J. You must provide authorized distributor(s) with a copy of this EUA and communicate to authorized distributor(s) any subsequent amendments that might be made to this EUA and its authorized accompanying materials (e.g., Fact Sheets).
- K. You may request changes to this EUA for your product, including to the Scope of Authorization (Section II in this letter) or to the authorized labeling, including requests to make available additional authorized labeling specific to an authorized distributor. Such additional labeling may use another name for the product but otherwise must be consistent with the authorized labeling, and not exceed the terms of authorization of this letter. Any request for changes to this EUA should be submitted to the Division of Microbiology (DMD)/Office of Health Technology 7 (OHT7)-Office of In Vitro Diagnostics and Radiological Health (OIR)/Office of Product Evaluation and Quality (OPEQ)/Center for Devices and Radiological Health (CDRH) and require appropriate authorization from FDA prior to implementation.
- L. You must comply with the following requirements pursuant to FDA regulations: 21 CFR 820 Subpart H (Acceptance Activities, 21 CFR 820.80 and 21 CFR 820.86), Subpart I (Nonconforming Product, 21 CFR 820.90), and Subpart O (Statistical Techniques, 21 CFR 820.250).

- M. You must have lot release procedures and the lot release procedures, including the study design and statistical power, must ensure that your product released for distribution has the clinical and analytical performance claimed in the authorized labeling.
- N. If requested by FDA, you must submit lot release procedures to FDA, including sampling protocols, testing protocols, and acceptance criteria, that you use to release lots of your product for distribution in the U.S. If such lot release procedures are requested by FDA, you must provide it within 48 hours of the request.
- O. You must evaluate the analytical limit of detection and assess traceability⁸ of your product with any FDA-recommended reference material(s). After submission to and concurrence with the data by FDA, you must update your labeling to reflect the additional testing. Such labeling updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- P. You must have a process in place to track adverse events, including any occurrence of false results with your product and report to FDA in accordance with 21 CFR Part 803.
- Q. You must evaluate the impact of SARS-CoV-2 viral mutations on your product's performance. Such evaluations must occur on an ongoing basis and must include any additional data analysis that is requested by FDA in response to any performance concerns you or FDA identify during routine evaluation. Additionally, if requested by FDA, you must submit records of these evaluations for FDA review within 48 hours of the request. If your evaluation identifies viral mutations that affect the stated expected performance of your device, you must notify FDA immediately (via email: CDRH-EUA-Reporting@fda.hhs.gov)
- R. If requested by FDA, you must update your labeling within 7 calendar days to include any additional labeling risk mitigations identified by FDA regarding the impact of viral mutations on test performance. Such updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.

Authorized Laboratories

- S. Authorized laboratories using your product must include with test result reports, all authorized Fact Sheets. Under exigent circumstances, other appropriate methods for disseminating these Fact Sheets may be used, which may include mass media.
- T. Authorized laboratories using your product must use your product as outlined in the authorized labeling. Deviations from the authorized procedures, including the authorized instruments, authorized extraction methods, authorized clinical specimen types, authorized control materials, authorized other ancillary reagents and authorized materials required to use your product are not permitted.

⁸ Traceability refers to tracing analytical sensitivity/reactivity back to an FDA-recommended reference material.

- U. Authorized laboratories that receive your product must notify the relevant public health authorities of their intent to run your product prior to initiating testing.
- V. Authorized laboratories using your product must have a process in place for reporting test results to healthcare providers and relevant public health authorities, as appropriate.
- W. Authorized laboratories must collect information on the performance of your product and report to DMD/OHT7-OIR/OPEQ/CDRH (via email: CDRH-EUA-Reporting@fda.hhs.gov) and you (techservices@thermofisher.com; 1 800 955 6288) any suspected occurrence of false positive or false negative results and significant deviations from the established performance characteristics of your product of which they become aware.
- X. All laboratory personnel using your product must be appropriately trained in RT-PCR techniques and use appropriate laboratory and personal protective equipment when handling this kit, and use your product in accordance with the authorized labeling.

Thermo Fisher Scientific, Inc. (You), Authorized Distributor(s) and Authorized Laboratories

- Y. You, authorized distributor(s), and authorized laboratories using your product must ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records must be made available to FDA for inspection upon request.

Conditions Related to Printed Materials, Advertising and Promotion

- Z. All descriptive printed matter, advertising, and promotional materials relating to the use of your product shall be consistent with the authorized labeling, as well as the terms set forth in this EUA and meet the requirements set forth in section 502(a), (q)(1), and (r) of the Act, as applicable, and FDA implementing regulations.
- AA. No descriptive printed matter, advertising, or promotional materials relating to the use of your product may represent or suggest that this test is safe or effective for the detection of SARS-CoV-2.
- BB. All descriptive printed matter, advertising, and promotional materials relating to the use of your product shall clearly and conspicuously state that:
 - This product has not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories;
 - This product has been authorized only for the detection of nucleic acid from SARS-CoV-2, not for any other viruses or pathogens; and
 - The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use

of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

The emergency use of your product as described in this letter of authorization must comply with the conditions and all other terms of this authorization.

V. Duration of Authorization

This EUA will be effective until the declaration that circumstances exist justifying the authorization of the emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 is terminated under Section 564(b)(2) of the Act or the EUA is revoked under Section 564(g) of the Act.

Sincerely,


Jacqueline A. O'Shaughnessy, Ph.D.
Acting Chief Scientist
Food and Drug Administration

Enclosure