

In sample pooling, nasal swab samples collected with either the Amazon COVID-19 Test Collection Kit DTC, Amazon COVID-19 Collection Kit or Amazon COVID-19 Test Collection Kit may be pooled with a maximum pooling ratio of 5:1 (5 samples for 1 test).¹ Individual samples are retested if results for pooled samples are positive for the presence of SARS-CoV-2 RNA or there is a test error (e.g., an invalid result).

An authorized Fact Sheet for Healthcare Providers is also available for download via the test report on the AmazonDx.com website and may be used by individuals to facilitate appropriate follow-up with their healthcare provider.

Table 2. Result reports displayed to individuals via the AmazonDx.com website

Reported Result	Explanation
SARS-CoV-2 Positive	<p>Your test result was POSITIVE, which means that the virus that causes COVID-19 was found in your sample.</p> <p>If you have a positive result, you should follow up with your healthcare provider who will work with you to determine how best to care for you based on the test results along with medical history and your symptoms. You should follow the latest CDC guidance to avoid spreading the virus to others, such as self-isolation to reduce the potential transmission of disease. There is still possibility that this test can give a positive result that is wrong (a false positive result).</p> <p>Please see the Fact Sheet for Individuals for more information about your test result.</p> <p>If needed you can provide the Fact Sheet for Healthcare Providers to your healthcare provider.</p>

¹ Pooling may be performed with a mixture of samples collected using the prescription use and non-prescription (DTC) collection devices that are authorized for use with the Amazon Test/Amazon DTC Test

