



July 28, 2022

Randy J. Prebula  
Partner  
Hogan Lovells US LLP  
555 Thirteen Street NW  
Washington, DC 20004  
Representing - Lucira Health, Inc.<sup>1</sup>

Re: EUA202920-S005  
Trade/Device Name: Lucira COVID-19 All-In-One Test Kit  
Dated: May 3, 2022  
Received: May 3, 2022

Dear Mr. Prebula:

This is to notify you that your request to update the device shelf-life stability claim from 12 months to 18 months when stored at  $30\pm2^\circ\text{C}$  is granted. Upon review, we concur that the data and information submitted in EUA202920-S005 supports the requested updates for use with the Lucira COVID-19 All-In-One Test Kit. FDA has updated the Fact Sheet for Healthcare Providers to reflect language used in more recent authorizations. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the Lucira COVID-19 All-In-One Test Kit issued on November 17, 2020.

Sincerely yours,

Uwe Scherf, M.Sc., Ph.D.  
Director, Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health

<sup>1</sup> Cc: David McGrath, John Beasley, and Ghazi Kashmolah from Lucira Health, Inc.  
1412 62nd Street  
Emeryville, CA 94608  
U.S. Food & Drug Administration  
10903 New Hampshire Avenue  
Silver Spring, MD 20903  
[www.fda.gov](http://www.fda.gov)