

Scope of Work

Laboratory Refrigerators

Background

The Microbiology section at the Food and Drug Administration (FDA) Northeast Food and Feed Laboratory is a facility designed specifically for testing samples regulatory samples for contamination. Processing these samples requires storage at refrigerator temperature. We need an additional refrigerator to meet our demands and replacing out of service and aging equipment.

Purpose

We are requesting this refrigerator specifically for storage of in progress samples and associated media and plates used in analyzing the sample. The additional refrigerator is meant to meet the growing demands of the laboratory.

Scope

The Contractor shall provide all resources necessary to accomplish the tasks and deliverables described in this Statement of Work (SOW). The Contractor shall provide a floor model, glass-door refrigerator.

I. Requirements

Minimum Technical Requirements:

The refrigerator:

- Shall have a temperature range which includes +2 to + 8°C
- Shall meet the electrical requirements of 115 V 60 Hz
- Shall have a capacity of at least 42 Cu Ft
- Shall have a key lock access (two keys are required)
- Shall be at least 30 inches deep and 60 inches wide
- Shall feature two glass doors of either the swing-type or sliding-type
- Shall have a cord length of at least eight feet

Trade and Service Specifications

1. The instrument must be a newly manufactured unit, not used and refurbished or previously used for demonstration.
2. FOB Point destination to include delivery to NFFL loading dock.
3. The entire system must be warranted for parts and labor for 12 months from the date of formal government acceptance. The vendor must also be capable of servicing the instrument through the covered warranty period. The system must include at least a one (1) year warranty and shall include at a minimum: coverage

on all non-consumable items and parts supplied including base instrument, factory-certified replacement parts, engineer labor and travel costs.

Deliverables

Table 1. Deliverables / Schedule

Refrigerator	3	Not later than September 1, 2022

Shipping Destination

POC: To Be Identified at time of award
Food and Drug Administration,
Northeast Food & Feed Laboratory
158-15 Liberty Avenue
Jamaica, NY 11433-1034

The delivery or services must be during regular business hours (Monday-Friday) during the times of 8:00 AM – 4:00 PM, excluding holidays.

Period of Performance

The period of performance begins the date of contract award execution and continues through the OEM warranty period (minimum of one year). The warranty period begins the date of formal Government acceptable. The equipment is to be delivered no later than 9/1/2022.

The anticipated Period of Performance is as follows:

Base Period: 09/01/2022 – 08/31/2023

Quotation Instructions

All quotes are due by e-mail to Nina Montgomery, Nina.Montgomery@fda.hhs.gov on or before August 9, 2022 at 10:00 am (Eastern Standard Time).