



August 29, 2022

Sung Jang  
Manager, Regulatory Affairs Division  
Access Bio, Inc.  
65 Clyde Road, Suite A  
Somerset, NJ 08873

Re: EUA210314/S011  
Trade/Device Name: *CareStart* COVID-19 Antigen Home Test  
Dated: July 12, 2022  
Received: July 12, 2022

Dear Sung Jang:

This is to notify you that your request to update the *CareStart* COVID-19 Antigen Home Test to extend the shelf-life expiration date to 15 months when stored at 1°C – 30°C, based on the results of your ongoing stability studies, is granted. Upon review, we concur that the data and information submitted in EUA210314/S011 support the requested update for the *CareStart* COVID-19 Antigen Home Test. By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the *CareStart* COVID-19 Antigen Home Test re-issued on November 22, 2021.

Sincerely yours,

---

Uwe Scherf, M.Sc., Ph.D.  
Director, Division of Microbiology Devices  
OHT7: Office of In Vitro Diagnostics and Radiological Health  
Office of Product Evaluation and Quality  
Center for Devices and Radiological Health