

***Rize Laboratory SARS nCoV-2019 Multiplexed Assay***  
***Test Summary: 11/2/2022***

**EMERGENCY USE AUTHORIZATION (EUA) TEST SUMMARY FOR THE**  
***Rize Laboratory - SARS nCoV-2019 Multiplexed Assay***

For *In vitro* Diagnostic Use

Rx Only

For use under Emergency Use Authorization (EUA) only

**The Rize Laboratory SARS nCoV-2019 Multiplexed Assay is an LDT that will be performed at the Rize Laboratory located at 545 SW 2nd Street, Suite 201, Corvallis, OR 97333, which is certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, and meets the requirements to perform high complexity tests.**

**INTENDED USE**

The Rize Laboratory SARS nCoV-2019 Multiplexed Assay is intended for the *in vitro* qualitative detection of RNA from SARS-CoV-2 in individual human anterior nasal swab samples collected by a healthcare provider (HCP) or self-collected under the supervision of an HCP from individuals, including individuals without symptoms or other reasons to suspect COVID-19, when tested at least once per week.

Testing is limited to Rize Laboratory located at 545 SW 2nd Street, Suite 201 Corvallis, OR 97333, which is certified under Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, and meets the requirements to perform high-complexity testing.

The Rize Laboratory SARS nCoV-2019 Multiplexed Assay is intended for use by qualified and trained clinical laboratory personnel specifically instructed and trained in the techniques of real-time PCR and *in vitro* diagnostic procedures. The Rize Laboratory SARS nCoV-2019 Multiplexed Assay is only for use under the Food and Drug Administration's Emergency Use Authorization.

Results are for the detection and identification of SARS-CoV-2 RNA. The SARS-CoV-2 nucleic acid is generally detectable in anterior nasal swab specimens during the acute phase of infection. Positive results are indicative of the presence of SARS-CoV-2 RNA; clinical correlation with patient history and other diagnostic information is necessary to determine patient infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease.

Negative results do not preclude SARS-CoV-2 infection and should not be used as the sole basis for patient management decisions. Negative results must be combined with clinical observations, patient history, and epidemiological information.

Laboratories within the United States and its territories are required to report all results to the appropriate public health authorities.

***Rize Laboratory SARS nCoV-2019 Multiplexed Assay***  
***Test Summary: 11/2/2022***

For serial testing programs, additional confirmatory testing for negative results may be necessary, if there is a high likelihood of COVID-19, such as an individual with a close contact with COVID-19 or with suspected exposure to COVID-19 or in communities with high prevalence of infection. Additional confirmatory testing for positive results may also be necessary, if there is a low likelihood of COVID-19, such as in individuals without known exposure to COVID-19 or residing in communities with low prevalence of infection.

1) **Special Conditions for Use Statements:**

For prescription use only

For *in vitro* diagnostic use

For Emergency Use Authorization (EUA) only

This test is authorized under the Umbrella EUA for SARS-CoV-2 Molecular Diagnostic Tests for Serial Testing (<https://www.fda.gov/media/154111/download>) for use in Rize Laboratory, that is certified under CLIA and meets requirements to perform high complexity tests, in which it was developed for qualitative detection of RNA from SARS-CoV-2 in individual human anterior nasal swab samples collected by a healthcare provider (HCP) or self-collected under the supervision of an HCP from individuals, including individuals without symptoms or other reasons to suspect COVID-19, when tested at least once per week using the test procedures validated in accordance with the requirements of the Umbrella EUA for SARS-CoV-2 Molecular Diagnostic Tests for Serial Testing.

#### **DEVICE DESCRIPTION AND TEST PRINCIPLE**

The Rize Laboratory SARS nCoV-2019 Multiplexed Assay is a reverse transcription polymerase chain reaction (RT -PCR) test. The SARS-CoV-2 primer and probe set(s) is designed to detect RNA from SARS-CoV-2 in anterior nasal swab specimens that were collected from individuals, including individuals without symptoms or other reasons to suspect COVID-19.

**Test Procedure:** Nucleic acids are isolated and purified from anterior nasal swabs using a nucleic acid extraction system. The purified nucleic acid is reverse transcribed into cDNA and amplified in one step by combining purified nucleic acid with the RT-qPCR master mix and qPCR primer & probes. In the process, the probe anneals to a specific N1 and N2 target sequence located between the forward and reverse primers. During the extension phase of the PCR cycle, the 5' nuclease activity of Taq polymerase degrades the probe, causing the reporter dye to separate from the quencher dye, generating a fluorescent signal. With each cycle, additional reporter dye molecules are cleaved from their respective probes, increasing the fluorescence intensity. Fluorescence intensity is monitored at each PCR cycle for each target for 45 cycles using Fam to detect the N1 gene, VIC to detect the N2 gene, and Rox to detect RNASE P.

*Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022*

**INSTRUMENTS USED WITH TEST**

**Instruments**

The Rize Laboratory SARS nCoV-2019 Multiplexed Assay test is to be used with the ABI 7500 Fast Thermocycler and the Kingfisher Flex.

**Reagents**

The primary reagents used in Rize Laboratory SARS nCoV-2019 Multiplexed Assay:

2x RT-qPCR OneStep Amplification Kit (catalog # RTQAK), containing 2x InhibiTaq Master Mix and Direct RT Mix. Primers and probes were ordered from LGC Biosearch Technologies or Integrated DNA Technologies and custom premixed by Empirical into a 20x working concentration (catalog #CNRDK).

Sample Type	Expiration
Swab Media	5 days or 120 hours

Swab Media	Top Color	Volume	Media Color
MicroTest M4RT	Orange	3 mL	Orange
Mantacc VTM	Red	10 mL	Red
Resolution Biomedical	White	5 mL	Clear
Pacific Biosupply	Red	3 mL	Clear

**CONTROL MATERIAL(s) TO BE USED WITH RIZE LABORATORY SARS nCoV-2019 MULTIPLEX ASSAY:**

Controls that are used with the test include:

- a) A **negative extraction control (NEC)** is needed to determine that the extraction of patient samples was completed successfully and is multiplexed in with every sample's results. This is nuclease free water that is used in place of a patient sample. Amplification of this extraction control from any of the three targets tested indicates contamination of the extraction step and all samples are re-extracted.

*Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022*

- b) A **positive template control (PTC)** is needed to determine the viability of the run to show potential positive patients. A positive plasmid control provided by the manufacturer is used at approximately 2X LOD and is included at the assay step.
  
- c) A **“no template” (negative) control (NTC)** is needed to detect contamination in the assay plating process. A NTC is included in the PCR plating step and analysis. Amplification of this NTC from any of the three targets tested indicates contamination on the PCR plate and the plate is rerun after the extraction step.

## **INTERPRETATION OF RESULTS**

All test controls must be examined prior to interpretation of patient results. If the controls are not valid, the patient results cannot be interpreted. Appropriate control interpretation criteria and result interpretation criteria are described here.

### **1. Examination and Interpretation of Control Results**

The following result evaluation rules apply to the controls of every nCoV-2019 plate run:

- **Negative Extraction Control (NEC)**
  - Expected result: No amplification of any target
  - If amplification is <40 for any target, all samples must be re-extracted.
  
- **Positive Template Control (PTC)**
  - Expected Result: Amplification of RNase P, N1, and N2
  - If no amplification, or amplification is > 40, the entire plate must be re-plated
  
- **No Template Control (NTC)**
  - Expected result: No amplification of any target
  - If amplification is <40 for any target, all elution plates must be re-plated

### **2. Examination and Interpretation of Patient Sample Results:**

Assessment of clinical specimen test results must be performed after the controls have been examined and determined to be valid and acceptable. If the controls are not valid, the patient results cannot be interpreted.

Every patient’s curve resulting from the qPCR process is checked by a member of the molecular staff. Samples that are <40 Ct for N1 and N2 are considered positive. Samples that have a Ct above 40 for N1 and N2 are considered negative. Samples that do not show amplification of N1 or N2 are considered negative. Curves that are not exponential or potential false positive results are re-extracted and rerun. Samples that do not show amplification for RNaseP are re-extracted and rerun. See the below chart for the result evaluation rule.

**Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022**

N1	N2	RP	Result Interpretation
+	+	+/-	Positive Sars-Cov-2
Only 1 of 2 targets positive		+/-	Inconclusive*
-	-	+	Not Detected
-	-	-	Invalid**

\* Inconclusive results should be repeated from extraction. If the result is still inconclusive the second time a new specimen should be collected if available. If it is not possible to collect a new specimen, it should be reported to the healthcare provider as inconclusive.

\*\* Invalid results should be repeated from extraction. If the result is still invalid the second time a new specimen should be collected if available. If it is not possible to collect a new specimen, it should be reported to the healthcare provider as invalid.

**PERFORMANCE EVALUATION**

1) **Limit of Detection (LoD) - Analytical Sensitivity:**

The LoD for the Rize Laboratory SARS nCoV-2019 Multiplexed Assay was evaluated and verified using ATCC VR-1986HK Heat inactivated nCoV-2019 Virus per the validation required by Appendix A of the Umbrella EUA for SARS-CoV-2 Molecular Diagnostic Tests for Serial Testing. Nucleic acid was extracted from the swabs using Applied Biosystems MagMAX™ Viral/Pathogen Nucleic Acid Isolation Kit and the reverse transcription RT-PCR was performed using the 7500 Fast Real-time PCR Thermocycler. Preliminary and Confirmation LoD results are included in the tables below.

Virus cp/uL	N1 Avg Ct	N2 Avg Ct	RP Avg Ct	Replicates
20	33.98	34.63	33.07	3
10	34.42	35.07	32.15	3
5	34.15	34.81	31.43	3
3	35.76	36.01	33.49	3
2	36.56	36.82	33.94	3
1	NA	NA	25.30	3
0.5	NA	43.20	27.13	3

***LoD Confirmation:***

**Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022**

Target	N1	N2
Virus Concentration	2 cp/uL	2 cp/uL
Positives/Total	18/20	19/20
% Detected	90%	95%
Mean Ct	38.21	37.48
Mean SD	0.59	1.05
CV	1.5%	2.8%

Data confirmed that LoD is **2 copies/uL**.

2) **Inclusivity (analytical reactivity):**

An alignment was performed with the oligonucleotide primer and probe sequences of the Rize Laboratory SARS nCoV-2019 Multiplexed Assay with 831,910 Global Genome Sequences and 295,775 U.S. Sequences publicly available SARS-CoV-2 sequences (including mutation variants of high prevalence, i.e., B.1.617.2 and sub-lineages on August 1<sup>st</sup>, 2022) from GISAID to demonstrate the predicted inclusivity of the assay.

Location	N1 Forward		N1 Rev.		N1 Probe								
	4	5	9	14	15	2	3	4	5	13	18	22	24
Mismatch Nucleotide	C>T	C>T	A>G	G>T	G>T	C>T	C>T	C>T	C>T	G>T	G>T	A>G	C>T
Global Sequence Mismatch Frequency (%)	0.11	0.14	0.39	0.7	0.05	0.03	1.55	0.24	0.29	13*	0.1	0.091	0.05
US Sequence Mismatch Frequency (%)	0.04	0.18	0.35	0.18	0.14	0.07	3.69	0.39	0.37	0.08	0.13	0.24	0.11

\*The only significantly prevalent mutation (N1 probe nucleotide 13) should not change the ability of the N1 probe to anneal to the target gene. We used an open source calculator to determine how the change affects the annealing temperature of the probe and it was minimally affected (less than 2 degrees C).

**Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022**

	N2 Forward			N2 Reverse		N2 Probe		
Location	4	8	16	4	13	7	10	13
Mismatch Nucleotide	C>T	C>T	G>T	C>A	G>A	T>C	C>T	C>T
Global Sequence Mismatch Frequency (%)	0.12	0.21	0.28	0.24	0.12	0.13	1.64	0.19
US Sequence Mismatch Frequency (%)	0.1	0.3	0.08	0.43	0.13	0.19	0.43	0.13

**3) Cross-reactivity (Analytical Specificity):**

Analytical specificity of the primer/probe combination for Rize Laboratory SARS nCoV-2019 Multiplexed Assay was evaluated by conducting sequence alignment of the primer/probe sequences of the test with publicly available genome sequences for potential cross-reacting microorganisms. The following organisms were tested with Rize Laboratory SARS nCoV-2019 Multiplexed Assay primer probe set and yielded no cross-reactivity:

Pathogen	Strain	N1 Target	N2 Target
Coronavirus	HKU1	Negative	Negative
Coronavirus	NL63	Negative	Negative
Coronavirus	OC43	Negative	Negative
Coronavirus	229E	Negative	Negative
Influenza A	H1-2009	Negative	Negative
Influenza A	A/H3	Negative	Negative
Influenza B	-	Negative	Negative
Parainfluenza 3	-	Negative	Negative
Rhinovirus	-	Negative	Negative

**Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022**

RSV A/B	-	Negative	Negative
---------	---	----------	----------

**4) Clinical Evaluation:**

Clinical evaluation of the Rize Laboratory SARS nCoV-2019 Multiplexed Assay was conducted with 60 individual anterior nasal swab clinical specimens collected from patients suspected of SARS-CoV-2 infection by a healthcare provider in COVID-19 disease endemic region(s). This study was performed using retrospective samples.

Nucleic acid was extracted from all specimens using the Applied Biosystems MagMAX™ Viral/Pathogen Nucleic Acid Isolation Kit on the Kingfisher Flex and reverse transcription Real-time-PCR was performed using the Empirical reagents on the 7500 Fast Thermocycler. Clinical results yielded 29 positives for SARS-CoV-2 and 31 negatives for SARS-CoV-2.

These same 60 clinical anterior nasal swab specimens were subsequently run on the FDA Emergency Use Authorized Thermo Fisher TaqPath Combo Kit results were compared.

**Data for the correlation study with the FDA Emergency Use Authorized Thermo Fisher TaqPath Combo Kit is summarized in the Table below:**

RIZE LAB Multiplex Test	TaqPath Combo Kit		Agreement		Performance
	Positive	Negative			
Positive	29	0	PPA = 29/30	97%	Sensitivity
Negative	1	30	NPA = 30/30	100%	Specificity

**LIMITATIONS**

The initial performance of this test was established based on the evaluation of a limited number of clinical specimens collected between April 2021 and August 2021 and were collected from various collection sites throughout the state of Oregon within the United States.

The clinical performance of this test has not been established in all circulating variants but is anticipated to be reflective of the variants in circulation at the time and location(s) of the clinical evaluation. As such, performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARS-CoV-2, and their prevalence, which change over time.

Clinical performance has been established in specimens collected from subjects suspected of COVID-19 by a healthcare provider. Performance of specimens collected from individuals without symptoms or other reasons to suspect COVID-19 has not been established. A study to determine the performance in individuals without symptoms or other reasons to suspect COVID-19 will be completed.



*Rize Laboratory SARS nCoV-2019 Multiplexed Assay  
Test Summary: 11/2/2022*

**WARNINGS:**

This product has not been FDA cleared or approved but has been authorized by FDA under an Emergency Use Authorization (EUA) for use by the laboratory that developed the test and which is certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, and meets the requirements to perform high complexity tests.

This product has been authorized only for the detection of nucleic acid from SARS-CoV-2, not for any other viruses or pathogens; and

The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated, or authorization is revoked sooner.