



Our STN: BL 125394/18

**SUPPLEMENT APPROVAL
February 17, 2023**

AVIOQ, Inc.
Attention: Linda D. Garner
76 T.W. Alexander Drive, Building A
Research Triangle Park, NC 27709

Dear Linda Garner:

We have approved your request received December 27, 2022, to supplement your Biologics License Application (BLA) submitted under section 351(a) of the Public Health Service Act for Human T-Lymphotropic Virus Types I & II (HTLV-I and HTLV-II/Enzyme Immuno Assay (EIA)/Lysate).

We hereby approve the draft Kit Component Labeling, Carton Labeling, and the Instructions for Use (IFU) for US-only Avioq HTLV I/II EIA/Lysate kit, submitted under amendment #2, dated February 10, 2023, and amendment #5 dated February 15, 2023. This is a reminder that as of September 24, 2014, medical devices that are licensed under the PHS Act are subject to certain provisions of the final Unique Device Identifier (UDI) rule. These provisions include the requirement to provide a UDI on the device label and packages (21 CFR 801.20), format dates on the device label in accordance with 21 CFR 801.18, and submit data to the Global Unique Device Identification Database (GUDID) (21 CFR 830 Subpart E). Additionally, please identify each device identifier implemented for the subject device, and the device identifiers that have been discontinued for the subject device as a labeling change in an annual report consistent with 21 CFR 601.12(f)(3). For more information on these requirements, please see the UDI website at <http://www.fda.gov/udi>.

Please submit all final printed labeling as PDF electronic copy (eCopy) at the time of use and include implementation information on Form FDA 356h as appropriate.

Two draft copies of the proposed introductory advertising or promotional labeling may be voluntarily submitted for advisory comment with a completed Form FDA 2253 to the Advertising and Promotional Labeling Branch at the following address:

Food and Drug Administration
Center for Biologics Evaluation and Research
Document Control Center
10903 New Hampshire Ave.
WO71-G112
Silver Spring, MD 20993-0002

Submissions may also be submitted electronically via email (under 150MB) at CBERDCC_eMailSub@fda.hhs.gov.

We will include the information contained in the above-referenced supplement in your BLA file.

Sincerely,

Hira L. Nakhasi, PhD
Director
Division of Emerging and
Transfusion Transmitted Diseases
Office of Blood Research and Review
Center for Biologics Evaluation and Research