



June 26, 2023

GPI USA, Inc.  
Attention: Kimberly Jimenez  
4900 Woodway Drive, Suite 680  
Houston, TX 77056

Re: BK220791  
Trade/Device Name: e-Delphyn Donor  
Regulation Number: 21 CFR 864.9165  
Regulation Name: Blood establishment computer software and accessories  
Regulatory Class: Class II  
Product Code: MMH  
Dated: November 29, 2022  
Received: December 6, 2022

Dear Kimberly Jimenez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act, or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and

listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Orieji Illoh, MD  
Director  
Division of Blood Components and Devices  
Office of Blood Research and Review  
Center for Biologics Evaluation and Research

Enclosure: Indications for Use

**Indications for Use**

**510(k) Number:** BK220791

**Device Name:** e-Delphyn Donor

**Indications for Use:**

e-Delphyn Donor is a web enabled modular computerized system intended to be used by trained personnel to aid in donor management. The functionalities include:

- Registering, deferring, merging, and qualifying donors.
- Registering bags.
- Recording the outcomes of the Donor History Questionnaire (DHQ).
- Recording medical screening data and phlebotomy information.
- Entering and evaluating test results in accordance with the decision rules.
- Processing units in components.
- Phenotype RBC/HLA searching and matching.
- Managing storage, inventory, and distribution of units.
- Managing therapeutic and autologous/directed donations requests.

Prescription Use   X   AND/OR  
(Part 21 CFR 801 Subpart D)

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CBER, Office of Blood Research and Review (OBRR)

---

Division Sign-Off, Office of Blood Research and Review