

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

# Select Updates for the Premarket Cybersecurity Guidance: Section 524B of the FD&C Act

---

## Draft Guidance for Industry and Food and Drug Administration Staff

***DRAFT GUIDANCE***

**This draft guidance document is being distributed for comment purposes only.**

**Document issued on March 13, 2024.**

You should submit comments and suggestions regarding this draft document within 60 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD 20852-1740. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions about this document regarding CDRH-regulated devices, contact [CDRHManufacturerShortage@fda.hhs.gov](mailto:CDRHManufacturerShortage@fda.hhs.gov). For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-8010, or by email at [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov).

**When final, this guidance will supersede “Cybersecurity in Medical Devices: Quality System Considerations and Content of Premarket Submissions,” issued September 27, 2023.**



U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Devices and Radiological Health  
Center for Biologics Evaluation and Research

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

## **Preface**

### **Additional Copies**

#### **CDRH**

Additional copies are available from the Internet. You may also send an email request to [CDRH-Guidance@fda.hhs.gov](mailto:CDRH-Guidance@fda.hhs.gov) to receive a copy of the guidance. Please include the document number GUI00001825 and complete title of the guidance in the request.

#### **CBER**

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov), or from the Internet at <https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>.

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

**Table of Contents**

I. Overview of Select Updates..... 1

II. Cyber Devices..... 2

    A. Who is Required to Comply with Section 524B of the FD&C Act ..... 2

    B. Devices Subject to Section 524B of the FD&C Act ..... 2

    C. Documentation Recommendations to Comply with 524B..... 3

        1. Plans and Procedures (Section 524B(b)(1)) ..... 4

        2. Design, Develop, and Maintain Processes and Procedures to Provide a Reasonable Assurance of Cybersecurity (Section 524B(b)(2)) ..... 5

        3. Software Bill of Materials (SBOM) (Section 524B(b)(3))..... 5

    D. Modifications ..... 5

        1. Changes That May Impact Cybersecurity ..... 6

        2. Changes Unlikely to Impact Cybersecurity ..... 6

    E. Reasonable Assurance of Cybersecurity of Cyber Devices..... 7

# Select Updates for the Premarket Cybersecurity Guidance: Section 524B of the FD&C Act

## Draft Guidance for Industry and Food and Drug Administration Staff

*This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.*

### I. Overview of Select Updates

The Food and Drug Administration (FDA or Agency) has developed this draft guidance to propose select updates to the FDA guidance “[Cybersecurity in Medical Devices: Quality System Considerations and Content of Premarket Submissions](#)” (hereafter referred to as the “Premarket Cybersecurity Guidance”).<sup>1</sup> The Premarket Cybersecurity Guidance, in its current form, remains the Agency’s current thinking on this topic until this draft guidance is finalized, at which time the finalized version of Section II. of this draft guidance will be added as Section VII. of the Premarket Cybersecurity Guidance. FDA intends to incorporate the updates proposed in this draft guidance into the Premarket Cybersecurity Guidance as one final guidance document after obtaining and considering public comment on these proposed select updates. The sections of the existing Premarket Cybersecurity Guidance that are unaffected by these proposed updates are not intended to be substantively changed, with the exception of technical edits for consistency.

Section 3305 of the Food and Drug Omnibus Reform Act of 2022 (“FDORA”), enacted on December 29, 2022, added section 524B “Ensuring Cybersecurity of Medical Devices” to the FD&C Act. Under section 524B(a) of the FD&C Act, a person who submits a 510(k), Premarket Approval Application (PMA), Product Development Protocol (PDP), De Novo, or Humanitarian Device Exemption (HDE) submission for a device that meets the definition of a “cyber device,” as defined under section 524B(c) of the FD&C Act, is required to submit information to ensure that cyber devices meet the cybersecurity requirements under section 524B(b) of the FD&C Act.

<sup>1</sup> Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/cybersecurity-medical-devices-quality-system-considerations-and-content-premarket-submissions>

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

35  
36 In general, FDA’s guidance documents do not establish legally enforceable responsibilities.  
37 Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only  
38 as recommendations, unless specific regulatory or statutory requirements are cited. The use of  
39 the word *should* in Agency guidances means that something is suggested or recommended, but  
40 not required.  
41

## 42 **II. Cyber Devices**

43 FDA is proposing to add a Section VII. to the [Premarket Cybersecurity Guidance](#) with the  
44 following language in this Section II. This section identifies the cybersecurity information FDA  
45 considers to generally be necessary to support obligations under section 524B of the FD&C Act.  
46

### 47 **A. Who is Required to Comply with Section 524B of the** 48 **FD&C Act**

49 Under section 524B(a) of the FD&C Act, a person, including a manufacturer,<sup>2</sup> who submits a  
50 premarket application or submission under any of the following pathways 510(k),<sup>3</sup> PMA,<sup>4</sup> PDP,  
51 De Novo, or HDE<sup>5</sup> for a device that meets the definition of a “cyber device,” as defined in  
52 section 524B(c) of the FD&C Act, is required to include such information as FDA may require  
53 to ensure that the cyber device meets the cybersecurity requirements under section 524B(b) of  
54 the FD&C Act.  
55

### 56 **B. Devices Subject to Section 524B of the FD&C Act**

57 Section 524B of the FD&C Act and its requirements apply to “cyber devices.” Section 524B(c)  
58 of the FD&C Act defines a “cyber device” as a device that “(1) includes software validated,  
59 installed, or authorized by the sponsor as a device or in a device; (2) has the ability to connect to  
60 the internet; and (3) contains any such technological characteristics validated, installed, or  
61 authorized by the sponsor that could be vulnerable to cybersecurity threats.”  
62

63 Informed in part by the definitions recognized by the National Institute for Standards and  
64 Technology (NIST) for the term “software,” FDA considers a “cyber device” to include devices

---

<sup>2</sup> Section 524B(a) of the FD&C Act places obligations on the “person” who submits a specific type of device marketing application. Section 524B(b) of the FD&C Act places obligations on a “sponsor.” For the purposes of this guidance, we assume that the manufacturer is the entity submitting the application and use the term accordingly throughout the guidance in lieu of the term “person” or “sponsor.” However, if another person submits the application or submission enumerated under section 524B(a) of the FD&C Act to the Agency, that person should follow the guidance for manufacturers herein. Whatever person submits the application for a cyber device is subject to the requirements of section 524B.

<sup>3</sup> For the purposes of this guidance “510(k)” refers to the original, special, and abbreviated 510(k) applications.

<sup>4</sup> For the purposes of this guidance “PMA” refers to the original PMA and supplement PMAs.

<sup>5</sup> For the purposes of this guidance “HDE” refers to the original HDE and supplement HDEs.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

65 that are or contain software, including software that is firmware or programmable logic.<sup>6</sup> FDA  
66 also considers the “ability to connect to the internet” to include devices that are able to connect  
67 to the internet, whether intentionally or unintentionally, through any means (including at any  
68 point identified in the evaluation of the threat surface<sup>7</sup> of the device and the environment of use).  
69 It is well-demonstrated that if a device has the ability to connect to the Internet, it is possible that  
70 it can be connected to the Internet, regardless of whether such connectivity was intended by the  
71 device sponsor.<sup>8</sup>

72  
73 FDA considers devices that include any of the following features to have the ability to connect to  
74 the internet. The list below is illustrative, not exhaustive:  
75

- 76 • Wi-Fi or cellular;
  - 77 • Network, server, or Cloud Service Provider connections;
  - 78 • Bluetooth or Bluetooth Low Energy;
  - 79 • Radiofrequency communications;
  - 80 • Inductive communications; and
  - 81 • Hardware connectors capable of connecting to the internet (e.g., USB, ethernet,  
82 serial port).
- 83

## **C. Documentation Recommendations to Comply with 524B**

84 For applicable premarket submission types, manufacturers must provide documentation to  
85 comply with the requirements under section 524B of the FD&C Act. Recommendations  
86 regarding the documentation to support each of the requirements is discussed in the sections  
87 below.  
88  
89  
90  
91

---

<sup>6</sup> NIST defines a programmable logic controller (PLC) as “[a] solid-state control system that has a user-programmable memory for storing instructions for the purpose of implementing specific functions such as I/O control, logic, timing, counting, three mode (PID) control, communication, arithmetic, and data and file processing.” A PLC is therefore a combination of two components: (1) the hardware controller, and (2) the “user-programmable memory,” or programmable logic, that instructs the hardware controller to execute specified functions. NIST defines software as, among other things, “computer programs and data stored in hardware – typically in read only memory or programmable read-only memory.” Programmable logic is therefore a specific type of computer program and/or data stored on hardware, and is thus a type of software. See <https://csrc.nist.gov/glossary> for more information on NIST’s definitions of these terms.

<sup>7</sup> For the purposes of this guidance, “threat surface” means the set of points on the boundary of a system, a system element, or an environment where a cyber threat can try to enter, cause an effect on, or extract data from, that system, system element, or environment” (definition is adapted from <https://csrc.nist.gov/glossary/>). For the purposes of this guidance “threat surface” is synonymous with the term “attack surface,” however, FDA uses the term “threat surface” rather than “attack surface,” because cyber threats need not necessarily be an “attack” to pose a risk to a medical device and its related system.

<sup>8</sup> See <https://www.trustdimension.com/wp-content/uploads/2015/02/MedJack.4-ilovepdf-compressed.pdf>; <https://www.forbes.com/sites/thomasbrewster/2017/05/17/wannacry-ransomware-hit-real-medical-devices/?sh=43561087425c>

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

92  
93  
94  
95  
96  
97  
98  
99  
100  
101  
102  
103  
104  
105  
106  
107  
108  
109  
110  
111  
112  
113  
114  
115  
116  
117  
118  
119  
120  
121  
122  
123  
124  
125  
126  
127  
128  
129

### **1. Plans and Procedures (Section 524B(b)(1))**

Section 524B(b)(1) of the FD&C Act requires manufacturers of cyber devices to submit to FDA “a plan to monitor, identify, and address, as appropriate, in a reasonable time, postmarket cybersecurity vulnerabilities and exploits, including coordinated vulnerability disclosure and related procedures” in their premarket submissions. We recommend that the plan contain the information recommended for the Cybersecurity Management Plan described in Section VI.B. of the [Premarket Cybersecurity Guidance](#). Additionally, such a plan should also address the additional items discussed below.

First, FDA considers that coordinated vulnerability disclosure (CVD) and related procedures could include:

- Coordinated disclosure of vulnerabilities and exploits identified by external entities (including third-party software suppliers and researchers);
- Disclosure of vulnerabilities and exploits identified by the manufacturer of cyber devices; and
- Manufacturer procedures to carry out disclosures of the vulnerabilities and exploits, as identified above.<sup>9</sup>

Second, plans required by section 524B(b)(1) of the FD&C Act should also describe the timeline, with associated justifications, to develop and release required updates and patches:

- Section 524B(b)(2)(A) of the FD&C Act requires manufacturers of cyber devices to make updates and patches for known unacceptable vulnerabilities available on a reasonably justified regular cycle.
- Section 524B(b)(2)(B) of the FD&C Act requires manufacturers of cyber devices to make available updates and patches to the device and related systems<sup>10</sup> to address as soon as possible out of cycle, critical vulnerabilities that could cause uncontrolled risks.

Third, we recommend that manufacturers of cyber devices anticipate and make appropriate updates to these plans,<sup>11</sup> as well as to the processes, and procedures discussed in section II.C.2. below,<sup>12</sup> as new information becomes available, such as when new risks, threats, vulnerabilities, assets, or adverse impacts are discovered throughout the total product lifecycle. To support such efforts, manufacturers should also create or update appropriate documentation (e.g., threat modeling) and maintain it throughout the device lifecycle. Doing so will allow manufacturers to quickly identify vulnerability impacts once a device is released and could also help satisfy the patching requirements of section 524B(b)(2)(A)-(B) of the FD&C Act.

---

<sup>9</sup> For the purposes of this guidance, manufacturer procedures to carry out disclosures of the vulnerabilities and exploits may include procedures to inform device users, customers, patients, and other relevant healthcare stakeholders.

<sup>10</sup> For the purposes of this guidance, we are considering “related systems” to the extent needed to determine that the device, as it interacts with related systems, remains cybersecurity. Related systems are further described in Section II.C.2, below.

<sup>11</sup> See section 524B(b)(1) of the FD&C Act.

<sup>12</sup> See section 524B(b)(2) of the FD&C Act.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

130 The required plans,<sup>13</sup> as well as the processes, and procedures discussed in section II.C.2.  
131 below,<sup>14</sup> also should, as appropriate, account for any differences in the risk management for  
132 fielded devices (e.g., differences between marketed devices and devices no longer marketed but  
133 still in use). For example, if an update is not applied automatically for all fielded devices, then  
134 there will likely be different risk profiles for the differing software configurations of the device.  
135 Vulnerabilities should be assessed for any differing impacts for all fielded versions to ensure  
136 patient risks are being accurately assessed.  
137

### **2. Design, Develop, and Maintain Processes and Procedures to Provide a Reasonable Assurance of Cybersecurity (Section 524B(b)(2))**

141 Manufacturers of cyber devices must “design, develop, and maintain processes and procedures to  
142 provide a reasonable assurance that the device and related systems are cybersecure . . .” (section  
143 524B(b)(2) of the FD&C Act). FDA considers related systems to include, among other things,  
144 manufacturer-controlled elements, such as other devices, software that performs “other  
145 functions” as described in FDA’s Guidance “[Multiple Function Device Products: Policy and  
146 Considerations](#),”<sup>15</sup> software/firmware update servers, and connections to health care facility  
147 networks. The documentation recommendations identified in the [Premarket Cybersecurity  
148 Guidance](#) and summarized in Appendix 4 of the same guidance should be considered and used to  
149 demonstrate reasonable assurance that the device and related systems are cybersecure as required  
150 by section 524B(b)(2) of the FD&C Act.  
151

### **3. Software Bill of Materials (SBOM) (Section 524B(b)(3))**

154 Section 524B(b)(3) of the FD&C Act requires manufacturers of cyber devices to provide an  
155 SBOM, including commercial, open-source, and off-the-shelf software components. To assist  
156 with complying with this requirement, we recommend that a cyber device provide SBOMs that  
157 contain the information recommended in Section V.A.4(b) of the [Premarket Cybersecurity  
158 Guidance](#).  
159

## **D. Modifications**

161 The new requirements under section 524B of the FD&C Act apply to a person who submits an  
162 application or submission under section 510(k), 513(De Novo), 515(c)(PMA), 515(f)(PDP) or  
163 520(m)(HDE) for a device that meets the definition of a cyber device. Therefore, a person  
164 required to submit an application under one of the enumerated provisions for a device  
165 modification would also need to comply with the requirements in section 524B of the FD&C

---

<sup>13</sup> See section 524B(b)(1) of the FD&C Act.

<sup>14</sup> See section 524B(b)(2) of the FD&C Act.

<sup>15</sup> Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/multiple-function-device-products-policy-and-considerations>



## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

166 Act. In keeping with least burdensome principles,<sup>16</sup> the information we recommend that  
167 manufacturers of cyber devices provide will generally differ based on the type of change and  
168 whether such change impacts the cybersecurity of the device. Overall, we recommend that  
169 manufacturers use the recommendations below to determine the information FDA recommends  
170 manufacturers of cyber devices provide to demonstrate they have met the new requirements  
171 under section 524B of the FD&C Act when submitting a premarket submission for a device  
172 modification.  
173

#### 174 **1. Changes That May Impact Cybersecurity**

175 In general, changes that may impact cybersecurity could include changes to authentication or  
176 encryption algorithms, new connectivity features, or changing software update  
177 process/mechanisms. For these types of changes, see Section II.C., above, for required and  
178 recommended documentation to be included with each premarket submission (see section 524B  
179 of the FD&C Act).

#### 180 **2. Changes Unlikely to Impact Cybersecurity**

181 In general, changes unlikely to impact cybersecurity could include material changes, sterilization  
182 method changes, or a change to an algorithm without change to architecture/software  
183 structure/connectivity.  
184

185 For these types of changes, FDA recommends that manufacturers of cyber devices provide the  
186 following information to meet their premarket submission requirements in section 524B of the  
187 FD&C Act:  
188

- 189 • 524B(b)(1)
  - 190 • If not previously provided, manufacturers must provide a plan as described in
  - 191 section 524B(b)(1) of the FD&C Act; we recommend that it contain the
  - 192 information as described in Section II.C.1., above.
  - 193 • If a plan described in Section II.C.1., above, was previously provided, the
  - 194 manufacturer should provide a reference to the prior submission, a summary of
  - 195 any changes to the plan, and summaries of any updates/patches made to address
  - 196 vulnerabilities or exploits.
- 197 • 524B(b)(2)
  - 198 • Instead of the full documentation described as required or recommended in
  - 199 Section II.C.2., above, manufacturers may provide summary information to
  - 200 provide that there is a reasonable assurance that the device and related systems are
  - 201 cybersecure and no uncontrolled vulnerabilities, as described in Sections III.J. and
  - 202 VII.B. of the FDA guidance “[Postmarket Management of Cybersecurity in](#)
  - 203 [Medical Devices](#),”<sup>17</sup> are present. FDA recommends that this information include a
  - 204 summary assessment documenting the statements made in the summary

---

<sup>16</sup> For more information on FDA’s least burdensome provisions, see FDA’s guidance [The Least Burdensome Provisions: Concept and Principles](#)

<sup>17</sup> Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/postmarket-management-cybersecurity-medical-devices>

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

205 assessment, including a summary assessment of any cybersecurity impact from  
206 changes made since the last authorization (i.e., Letter to File, Annually  
207 Reportable) and a summary assessment of any vulnerabilities identified since the  
208 last authorization.

- 209 • If there are any limitations to updating the cybersecurity of the cyber device and  
210 related systems, the manufacturer should provide a description of the limitations  
211 of the system which prevent further cybersecurity controls, an assessment of the  
212 residual cybersecurity risk, and an assessment of the benefits and risks of the  
213 system.
- 214 • 524B(b)(3)
  - 215 • Section 524B(b)(3) of the FD&C Act requires manufacturers of cyber devices to  
216 provide an SBOM, including commercial, open-source, and off-the-shelf software  
217 components. To assist with complying with this requirement, we recommend that  
218 a cyber device provide SBOMs that contain the information recommended in  
219 Section V.A.4(b) of the [Premarket Cybersecurity Guidance](#).

220  
221 In general, in its cybersecurity review, FDA intends to focus substantive review on modifications  
222 to cybersecurity controls or modifications that are likely to affect cybersecurity. However,  
223 regardless of the type of change being proposed to the device in the premarket submission, FDA  
224 intends to take into account known cybersecurity concerns that are applicable to such device  
225 when conducting its premarket reviews and in determining whether the device has a reasonable  
226 assurance of cybersecurity.

## **E. Reasonable Assurance of Cybersecurity of Cyber 229 Devices**

230 Section 3305(c) of FDORA provides that nothing in section 524B of the FD&C Act “shall be  
231 construed to affect the Secretary’s authority related to ensuring that there is a reasonable  
232 assurance of the safety and effectiveness of devices, which may include ensuring that there is a  
233 reasonable assurance of the cybersecurity of certain cyber devices . . .” FDA interprets this  
234 provision to mean that a “reasonable assurance of cybersecurity” can be part of FDA’s  
235 determination of a device’s safety and effectiveness. Moreover, a determination that there is a  
236 reasonable assurance of cybersecurity is relevant to the various premarket pathways and  
237 authorization under them, specifically, FDA’s review of a PMA, PDP, De Novo, HDE, and  
238 510(k). With the exponential growth of interconnected devices on the market over the past few  
239 years (see Section I. of the [Premarket Cybersecurity Guidance](#)), ensuring cybersecurity has  
240 become essential to FDA’s ability to protect the public health and provide reasonable assurance  
241 of safety and effectiveness of devices.

242  
243 When evaluating a 510(k) submission, FDA considers changes to the environment of use (e.g.,  
244 changes in technology the subject device will interact with or operate within, and any new risks  
245 or vulnerabilities the device will be exposed to), new risks or vulnerabilities in the technological  
246 characteristics compared to the predicate device submission (e.g., changes to level of support for  
247 component software, vulnerabilities in communication protocols or technology used by the  
248 subject device), and how the subject device design and/or performance testing address these new

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

249 risks or vulnerabilities.<sup>18</sup> For example, if in reviewing the 510(k) for an alarm for a central  
250 nursing station software, FDA identifies that the device has increased risks compared to its  
251 predicate because it does not have the necessary encryption to protect against a recently  
252 identified cyber threat, FDA may ask for additional performance data. If the data provided is  
253 inadequate, FDA would likely make a determination that the new device is not substantially  
254 equivalent (NSE) to the predicate device because this threat, if exploited, could negatively  
255 impact the safety and effectiveness of the device because alarm accuracy is essential for health  
256 care providers to effectively monitor the health of patients in a hospital.

DRAFT

---

<sup>18</sup> For more information about current review practices for 510(k) submission, see FDA’s guidance [The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \[510\(k\)\]](#)