

FACT SHEET FOR HEALTHCARE PROVIDERS

Roche Molecular Systems, Inc.

cobas® liat SARS-CoV-2, Influenza A/B & RSV nucleic acid test

Updated: August 09, 2024

Coronavirus
Disease 2019
(COVID-19)

This Fact Sheet informs you of the significant known and potential risks and benefits of the emergency use of the **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test.

The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test is authorized for use with anterior nasal (nasal) swab and nasopharyngeal swab specimens collected from individuals with signs and symptoms of respiratory tract infection consistent with COVID-19 by their healthcare provider. Clinical signs and symptoms of respiratory viral infection due to SARS-CoV-2, Respiratory Syncytial Virus (RSV), and influenza can be similar.

All patients whose specimens are tested with this assay will receive the Fact Sheet for Patients: Roche Molecular Systems, Inc. - cobas® liat SARS-CoV-2, Influenza A/B & RSV nucleic acid test.

What are the symptoms of COVID-19?

Many patients with COVID-19 have developed fever and/or symptoms of acute respiratory illness (e.g., cough, dyspnea), although some individuals experience only mild symptoms or no symptoms at all. The current information available to characterize the spectrum of clinical illness associated with COVID-19 suggests that, when present, symptoms include cough, shortness of breath or dyspnea, fever, chills, myalgias, headache, sore throat, new loss of taste or smell, nausea or vomiting or diarrhea. Signs and symptoms may appear any time from 2 to 14 days after exposure to the virus, and the median time to symptom onset is approximately 5 days. For further information on the symptoms of COVID-19 please see the link provided in "Where can I go for updates and more information?" section.

Public health officials have identified cases of COVID-19 throughout the world, including the United States. Please check the CDC COVID-19 webpage (see link provided in "Where can I go for updates and more information?" section at the end of this document) or your local jurisdictions website for the most up to date information.

This test is to be performed only using anterior nasal (nasal) swab and nasopharyngeal swab specimens collected from individuals with signs and symptoms of respiratory tract infection consistent with COVID-19 by their healthcare provider.

What do I need to know about COVID-19 testing?

Current information on COVID-19 for healthcare providers is available at CDC's webpage, *Information for Healthcare Providers* (see links provided in "Where can I go for updates and more information?" section).

- The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test can be used to test anterior nasal (nasal) swab and nasopharyngeal swab specimens.
- The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test should be ordered for the detection and differentiation of SARS-CoV-2, influenza A, influenza B, and/or respiratory syncytial virus (RSV) in individuals with signs and symptoms of respiratory tract infection consistent with COVID-19 by their healthcare provider.
- The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test is authorized for use in laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform high, moderate, or waived complexity tests.
- The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test is authorized for use at the Point of Care (POC), i.e., in patient care settings operating under a CLIA Certificate of Waiver, Certificate of Compliance, or Certificate of Accreditation.

Specimens should be collected with appropriate infection control precautions. Current guidance is available at the CDC's website (see links provided in "Where can I go for updates and more information?" section).

When collecting and handling specimens from individuals suspected of being infected with the virus that causes COVID-19, appropriate personal protective equipment should be used as outlined in the CDC

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling 1-800-FDA-1088

FACT SHEET FOR HEALTHCARE PROVIDERS

Roche Molecular Systems, Inc.

cobas® liat SARS-CoV-2, Influenza A/B & RSV nucleic acid test

Updated: August 09, 2024

Coronavirus
Disease 2019
(COVID-19)

Interim Guidelines for Biosafety and COVID-19. For additional information, refer to the *CDC Interim Guidelines for Collecting and Handling of Clinical Specimens for COVID-19 Testing* (see links provided in “Where can I go for updates and more information?” section).

What does it mean if the specimen tests positive for the virus that causes COVID-19?

A positive test result for COVID-19 indicates that RNA from SARS-CoV-2 was detected, and therefore the patient is infected with the virus and presumed to be contagious. Laboratory test results should always be considered in the context of clinical observations and epidemiological data (such as local prevalence rates and current outbreak/epicenter locations) in making a final diagnosis and patient management decisions. Patient management should be made by a healthcare provider and follow current CDC guidelines.

The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test has been designed to minimize the likelihood of false positive test results. However, it is still possible that this test can give a false positive result, even when used in locations where the prevalence is below 5%. In the event of a false positive result, risks to patients could include the following: a recommendation for isolation of the patient, monitoring of household or other close contacts for symptoms, patient isolation that might limit contact with family or friends and may increase contact with other potentially COVID-19 patients, limits in the ability to work, delayed diagnosis and treatment for the true infection causing the symptoms, unnecessary prescription of a treatment or therapy, or other unintended adverse effects.

All laboratories using this test must follow the standard testing and reporting guidelines according to their appropriate public health authorities.

What does it mean if the specimen tests negative for the virus that causes COVID-19?

A negative test result for this test means that SARS-CoV-2 RNA was not present in the specimen above the limit of detection. However, a negative result does not rule out COVID-19 and should not be used as the sole basis for treatment or patient management decisions. It

is possible to test a person too early or too late during SARS-CoV-2 infection to make an accurate diagnosis via the **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test.

When diagnostic testing is negative, the possibility of a false negative result should be considered in the context of a patient's recent exposures and the presence of clinical signs and symptoms consistent with COVID-19. The possibility of a false negative result should especially be considered if the patient's recent exposures or clinical presentation indicate that COVID-19 is likely, and diagnostic tests for other causes of illness (e.g., other respiratory illness) are negative.

If COVID-19 is still suspected based on exposure history together with other clinical findings, re-testing with an alternative method should be considered by healthcare providers in consultation with public health authorities. Additional testing may be helpful to ensure testing was not conducted too early.

Risks to a patient of a false negative test result include: delayed or lack of supportive treatment, lack of monitoring of infected individuals and their household or other close contacts for symptoms resulting in increased risk of spread of COVID-19 within the community, or other unintended adverse events.

The performance of this test was established based on the evaluation of a limited number of clinical specimens. The clinical performance has not been established in all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation. Performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARS-CoV-2 and their prevalence, which change over time.

What does it mean if the specimen tests positive for influenza A, influenza B, and/or RSV?

A positive test result for influenza A virus, influenza B virus, and/or respiratory syncytial virus (RSV) indicates that RNA from one or more of these viruses was detected, the patient is infected with the virus(es) and is presumed to be contagious. Laboratory test results should always be considered in the context of clinical

FACT SHEET FOR HEALTHCARE PROVIDERS

Roche Molecular Systems, Inc.

cobas® liat SARS-CoV-2, Influenza A/B & RSV nucleic acid test

Updated: August 09, 2024

Coronavirus
Disease 2019
(COVID-19)

findings and observations and epidemiological data in making a final diagnosis. Patient management decisions should be made by a healthcare provider and follow current CDC guidelines. Results (positive and negative) for influenza and/or RSV should be interpreted with caution. If a result is inconsistent with clinical presentation and/or other clinical and epidemiological information, FDA-cleared influenza and/or RSV NAATs are available for confirmation if clinically indicated.

The **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test has been designed to minimize the likelihood of false-positive test results. However, in the event of a false-positive result, risks to individuals could include the following: a recommendation for isolation of the patient, monitoring of household or other close contacts for symptoms, patient isolation that might limit contact with family of friends, limits in the ability to work, delayed diagnosis and treatment for the true infection causing the symptoms, unnecessary prescription of an antiviral medication or other therapy, or other unintended adverse effects.

What does it mean if the specimen tests negative for influenza A, influenza B, and/or RSV?

A negative test result for influenza viruses and/or RSV means that influenza A, B, and/or RSV RNA was not present in the specimen above the limit of detection. However, a negative result does not rule out influenza A, influenza B, and/or RSV infection and should not be used as the sole basis for treatment or patient management decisions.

When diagnostic testing results are negative, the possibility of a false-negative result should be considered in the context of a patient's recent exposures and the presence of clinical signs and symptoms consistent with influenza. The possibility of a false-negative result should especially be considered if the patient's recent exposures or clinical presentation indicate that influenza A, influenza B, and/or RSV is likely, and diagnostic test results for other causes of illness (e.g., other respiratory illness) are negative. If influenza A, influenza B, and/or RSV is still suspected based on exposure history and clinical findings, retesting should be considered by healthcare providers in consultation with public health authorities.

Laboratory test results should always be considered in the context of clinical findings and observations and/or epidemiological data in making a final diagnosis. Patient management decisions should be made by a healthcare provider and follow current CDC guidelines. Results (positive and negative) for influenza or RSV should be interpreted with caution. If an influenza or RSV result is inconsistent with clinical presentation and/or other clinical or epidemiological information, FDA-cleared influenza and RSV NAATs are available for confirmation if clinically indicated.

Risks to individuals from a false-negative **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test result for influenza A, influenza B, and/or RSV include: delayed or lack of supportive treatment; lack of monitoring of infected patients and their household or other close contacts for symptoms, resulting in increased risk of spread of influenza A, influenza B, and/or RSV within the community; or other unintended adverse events.

What does it mean if the specimen tests positive for SARS-CoV-2, influenza A, influenza B, and/or RSV viruses? Is co-infection possible?

Yes, it is possible for an individual to be infected with more than one virus simultaneously. A positive test result for the viruses that cause COVID-19, influenza A, influenza B, and/or RSV indicates that RNA from these viruses was detected, the patient may be co-infected, and is presumed to be contagious. Laboratory test results should always be considered in the context of clinical findings and observations and epidemiological data in making a final diagnosis. Patient management decisions should be made with a healthcare provider and follow current CDC guidelines.

What is an EUA?

The United States FDA has made this test available under an emergency access mechanism called an Emergency Use Authorization (EUA). The EUA is supported by the Secretary of Health and Human Service's (HHS's) declaration that circumstances exist to justify the emergency use of in vitro diagnostics (IVDs) for the detection and/or diagnosis of the virus that causes COVID-19.

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling 1-800-FDA-1088

FACT SHEET FOR HEALTHCARE PROVIDERS

Roche Molecular Systems, Inc.

cobas® liat SARS-CoV-2, Influenza A/B & RSV nucleic acid test

Updated: August 09, 2024

Coronavirus
Disease 2019
(COVID-19)

An IVD made available under an EUA has not undergone the same type of review as an FDA-approved or cleared IVD. FDA may issue an EUA when certain criteria are met, which includes that there are no adequate, approved, available alternatives, and based on the totality of scientific evidence available, it is reasonable to believe that this IVD may be effective in diagnosing COVID-19.

The EUA for this test is in effect for the duration of the COVID-19 declaration justifying emergency use of IVDs, unless terminated or revoked (after which the test may no longer be used).

What are the approved available alternatives?

FDA has approved/cleared certain influenza and RSV tests. Any tests that have received full marketing status (e.g., cleared, approved), as opposed to an EUA, by FDA can be found by searching the medical device databases here: <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>. A cleared or approved test should be used instead of a test made available under an EUA, when appropriate and available. FDA has issued EUAs for other tests that can be found at:

<https://www.fda.gov/emergency-preparedness-and-response/mcm-legal-regulatory-and-policy-framework/emergency-use-authorization>.

Where can I go for updates and more information?

CDC webpages:

General: <https://www.cdc.gov/covid/>

Symptoms:

<https://www.cdc.gov/covid/signs-symptoms/>

Healthcare Providers:

<https://www.cdc.gov/covid/hcp/>

Information for Laboratories: <https://www.cdc.gov/covid/php/lab/>

Isolation Precautions in Healthcare Settings:

<https://www.cdc.gov/infection-control/hcp/isolation-precautions/>

Specimen Collection: <https://www.cdc.gov/covid/hcp/clinical-care/clinical-specimen-guidelines.html>

Infection Control: <https://www.cdc.gov/coronavirus/2019-ncov/hcp/infection-control-recommendations.html>

FDA webpages:

General: www.fda.gov/novelcoronavirus

EUAs: (includes links to patient fact sheet and manufacturer's instructions) <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas>

Roche Molecular Systems, Inc.:

4300 Hacienda Drive
Pleasanton, CA 94588

Customer and Technical Support:

Phone: 1-800-800-5973 (toll free)

Email: indianapolis_usa.liatsupport@roche.com

Website: https://www.roche.com/about/business/roche_worldwide.htm

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling 1-800-FDA-1088