



March 27, 2025

Riazur Rahman, MS
Regulatory Affairs Manager
Healgen Scientific, LLC
3818 Fuqua Street
Houston, TX 77047

Re: EUA240010/S002

Trade/Device Name: Healgen COVID-19/Flu A&B Ag Combo Rapid Test Cassette (Swab)

Dated: March 13, 2025

Received: March 13, 2025

Dear Riazur Rahman:

This is to notify you that your request to update the Healgen COVID-19/Flu A&B Ag Combo Rapid Test Cassette (Swab)¹ to extend the shelf-life expiration date to 24 months when stored at 2°C – 30°C, based on the results of your stability studies, is granted. Upon review, we concur that the data and information submitted in EUA240010/S002 supports the requested update for the Healgen COVID-19/Flu A&B Ag Combo Rapid Test Cassette (Swab). By submitting this EUA revision for review by the Food and Drug Administration (FDA), you have complied with the Conditions of Authorization stated in the letter authorizing the emergency use of the Healgen COVID-19/Flu A&B Ag Combo Rapid Test Cassette (Swab) issued on June 10, 2024.

Sincerely yours,

Joseph Briggs, Ph.D.
Deputy Director, Division of Microbiology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

¹ Note the shelf-life for the External Controls Healgen COVID-19/Flu A&B Ag Combo Rapid Test Cassette (Swab) was not evaluated as part of this study and remains at 15 months when stored at 2°C – 30°C.
U.S. Food & Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20903
www.fda.gov