

Individuals using assistive technology may not be able to fully access the information contained in this file. For assistance, please call 800-835-4709 or 240-402-8010, extension 1. CBER Consumer Affairs Branch or send an e-mail to: ocod@fda.hhs.gov and include 508 Accommodation and the title of the document in the subject line of your e-mail.

BLOOD GROUPING REAGENTS

Anti-S (Monoclonal) Anti- \bar{s} (Monoclonal)

IVD

In vitro diagnostic medical device



INTENDED USE

American Red Cross (Red Cross) Anti-S (Monoclonal) and Anti- \bar{s} (Monoclonal) are used for the *in vitro* detection of the S (MNS3) and \bar{s} (MNS4) antigens respectively, on human red blood cells by direct agglutination using the manual tube testing method.

SUMMARY AND EXPLANATION

The S (MNS3) and \bar{s} (MNS4) antigens were first described in 1947 by Walsh and Montgomery and by Race and Sanger.¹ The MNS system is highly complex consisting of 50 antigens.² S and \bar{s} are a pair of polymorphic antithetical antigens of the MNS system, carried on glycophorin B (GPB)³. The associated antibodies are responsible for transfusion reactions and Hemolytic Disease of the Fetus and Newborn (HDFN) and therefore are important in pre-transfusion testing and the prediction of HDFN.

The frequency of the S and \bar{s} antigens vary between different populations, as shown in Table 1:

Reagent		Phenotype	Prevalence ³ (%)	
Anti-S	Anti- \bar{s}		Whites	US Blacks
+	0	S+ \bar{s} -	11	6
+	+	S+ \bar{s} +	44	25
0	+	S- \bar{s} +	45	68
0	0	S- \bar{s} -	0	1.5

PRINCIPLE OF PROCEDURE

When the Directions for Use (DFU) are followed, these reagents will detect the presence of the associated antigen by causing visible agglutination of red blood cells expressing the specific antigen (positive test) after incubation at 18°C to 28°C. The absence of the associated antigen is demonstrated by the lack of agglutination of the red blood cells (negative test), after incubation at 18°C to 28°C, within the limits of the procedure.

REAGENT

Red Cross Anti-S (Monoclonal) and Anti- \bar{s} (Monoclonal) are prepared from monoclonal antibodies derived from cell culture supernatant from human/murine heterohybridoma cell lines (Anti-S MS-94 and Anti- \bar{s} P3BER). The antibodies are immunoglobulin class IgM and provide potent and specific reagents that meet the requirements of the Food and Drug Administration (FDA). Each lot is optimized for tube testing and standardized for pH and total protein concentration in a buffered diluent containing macromolecular chemical potentiators. The bovine albumin component of these products is derived exclusively from United States sources of disease-free cattle,

inspected, and certified by the U.S. Veterinary Services. This ruminant-based product is deemed to have low Transmissible Spongiform Encephalopathy (TSE) risk. The reagents contain sodium azide (0.1% final concentration) as a preservative.

These monoclonal antibodies are manufactured using intermediate products produced for the Red Cross in a shared manufacturing agreement with Millipore (UK) Ltd., Fleming Road, Kirkton Campus, EH547BN, Livingston, UK; U.S. License Number 1761.

Red Cross Anti-S (Monoclonal) and Anti- \bar{s} (Monoclonal) are for *in vitro* diagnostic use and are supplied ready for use. Use as furnished, do not dilute.

Red Cross Monoclonal Blood Grouping Reagents meet FDA potency requirements.

CAUTION STATEMENTS Rx only

CAUTION: All blood products should be treated as potentially infectious. Source material from which these products were derived was found negative when tested in accordance with current FDA required tests. No known test methods can offer assurance that products derived from hybridoma cell culture supernatant will not transmit infectious agents. The absence of murine virus has not been determined.

CAUTION: This Product Contains Natural Rubber Latex (Dropper Bulbs) Which May Cause Allergic Reactions.

WARNING: Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. If discarded into sinks, flush with a large volume of water to prevent azide build-up.

STORAGE

Store at 2°C to 8°C when not in use.

Do not use beyond the expiration date. The format for the expiration date is expressed as YYYY-MM-DD (year-month-day).

Do not use if turbid.

SPECIMEN COLLECTION AND PREPARATION

No special preparation of the patient or donor is required prior to specimen collection. Blood should be collected by an acceptable phlebotomy technique. Samples may be drawn into EDTA, CPD, CPDA-1, CPD with AS-1, CP2D with AS-3, and CPD with AS-5, or may be drawn without anticoagulant (non-barrier red top). Do not use samples drawn into tubes with neutral gel separators.

Samples should be tested as soon as possible after collection. Samples should be stored at 2°C to 8°C when not required for testing.

Blood drawn into EDTA, and non-barrier red top (clotted) tubes should be tested less than or equal to 10 days from collection. Donor samples drawn into CPD, CPDA-1, CPD with AS-1, CP2D with AS-3, and CPD with AS-5 may be tested up to the expiration date of the donor unit.

MATERIALS

Materials provided:

1. Red Cross Anti-S (Monoclonal) or Anti- \bar{s} (Monoclonal)

Materials required but not supplied:

1. Test tubes, 10 x 75mm or 12 x 75mm
2. Test tube racks
3. Pipettes
4. Centrifuge calibrated for serological use
5. Isotonic Saline (pH 6.5 – 7.5)
6. Antigen positive and negative control cells
7. Timer
8. Optical aid (optional)

PROCEDURE

1. Prepare a 2-4% suspension of red blood cells washed at least once with isotonic saline.
2. Add 1 drop of Red Cross Anti-S (Monoclonal) or Anti- \bar{s} (Monoclonal) reagent to an appropriately labeled test tube.
3. Add 1 drop of the previously prepared 2-4% red blood cell suspension.
4. Mix well.
5. Incubate at room temperature (18°C to 28°C) for 5 minutes.
6. Mix well and centrifuge the tube for 15 seconds at 3400 rpm (900-1000 rcf*) or equivalent, as indicated on the quality control calibration. Resuspend the red blood cells by gentle agitation.
7. Read macroscopically for agglutination and record results. An optical aid may be used if desired.

*rcf=0.00001118 x radius (cm) x (rpm)²

STABILITY OF REACTION

Following centrifugation, all tube tests should be read/recorded without delay. Time delays may cause a dissociation of the antigen-positive complexes resulting in false negative or weaker than expected reactions.

QUALITY CONTROL

The reactivity of these Blood Grouping Reagents should be confirmed on each day of use by testing with known antigen-positive (preferably heterozygous expression) and with known antigen-negative red blood cells. Each reagent is acceptable for use if it reacts only with the antigen-positive red blood cells.

INTERPRETATION OF RESULTS

Positive (+) test result: visible agglutination of the red blood cells after the 18°C to 28°C incubation and centrifugation.

NOTE: Hemolysis, if obtained, should not be interpreted as a positive result since the conditions for complement activation due to a red cell antibody-antigen reaction do not exist.

Negative (-) test result: no agglutination of red blood cells after the 18°C to 28°C incubation and centrifugation.

LIMITATIONS

All serological tests have limitations. To maximize success in obtaining valid results, follow the DFU carefully. Deviations from manufacturer's instructions without appropriate validation and controls may produce erroneous results.

Samples with a positive direct antiglobulin test, cold agglutinins, or rouleaux formation may show false positive results in testing with monoclonal antibodies. Results from these samples must be interpreted with caution. If false positive results are suspected, an appropriate control, for example 6% albumin solution, should be included in any investigatory process.

False positive or false negative test results may occur from bacterial or chemical contamination of test materials, improper incubation temperature or time, centrifugation errors, storage temperatures excursions, or omission of test reagents.

Enzyme treated red blood cells may give false negative reactions with Red Cross Anti-S (Monoclonal) and Red Cross Anti- \bar{s} (Monoclonal). Enzyme treated red blood cells may give false positive reactions with Red Cross Anti- \bar{s} (Monoclonal). Use of enzyme treated cells with these reagents is not recommended.

The resuspension of the button in serological reactions in the tube test procedure must be carried out by using gentle agitation. Shaking too strongly may cause agglutinates to be dispersed.

Suppressed or weakened expression of blood group antigens may give rise to false negative reactions or non-concordance with similar reagents and/or alternate methodologies. Similarly, variants may give rise to unexpected positive or negative reactions especially in comparison with previously typed with polyclonal blood grouping reagents or reagents made with other cell line-derived monoclonal sources.

SPECIFIC PERFORMANCE CHARACTERISTICS

Red Cross Anti-S (Monoclonal) and Anti- \bar{s} (Monoclonal) have been manufactured to meet FDA potency requirements. Each lot is tested against a panel of antigen-positive red blood cells (heterozygous expression and/or weakened expression if possible) to ensure appropriate reactivity when used by the recommended test procedure in the DFU. The specificity of each lot is verified by the recommended tube testing method using a panel of red cells that lack the reagent antigen.

PERFORMANCE CHARACTERISTICS BY MANUAL TUBE METHOD

The performance of Red Cross Anti-S (Monoclonal) and Anti- \bar{s} (Monoclonal) was tested against FDA licensed reference reagents (comparators) in a multi-center field trial, representing blood collection establishments, transfusion services and/or immunohematology reference laboratories which tested donor, patient, and neonatal samples. The testing sites were selected to cover diverse geographic locations and patient/donor populations.

FIELD COMPARISON STUDY RESULTS

Table 2

N=1517		Comparator Reagent			
		Positive	Negative		
Red Cross Anti-S Lot S-2	Positive	796	0	Positive Percent Agreement	100.00%
	Positive			PPA (95% 1-Sided LCI)	99.62%
	Negative	0	721	Negative Percent Agreement	100.00%
				NPA (95% 1-Sided LCI)	99.59%

Table 3

N=1517		Comparator Reagent			
		Positive	Negative		
Red Cross Anti-S Lot S-3	Positive	796	0	Positive Percent Agreement	100.00%
				PPA (95% 1-Sided LCI)	99.62%
	Negative	0	721	Negative Percent Agreement	100.00%
				NPA (95% 1-Sided LCI)	99.59%

Table 4

N=1699		Comparator Reagent			
		Positive	Negative		
Red Cross Anti-S Lot LS-2	Positive	1271	0	Positive Percent Agreement	100.00%
				PPA (95% 1-Sided LCI)	99.76%
	Negative	0	428	Negative Percent Agreement	100.00%
				NPA (95% 1-Sided LCI)	99.30%

Table 5

N=1699		Comparator Reagent			
		Positive	Negative		
Red Cross Anti-S Lot LS-3	Positive	1271	0	Positive Percent Agreement	100.00%
				PPA (95% 1-Sided LCI)	99.76%
	Negative	0	428	Negative Percent Agreement	100.00%
				NPA (95% 1-Sided LCI)	99.30%

PRECISION STUDY RESULTS

As part of the performance evaluation, a precision and reproducibility study was performed using multiple operators at multiple sites on 5 non-consecutive days over a 20-day period. Testing confirmed repeatable and reproducible results within each run, run-to-run, day-to-day, operator-to-operator, and site-to-site. There were no discordant results.

A lot-to-lot study was also performed using three lots of each specificity on 5 non-consecutive days over a 20-day period which confirmed repeatable and reproducible results within each test run, run-to-run, and day-to-day. There were no discordant results.

Specificity tests submitted to the FDA for release of an individual lot of product will be furnished upon request.

For technical questions, contact the American Red Cross Diagnostic Manufacturing Division at 1-800-882-3737.

BIBLIOGRAPHY

1. Issitt PD, Anstee DJ. Applied Blood Group Serology. 4th ed. Durham: Montgomery Scientific, 1998.
2. Cohn CS, ed. Technical Manual. 21st ed. Bethesda MD: AABB, 2023.
3. Reid ME, Lomas-Francis C. The Blood Group Antigen Facts Book. 3rd ed. San Diego: Elsevier Academic Press, 2012.

The American National Red Cross
Washington, DC 20006
U.S. License 190
REF-0001671 [1]
Version 1, Date of Issuance 07/03/2025