

*Contains Nonbinding Recommendations*

# **Guidance for Industry and FDA Staff**

---

## **The Review and Inspection of Premarket Approval Application Manufacturing Information and Operations**

**Document issued on: January 8, 2008**

**The draft of this document was issued on June 19, 2006.**

For questions regarding this document contact Timothy A. Ulatowski at 240-276-0100 ([timothy.ulatowski@fda.hhs.gov](mailto:timothy.ulatowski@fda.hhs.gov)) or James L. Woods at 240-276-0443 ([james.woods@fda.hhs.gov](mailto:james.woods@fda.hhs.gov)).



**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Devices and Radiological Health  
Office of Compliance  
Office of In Vitro Diagnostic Device Evaluation and Safety**

## **Preface**

## **Public Comment**

Written comments and suggestions may be submitted at any time for Agency consideration to the Division of Dockets Management, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD, 20852.

Alternatively, electronic comments may be submitted to <http://www.fda.gov/dockets/ecomments>. When submitting comments, please refer to Docket No. 2006D-0063. Comments may not be acted upon by the Agency until the document is next revised or updated.

## **Additional Copies**

Additional copies are available from the Internet at: <http://www.fda.gov/cdrh/comp/guidance/1566.pdf>. You may also send an e-mail request to [dsmica@fda.hhs.gov](mailto:dsmica@fda.hhs.gov) to receive an electronic copy of the guidance or send a fax request to 240-276-3151 to receive a hard copy. Please use the document number 1566 to identify the guidance you are requesting.

# Guidance for Industry and FDA Staff

---

## The Review and Inspection of Premarket Approval Application Manufacturing Information and Operations

---

*This guidance represents the Food and Drug Administration's (FDA's) current thinking on this topic. It does not create or confer any rights for or on any person and does not operate to bind FDA or the public. You can use an alternative approach if the approach satisfies the requirements of the applicable statutes and regulations. If you want to discuss an alternative approach, contact the FDA staff responsible for implementing this guidance. If you cannot identify the appropriate FDA staff, call the appropriate number listed on the title page of this guidance.*

### Introduction

The Medical Device User Fee and Modernization Act of 2002 (MDUFMA) amended the Federal Food, Drug, and Cosmetic Act (the Act) (21 U.S.C. 321 et seq.) to allow for the collection of user fees for the review of certain marketing applications. A portion of the fee collected for premarket approval applications (PMAs) will help cover the costs associated with the review of the PMA manufacturing section information and the inspection of the manufacturing facilities. In a letter to Congress that accompanied the user fee legislation, the Secretary of Health and Human Services committed to “improve the scheduling and timeliness of preapproval inspections.”<sup>1</sup>

This guidance explains for applicants the process involved with the review of a PMA manufacturing section and inspection of the manufacturing operations described in the manufacturing section. This guidance is also generally applicable to the process involved

---

<sup>1</sup> This letter can be found at: [www.fda.gov/cdrh/mdufma/pgoals.html](http://www.fda.gov/cdrh/mdufma/pgoals.html)

## *Contains Nonbinding Recommendations*

with the review of manufacturing information in certain PMA supplements. This guidance does not address premarket notification (510(k)) submissions because a premarket inspection is not ordinarily conducted for this type of premarket submission. FDA believes that the procedural information outlined in this document should help applicants and FDA schedule and complete their work in a timely manner.

The following will be addressed in this guidance:

- The sequence of events as the Office of Compliance (OC) or the Office of *In Vitro* Diagnostic Device Evaluation and Safety (OIVD) reviews the manufacturing section of a PMA;
- The administrative process and projected timeframes involved with each step; and
- How the inspection of a manufacturing facility fits into the approval process.

FDA's guidance documents, including this guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency's current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidances means that something is suggested or recommended, but not required.

### **The Least Burdensome Approach**

We believe we should consider the least burdensome approach in all areas of medical device regulation. This guidance reflects our careful review of the relevant scientific and legal requirements and what we believe is the least burdensome way for you to comply with those requirements. However, if you believe that an alternative approach would be less burdensome, please contact us so we can consider your point of view. You may send your written comments to the contact person listed in the preface to this guidance or to the CDRH Ombudsman. Comprehensive information on CDRH's Ombudsman, including ways to contact him, can be found on the Internet at <http://www.fda.gov/cdrh/ombudsman/>

### **Scope**

This guidance explains the administrative process used by OC or OIVD to review a PMA's Quality System (QS) regulation (21 CFR 820) information. OC has already published guidance, entitled "Quality System Information for Certain Premarket Application Reviews," which identifies the QS regulation information an applicant should include in the PMA manufacturing section.<sup>2</sup> In addition, two corollary documents discuss generally the FDA review procedures and clock for PMAs, including the review of QS regulation information and the inspection process. One document is entitled, "FDA and Industry Actions on Premarket Approval Applications (PMAs): Effect on FDA Review Clock and Performance

---

<sup>2</sup> This guidance is available at: <http://www.fda.gov/cdrh/comp/guidance/1140.html>

## *Contains Nonbinding Recommendations*

Assessment.”<sup>3</sup> The other guidance is entitled, “Premarket Approval Application Modular Review.”<sup>4</sup>

### **Premarket Approval Applications**

The usual path to approval of a device is submission of a PMA, which may be traditional or modular, and expedited or non-expedited.<sup>5</sup> When seeking premarket approval for your device, you should select the appropriate type of PMA submission based on the following:

#### 1. Traditional PMA

In this PMA format, you would submit all the elements required for a PMA, e.g., complete scientific and technical information about the device, manufacturing information, non-clinical study information, and statistically valid and reliable data from clinical studies, at the same time in a single application, so we can determine whether there is a reasonable assurance that the device is safe and effective for its intended use. For guidance on the type of information needed for FDA to file your PMA, see “Premarket Approval Application Filing Review.”<sup>6</sup>

#### 2. Modular PMA

This PMA format consists of sections or modules submitted separately that together become a complete application. Each module includes elements, tests, or other information that constitute a component of a complete PMA, such as manufacturing information or clinical data. For more information on the Modular PMA Program, see the guidance entitled “Premarket Approval Application Modular Review.”<sup>4</sup>

#### 3. Expedited PMA (Traditional and Modular)

We give priority to PMAs for devices under certain circumstances. The Office of Device Evaluation (ODE) and OIVD determine, using criteria defined in section 515(d)(5) of the Act, whether a PMA qualifies for expedited status. For more information on expedited PMAs, see the guidance entitled, “Expedited Review of Premarket Submissions for Devices.”<sup>7</sup>

### **B. Types of PMA Supplements**

---

<sup>3</sup> This guidance can be found at: [www.fda.gov/cdrh/mdufma/guidance/1218.html](http://www.fda.gov/cdrh/mdufma/guidance/1218.html)

<sup>4</sup> This guidance can be found at: [www.fda.gov/cdrh/mdufma/guidance/835.html](http://www.fda.gov/cdrh/mdufma/guidance/835.html)

<sup>5</sup> The Product Development Protocol and, for devices that meet narrow criteria, the Humanitarian Device Exemption provide alternate approval mechanisms. This guidance does not apply to these application types.

<sup>6</sup> This guidance can be found at: [www.fda.gov/cdrh/ode/guidance/297.html](http://www.fda.gov/cdrh/ode/guidance/297.html)

<sup>7</sup> This guidance can be found at: <http://www.fda.gov/cdrh/mdufma/guidance/108.html>

## *Contains Nonbinding Recommendations*

You must submit a PMA supplement for review and approval if you make a change affecting the safety or effectiveness of a device for which you have an approved PMA. (21 CFR 814.39(a).) Some changes do not require a supplement and some changes may be made using alternative forms of submission, as specified in FDA regulations. MDUFMA added definitions to the FDC Act of several types of PMA supplements, including: panel-track supplements, 180-day supplements, and real-time supplements. See FDC Act § 737, 21 U.S.C. § 379.

Panel-track and 180-day supplements may include manufacturing information. If so, OC or OIVD will review the information and an inspection may be required, depending on the proposed change. When both a review of manufacturing information and an inspection are needed, the review process and timelines described in this guidance generally apply.

Typically, there are no inspections associated with the following types of PMA submissions so they will not be addressed in this guidance:

- Real-time supplements;
- 30-day notices;
- 135-day supplements;
- Special PMA Supplements-Changes Being Affected;
- Express PMA supplements; and
- PMA annual reports.

### 1. Panel-track Supplements

Section 737(4)(B) of the Act, which was added by section 102 of MDUFMA, defines a "panel-track supplement" as "a supplement to an approved premarket application or premarket report under section 515 that requests a significant change in design or performance of the device, or a new indication for use of the device, and for which substantial clinical data are necessary to provide a reasonable assurance of safety and effectiveness." (21 U.S.C. 379i (4)(B)).

Panel-track supplements for changes in device design or performance that may significantly affect clinical outcome may require the submission of manufacturing information and an inspection.

We give priority review to panel-track supplements meeting the criteria defined in section 515(d)(5) of the Act. Therefore, in the discussion below, references to expedited and non-expedited PMAs include expedited and non-expedited panel-track supplements.

### 2. 180-Day Supplements

### *Contains Nonbinding Recommendations*

Under section 737(4)(C) of the Act, which was added by section 102 of MDUFMA, a "180-day supplement" is defined as:

"a supplement to an approved premarket application or premarket report under section 515 that is not a panel-track supplement and requests a significant change in components, materials, design, specification, software, color additives, or labeling." (21 U.S.C. 379i(4)(C))

Read in conjunction with section 515(d)(6) of the Act, this language means that submission of a 180-Day Supplement is required for certain types of significant changes to the approved device that affect safety and effectiveness of the device. In general, in order for a change to be submitted as a 180-Day Supplement, the clinical data provided in support of the original device approval should still be applicable in supporting the approval of the modified device. In most cases, for such modifications, only new pre-clinical testing is needed to demonstrate reasonable assurance of safety and effectiveness of the modified device. In some instances, however, additional limited confirmatory clinical data may be necessary to provide a bridge between the clinical data set for the original device and the expected clinical performance of the modified device. Although additional clinical data may be necessary, the data collected are usually from a limited number of patients. Changes to devices that may require a 180-Day Supplement include:

- the principle of operation
- the control mechanism
- the device design or performance
- the labeling
- new testing requirements or acceptance criteria.

A 180-day supplement involving changes to the design or manufacturing may require submission of manufacturing information and an inspection.

#### **C. The Review Process in Brief**

The premarket review process begins when the applicant submits six copies of a fileable PMA<sup>8</sup> or three copies of a fileable PMA supplement<sup>9</sup> to the CDRH Document Mail Center (DMC) in ODE. Upon receipt, OC or OIVD will review the PMA's manufacturing section. OC or OIVD will request an inspection of the facility if one is needed. There may be more than one manufacturing facility for an original PMA and a facility may be a domestic or foreign site.

---

<sup>8</sup> See 21 CFR 814.20(b)(2).

<sup>9</sup> See 21 CFR 814.39(c).

## *Contains Nonbinding Recommendations*

Following an inspection, OC or OIVD will receive and analyze the establishment inspection report for the manufacturing facility. OC or OIVD will then make a recommendation to ODE on the status of the quality system for the facility.

This guidance describes the timeframes within which FDA's review of the submission should be completed based on the goals set forth in the *Congressional Record*. Significant deficiencies in the manufacturing information and/or manufacturing operations may result in more than one cycle of review. Flow charts depicting the PMA manufacturing section review and inspection timelines are provided in Attachment A.

## **FDA Review of the PMA Manufacturing Section, and the Inspection Process**

### **A. What procedure does FDA use to review a PMA manufacturing section?**

When a fileable PMA arrives at the ODE/DMC, the DMC processes the application as follows:

1. Logs in and tracks the submission.
2. Assigns due dates for the CDRH offices based on the date of receipt, e.g., 180 days after receipt.
3. Alerts the ODE Program Operations Staff (POS) of the incoming PMA.
4. Sends one complete copy of the PMA to OC Field Operations Branch (OC/FOB) within 7 calendar days of receipt.
5. Files one copy in the DMC.
6. Further processes the application administratively, prepares a transmittal coversheet, and forwards the document to the appropriate ODE review division or OIVD.

When OC/FOB receives its copy of the PMA, it processes the document as follows:

1. OC/FOB assigns an internal tracking number for the PMA, establishes a due date (see Table 1) for the OC or OIVD review of the manufacturing information, and generates a transmittal coversheet.
2. The PMA with its transmittal coversheet are delivered to the appropriate enforcement division in OC, or to OIVD. If clinical data are submitted in the PMA, OC/FOB forwards the clinical section to the OC Division of Bioresearch Monitoring (DBM) for review.

Note: DBM's review and processing of the clinical and preclinical sections are addressed in a separate MDUFMA guidance, "Guidance for Industry and Food and Drug Administration Staff: the Review and Inspection of Premarket Approval Applications Under the Bioresearch Monitoring Program."

3. OC/FOB alerts the district office associated with a domestic manufacturing site identified in the PMA as to the receipt of the application. OC/FOB requests a

***Contains Nonbinding Recommendations***

response from the appropriate FDA district office within 7 calendar days regarding the inspection history of the manufacturing site. If there are multiple facilities involved in the manufacturing process, OC/FOB will notify all relevant districts.

The district will determine the following information concerning the manufacturing site from its records:

- a. the date of last inspection;
- b. classification of the last inspection report; and
- c. the similarity of the products inspected to the PMA device and similarity of the manufacturing operations inspected by FDA to those used for the PMA device under review.

Once the information is received from the district office, OC/FOB forwards the information to the assigned OC enforcement division, or to OIVD.

- 4. OC serves as the district office for any foreign facility. OC, with input from OIVD when necessary, determines whether a foreign manufacturing facility requires an inspection based on the same information noted in item 3 above.

For foreign sites, OC/FOB notifies the Office of Regional Operations (ORO), Division of Field Investigations (DFI), International Operations Branch (IOB) of a pending assignment. ORO/DFI/IOB coordinates the foreign site inspections for the FDA field staff. A provisional inspection assignment for a foreign facility is transmitted from OC/FOB to ORO/DFI/IOB early in the review process (see Attachment A). This early assignment is done before completion of the manufacturing section review because foreign inspections take longer than domestic inspections to organize and complete due to scheduling, travel logistics, and coordinating with foreign governments and the U.S. State Department. The longer organizational time for a foreign inspection enables the assigned OC enforcement division or OIVD to substantially review the PMA manufacturing section and to cancel the inspection if there are major deficiencies.

- 5. The time frames indicated below in Table 1 are the performance goals for OC or OIVD to complete review of the PMA manufacturing section.<sup>10</sup>

**Table 1**

<b>TYPE OF PMA APPLICATION</b>	<b>MANUFACTURING SECTION REVIEW TIMEFRAME</b>
Non-Expedited Traditional	30 calendar days
Expedited Traditional	20 calendar days
Non-Expedited Modular	30 calendar days when submitted

---

<sup>10</sup> Time frames for the QS/GMP review begin one day after FOB receives the PMA document from the DMC.

*Contains Nonbinding Recommendations*

	with final module; up to 90 days if submitted prior to final module
Expedited Modular	20 calendar days when submitted with final module; up to 90 days if submitted prior to final module
180-Day Supplement	30 calendar days
Amendment	20 or 30 calendar days depending on type of PMA amended

**B. What happens when OC or OIVD completes its review of the PMA manufacturing section?**

The OC or OIVD reviewer (with supervisory concurrence) may: (1) recommend that the manufacturing section is acceptable or (2) prepare a deficiency letter identifying significant deficiencies and requesting additional information from the applicant. The OC or OIVD reviewer will also determine, with supervisory concurrence, whether an inspection is required.

1. If OC or OIVD determines that the PMA’s manufacturing section is significantly deficient due to numerous omissions of information needed to assess the manufacturing operations, the applicant should expect to receive a deficiency letter identifying the specific information to submit to allow FDA to complete the review of the manufacturing section. The OC or OIVD reviewer should communicate in real-time with the applicant to clarify the number and type of deficiencies in a deficiency letter. If there are only a few minor deficiencies, these will likely be handled by real time communication (e.g., phone, email and fax). An extended response time by the applicant to a deficiency letter or to a real time communication for more information, or a delay in the submission of substantial additional manufacturing data, may significantly delay the completion of the initial review. Submission of a response to a deficiency letter will begin a new 20, 30, or up to 90 day manufacturing section review timeframe depending on the type of application. (See Table 1.)
2. If the review of the manufacturing section is substantially complete and there are no deficiencies in the manufacturing section, or if the deficiencies have been adequately addressed through phone calls or additional submissions, OC or OIVD will evaluate the district office’s inspection history of the domestic facility. If OC or OIVD determines that FDA should conduct a QS/GMP inspection, OC/FOB will initiate an assignment with ORO for a domestic inspection. OC/FOB will also confirm that the already initiated foreign inspection assignment should proceed.
3. The PMA applicant should be ready for inspection at the time of the filing of the PMA application unless the applicant states in the PMA that it is not ready. The District (or OC for a foreign facility) should contact the facility’s management to alert them to the pending inspection. At that time, the applicant should confirm that the facility is ready for inspection. The manufacturing process should preferably be in operation by the time the inspection is conducted. In any case, all process

### *Contains Nonbinding Recommendations*

validations should be completed and the design successfully transferred into production.

4. OC/FOB plans to cancel the assigned or scheduled inspection under the following circumstances: if the PMA is determined to be not fileable; ODE or OIVD issues a major deficiency letter, denial letter, withdrawal letter, or not approvable letter; or the applicant communicates to the FDA that the manufacturing site is not ready for inspection.

If the facility is not ready for inspection, the District (or OC for a foreign facility) will ask the applicant to send a letter to the District or OC, signed by the most responsible person at the firm, indicating that they are not ready for inspection, the reason for not being ready, and when they expect to be ready. This letter will be made part of the PMA record.

The applicant should alert CDRH in writing or by real time communication in advance of when the facility will be ready for inspection. OC/FOB will then reissue an inspection assignment and the inspection will be completed as stated in Section C.

#### **C. How much time does the FDA allow for a PMA QS/GMP inspection?**

##### Domestic Inspections:

FDA field staff should complete the inspection within 45 calendar days after receipt of the inspection assignment from OC or OIVD.

##### Foreign Inspections:

FDA field staff should complete the inspection within 60 calendar days after receipt of the inspection assignment. Please note, foreign travel requires additional time for scheduling, travel logistics, and for coordinating with foreign governments and the U.S. State Department.

An additional 30 calendar days are allotted for the field staff to write and forward a PMA application establishment inspection report to FOB for both domestic and foreign inspections.

#### **D. What happens after the establishment inspection report (EIR) is completed by the field staff?**

There are several potential outcomes from an inspection:

- If FDA concludes that there are no objectionable conditions or practices with the manufacturing operations, or any that do not justify further regulatory action, i.e., a No Action Indicated (NAI) result, then OC/FOB will notify ODE POS that the PMA approval can proceed from a manufacturing operations perspective.
- If FDA concludes that there are objectionable conditions or practices, but these do not warrant administrative or regulatory action, i.e., a Voluntary Action Indicated

### *Contains Nonbinding Recommendations*

(VAI) result, then FDA may communicate the objectionable conditions or practices to the applicant and recommend that the applicant voluntarily correct the deficiencies. FDA may evaluate the corrections during a subsequent inspection. As above, OC/FOB will notify ODE POS that the PMA approval can proceed from a manufacturing operations perspective.

- If FDA concludes that there are objectionable conditions or practices that warrant regulatory or administrative sanctions, i.e., an Official Action Indicated (OAI) result, then there are two possible outcomes:
  1. If the conditions or practices pertain only to the pending PMA device, then OC or OIVD will issue a deficiency letter within 30 calendar days for a non-expedited PMA and within 20 days for an expedited PMA.
  2. If the conditions or practices pertain not only to the pending PMA device but also to marketed products, then FDA may issue a Warning Letter or Untitled Letter.

In either case, the letter will inform the addressee that no approval for a pending PMA will issue until the objectionable conditions or practices are corrected.

OC/FOB is informed of the issuance of a Warning or Untitled Letter by the issuing office. Based upon the deficiencies noted in the Warning or Untitled Letter, OC/FOB will notify ODE POS when the manufacturing operation is not satisfactory and recommend that the PMA approval be withheld. ODE POS may then make the appropriate entry into the CDRH document tracking database. If ODE or OIVD has completed its review of the PMA and concludes the PMA warrants an approval or approvable action, except for the unsatisfactory manufacturing aspects, ODE or OIVD will issue an approvable letter notifying the manufacturer of the pending GMP issues. This letter stops the review clock for the PMA.

The applicant should provide a complete response to the FDA office that issued the letter within the timeframe noted. A complete response to a deficiency, Warning or Untitled Letter begins a new 20 or 30 day FDA manufacturing section review period (See Table 1.) More than one cycle of correspondence may be needed to resolve deficiencies. Once deficiencies or violations appear to have been corrected, OC/FOB may issue a new inspection assignment, if needed, beginning a new cycle of review including the 45 or 60 day inspection as well as EIR completion and review by the District and OC or OIVD.

#### **E. In summary, what is the general review cycle time for the manufacturing section and operations?**

Attachment A lists the internal FDA process phases and times involved with domestic and foreign inspections. The inspection assignment phase and manufacturing review phase are concurrent. In order to help meet the MDUFMA PMA performance goals, FDA plans to

*Contains Nonbinding Recommendations*


complete its review of the manufacturing information and the inspection within a total cycle time of approximately 140 days for a nonexpedited PMA and 120 days for an expedited PMA.

**F. What are the most common factors that delay the review of a PMA manufacturing section or delay the inspection process?**

- When OC or OIVD issues a manufacturing section deficiency letter to a PMA applicant, a timely and thorough response facilitates the review process. Until OC or OIVD receives the requested information, the review of the manufacturing section remains on hold and no inspection will be scheduled.
- When a timely response is not received to real-time questions during the manufacturing section review and the milestone is exceeded (see Attachment A), FDA may issue a manufacturing section deficiency letter and/or an approvable pending GMP letter.
- The manufacturing process should be in operation as soon as possible after PMA submission. As noted, all process validations should be completed and the design successfully transferred into production. Rescheduling a PMA inspection may result in further delays of the FDA review of the PMA's manufacturing section and operations and result in an approvable pending GMP letter.

Attachment A  
Process Flow Charts

**OC MDUFMA PMA Review Milestones in Total  
FDA Calendar Days - Domestic Inspections**



**OC MDUFMA PMA Review Milestones in Total  
FDA Calendar Days - Foreign Inspections**

