

On February 2, 2024, FDA published the final rule to amend the Quality System (QS) regulation in 21 CFR part 820 ([89 FR 7496](#), effective February 2, 2026). The revised 21 CFR part 820 is now titled the Quality Management System Regulation (QMSR). The QMSR harmonizes quality management system requirements by incorporating by reference the international standard specific for medical device quality management systems set by the International Organization for Standardization (ISO), ISO 13485:2016. The FDA has determined that the requirements in ISO 13485 are, when taken in totality, substantially similar to the requirements of the QS regulation, providing a similar level of assurance in a firm's quality management system and ability to consistently manufacture devices that are safe and effective and otherwise in compliance with the Federal Food, Drug, and Cosmetic Act (FD&C Act).

This guidance document was issued prior to the effective date of the final rule. FDA encourages manufacturers to review the current QMSR to ensure compliance with the relevant regulatory requirements.

# Refuse to Accept Policy for 510(k)s

---

## Guidance for Industry and Food and Drug Administration Staff

**Document issued on: April 21, 2022.**

**Document originally issued on May 20, 1994.**

**This document supersedes “Refuse to Accept Policy for 510(k)s” issued  
September 13, 2019.**

For questions about this document regarding CDRH-regulated devices, contact ORP: Office of Regulatory Programs, DRP1: Division of Submission Support, Premarket Notification and Classification Team by email at [510K\\_Program@fda.hhs.gov](mailto:510K_Program@fda.hhs.gov) or at 301-796-5640.

For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) by email at [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov) or at 800-835-4709 or 240-402-8010.

**OMB Control No. 0910-0120**  
**Current expiration date available at <https://www.reginfo.gov>**  
**See additional PRA statement in Section IX of this guidance**



**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Devices and Radiological Health  
Center for Biologics Evaluation and Research**

## Preface

### Public Comment

You may submit electronic comments and suggestions at any time for Agency consideration to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD, 20852. Identify all comments with the docket number FDA-2012-D-0523. Comments may not be acted upon by the Agency until the document is next revised or updated.

### Additional Copies

Additional copies are available from the Internet. You may also send an e-mail request to [CDRH-Guidance@fda.hhs.gov](mailto:CDRH-Guidance@fda.hhs.gov) to receive a copy of the guidance. Please include the document number 1793 and complete title of the guidance in the request.

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., WO71, Room 3128, Silver Spring, MD 20903, or by calling 800-835-4709 or 240-402-8010, by email, [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov), or from the Internet at <https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>.

## Table of Contents

I. Purpose .....	4
II. Background .....	5
III. Scope.....	6
IV. Q-submission Interaction.....	6
V. 510(k) Refuse to Accept Policies and Procedures .....	7
VI. Refuse to Accept Principles.....	10
VII. The Checklist – Preliminary Questions.....	12
VIII. The Checklists – Acceptance Review .....	14
IX. Paperwork Reduction Act of 1995 .....	19
<b>Appendix A. Acceptance Checklist for Traditional 510(k)s .....</b>	<b>20</b>
<b>Appendix B. Acceptance Checklist for Abbreviated 510(k)s.....</b>	<b>55</b>
<b>Appendix C. Acceptance Checklist for Special 510(k)s.....</b>	<b>91</b>

# Refuse to Accept Policy for 510(k)s

## Guidance for Industry and Food and Drug Administration Staff

*This guidance represents the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.*

### I. Purpose

The purpose of this document is to explain the procedures and criteria FDA intends to use in assessing whether a premarket notification (510(k)) submission meets a minimum threshold of acceptability and should be accepted for substantive review.

Focusing FDA's review resources on complete submissions will provide a more efficient approach to ensuring that safe and effective medical devices reach patients as quickly as possible. Moreover, with the enactment of the Medical Device User Fee and Modernization Act of 2002 (MDUFMA), the Medical Device User Fee Amendments of 2007 (MDUFA II), the Medical Device User Fee Amendments of 2012 (MDUFA III), and the Medical Device User Fee Amendments of 2017 (MDUFA IV),<sup>1</sup> FDA agreed to performance goals based on the timeliness of reviews. Acceptance review therefore takes on additional importance in both encouraging quality submissions from submitters of 510(k) notifications and allowing FDA to appropriately concentrate resources on complete submissions.

Therefore, the current 510(k) Refuse to Accept (RTA) policy includes an early review against specific acceptance criteria and to inform the submitter within the first 15 calendar days after receipt of the submission if the submission is administratively complete, or if not, to identify the missing element(s). In order to enhance the consistency of our acceptance decisions and to help submitters better understand the types of information FDA needs to conduct a substantive review, this guidance, including the checklists included in the appendices, clarify the necessary elements and contents of a complete 510(k) submission. The process we outline is applicable to all devices reviewed through the 510(k) notification process and has been compiled into checklists for use by FDA review staff.

It is critical to distinguish between the completeness of the regulatory submission, and the

<sup>1</sup> See Title II of the FDA Reauthorization Act of 2017 (Public Law 115-52).

## ***Contains Nonbinding Recommendations***

quality of the data provided and any studies conducted in support of the submission. The assessment of the completeness of the 510(k) occurs during the acceptance review, while the assessment of the quality of the submitted information occurs during the substantive review. FDA will base acceptance on the objective criteria outlined in the associated Acceptance Checklist and not on the quality of the data.

The contents of this document do not have the force and effect of law and are not meant to bind the public in any way, unless specifically incorporated into a contract. This document is intended only to provide clarity to the public regarding existing requirements under the law. FDA guidance documents, including this guidance, should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidance means that something is suggested or recommended, but not required.

## **II. Background**

The purpose of the 510(k) acceptance review is to assess whether a submission is administratively complete, in that it includes all of the information necessary for FDA to conduct a substantive review and to reach a determination regarding substantial equivalence under section 513(i) of the Federal Food, Drug, and Cosmetic (FD&C) Act (21 United States Code (U.S.C.) § 360c(i)). To find a device substantially equivalent under section 513(i) of the FD&C Act, FDA must find that it has the same intended use as the predicate device, and either (1) has the same technological characteristics as the predicate device, or (2) has different technological characteristics, as defined at section 513(i)(1)(B), and the submission contains information, including appropriate clinical or scientific data if necessary, that demonstrates the device is as safe and effective as the predicate and does not raise different questions of safety and effectiveness than the predicate.

The 510(k) regulations at 21 CFR 807.87 to 807.100 provide greater detail regarding the specific information that each premarket notification submission must contain. For example, the submission must include proposed labeling (807.87(e)), a statement regarding the similarities and differences between the device and others of comparable type (807.87(f)), supporting data (807.87(f) and 807.100(b)(2)(ii)(B)), and FDA may request any additional information necessary to determine whether the device is substantially equivalent when the information provided is insufficient to enable such a determination (807.87(m)). Please also refer to our guidance document entitled, “[Format for Traditional and Abbreviated 510\(k\)s](#).<sup>2</sup>

Prior guidances and checklists relating to 510(k) RTA policy (i.e., 510(k) Refuse to Accept Policy, dated June 30, 1993, and 510(k) Refuse to Accept Procedures (K94-1) blue book memo, dated May 20, 1994) focused on defining broad issues or principles. Additionally, the checklists associated with these guidances dealt largely with administrative elements but did not address specific content that is essential for 510(k) review. As a result, FDA had accepted many inadequate submissions for review, and FDA staff invested significant time in

---

<sup>2</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/format-traditional-and-abbreviated-510ks>.

## ***Contains Nonbinding Recommendations***

constructing extensive letters requesting all of the additional information needed to conduct a substantive review. This approach was an inefficient use of resources and frequently lengthened review times. For additional information see CDRH's "[Analysis Of Premarket Review Times Under The 510\(k\) Program](#)."<sup>3</sup>

The goal of the guidance titled "Refuse to Accept Policy for 510(k)s," dated December 31, 2012 was to clarify the content needed in Traditional, Abbreviated, and Special 510(k) submissions to allow FDA to conduct a substantive review, thereby enhancing the quality of received 510(k) submissions and improving overall review time. The review process presented in this document is captured in the Acceptance Checklists for Traditional, Abbreviated, and Special 510(k) submissions, which FDA staff will use during the acceptance review process.

## **III. Scope**

The information presented in this document is intended to provide FDA staff with a clear, consistent approach for acceptance review for Traditional, Abbreviated, and Special 510(k) notifications and to outline the RTA policy on 510(k)s. It will also help submitters prepare 510(k) notifications that are administratively complete for FDA to conduct a substantive review.

The acceptance policy does not alter the substantial equivalence decision-making process once the submission has been accepted for review; however, it does alter the start of the FDA review clock for purposes of MDUFA performance goals for those submissions that are not accepted for review. More information regarding FDA's review clock is provided in Section V of this document.

This document does not address the monetary aspects or the MDUFA goals associated with 510(k)s. Information pertaining to the fees and payment procedures for submission of a 510(k) notification can be found in FDA's guidance document "[User Fees and Refunds for Premarket Notification Submissions \(510\(k\)s\)](#)."<sup>4</sup>

## **IV. Q-submission Interaction**

For general information regarding the 510(k) regulations under 21 CFR Part 807, submitters should consult CDRH's Division of Industry and Consumer Education (DICE) or CBER's Manufacturers Assistance and Technical Training Branch. Before submitting a 510(k) notification, we encourage submitters, especially those who are less familiar with the 510(k) review program or who have novel issues to address, to interact with the appropriate FDA review staff. Such Q-submission interaction is an important way of improving the quality and completeness of a 510(k). For additional information regarding the Q-Submission process, please refer to the guidance titled "[Requests for Feedback and Meetings for Medical Device](#)

---

<sup>3</sup> <https://www.fda.gov/media/80972/download>.

<sup>4</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/user-fees-and-refunds-premarket-notification-submissions-510ks>.

**Submissions: The Q-Submission Program.<sup>5</sup>**

In addition, other FDA guidance documents and resources provide valuable information for preparing 510(k)s, including:

- “[Format for Traditional and Abbreviated 510\(k\)s](#);”<sup>6</sup>
- “[The Special 510\(k\) Program](#);”<sup>7</sup>
- “[The Abbreviated 510\(k\) Program](#);”<sup>8</sup>
- “[The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \[510\(k\)\]](#);”<sup>9</sup>
- “[eCopy Program for Medical Device Submissions](#);”<sup>10</sup>
- “[Types of Communication During the Review of Medical Device Submissions](#);”<sup>11</sup>
- “[Intent to Exempt Certain Unclassified Medical Devices from Premarket Notification Requirements](#);”<sup>12</sup>
- Other applicable [device-specific and cross-cutting guidance documents](#),<sup>13</sup> and
- [CDRH Device Advice](#).<sup>14</sup>

## **V. 510(k) Refuse to Accept Policies and Procedures**

FDA staff conduct acceptance reviews of Traditional, Abbreviated, or Special 510(k)s based on objective criteria using the applicable Acceptance Checklist (see Appendices A-C) to ensure that the 510(k) is administratively complete. In order for the submission to be accepted, all administrative elements identified as RTA items should be present or a rationale should be provided for those elements determined by the submitter to be not applicable. To aid in the administrative review, it is recommended that submitters complete and submit acceptance checklists with their submissions that identify the location of supporting information for each RTA element.

The acceptance review occurs prior to the substantive review and should be conducted and completed within 15 calendar days of FDA receiving the 510(k) notification. An acceptance review will only begin for 510(k) submissions for which the applicable user fee has been

---

<sup>5</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>.

<sup>6</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/format-traditional-and-abbreviated-510ks>.

<sup>7</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/special-510k-program>.

<sup>8</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/abbreviated-510k-program>.

<sup>9</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k>.

<sup>10</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions>.

<sup>11</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/types-communication-during-review-medical-device-submissions>.

<sup>12</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/intent-exempt-certain-unclassified-medical-devices-premarket-notification-requirements>.

<sup>13</sup> <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/guidance-documents-medical-devices-and-radiation-emitting-products>.

<sup>14</sup> <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>.

## ***Contains Nonbinding Recommendations***

paid and a valid eCopy has been received.<sup>15</sup>

The staff will select the applicable checklist based on the 510(k) type (i.e., Traditional, Abbreviated, or Special). The acceptance review will be conducted on original 510(k) submissions and responses to RTA communications, but not supplements or amendments submitted in response to requests for additional information after a submission has been accepted. The staff should assess whether the submission should be accepted by first answering the preliminary questions below, and then verifying that the submission contains all of the information identified as RTA items in the checklist.

The purpose of the 510(k) acceptance review is to assess whether a submission is administratively complete, in that it includes all of the information necessary for FDA to conduct a substantive review. Therefore, the submission generally should not be accepted and should receive an RTA designation if one or more of the items noted as RTA items in the checklist are not present and no explanation is provided for the omission(s). However, during the RTA review, FDA staff has discretion to determine whether missing checklist items are needed to ensure that the submission is administratively complete to allow the submission to be accepted. FDA staff also has discretion to request missing checklist items interactively from submitters during the RTA review. Interaction during the RTA review is dependent on FDA staff's determination that outstanding issues are appropriate for interactive review and that adequate time is available for the submitter to provide supporting information and for FDA staff to assess responses before the Acceptance deadline of 15 days.

If one or more items noted as RTA items on the Acceptance Checklist are not present and FDA staff conducting the acceptance review determine that the submission should be refused, FDA staff should obtain management concurrence and notify the designated 510(k) contact person electronically<sup>16</sup> that the submission has not been accepted.<sup>17</sup> FDA staff should also provide the submitter with a copy of the completed checklist indicating which item(s) are the basis for the RTA designation.

The 510(k) submitter may respond to the RTA notification by providing the missing information identified in the checklist. When providing the missing information, the submitter should submit this information to be included in the file under the originally assigned 510(k) number. A new submission and new user fee are not necessary. Nor is it

---

<sup>15</sup> See sections 738(a)(2)(A)(viii) and 738(f)(1) of the FD&C Act, section 745A(b) of the FD&C Act, and the FDA guidance, “[eCopy Program for Medical Device Submissions](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions>. Additional information is also provided in the guidance “[FDA and Industry Actions on Premarket Notification \(510\(k\)\) Submissions: Effect on FDA Review Clock and Goals](#),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/fda-and-industry-actions-premarket-notification-510k-submissions-effect-fda-review-clock-and-goals>.

<sup>16</sup> For additional information about email communications with CBER, please see [SOPP 8119: Use of Email for Regulatory Communications](#), available at <https://www.fda.gov/media/108992/download>.

<sup>17</sup> As outlined in the commitment letter for MDUFA III [158 CONG. REC. S8277-S8281 (daily ed. Corrected December 20, 2012) (Letters from the Secretary of Health and Human Services Re: Medical Device User Fee Program), also available at <https://www.fda.gov/media/83244/download>], the review clock will not start until the 510(k) submission is accepted for review.

## ***Contains Nonbinding Recommendations***

necessary to re-send the entire 510(k) submission, unless FDA notes otherwise (e.g., because the majority of the submission is not in English, or the submission is missing the majority of the items on the checklist). It is sufficient to submit and address only the information requested in the RTA notification. If a complete response to the RTA notification is not received within 180 days of the date of RTA notification, FDA will consider the 510(k) to be withdrawn and the submission will be closed in the system.

Upon receipt of newly submitted information in response to an RTA notification, FDA staff should conduct the acceptance review again following the same procedure within 15 calendar days of receipt of the new information. This acceptance review will assess whether the new information makes the submission complete according to the checklist criteria for completeness. If the submission is still found to be incomplete, FDA staff should notify the contact person and provide a copy of the new checklist indicating the missing item(s).

When a submission is accepted, FDA staff should electronically notify the submission contact person that the 510(k) has been accepted and begin a substantive review of the submission to determine substantial equivalence. Should FDA fail or choose not to complete the acceptance review within the acceptance review period (i.e., within 15 calendar days of receipt), the submitter should be electronically notified that the acceptance review was not completed and the submission is under substantive review. FDA may request any information that may have resulted in an RTA designation during the substantive review.<sup>18</sup> Once a submission has been accepted, FDA may ask for any information during the substantive review that may have been unintentionally overlooked during the acceptance review.

### *FDA Review Clock*

As explained in the commitment letter for MDUFA IV referenced in Title II of the FDA Reauthorization Act of 2017 (Public Law 115-52), “FDA days begin on the date of receipt of the submission or of the amendment to the submission that enables the submission to be accepted (510(k)) or filed (PMA).”<sup>19</sup> As stated above, an acceptance review will only begin for 510(k) submissions for which the applicable user fee has been paid and a valid eCopy has been received. Thus, the FDA review clock does not start when a submission is placed on eCopy or User Fee hold or designated RTA.

510(k) submissions and additional information submitted in response to a RTA designation are received by the respective Center’s Document Control Center (DCC). The FDA review clock start date is the DCC receipt date of the most recent submission or additional information that resulted in an acceptance designation for the 510(k), provided the submission user fee has been paid and a validated eCopy has been provided. For example, if the

---

<sup>18</sup> In the case of a government closure during the 15-day acceptance review period, the review period may be extended by a comparable number of business days that the FDA buildings are closed. If the submitter receives an automated notice that the acceptance review was not completed because the screening period has exceeded 15 days, FDA may send a correction notice to the submitter.

<sup>19</sup> See 163 CONG. REC. S4729-S4736 (daily ed. August 2, 2017) (Food and Drug Administration User Fee Reauthorization), also available at <https://www.fda.gov/media/102699/download>.

### ***Contains Nonbinding Recommendations***

submission is accepted for substantive review on the first acceptance review, the FDA review clock start date is the DCC receipt date of the submission. However, if the submission is designated RTA, the FDA review clock start date is not yet known. In such cases, the clock start date will be the DCC receipt date of the submission including the additional information that results in an acceptance designation (even if FDA later requests information that should have been requested during acceptance review). In the event the acceptance review was not completed within 15 calendar days, the submission will be considered to be under substantive review, and the FDA review clock start date will be the DCC receipt date of the most recently received information that was the subject of the acceptance review for the submission. Once the submission is under substantive review the calendar days used to conduct the acceptance review (i.e., up to 15 days) are included within the 60 calendar days to reach the Substantive Interaction goal as described in the aforementioned commitment letter for MDUFA IV.

#### *Notification of Acceptance Review Result*

The submitter should receive an electronic notification of the acceptance review result within 15 calendar days of DCC receipt (i.e., that the submission has been accepted for substantive review, that the submission is not accepted for review (RTA), or that the submission is now under substantive review because the acceptance review was not completed). The notification will be sent only to the designated contact person identified in the submission. This notification will also serve to identify the FDA 510(k) lead reviewer<sup>20</sup> assigned to the submission. The notification of either the acceptance or RTA designation will be made only with supervisory concurrence of the reviewer's acceptance review determination. The notification of acceptance or RTA designation may occur on any day prior to the 15<sup>th</sup> calendar day of DCC receipt. However, in the event the acceptance review was not completed within 15 calendar days, a notification that an RTA review was not completed will be sent on the 16<sup>th</sup> day. In the case of RTA designation, the notification should be accompanied by the completed checklist indicating the missing elements that resulted in the RTA designation. The completed checklists are considered part of the submission's administrative file and are not posted publicly when FDA makes the RTA designation. Therefore, it is imperative that the submission identify complete contact information, including the email address to which the notification should be sent.<sup>21</sup>

## **VI. Refuse to Accept Principles**

In order to use this guidance appropriately, FDA staff should review the following basic principles regarding FDA's review policies and procedures.

Acceptance should not be based on a substantive review of the information provided in the 510(k) notification.

---

<sup>20</sup> In the case of 510(k)s submitted to CBER, whenever the term lead reviewer is used in this guidance, the equivalent CBER contact person is the regulatory project manager (RPM).

<sup>21</sup> CBER will accommodate the use of faxes; submitters may also wish to provide a fax number.

### ***Contains Nonbinding Recommendations***

It is important to make the distinction between the acceptance review and the substantive review. The acceptance review is conducted to assess whether the submission contains all of the appropriate elements, as identified in the applicable checklist, in order to begin a substantive review. In assessing whether a 510(k) notification should be accepted, submitted information is not evaluated for adequacy to support a finding of substantial equivalence. The checklist is a tool to ensure that the submission contains the necessary information in order to conduct a substantive review (i.e., FDA should not refuse to accept a submission if information is present but inadequate to support a finding of substantial equivalence). The evaluation of the quality of the content and the substantial equivalence decision making process occur within the substantive review once the file has been accepted.

FDA staff should determine whether the submitter provided a justification for any alternative approach.

The submitter may provide a rationale for why any criteria in the checklist are not applicable to the device. Likewise, the submitter may provide a rationale for any deviation from a device- specific or cross-cutting guidance document or FDA-recognized standard. It is FDA’s expectation that each item in the checklist will be addressed either by including the requested information or providing a rationale for why it is not applicable or why there is a deviation.

FDA will not consider a given criterion in the checklist to be “Present” if the submission fails to include either the information identified or a rationale for omission or deviation. If a justification to omit certain information or for taking an alternative approach is provided, FDA will consider the adequacy of that justification or alternative approach during substantive review of the submission. See Acceptance Review section below for examples and further explanation.

Device-specific and cross-cutting guidance documents, applicable recognized standards, and applicable regulations will be considered when making an RTA determination.

Before submitting a 510(k), the submitter should consider the currently available guidance documents and standards, as well as applicable regulations for the proposed device in the preparation of the submission. FDA staff and industry are encouraged to refer to the [product classification database](#)<sup>22</sup> to assist in identifying any applicable recognized consensus standards. If citing voluntary consensus standards, the submitter should consider FDA’s guidance “[Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices](#).”<sup>23</sup>

Specifically, the checklist includes questions regarding whether the submission has addressed recommendations regarding the device description, labeling, and performance testing as outlined in a device-specific guidance, special controls or another specific regulation. Note that “addressed” means that the submission includes information pertinent to those

---

<sup>22</sup> <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPCD/classification.cfm>.

<sup>23</sup> <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices>.

### ***Contains Nonbinding Recommendations***

recommendations or requirements; assessment of the adequacy of that information in meeting those recommendations or requirements should be assessed during review.

If there is a device-specific guidance, other than a special controls guidance document, the submission includes information to establish that the submitter has addressed the recommendations or otherwise provided an alternative approach intended to address the applicable statutory and/or regulatory criteria.

If there are special controls applicable to the device, the submission includes information addressing the particular mitigation measures set forth in the special controls, or uses alternative mitigation measures and provides a rationale to demonstrate that those alternative measures identified by the submitter will provide at least an equivalent assurance of safety and effectiveness.

## **VII. The Checklist – Preliminary Questions**

Within 15 calendar days of receipt of the 510(k), FDA staff should answer the preliminary questions below, which are included on the first page of the Acceptance Checklists for Traditional and Abbreviated 510(k)s. The preliminary questions are intended to be answered by the 510(k) lead reviewer as an initial screening of the submission. FDA does not intend for the submitter to have addressed these items in their submission. Depending upon the answers to these preliminary questions, the remainder of the acceptance review may or may not be necessary.

If the responses to the preliminary questions and subsequent consultation with the Center personnel identified below indicate that the 510(k) acceptance review should not continue<sup>24</sup> the 510(k) lead reviewer should promptly:

- inform the 510(k) review team (including consulting reviewers); and
- notify the submitter using proper administrative procedures.

The preliminary questions are:

**1. Is the product a device (per section 201(h) of the FD&C Act) or a combination product (per 21 CFR 3.2(e)) with a device constituent part subject to review in a 510(k)?**

If the product does not appear to meet the definition of a device under section 201(h) of the FD&C Act, or does not appear to be a combination product with a device constituent part, then the 510(k) lead reviewer should consult with the CDRH Product Jurisdiction

---

<sup>24</sup> FDA will not process a 510(k) unless it meets the following requirements: i) the submission must be sent with the user fee required by section 738 of the FD&C Act, and ii) a valid eCopy is provided. See section 738(a)(2)(A)(viii) and 738(f)(1) of the FD&C Act, section 745A(b) of the FD&C Act, and the FDA guidance, “[eCopy Program for Medical Device Submissions](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions>. Because any 510(k) not meeting these two requirements will not be processed by the respective Center’s DCC, they are not included in the checklist.

***Contains Nonbinding Recommendations***

Officer or the CBER Product Jurisdiction Officer to determine the appropriate action, and inform management. If FDA staff determines that the product is not a device and is not a combination product with a device constituent part, the 510(k) lead reviewer should stop the review and notify the submitter.

**2. Is the submission with the appropriate Center?**

If the submission is for a single-entity device and appears to be subject to review in a Center different from the one to which it was submitted, or if it is for a combination product with a device constituent part and it appears that a Center different from the one to which it was submitted has the lead, the 510(k) lead reviewer should consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action and inform management. If the 510(k) is submitted to CDRH and CDRH staff determines that the submission is not subject to CDRH review, or the 510(k) is submitted to CBER and CBER staff determines that the submission is not subject to CBER review, the 510(k) review team should stop the review and notify the submitter.

**3. If a Request for Designation (RFD) was submitted for the device or combination product with a device constituent part and assigned to your center, identify the RFD # and confirm the following:**

- (a) Is the device or combination product the same (e.g., design, formulation) as that presented in the RFD submission?**
- (b) Are the indications for use for the device or combination product identified in the 510(k) the same as those identified in the RFD submission?**

An RFD determination is specific to the device or combination product and indications for use for the device or combination product described in the RFD submission. If the device or combination product has been modified or the indications for use have been modified since the RFD, the RFD determination may no longer be applicable and jurisdiction may need to be reevaluated by the Office of Combination Products (OCP). The 510(k) lead reviewer should consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action and inform management.

**4. Is the submission for a combination product that contains as a constituent part a drug that has the same active moiety as an approved drug with exclusivity as described in section 503(g)(5)(C)(ii)-(v) of the FD&C Act?**

If the submission is for a combination product and contains as a constituent a drug that has the same active moiety as an approved drug with exclusivity as described in section 503(g)(5)(C)(ii)-(v) of the FD&C Act, the 510(k) lead reviewer should contact the CDRH Product Jurisdictional Officer or CBER Product Jurisdiction Officer to determine the appropriate action and inform management.

**5. Is this device type eligible for a 510(k) submission?**

FDA staff should determine whether the 510(k) submission is for a device type for which 510(k) is known to be an inappropriate regulatory approach, such as when the device type is class III type and a PMA is required, or the device type is class I or II and 510(k)-exempt. If a 510(k) is not appropriate, FDA staff should make this determination during the acceptance review and notify the submitter of the determination. This preliminary question is not intended to identify submissions for which a substantive review is required in order to determine if 510(k) is an inappropriate approach (e.g., device has a new intended use or device has different technological characteristics that raise different questions of safety and effectiveness).

**6. Is there a pending PMA for the same device with the same indications for use?**

If the submitter has a PMA for the same device with the same indications for use pending, the review team should stop the review. The 510(k) review team should consult management and other Center resources to determine which premarket review pathway applies to the device and the appropriate processes for addressing the situation. FDA staff should also consult management and other Center resources if a 510(k) and PMA have been submitted for the same device type by different submitters.

**7. If clinical studies have been submitted, is the submitter the subject of the Application Integrity Policy (AIP)?<sup>25</sup>**

The 510(k) lead reviewer should refer to the AIP list.<sup>26</sup> If the submitter is on the list, the reviewer should consult the CDRH Office of Product Evaluation and Quality/Office of Clinical Evidence and Analysis/Division of Clinical Science and Quality (OPEQ/OCEA/DCEA1) or CBER Office of Compliance and Biologics Quality/Division of Inspections and Surveillance/ Bioresearch Monitoring Branch (OCBQ/DIS/BMB) to determine the appropriate action.

## **VIII. The Checklists – Acceptance Review**

### *Organizational Elements*

Although missing one or more of the items in the table of Organizational Elements in the Acceptance Checklists, such as a Table of Contents or page numbers, generally will not lead to an RTA decision, we strongly encourage submitters to incorporate these elements in their submissions to facilitate FDA review and decision-making. If, however, the submission is so disorganized that FDA cannot locate the information needed to assess substantial

---

<sup>25</sup> When data in a pending submission have been called into question by certain wrongful acts (fraud, untrue statements of material facts, bribery, or illegal gratuities), FDA intends to defer substantive scientific review of such data until completion of a validity assessment and questions regarding reliability of the data are resolved. (See FDA Guide 7150.09 Compliance Policy Guide, Chapter 50 – General Policy – Subject: Fraud, Untrue Statements of Material Facts, Bribery, and Illegal Gratuities, 56 FR 46191).

<sup>26</sup> <https://www.fda.gov/inspections-compliance-enforcement-and-criminal-investigations/application-integrity-policy/application-integrity-policy-list>.

## ***Contains Nonbinding Recommendations***

equivalence, or if the submission is so poorly written (e.g., in broken English) that the information submitted to support substantial equivalence cannot be understood, the submission should receive an RTA decision.

### *Elements of a Complete Submission (RTA Items)*

The objective criteria in these checklists outline those elements that are explicitly required by regulation or that are essential to FDA's substantive review of the submission and determination of substantial equivalence under section 513(i) of the FD&C Act. For example, proposed labels, labeling, and instructions are required by 21 CFR 807.87(e)), while a description of the materials, design, and other features of the device is essential to determining whether its technological characteristics are the same as those of the predicate and whether any differences raise different questions of safety and effectiveness under section 513(i) of the FD&C Act.

We have also identified several categories and subcategories of data and information that, when applicable, are critical to supporting a statement indicating the device is similar to and/or different from other products of comparable type under 21 CFR 807.87(f) and the substantial equivalence determination. For example, if the new device has direct or indirect tissue-contacting components, a biocompatibility assessment will be essential to evaluating whether the new device is as safe as the predicate with respect to the risk of toxicity it poses to the patient. While testing and data would usually be necessary for such an assessment, this is not always the case (for example if the device under review and the predicate are identical in all relevant respects), and acceptance should be based only on the presence of an item or an explanation why the item is not applicable, not the adequacy of such explanation. If the device has no direct or indirect tissue-contacting components, no biocompatibility assessment would be necessary and the biocompatibility items on the checklist would be not applicable.

Because the applicability of these categories is also critical to the substantial equivalence determination, in order to be accepted, all submissions should include a statement indicating whether these categories apply, as outlined in the Acceptance Checklist (e.g., materials, presence of software, whether the device is intended to be used sterile). When performance data are provided, the submission of full test reports describing how the testing was conducted is crucial to FDA's assessment of whether the data support a finding of substantial equivalence.

Where a device-specific guidance document exists for the subject device, the submitter should follow the recommendations included in that document, or the submitter should provide a rationale for addressing the scientific issues discussed in the guidance document using an alternative approach intended to address the applicable statutory and/or regulatory criteria. In the absence of the recommended information and without a rationale for an alternative approach, the submission should be considered incomplete and not accepted. If special controls have been identified, those controls should be addressed in order for the submission to be accepted, or alternative mitigation measures providing a rationale to demonstrate that those alternative measures will provide at least an equivalent assurance of

## ***Contains Nonbinding Recommendations***

safety and effectiveness should be identified.

### *Applying the Checklist of RTA Items*

Using the Acceptance Checklist appropriate to the submission type (traditional, abbreviated, or special), within 15 calendar days of receipt of the 510(k), FDA staff should answer each question for the elements identified as RTA items. For those items that have an option of “yes,” “no,” or “not applicable (N/A)” as an answer, the item should receive an answer of “yes” or “N/A” for the 510(k) submission to be accepted for substantive review.

For the first question in each section related to the need for certain performance data (such as biocompatibility, sterilization, software), FDA staff should indicate whether the submission has addressed one of the options for the 510(k) submission to be accepted for substantive review. For example, the submission should state explicitly that either there are or are not direct or indirect (e.g., through fluid infusion) tissue-contacting components in order for the submission to be considered complete and accepted for substantive review.

### *Elements Marked “Not applicable”*

In developing the checklists, the Agency has considered the general categories and respective subcategories of information that are necessary to conduct a substantive review for the wide range of medical devices that are appropriate for review under 510(k). All such criteria may not be pertinent to a particular device. FDA staff should select “N/A” for those elements that do not apply to the subject device. For example, the requirements for financial certification and disclosure statements (21 CFR 807.87(i)) and statements of compliance for clinical investigations (21 CFR 807.87(j)) only apply to submissions with clinical data. If the submission contains no clinical data, FDA staff should select “N/A.”

### *Adequacy of information*

In order to make the checklist criteria objective, for each RTA item, FDA should consider only the presence or omission of the element or a rationale for the omission of the element or use of an alternative approach during acceptance review. It is likely that FDA staff will encounter scenarios where information is provided, but is incomplete or inadequate. In such instances, FDA staff should answer the question for the respective item as “Yes,” but may communicate the inadequacy or request additional information in the course of the substantive review. For example, the submitter may have provided full test reports for all performance testing; however, during the acceptance review, the reviewer may note that the *results* of a particular test may not be sufficient to support a finding of substantial equivalence and additional justification would be needed. The performance testing criterion would be marked “Yes” in the checklist, and the full assessment of the results and communication to the submitter that additional justification is needed should occur during the substantive review.

During RTA review, issues may be identified that do not determine the acceptability of a submission, but are issues that should be resolved prior to a final decision. This does not

## ***Contains Nonbinding Recommendations***

mean that a complete review of the submission has been conducted. These identified issues are referred to as “observations.” If observations are identified, they will be attached to the RTA checklist that is sent to the official contact. This attachment is called an Addendum. You do not need to provide a response to the observation(s) in your RTA response in order for your file to be considered administratively complete and accepted for substantive review; however, addressing these questions could help to facilitate the substantive review of your submission.

### *Elements Marked “No”*

For any acceptance criterion designated as “No,” FDA intends to provide an explanation to describe the missing element(s), if needed. This explanation is particularly important for a criterion in which it may not be immediately apparent to the submitter what necessary information, specifically, is not present. For example, the Device Description section includes an element that states, “submission addresses device description recommendations outlined in the device- specific guidance document” and a notation of “No” alone may not be sufficient to inform the submitter of what specific piece(s) of information is missing. FDA staff should include a list or statement of the additional information that is necessary to meet the acceptance criteria. This list or statement can be communicated in the “comment” section on the checklist beside each specific criterion.

### *Prior Submissions Relevant to the Submission Under Review*

For certain submissions, the submitter may have made prior submissions for the same device for which FDA provided feedback related to the data or information needed to support substantial equivalence (e.g., a Q-Submission, investigational device exemption (IDE) application, prior NSE determination, prior 510(k) that was deleted or withdrawn). When such prior feedback is relevant to determining whether substantial equivalence of the subject device exists, the new submission should include information to address this prior feedback and the checklists should include criteria related to this issue. To address the criterion regarding whether a prior submission (or no prior submission) exists, FDA recommends that submitters provide this information in Section F (prior related submissions section) of the CDRH Premarket Review Submission Cover Sheet form ([Form 3514](#)).<sup>27</sup> Submitters should list prior submission numbers in Section F of this form or state that there were no prior submissions to address this criterion. Please be advised that leaving this section of the form blank will not be considered a statement that there were no prior submissions. This information may also be included in the Cover Letter (i.e., as a statement that there were no prior submissions for the device or a listing of the numbers(s) of the prior submission(s)). Where one or more prior submissions do exist, FDA suggests designating a separate section of the submission that identifies the prior submission(s) by number, includes a copy of the FDA feedback (e.g., letter, meeting minutes), and states how or where in the submission this prior feedback was addressed. Note that the adequacy of how the feedback was addressed should be assessed during the substantive review.

---

<sup>27</sup> <https://www.fda.gov/media/72421/download>.

## ***Contains Nonbinding Recommendations***

### *Combination Product Administrative Items*

The 21st Century Cures Act, which amended section 503(g) of the FD&C Act, requires submitters seeking action on a combination product to identify the product as such (section 503(g)(8)(C)(v) of the FD&C Act). Additionally, per the amended section 503(g)(5), submissions for device-led, device-drug combination products must include the patent certification or statement as described in section 505(b)(2) and provide notice as described in section 505(b)(3) if the combination product contains as a constituent part an approved drug (see section 503(g)(5)(A) of the FD&C Act). Submitters of products that are not combination products, as defined in 21 CFR 3.2(e), should mark “N/A” and omit this section pertaining to combination products.

### *Submitters of Combination Products That Do Not Contain as a Constituent Part an Approved Drug*

If the combination products do not include as a constituent part an approved drug as defined in section 503(g)(5)(B), submitters of device-led, device-drug combination products should mark “N/A” for question 10 (question 9 in the Special RTA Checklist).

### *Submitters of Combination Products That Contain as a Constituent Part an Approved Drug*

Submitters of combination products containing as a constituent part an approved drug should address question 11 (question 10 in the Special RTA Checklist) by including patent information. For each relevant patent, the submitter should include certification to one of the following certifications:

- i. That such patent information has not been filed (section 505(b)(2)(A)(i)).
- ii. That such patent has expired (section 505(b)(2)(A)(ii)).
- iii. The date on which the patent will expire (section 505(b)(2)(A)(iii)).
- iv. That such patent is invalid or will not be infringed by the manufacture, use, or sale of the drug constituent part for which this submission is made (section 505(b)(2)(A)(iv)).

However, for a method of use patent which does not claim a use for which the submitter is seeking approval, the submitter should include a statement per section 505(b)(2)(B) that the method of use patent does not claim such a use.

Submitters including a certification under paragraph (iv) (section 505(b)(2)(A)(iv)) should also certify that they will provide notice to the owner of the patent(s) and the holder of the approved application that lists the patent(s) that is/are being challenged. The process for giving notice is provided in section 505(b)(3) of the FD&C Act. Submitters should submit to FDA documentation of the date of receipt of notice by holder of the approved application and patent(s) owner.

### *Conversion of Special 510(k) to Traditional 510(k)*

FDA has developed separate checklists to address the differences in content for Special and

### ***Contains Nonbinding Recommendations***

Traditional 510(k) submissions. FDA staff will utilize the appropriate checklist based on the file type as designated by the submitter. In the event that the submitter has submitted a Special 510(k), but FDA determines that the file should be converted to a Traditional 510(k)<sup>28</sup> FDA will notify the contact person designated in the 510(k) submission of the conversion and the rationale for the conversion. If the file is converted from a Special to a Traditional within the 15 calendar day acceptance review period, the Traditional 510(k) Acceptance Checklist will be used to conduct the acceptance review and the review clock start date will be assigned as outlined in the 510(k) Refuse to Accept Policies and Procedures section above. Given the differences in content for Special and Traditional 510(k)s, it is likely that the converted submission will result in an RTA designation using the Traditional Acceptance Checklist. FDA staff should provide the completed Acceptance Checklist for Traditional submissions indicating which elements are missing. The submitter may respond by providing the identified information and the subsequent acceptance review will proceed with the Traditional checklist. If the file is converted from a Special to a Traditional after the 15 calendar day acceptance review period, any missing information that would have resulted in RTA designation should be obtained during the substantive review.

## **IX. Paperwork Reduction Act of 1995**

This guidance contains information collection provisions that are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501-3521).

The time required to complete this information collection is estimated to be 79.25 hours for preparation of a 510(k) submission. Send comments regarding this burden estimate or suggestions for reducing this burden to: FDA PRA Staff, Office of Operations, Food and Drug Administration, [PRAStaff@fda.hhs.gov](mailto:PRAStaff@fda.hhs.gov)

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. The OMB control number for this information collection is 0910-0120 (To find the current expiration date, search for this OMB control number available at <https://www.reginfo.gov>).

---

<sup>28</sup> Please see “Special 510(k) Factors,” items 1-4 of the Acceptance Checklist for Special 510(k)s for potential reasons for conversion.

## Appendix A. Acceptance Checklist for Traditional 510(k)s

**(Should be completed within 15 days of DCC receipt)**

The following information is not intended to serve as a comprehensive review. FDA recommends that the submitter include this completed checklist as part of the application.

**510(k)#:**

**Date Received by DCC:**

**510(k) Lead Reviewer:**

**Center:**

**Office:**

**Division:**

**Decision: Accept \_\_\_\_\_ Refuse to Accept \_\_\_\_\_**

**If Accept, notify the submitter.**

**If Refuse to Accept, notify submitter electronically and include a copy of this checklist.**

**Is an Addendum attached?: Yes No**

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete; it means the reviewer did not assess the element during the RTA review and that the element will be assessed during substantive review.

Preliminary Questions				
Answers in the shaded blocks indicate consultation with an identified Center advisor is needed. (Boxes checked in this section represent FDA's preliminary assessment of these questions at the time of administrative review.)		Yes	No	N/A
1. <b>Is the product a device (per section 201(h) of the FD&amp;C Act) or a combination product (per 21 CFR 3.2(e)) with a device constituent part subject to review in a 510(k)?</b>  If it appears not to be a device (per section 201(h) of the FD&C Act) or such a combination product (per 21 CFR 3.2(e)), or you are unsure, consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action, and inform management. <i>Provide a summary of the Product Jurisdiction Officer's determination/recommendation/action in the comment section below.</i>		<input type="checkbox"/>	<input type="checkbox"/>	

***Contains Nonbinding Recommendations***

If the product does not appear to be a device or such a combination product, mark “No.”	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>Comments:</b>			
<b>2. Is the submission with the appropriate Center?</b>  If the product is a device or a combination product with a device constituent part, is it subject to review by the Center in which the submission was received? If you believe the submission is not with the appropriate Center or you are unsure, consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action and inform your management. <i>Provide a summary of the Product Jurisdiction Officer's determination/recommendation/action in the comment section below.</i>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
If submission should not be reviewed by your Center mark “No.”			
<b>Comments:</b>			
<b>3. If a Request for Designation (RFD) was submitted for the device or combination product with a device constituent part and assigned to your center, identify the RFD # and confirm the following:</b>  (a) Is the device or combination product the same (e.g., design, formulation) as that presented in the RFD submission? (b) Are the indications for use for the device or combination product identified in the 510(k) the same as those identified in the RFD submission?  If you believe the product or the indications presented in the 510(k) have changed from the RFD, or you are unsure, consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action and inform your management. <i>Provide a summary of Product Jurisdiction Officer's determination/recommendation/action in the comment section below.</i>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
If the answer to either question above is no, mark “No.” If there was no RFD, mark “N/A.”			
<b>Comments:</b>			
<b>4. Is the submission for a combination product that contains as a constituent part a drug that has the same active moiety as an approved drug with exclusivity as described in section 503(g)(5)(C)(ii)-(v) of the FD&amp;C Act?</b>  If “Yes,” then contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer to determine the appropriate action and inform your management. <i>Provide the summary of the Product Jurisdiction Officer's determination/recommendation/action in the comment section below.</i>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<b>Comments:</b>			

*Contains Nonbinding Recommendations*

<p><b>5. Is this device type eligible for a 510(k) submission?</b></p> <p>If a 510(k) does not appear to be appropriate (e.g., class III type and PMA required, or class I or II type and 510(k)-exempt), consult with the appropriate CDRH or CBER staff during the acceptance review, provide a summary of the discussion with them, and indicate their recommendation/action in the comment section below. If 510(k) is not the appropriate regulatory submission, mark "No."</p>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>Comments:</b>			
<p><b>6. Is there a pending PMA for the same device with the same indications for use?</b></p> <p>If "Yes," consult your management and CDRH Office of Product Evaluation and Quality/Office of Regulatory Programs/Division of Regulatory Programs 1 (Submission Support) (OPEQ/ORP/DRP1) or appropriate CBER staff to determine the appropriate action.</p>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>Comments:</b>			
<p><b>7. If clinical studies have been submitted, is the submitter the subject of an Application Integrity Policy (AIP)?</b></p> <p>If "Yes," consult with the CDRH Office of Product Evaluation and Quality/Office of Clinical Evidence and Analysis/Division of Clinical Science and Quality (OPEQ/OCEA/DCEA1) or CBER Office of Compliance and Biologics Quality/Division of Inspections and Surveillance/Bioresearch Monitoring Branch (OCBQ/DIS/BMB) to determine the appropriate action, provide a summary of the discussion with them, and indicate their recommendation/action.</p> <p>If no clinical studies have been submitted, mark "N/A." Check on the AIP list at <a href="https://www.fda.gov/inspections-compliance-enforcement-and-criminal-investigations/application-integrity-policy/application-integrity-policy-list">https://www.fda.gov/inspections-compliance-enforcement-and-criminal-investigations/application-integrity-policy/application-integrity-policy-list</a>.</p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
<b>Comments:</b>			

- If the answer to 1 or 2 appears to be "No," then stop review of the 510(k) and contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.
- If the answer to 3a or 3b appears to be "No," then stop the review and contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.
- If the answer to 4 is "Yes," then contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer, provide a summary of the discussion with them, and indicate their recommendation/action.
- If the answer to 5 is "No", the lead reviewer should consult division management and other Center resources to determine the appropriate action. Note that, for a device which is clearly ineligible for a 510(k) submission (such as a device type which is class III requiring PMA or

*Contains Nonbinding Recommendations*

class I/II and 510(k) exempt), this may be considered a basis for a refusal to accept the submission. A 510(k) submitted for a class I/II, 510(k)-exempt device that trips the limitations of the exemption would not be refused on this basis.

- If the answer to 6 is “Yes,” then stop review of the 510(k), contact CDRH/OPEQ/ORP/DRP1, or appropriate CBER staff.
- If the answer to 7 is “Yes,” then contact CDRH/OPEQ/OCEA/DCEA1 or CBER/OCBQ/DIS/BMB, provide a summary of the discussion with DCEA1 or BMB Staff, and indicate their recommendation/action.

<b>Organizational Elements</b>			
Failure to include these items should not result in an RTA designation.			
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.		Yes	No
*Page #			
1.	Submission contains a Table of Contents.	<input type="checkbox"/>	<input type="checkbox"/>
2.	Each section is labeled (e.g., headings or tabs designating Device Description section, Labeling section, etc.).	<input type="checkbox"/>	<input type="checkbox"/>
3.	All pages of the submission are numbered. <i>All pages should be numbered in such a manner that information can be referenced by page number. This may be done either by consecutively numbering the entire submission, or numbering the pages within a section (e.g., 12-1, 12-2...).</i>	<input type="checkbox"/>	<input type="checkbox"/>
4.	Type of 510(k) is identified (i.e., Traditional, Abbreviated, or Special) <i>If type of 510(k) is not designated, review as a Traditional 510(k).</i>	<input type="checkbox"/>	<input type="checkbox"/>
Comments:			

<b>Elements of a Complete Submission (RTA Items)</b> <b>(21 CFR 807.87 unless otherwise indicated)</b>
Submission should be designated RTA if not addressed.
<ul style="list-style-type: none"><li>• Any “No” answer will result in a “Refuse to Accept” decision; however, FDA staff has discretion to determine whether missing items are needed to ensure that the submission is administratively complete to allow the submission to be accepted or to request missing checklist items interactively from submitters during the RTA review.</li><li>• Each element on the checklist should be addressed within the submission. The submitter may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criterion is considered present (Yes). An assessment of the rationale will be considered during the review of the submission.</li></ul>

*Contains Nonbinding Recommendations*

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>			Yes	No	N/A	*Page #
<b>A. Administrative</b>						
1.	All content used to support the submission is written in English (including translations of test reports, literature articles, etc.).		<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:					
2.	Submission identifies the following (FDA recommends use of the CDRH Premarket Review Submission Cover Sheet form ( <a href="#">Form 3514</a> , available at <a href="https://www.fda.gov/media/72421/download">https://www.fda.gov/media/72421/download</a> )):					
	a. Device trade/proprietary name		<input type="checkbox"/>	<input type="checkbox"/>		
	b. Device class and panel OR Classification regulation OR Statement that device has not been classified with rationale for that conclusion		<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:					
3.	Submission contains an Indications for Use Statement with Rx and/or OTC designated (see also 21 CFR 801.109, and FDA’s final rule, “ <a href="#">Use of Symbols in Labeling</a> ” (81 FR 38911), available at <a href="https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling">https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling</a> ). <i>See recommended format</i> ( <a href="https://www.fda.gov/media/86323/download">https://www.fda.gov/media/86323/download</a> ).		<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:					
4.	Submission contains a 510(k) Summary or 510(k) Statement. <i>Refer to 21 CFR 807.92 and 21 CFR 807.93 for contents of 510(k) Summary and Statement, respectively. Adequacy of the content will be assessed during substantive review.</i>		<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:					

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.			Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>						
	5.	Submission contains a Truthful and Accuracy Statement per 21 CFR 807.87(l). <i>See recommended format (<a href="https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-truthful-and-accurate-statement">https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-truthful-and-accurate-statement</a>).</i>	<input type="checkbox"/>	<input type="checkbox"/>		
Comments:						
	6.	Submission is a class III 510(k) Device. <i>Select “N/A” only if submission is not a class III 510(k).</i>	<input type="checkbox"/>		<input type="checkbox"/>	
	a.	Contains class III Summary and Certification per 21 CFR 807.87(k). <i>See recommended content (<a href="https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-class-iii-certification-and-summary">https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-class-iii-certification-and-summary</a>). Select “N/A” only if submission is not a class III 510(k).</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:						
	7.	Submission contains clinical data. <i>Select “N/A” if the submission does not contain clinical data. If “N/A” is selected, parts a, b, and c below are omitted from the checklist.</i>	<input type="checkbox"/>		<input type="checkbox"/>	
	a.	Submission includes completed Financial Certification ( <a href="https://www.fda.gov/media/70465/download">FDA Form 3454</a> , available at <a href="https://www.fda.gov/media/70465/download">https://www.fda.gov/media/70465/download</a> ) or Disclosure ( <a href="https://www.fda.gov/media/69872/download">FDA Form 3455</a> , available at <a href="https://www.fda.gov/media/69872/download">https://www.fda.gov/media/69872/download</a> ) information for each covered clinical study included in the submission. <i>Select “N/A” if the submitted clinical data is not a “covered clinical study” as defined in the guidance entitled “<a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators">Financial Disclosures by Clinical Investigators</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators</a>.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

***Contains Nonbinding Recommendations***

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>				Yes	No	N/A	*Page #
		b.	<p>Submission includes completed Certification of Compliance with requirements of ClinicalTrials.gov Data Bank (see FDA Form 3674 which can be obtained at <a href="https://www.fda.gov/about-fda/reports-manuals-forms/forms">https://www.fda.gov/about-fda/reports-manuals-forms/forms</a>) (42 U.S.C. 282(j)(5)(B)) for each applicable device clinical trial included in the submission.</p> <p><i>Select “N/A” if the submitted clinical data is not an “applicable device clinical trial” as defined in <a href="#">Title VIII of FDAAA, Sec. 801(j)</a>.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		c.	<p>Statements of Compliance for Clinical Investigations</p> <p><i>Select “N/A” if the submission does not contain any clinical data from investigations (as defined in 21 CFR 812.3(h)) to demonstrate substantial equivalence.</i></p> <p><i>For multicenter clinical investigations involving both United States (US) and outside United States (OUS) sites, part (i) should be addressed for the US sites and part (ii) should be addressed for the OUS sites. 21 CFR 812.28 applies to all OUS clinical investigations that enroll the first subject on or after February 21, 2019.</i></p> <p><i>Please refer to the guidance document entitled “<a href="#">Acceptance of Clinical Data to Support Medical Device Applications and Submissions - Frequently Asked Questions</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/acceptance-clinical-data-support-medical-device-applications-and-submissions-frequently-asked">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/acceptance-clinical-data-support-medical-device-applications-and-submissions-frequently-asked</a> for more information.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
		i.	<p>For each clinical investigation conducted in the US, the submission includes a statement that the investigation was conducted in compliance with 21 CFR parts 50, 56, and 812 (or, with respect to part 56, that it was not subject to the regulations under 21 CFR 56.104 or 56.105).</p> <p><b>OR</b></p> <p>The submission includes a brief statement of the reason for noncompliance with 21 CFR parts 50, 56, and/or 812.</p> <p><i>Select “N/A” if the clinical investigations were conducted solely OUS.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		ii.	<p>For each clinical investigation conducted OUS, the submission includes a statement that the clinical investigations were conducted in accordance with good clinical practice (GCP) as described in 21 CFR 812.28(a)(1).</p> <p><b>OR</b></p> <p>The submission includes a waiver request in accordance with 21 CFR 812.28(c).</p> <p><b>OR</b></p> <p>The submission includes a brief statement of the reason for not conducting the investigation in accordance with GCP and a description of steps taken to ensure that the data and results are credible and accurate and that the rights, safety, and well-being of subjects have been adequately protected.</p> <p><i>Select “N/A” if the clinical investigations were conducted solely inside the US.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
			Comments:				

***Contains Nonbinding Recommendations***

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>			Yes	No	N/A	*Page #	
	8.	<p>The submission identifies prior submissions for the same device included in the current submission (e.g., submission numbers for a prior not substantially equivalent [NSE] determination, prior deleted or withdrawn 510(k), Q-Submission, IDE, PMA, etc.).</p> <p><b>OR</b></p> <p>States that there were no prior submissions for the subject device.</p> <p><i>Prior submissions (or no prior submissions) for this device should be included in Section F (prior related submissions) of the CDRH Premarket Review Submission Cover Sheet form (<a href="#">Form 3514</a>, available at <a href="https://www.fda.gov/media/72421/download">https://www.fda.gov/media/72421/download</a>). This information may also be included in the Cover Letter (i.e., as a statement that there were no prior submissions for the device or a listing of the number(s) of the prior submissions).</i></p>	<input type="checkbox"/>	<input type="checkbox"/>			
	a.	<p>If there were prior submissions, the submitter has identified where in the current submission any issues related to a determination of substantial equivalence from prior submissions for this device are addressed.</p> <p><i>To address this criterion, it is recommended that the submission include a separate section with the prior submission number(s), a copy of the FDA feedback (e.g., letter, meeting minutes), and a statement of how or where in the submission this prior feedback was addressed. Note that adequacy of how the feedback was addressed will be assessed during the substantive review.</i></p> <p><i>Select “N/A” if the submitter states there were no prior submissions.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	9.	<p>The submission utilizes voluntary consensus standard(s) (See section 514(c) of the FD&amp;C Act). <i>This includes both FDA-recognized and non-recognized consensus standards. Select “N/A” if the submission does not utilize voluntary consensus standards.</i></p>	<input type="checkbox"/>		<input type="checkbox"/>		
	a.	The submission cites FDA-recognized voluntary consensus standard(s).	<input type="checkbox"/>		<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
		i.	<p>The submission includes a Declaration of Conformity (DOC) as outlined in FDA’s guidance “<a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices</a>.</p> <p><b>OR</b></p> <p>If citing general use of a standard as noted in FDA’s guidance “<a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a>,” the basis of such use is included along with the underlying information or data that supports how the standard was used.</p>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	The submission cites non-FDA-recognized voluntary consensus standard(s).	<input type="checkbox"/>		<input type="checkbox"/>	
		i.	The basis of use is included along with the underlying information or data that supports how the standard was used.	<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:						
	<p><b>Combination Product Provisions – Per 503(g) of the FD&amp;C Act.</b>  Select “N/A” if the product is not a combination product. 21 CFR 3.2(e). The remaining criteria in this section will be omitted from the checklist if “N/A” is selected. If you are unsure if the product is a combination product, consult with the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.</p>					<input type="checkbox"/>	
10.	Submission identifies the product as a combination product.				<input type="checkbox"/>	<input type="checkbox"/>	
11.	The combination product contains as a constituent part an approved drug as defined in section 503(g)(5)(B) of the FD&C Act. Select “N/A” if the combination product does not contain as a constituent part an approved drug. Please also select “N/A” if a right of reference or use for the drug constituent part(s) is included with the submission. If “N/A” is selected, part a below is omitted from the checklist.				<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		a.	The submission includes appropriate patent statement or certification and a statement that the submitter will give notice, as applicable. See section 503(g)(5)(A)&(C) of the FD&C Act.	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:					
<b>B.</b>	<b>Device Description</b>							
	12.	The device has a device-specific guidance document, special controls, and/or requirements in a device-specific classification regulation regarding the device description that is applicable to the subject device. <i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>			<input type="checkbox"/>	<input type="checkbox"/>		
		a.	<p>The submission addresses device description recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.			Yes	No	N/A	*Page #	
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>							
	b.	<p>The submission includes device description information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	13.	Descriptive information is present and consistent within the submission (e.g., the device description section is consistent with the device description in the labeling).	<input type="checkbox"/>	<input type="checkbox"/>			
		Comments:					
	14.	The submission includes descriptive information for the device, including the following:					
	a.	A description of the principle of operation or mechanism of action for achieving the intended effect.	<input type="checkbox"/>	<input type="checkbox"/>			
	b.	A description of proposed conditions of use, such as surgical technique for implants; anatomical location of use; user interface; how the device interacts with other devices; and/or how the device interacts with the patient.	<input type="checkbox"/>	<input type="checkbox"/>			
	c.	<p>A list and description of each device for which clearance is requested.</p> <p><i>Select “N/A” if there is only one device or model. “Device” may refer to models, part numbers, various sizes, etc.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
		d.	Submission contains representative engineering drawing(s), schematics, illustrations, photos and/or figures of the device. <b>OR</b> Submission includes a statement that engineering drawings, schematics, etc. are not applicable to the device (e.g., device is a reagent and figures are not pertinent to describe the device). <i>In lieu of engineering drawings, schematics, etc. of each device to be marketed, “representative” drawings, etc. may be provided, where “representative” is intended to mean that the drawings, etc. provided capture the differences in design, size, and other important characteristics of the various models, sizes, or versions of the device(s) to be marketed.</i>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:				
	15.		Device is intended to be marketed with accessories and/or as part of a system. <i>Select “N/A” if the device is not intended to be marketed with accessories and/or as part of a system. If “N/A” is selected, parts a-c below are omitted from the checklist.</i>	<input type="checkbox"/>		<input type="checkbox"/>	
		a.	Submission includes a list of all accessories to be marketed with the subject device.	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	Submission includes a description (as detailed in item 14a., 14b., and 14d. above) of each accessory. <i>Select “N/A” if the accessory(ies) has been previously cleared, or is exempt, and the proposed indications for use are consistent with the cleared indications.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		c.	A 510(k) number is provided for each accessory that received a prior 510(k) clearance. <b>AND</b> A statement is provided that identifies accessories that have not received prior 510(k) clearance.	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:				

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
<b>C. Substantial Equivalence Discussion</b>							
	16.	Submitter has identified a predicate device(s), including the following information:					
	a.	<p>Predicate device identifier provided (e.g., 510(k) number, De Novo number, reclassified PMA number, classification regulation reference, if exempt (e.g., 21 CFR 872.3710), or statement that the predicate is a preamendment device).</p> <p>For predicates that are preamendments devices, information is provided to document preamendments status.</p> <p><i>Information regarding <a href="#">documenting preamendment status is available online (<a href="https://www.fda.gov/medical-devices/quality-and-compliance-medical-devices/preamendment-status">https://www.fda.gov/medical-devices/quality-and-compliance-medical-devices/preamendment-status</a>).</a></i></p>		<input type="checkbox"/>	<input type="checkbox"/>		
	b.	The identified predicate(s) is consistent throughout the submission (e.g., the predicate(s) identified in the Substantial Equivalence section is the same as that listed in the 510(k) Summary (if applicable) and that used in comparative performance testing.		<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	17.	<p>Submission includes a comparison of the following for the predicate(s) and subject device and a discussion why any differences between the subject and predicate(s) do not impact safety and effectiveness [see section 513(i)(1)(A) of the FD&amp;C Act and 21 CFR 807.87(f)].</p> <p><i>See the FDA guidance document “<a href="#">The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)]</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k</a> for more information on comparing intended use and technological characteristics.</i></p>					

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.							
		a.	Indications for use <i>If there are no differences between the subject device and the predicate(s) with respect to indications and intended use, this should be explicitly stated.</i>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	Technology, including technical specifications, features, materials, and principles of operation <i>Examples of technological characteristics include, but are not limited to design, features, materials, energy source, and principle of operation.</i>  <i>FDA recommends a tabular format for comparing technological characteristics. Any characteristic that is the same as the predicate(s) should be explicitly stated. Differences in technological characteristics should be identified and a rationale provided why they do not raise different questions of safety and effectiveness.</i>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:				
D.	<b>Proposed Labeling (see also 21 CFR parts 801 and 809 as applicable)</b>						
	18.		Submission includes proposed package labels and labeling (e.g., instructions for use, package insert, operator's manual).	<input type="checkbox"/>	<input type="checkbox"/>		
		a.	Indications for use are stated in labeling and are identical to Indications for Use form and 510(k) Summary (if 510(k) Summary provided).	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	Labeling includes: <ul style="list-style-type: none"> <li>- Statements of conditions, purposes or uses for which the device is intended (e.g., hazards, warnings, precautions, contraindications) (21 CFR 801.5)</li> <li><b>AND</b></li> <li>- Includes adequate directions for use (see 21 CFR 801.5)</li> <li><b>OR</b></li> <li>- Submission states that device qualifies for exemption per 21 CFR 801 Subpart D</li> </ul>	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
		Comments:					
	19.	Labeling includes name and place of business of the manufacturer, packer, or distributor (21 CFR 801.1).		<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	20.	Labeling includes the prescription statement (see 21 CFR 801.109(b)(1)) or Rx Only symbol (see also Section 502(a) of the FD&C Act and FDA’s final rule, “ <a href="#">Use of Symbols in Labeling</a> ” (81 FR 38911), available at <a href="https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling">https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling</a> ). <i>Select “N/A” if not indicated for prescription use.</i>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:					
	21.	The device has a device-specific guidance document, special controls, and/or requirements in a device-specific classification regulation regarding labeling that is applicable to the subject device. <i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>		<input type="checkbox"/>	<input type="checkbox"/>		
	a.	<p>The submission addresses labeling recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #			
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.										
		b.	<p>The submission includes labeling information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
			Comments:							
	22.		<p>If the device is an in vitro diagnostic device, provided labeling includes all applicable information required per 21 CFR 809.10.</p> <p><i>Select “N/A” if not an in vitro diagnostic device.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
			Comment:							
E.	<p><b>Sterilization</b></p> <p><i>If an in vitro diagnostic (IVD) device and sterilization is not applicable, select “N/A.” The criteria in this section will be omitted from the checklist if “N/A” is selected.</i></p>					<input type="checkbox"/>				

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.	Yes	No	N/A	*Page #
<p>Submission states that the device and/or accessories, if applicable, are: <i>(one of the below must be checked)</i></p> <p><input type="checkbox"/> Provided sterile, intended to be single-use</p> <p><input type="checkbox"/> Requires processing during its use-life</p> <p><input type="checkbox"/> Non-sterile when used (and no processing required)</p> <p><input type="checkbox"/> Information regarding the sterility status of the device is not provided (if this box is checked, please also check one of the two boxes below)</p> <p style="padding-left: 20px;"><input type="checkbox"/> Sterility status not needed for this device (e.g., software-only device)</p> <p style="padding-left: 20px;"><input type="checkbox"/> Sterility status needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “non-sterile when used” or “not provided and not needed” is selected, the sterility-related criteria below are omitted from the checklist.</i></p> <p><i>If information on sterility status is not provided, and it is needed or the need for this information is unclear, select “No.”</i></p> <p><i>The “Requires processing during its use-life” option refers to devices falling into one of the four categories below:</i></p> <ul style="list-style-type: none"> <li>• <i>Supplied sterile and requires reprocessing prior to subsequent patient use</i></li> <li>• <i>Supplied non-sterile and requires user to process the device for initial use, as well as to reprocess the device after each use</i></li> <li>• <i>Reusable medical device (single-user) reprocessed between each use</i></li> <li>• <i>Single-use medical devices initially supplied as non-sterile to the user, and requiring the user to process the device prior to its use</i></li> </ul> <p><i>Please refer to the FDA guidance document “<a href="#">Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/reprocessing-medical-devices-health-care-settings-validation-methods-and-labeling">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/reprocessing-medical-devices-health-care-settings-validation-methods-and-labeling</a>, for additional information.</i></p>	<input type="checkbox"/>			
Comments:				
23.	Assessment of the need for cleaning and subsequent disinfection or sterilization information.			

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
		a.	Identification of device and/or accessories, if applicable, that are provided sterile. <i>Select “N/A” if no part of the device or accessories are provided sterile.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		b.	Identification of device and/or accessories, if applicable, that are end user sterilized or disinfected. <i>Select “N/A” if no part of the device or accessories are end user sterilized or disinfected.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		c.	Identification of device and/or accessories, if applicable, that are reusable. <i>Select “N/A” if no part of the device or accessories, are reusable.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
			Comments:				
	24.		If the device and/or accessories, if applicable, are provided sterile. <i>Select “N/A” if no part of the device or accessories are provided sterile, otherwise complete a-f below.</i>			<input type="checkbox"/>	
		a.	Sterilization method is stated for each device (including dose for radiation sterilization)	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	A description of method to validate the sterilization parameters is provided for each proposed sterilization method (e.g., half-cycle method and full citation of FDA-recognized standard, including date). <i>Note: the sterilization validation report is not required.</i>	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	For devices sterilized using chemical sterilants such as ethylene oxide (EO) and hydrogen peroxide, submission states maximum levels of sterilant residuals remaining on the device and sterilant residual limits. <i>Select “N/A” if not sterilized using chemical sterilants.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		d.	Sterility Assurance Level (SAL) stated	<input type="checkbox"/>	<input type="checkbox"/>		
		e.	Submission includes description of packaging	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		f.	For products labeled “non-pyrogenic,” a description of the method used to make the determination stated (e.g., limulus amebocyte lysate [LAL]). <i>Select “N/A” if not labeled “non-pyrogenic.”</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:					
	25.		If the device and/or accessory, if applicable, is reusable or end user sterilized or disinfected: <i>Select “N/A” if no part of the device or accessories are reusable or end user sterilized or disinfected, otherwise complete a-d below.</i>			<input type="checkbox"/>		
		a.	Cleaning method is provided in labeling for each device and/or accessory, if applicable. <i>Select “N/A” if not reusable and does not need cleaning prior to disinfection or sterilization.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	Disinfection method is provided in labeling for each device and/or accessory, if applicable. <i>Select “N/A” if not disinfected (i.e., undergoes terminal sterilization) prior to use.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	Sterilization method is provided in labeling for each device and/or accessory, if applicable. <i>Select “N/A” if not sterilized (i.e., undergoes disinfection) prior to use.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
		d.	Device types in this submission are listed in the Federal Register (FR) Notice entitled “ <a href="#">Validated Instructions for Use and Validation Data Requirements for Certain Reusable Medical Devices in Premarket Notifications</a> ” (Reprocessing FR Notice, available at <a href="https://www.federalregister.gov/documents/2017/06/09/2017-12007/medical-devices-validated-instructions-for-use-and-validation-data-requirements-for-certain-reusable">https://www.federalregister.gov/documents/2017/06/09/2017-12007/medical-devices-validated-instructions-for-use-and-validation-data-requirements-for-certain-reusable</a> ). <i>Device types identified in the Reprocessing FR Notice represent devices posing a greater likelihood of microbial transmission and represent a high risk of infection. Select “N/A” if the device type in the submission is not included in the Reprocessing FR Notice.</i>	<input type="checkbox"/>		<input type="checkbox"/>	
		i.	If device types in this submission are included in the Reprocessing FR Notice, the submission includes protocols and test reports for validating the reprocessing instructions. <i>Select “N/A” if the device type in the submission is not included in the Reprocessing FR Notice.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:					
	26.	The device has a device-specific guidance document, special controls, and/or requirement in a device-specific classification regulation regarding sterility and/or reprocessing that is applicable to the subject device. <i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>			<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #	
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.								
		a.	<p>The submission addresses sterility and/or reprocessing recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	<p>The submission includes sterility and/or reprocessing information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:					
F.	<b>Shelf-Life</b>							
	27.	<p>Proposed shelf life/ expiration date stated</p> <p><b>OR</b></p> <p>Statement that shelf-life is not applicable because of low likelihood of time-dependent product degradation.</p>		<input type="checkbox"/>	<input type="checkbox"/>			

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
		Comments:					
	28.	For a sterile device, submission includes summary of methods used to establish that device packaging will maintain a sterile barrier for the entirety of the proposed shelf-life. <i>Select “N/A” if the device is not provided sterile.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	29.	Submission includes summary of methods used to establish that device performance is maintained for the entirety of the proposed shelf-life (e.g., mechanical properties, coating integrity, pH, osmolality, etc.). <b>OR</b> Statement why performance data is not needed to establish maintenance of device performance characteristics over the shelf-life period.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
<b>G.</b>	<b>Biocompatibility</b>	<i>If an in vitro diagnostic (IVD) device, select “N/A.” The criteria in this section will be omitted from the checklist if “N/A” is selected.</i>					

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.	Yes	No	N/A	*Page #
<p>Submission states that there: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Are direct or indirect tissue-contacting components</p> <p><input type="checkbox"/> Are no direct or indirect tissue-contacting components</p> <p><input type="checkbox"/> Information regarding tissue contact status of the device is not provided (if this box checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> Tissue contact information not needed for this device (e.g., software-only device)</p> <p><input type="checkbox"/> Tissue contact information is needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “are no” or “not provided and not needed” is selected, the biocompatibility-related criteria below are omitted from the checklist. If information on the tissue-contact status is not provided, and contact information is needed or its contact status is unclear, select “No.”</i></p> <p><i>An example of a direct tissue-contacting device would be an implant that has direct contact with tissues during use. An example of an indirect tissue-contacting device would be fluid entering the body following passing through device/device components not in direct contact with the tissue.</i></p>	<input type="checkbox"/>			
Comments:				
30.	Submission includes a list identifying each tissue-contacting device component (e.g., implant, delivery catheter) and associated materials of construction for each component, including identification of color additives, if present.	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:			
31.	Submission identifies contact classification (e.g., surface-contacting, less than 24 hour duration) for each tissue-contacting device component (e.g., implant, delivery catheter).	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:			

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #	
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>						
32.	<p>For a biocompatibility assessment of tissue-contacting components, submission includes:</p> <ul style="list-style-type: none"> <li>Each relevant endpoint for the device (as identified in device-specific guidance, or Attachment A of the FDA guidance document entitled “<a href="#">Use of International Standard ISO 10993-1, ‘Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process,’</a>” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-international-standard-iso-10993-1-biological-evaluation-medical-devices-part-1-evaluation-and">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-international-standard-iso-10993-1-biological-evaluation-medical-devices-part-1-evaluation-and</a>), has been addressed.</li> <li>For any testing performed, test protocol (including identification and description of test article including whether the test article is the device in its final finished form using the recommended approach in Attachment F of “<a href="#">Use of International Standard ISO 10993-1, ‘Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process,’</a>” methods, and pass/fail criteria), and analysis of results (including tables with data points and statistical analyses, where appropriate), as described in Attachment E of the guidance document entitled “<a href="#">Use of International Standard ISO 10993-1, ‘Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process’</a>” provided for each completed test.</li> </ul> <p><b>OR</b></p> <p>A statement that biocompatibility testing is not needed with a rationale that considers all relevant endpoints (e.g., materials and manufacturing/processing are identical to the predicate).</p>	<input type="checkbox"/>	<input type="checkbox"/>			
	Comments:					
<b>H.</b>	<b>Software</b>					

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.	Yes	No	N/A	*Page #
<p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does contain software/firmware</p> <p><input type="checkbox"/> Does not contain software/firmware</p> <p><input type="checkbox"/> Information on whether device contains software/firmware is not provided (if this box checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> Software/firmware information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> Software/firmware information is needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not contain” or “not provided and not needed” is selected, the software-related criteria below are omitted from the checklist. If information on software is not provided, and this information is needed or the need is unclear, select “No.”</i></p>	<input type="checkbox"/>			
Comments:				
33. Submission includes a statement of software level of concern and rationale for the software level of concern	<input type="checkbox"/>	<input type="checkbox"/>		
Comments:				
34. All applicable software documentation provided based on level of concern identified by the submitter, as described in “ <a href="#">Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices</a> ,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-content-premarket-submissions-software-contained-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-content-premarket-submissions-software-contained-medical-devices</a> , or the submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through an alternative approach (i.e., the submitter has identified an alternate approach with a rationale).  <i>Note: This element is also applicable to non-internally generated or off-the-shelf (OTS) software used in the device.</i>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>		Yes	No	N/A	*Page #	
	Comments:					
I.	<b>Cybersecurity</b>					
	<p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does contain any external wired and/or wireless communication interfaces (Wired: USB, ethernet, SD, CD, RGA, etc. or Wireless: Wi-Fi, Bluetooth, RF, inductive, Cloud, etc.)</p> <p><input type="checkbox"/> Does not contain external interfaces as described above</p> <p><input type="checkbox"/> Information on whether device has external interfaces is not provided (if this box is checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> Cybersecurity information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> Cybersecurity information is needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not contain” or “not provided and not needed” is selected, the cybersecurity criteria below are omitted from the checklist. If information on cybersecurity is not provided, and this information is needed or the need is unclear, select “No.”</i></p>	<input type="checkbox"/>				
	35. All applicable documentation identified by the submitter, as described in “ <a href="#">Guidance for the Content of Premarket Submissions for Management of Cybersecurity in Medical Devices</a> ,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/content-premarket-submissions-management-cybersecurity-medical-devices-0">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/content-premarket-submissions-management-cybersecurity-medical-devices-0</a> .	<input type="checkbox"/>	<input type="checkbox"/>			
	<b>OR</b> Submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through an alternative approach (i.e., the submitter has identified an alternate approach with a rationale).					
	Comments:					
J.	<b>Electrical Safety and EMC</b>					

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.	Yes	No	N/A	*Page #
<p>Electrical Safety:</p> <p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does require electrical safety evaluation</p> <p><input type="checkbox"/> Does not require electrical safety evaluation</p> <p><input type="checkbox"/> Information on whether device requires electrical safety evaluation is not provided (if this box checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> Electrical safety information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> Electrical safety information needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not require” or “not provided and not needed” is selected, the electrical safety criteria below are omitted from the checklist. If information on electrical safety is not provided, and it is needed or the need for this information is unclear, select “No.”</i></p>	<input type="checkbox"/>			
<p>36. Submission includes evaluation of electrical safety (e.g., per IEC 60601-1, or equivalent FDA-recognized standard, and if applicable, a device-specific standard).</p> <p><b>OR</b></p> <p>Submission includes electrical safety evaluation using methods or standards that are not FDA-recognized and submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through this alternative approach (i.e., the submitter has identified alternate methods or standards with a rationale).</p>	<input type="checkbox"/>	<input type="checkbox"/>		
Comments:				

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #	
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>						
	<p>EMC:</p> <p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does require EMC evaluation</p> <p><input type="checkbox"/> Does not require EMC evaluation</p> <p><input type="checkbox"/> Information on whether device requires EMC evaluation not provided (if this box checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> EMC information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> EMC information needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not require” or “not provided and not needed” is selected, the EMC criteria below are omitted from the checklist. If information on EMC is not provided, and it is needed or the need for this information is unclear, select “No.”</i></p>	<input type="checkbox"/>				
	Comments:					
	37.	<p>Submission includes evaluation of electromagnetic compatibility (e.g., per IEC 60601-1-2 or equivalent FDA-recognized standard and if applicable, a device-specific standard).</p> <p><b>OR</b></p> <p>Submission includes electromagnetic compatibility evaluation using methods or standards that are not FDA-recognized and submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through this alternative approach (i.e., the submitter has identified alternate methods or standards with a rationale).</p>				
		Comments:				

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>					
<b>K.</b>	<b>Performance Data General</b> <i>If an in vitro diagnostic (IVD) device, select “N/A.” The criteria in this section will be omitted from the checklist if “N/A” is selected.</i> <i>Performance data criteria relating to IVD devices is addressed in Section L.</i>			<input type="checkbox"/>	
	Comments:				
	38. Summaries of the non-clinical laboratory studies and full test reports* are provided.  *Summary and full test report content recommendations can be found in FDA’s guidance “ <a href="#">Recommended Content and Format of Non-Clinical Bench Performance Testing Information in Premarket Submissions</a> ,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/recommended-content-and-format-non-clinical-bench-performance-testing-information-premarket">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/recommended-content-and-format-non-clinical-bench-performance-testing-information-premarket</a> .  If a submitter chooses to declare conformity to a voluntary consensus standard that FDA has recognized, submission of a full test report may not be necessary. Refer to 9a. See FDA’s guidance “ <a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a> ,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices</a> . <i>Select “N/A” if the submission appropriately does not include performance data or there are no completed tests without a Declaration of Conformity.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	a. Submission includes an explanation of how the data generated from each test supports a finding of substantial equivalence (e.g., comparison to predicate device testing, dimensional analysis, etc.).  <i>Select “N/A” if the submission does not include performance data.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:				

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>			<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
	39.	<p>The device has a device-specific guidance document, special controls, and/or requirement in a device-specific classification regulation regarding performance data that is applicable to the subject device.</p> <p><i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		<p>a. The submission addresses performance data recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		<p>b. The submission includes performance data that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:				

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>			<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
	40.	<p>If literature is referenced in the submission, submission includes:</p> <p><i>Select “N/A” if the submission does not reference literature. If “N/A” is selected, parts a and b below are omitted from the checklist.</i></p> <p><i>Note that the applicability of the referenced article to support a substantial equivalence finding should be assessed during the substantive review; only the presence of a discussion is required to support acceptance.</i></p>			<input type="checkbox"/>		
		a. Legible reprints or a summary of each article.	<input type="checkbox"/>	<input type="checkbox"/>			
		b. Discussion of how each article is applicable to support the substantial equivalence of the subject device to the predicate.	<input type="checkbox"/>	<input type="checkbox"/>			
		Comments:					
	41.	<p>For each completed animal study, the submission provides the following:</p> <p><i>Select “N/A” if no animal study was conducted. If “N/A” is selected, parts a-c below are omitted from the checklist. Note that this section does not address biocompatibility evaluations, which are assessed in Section G of the checklist.</i></p>			<input type="checkbox"/>		
		a. Submission includes a study protocol which includes all elements as outlined in 21 CFR 58.120	<input type="checkbox"/>	<input type="checkbox"/>			
		b. Submission includes final study report which includes all elements outlined in 21 CFR 58.185	<input type="checkbox"/>	<input type="checkbox"/>			
		c. Submission contains a statement that the study was conducted in compliance with applicable requirements in the GLP regulation (21 CFR Part 58), OR, if the study was not conducted in compliance with the GLP regulation, the submission explains why the noncompliance would not impact the validity of the study data provided to support a substantial equivalence determination.	<input type="checkbox"/>	<input type="checkbox"/>			
		Comments:					

*Contains Nonbinding Recommendations*

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>				Yes	No	N/A	*Page #	
L.	<b>Performance Characteristics – In Vitro Diagnostic Devices Only</b> (see also 21 CFR 809.10(b)(12))							
	Submission indicates that device: ( <i>one of the below must be checked</i> ) <input type="checkbox"/> Is an in vitro diagnostic device <input type="checkbox"/> Is not an in vitro diagnostic device <i>If “is not” is selected, the performance data-related criteria below are omitted from the checklist.</i>							
	42.	Submission includes the following studies, as appropriate for the device type, including associated protocol descriptions, study results and line data:						
		a.	Precision/reproducibility			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.	Accuracy (includes as appropriate linearity; calibrator or assay traceability; calibrator and/or assay stability protocol and acceptance criteria; assay cut-off; method comparison or comparison to clinical outcome; matrix comparison; and clinical reference range or cutoff).			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	Sensitivity (detection limits, LoB, LoD, LoQ where relevant for the device type).			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		d.	Analytical specificity			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			Comments:					
	43.	The device has a device-specific guidance document, special controls, and/or requirement in a device-specific classification regulation regarding performance data that is applicable to the subject device.			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		<i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>						

***Contains Nonbinding Recommendations***

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>				Yes	No	N/A	*Page #		
		a.	<p>The submission addresses performance data recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>			
		b.	<p>The submission includes performance data that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:						

*Contains Nonbinding Recommendations*

<b>Digital Signature Concurrence Table</b>	
Reviewer Sign-Off	
Management Sign-Off (digital signature optional)*	

\*Management review of checklist and concurrence with decision required.

## Appendix B. Acceptance Checklist for Abbreviated 510(k)s

**(Should be completed within 15 days of DCC receipt)**

The following information is not intended to serve as a comprehensive review. FDA recommends that the submitter include this completed checklist as part of the application.

**510(k)#:**

**Date Received by DCC:**

**510(k) Lead Reviewer:**

**Center:**

**Office:**

**Division:**

**Decision: Accept \_\_\_\_\_ Refuse to Accept \_\_\_\_\_**

**If Accept, notify the submitter**

**If Refuse to Accept, notify submitter electronically and include a copy of this checklist.**

**Is an Addendum attached?: Yes No**

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete; it means the reviewer did not assess the element during the RTA review and that the element will be assessed during substantive review.

Preliminary Questions			
Answers in the shaded blocks indicate consultation with an identified Center advisor is needed. (Boxes checked in this section represent FDA's preliminary assessment of these questions at the time of administrative review.)	Yes	No	N/A
<p><b>1. Is the product a device (per section 201(h) of the FD&amp;C Act) or a combination product (per 21 CFR 3.2(e)) with a device constituent part subject to review in a 510(k)?</b></p> <p>If it appears not to be a device (per section 201(h) of the FD&amp;C Act) or such a combination product (per 21 CFR 3.2(e)), or you are unsure, consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action, and inform management. <i>Provide a summary of the Product Jurisdiction Officer's determination/recommendation/action in the comment section below.</i></p> <p>If the product does not appear to be a device or such a combination product, mark</p>	<input type="checkbox"/>	<input type="checkbox"/>	

***Contains Nonbinding Recommendations***

“No.”			
<b>Comments:</b>			
<b>2. Is the submission with the appropriate Center?</b>  If the product is a device or a combination product with a device constituent part, is it subject to review by the Center in which the submission was received? If you believe the submission is not with the appropriate Center or you are unsure, consult with the CDRH Product Jurisdiction or the CBER Product Jurisdiction Officer to determine the appropriate action and inform your management. <i>Provide a summary of the Product Jurisdiction Officer’s determination/recommendation/action in the comment section below.</i>	<input type="checkbox"/>	<input type="checkbox"/>	
If submission should not be reviewed by your Center mark “No.”			
<b>Comments:</b>			
<b>3. If a Request for Designation (RFD) was submitted for the device or combination product with a device constituent part and assigned to your center, identify the RFD # and confirm the following:</b>  <b>(a) Is the device or combination product the same (e.g., design, formulation) as that presented in the RFD submission?</b>  <b>(b) Are the indications for use for the device or combination product identified in the 510(k) the same as those identified in the RFD submission?</b>  If you believe the product or the indications presented in the 510(k) have changed from the RFD, or you are unsure, consult with the CDRH Product Jurisdiction Officer or the CBER Product Jurisdiction Officer to determine the appropriate action and inform your management. <i>Provide summary of Product Jurisdiction Officer’s determination/recommendation/action in the comment section below.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
If the answer to either question above is no, mark “No.” If there was no RFD, mark “N/A.”			
<b>Comments:</b>			
<b>4. Is the submission for a combination product that contains as a constituent part a drug that has the same active moiety as an approved drug with exclusivity as described in section 503(g)(5)(C)(ii)-(v) of the FD&amp;C Act?</b>  If “Yes,” then contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer to determine the appropriate action and inform your management. Provide the summary of the Product Jurisdiction Officer’s determination/recommendation/action in the comment section below.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

*Contains Nonbinding Recommendations*

<b>Comments:</b>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
<b>5. Is this device type eligible for a 510(k) submission?</b>  If a 510(k) does not appear to be appropriate (e.g., class III type and PMA required, or class I or II type and 510(k)-exempt), consult with the appropriate CDRH or CBER staff during the acceptance review, provide a summary of the discussion with them, and indicate their recommendation/action in the comment section below. If 510(k) is not the appropriate regulatory submission, mark "No."	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
<b>Comments:</b>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
<b>6. Is there a pending PMA for the same device with the same indications for use?</b>  If "Yes," consult your management and CDRH Office of Product Evaluation and Quality/Office of Regulatory Programs/Division of Regulatory Programs 1 (Submission Support) (OPEQ/ORP/DRP1) or appropriate CBER staff to determine the appropriate action.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>Comments:</b>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
<b>7. If clinical studies have been submitted, is the submitter the subject of an Application Integrity Policy (AIP)?</b>  If "Yes," consult with the CDRH Office of Product Evaluation and Quality/Office of Clinical Evidence and Analysis/Division of Clinical Science and Quality (OPEQ/OCEA/DCEA1) or CBER Office of Compliance and Biologics Quality/Division of Inspections and Surveillance/Bioresearch Monitoring Branch (OCBQ/DIS/BMB) to determine the appropriate action, provide a summary of the discussion with them, and indicate their recommendation/action.  If no clinical studies have been submitted, mark "N/A." Check on the AIP list at <a href="https://www.fda.gov/inspections-compliance-enforcement-and-criminal-investigations/application-integrity-policy/application-integrity-policy-list">https://www.fda.gov/inspections-compliance-enforcement-and-criminal-investigations/application-integrity-policy/application-integrity-policy-list</a> .	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>Comments:</b>			

- **If the answer to 1 or 2 appears to be "No," then stop review of the 510(k) and contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.**
- **If the answer to 3a or 3b appears to be "No," then stop the review and contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.**
- **If the answer to 4 is "Yes," then contact the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer, provide a summary of the discussion with them, and indicate their recommendation/action.**
- **If the answer to 5 is "No", the lead reviewer should consult division management and other Center resources to determine the appropriate action. Note that, for a device which is clearly ineligible for a 510(k) submission (such as a device type which is class III requiring PMA or class I/II and 510(k) exempt), this may be**

***Contains Nonbinding Recommendations***

considered a basis for a refusal to accept the submission. A 510(k) submitted for a class I/II, 510(k)-exempt device that trips the limitations of the exemption would not be refused on this basis.

- If the answer to 6 is “Yes,” then stop review of the 510(k), contact CDRH/OPEQ/ORP/DRP1, or appropriate CBER staff.
- If the answer to 7 is “Yes,” then contact CDRH/OPEQ/OCEA/DCEA1 or CBER/OCBQ/DIS/BMB, provide a summary of the discussion with DCEA1 or BMB Staff, and indicate their recommendation/action.

**Abbreviated 510(k) Criteria**

(See “[The Abbreviated 510\(k\) Program](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/abbreviated-510k-program),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/abbreviated-510k-program> and “[Format for Traditional and Abbreviated 510\(k\)s](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/format-traditional-and-abbreviated-510ks),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/format-traditional-and-abbreviated-510ks>)

In order to qualify for review as an Abbreviated 510(k), one of the following criteria (1 or 2 or 3) should be met. Submission should be converted and reviewed as a Traditional 510(k) if one of these criteria is not met. Complete the Refuse to Accept Checklist for a Traditional 510(k) if submission is converted.

		Yes	No	N/A
1. <b>Submission relies on a guidance document and a summary report is provided that:</b> <i>Select “N/A” if submission does not rely on any guidance document(s). If “Yes,” address parts a and b below.</i>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
a.	Includes the device description, the manufacturer’s device design requirements, risk management information, and a description of test methods used to address performance characteristics.	<input type="checkbox"/>	<input type="checkbox"/>	
b.	Includes a description of how the guidance document was used to satisfy the requirements of 21 CFR 807.87 (e.g., data to support substantial equivalence) and lists any deviations, that is any alternative methods used to demonstrate substantial equivalence that is not described in the guidance. <i>Select “No” if the sponsor does not address whether there were deviations.</i>	<input type="checkbox"/>	<input type="checkbox"/>	

**Comments:**

2. <b>Submission relies on special control(s) to demonstrate substantial equivalence and a summary report is provided that:</b> <i>Select “N/A” if submission does not rely on any special controls. If “Yes,” address parts a-d below.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
--	--------------------------	--------------------------	--------------------------

***Contains Nonbinding Recommendations***

	a.	Includes the device description, the manufacturer's device design requirements, risk management information, and a description of test methods used to address performance characteristics.	<input type="checkbox"/>	<input type="checkbox"/>	
	b.	Includes a description of how the special control(s) was satisfied.	<input type="checkbox"/>	<input type="checkbox"/>	

**Comments:**

3.	<b>Submission relies on voluntary consensus standard(s) (See section 514(c) of the FD&amp;C Act). This includes both FDA-recognized and non-recognized consensus standards.</b>			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	<i>Select "N/A" if submission does not rely on any voluntary consensus standard(s). If "Yes," address part a below.</i>					
	a.	The submission cites FDA-recognized voluntary consensus standard(s).	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	i.	The submission includes a Declaration of Conformity (DOC) as outlined in FDA's guidance " <a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a> ," available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices</a> . <b>OR</b> If citing general use of a standard as noted in FDA's guidance " <a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a> ," the basis of such use is included along with the underlying information or data that supports how the standard was used.	<input type="checkbox"/>	<input type="checkbox"/>		
	b.	The submission cites non-FDA-recognized voluntary consensus standard(s).	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	i.	The basis of use is included along with the underlying information or data that supports how the standard was used.	<input type="checkbox"/>	<input type="checkbox"/>		

**Comments:**

**Does the submission meet one of the criteria above?**

- Yes, submission meets criteria for an Abbreviated 510(k). Continue with the remainder of this checklist below.
- No, submission does not meet criteria for an Abbreviated 510(k). Discontinue this RTA checklist, convert to a Traditional and apply the Traditional checklist.

<b>Organizational Elements</b>			
Failure to include these items should not result in an RTA designation.			
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>		<b>Yes</b>	<b>No</b>
<b>*Page #</b>			
1.	Submission contains a Table of Contents.	<input type="checkbox"/>	<input type="checkbox"/>
2.	Each section is labeled (e.g., headings or tabs designating Device Description section, Labeling section, etc.).	<input type="checkbox"/>	<input type="checkbox"/>
3.	All pages of the submission are numbered. <i>All pages should be numbered in such a manner that information can be referenced by page number. This may be done either by consecutively numbering the entire submission, or numbering the pages within a section (e.g., 12-1, 12-2...).</i>	<input type="checkbox"/>	<input type="checkbox"/>
4.	Type of 510(k) is identified (i.e., Traditional, Abbreviated, or Special) <i>If type of 510(k) is not designated, review as a Traditional 510(k).</i>	<input type="checkbox"/>	<input type="checkbox"/>
Comments:			

<b>Elements of a Complete Submission (RTA Items)</b> <b>(21 CFR 807.87 unless otherwise indicated)</b>
Submission should be designated RTA if not addressed
<ul style="list-style-type: none"><li>Any “No” answer will result in a “Refuse to Accept” decision; however, FDA staff has discretion to determine whether missing items are needed to ensure that the submission is administratively complete to allow the submission to be accepted or to request missing checklist items interactively from submitters during the RTA review.</li><li>Each element on the checklist should be addressed within the submission. The submitter may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criterion is considered present (Yes). An assessment of the rationale will be considered during the review of the submission.</li></ul>

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
<b>A. Administrative</b>							
1.	All content used to support the submission is written in English (including translations of test reports, literature articles, etc.).			<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:						
2.	Submission identifies the following (FDA recommends use of the CDRH Premarket Review Submission Cover Sheet form ( <a href="#">Form 3514</a> , available at <a href="https://www.fda.gov/media/72421/download">https://www.fda.gov/media/72421/download</a> )):						
	a.	Device trade/proprietary name			<input type="checkbox"/>	<input type="checkbox"/>	
	b.	Device class and panel OR Classification regulation OR Statement that device has not been classified with rationale for that conclusion			<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:						
3.	Submission contains an Indications for Use Statement with Rx and/or OTC designated (see also 21 CFR 801.109, and FDA’s final rule, “ <a href="#">Use of Symbols in Labeling</a> ” (81 FR 38911), available at <a href="https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling">https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling</a> ). <i>See recommended <a href="#">format</a> (<a href="https://www.fda.gov/media/86323/download">https://www.fda.gov/media/86323/download</a>).</i>			<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:						
4.	Submission contains a 510(k) Summary or 510(k) Statement. <i>Refer to 21 CFR 807.92 and 21 CFR 807.93 for contents of 510(k) Summary and Statement, respectively. Adequacy of the content will be assessed during substantive review.</i>			<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:						

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.			Yes	No	N/A	*Page #	
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>							
	5.	Submission contains a Truthful and Accuracy Statement per 21 CFR 807.87(l). <i>See recommended format (<a href="https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-truthful-and-accurate-statement">https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-truthful-and-accurate-statement</a>).</i>	<input type="checkbox"/>	<input type="checkbox"/>			
		Comments:					
	6.	Submission is a class III 510(k) Device. <i>Select “N/A” only if submission is not a class III 510(k).</i>	<input type="checkbox"/>		<input type="checkbox"/>		
	a.	Contains Class III Summary and Certification per 21 CFR 807.87(k). <i>See recommended content (<a href="https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-class-iii-certification-and-summary">https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-class-iii-certification-and-summary</a>).</i> Select “N/A” only if submission is not a Class III 510(k).	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	7.	Submission contains clinical data. <i>Select “N/A” if the submission does not contain clinical data. If “N/A” is selected, parts a, b, and c below are omitted from the checklist.</i>	<input type="checkbox"/>		<input type="checkbox"/>		
	a.	Submission includes completed Financial Certification ( <a href="https://www.fda.gov/media/70465/download">FDA Form 3454</a> , available at <a href="https://www.fda.gov/media/70465/download">https://www.fda.gov/media/70465/download</a> ) or Disclosure ( <a href="https://www.fda.gov/media/69872/download">FDA Form 3455</a> , available at <a href="https://www.fda.gov/media/69872/download">https://www.fda.gov/media/69872/download</a> ) information for each covered clinical study included in the submission. <i>Select “N/A” if the submitted clinical data is not a “covered clinical study” as defined in the guidance entitled “<a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators">Financial Disclosures by Clinical Investigators</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators</a>.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>			Yes	No	N/A	*Page #
	b.	<p>Submission includes completed Certification of Compliance with requirements of ClinicalTrials.gov Data Bank (see FDA Form 3674 which can be obtained at <a href="https://www.fda.gov/about-fda/reports-manuals-forms/forms">https://www.fda.gov/about-fda/reports-manuals-forms/forms</a>) (42 U.S.C. 282(j)(5)(B)) for each applicable device clinical trial included in the submission.</p> <p><i>Select “N/A” if the submitted clinical data is not an “applicable device clinical trial” as defined in <a href="#">Title VIII of FDAAA, Sec. 801(j)</a></i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	c.	<p>Statements of Compliance for Clinical Investigations</p> <p><i>Select “N/A” if the submission does not contain any clinical data from investigations (as defined in 21 CFR 812.3(h)) to demonstrate substantial equivalence.</i></p> <p><i>For multicenter clinical investigations involving both United States (US) and outside United States (OUS) sites, part (i) should be addressed for the US sites and part (ii) should be addressed for the OUS sites. 21 CFR 812.28 applies to all OUS clinical investigations that enroll the first subject on or after February 21, 2019.</i></p> <p><i>Please refer to the guidance document entitled “<a href="#">Acceptance of Clinical Data to Support Medical Device Applications and Submissions - Frequently Asked Questions</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/acceptance-clinical-data-support-medical-device-applications-and-submissions-frequently-asked">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/acceptance-clinical-data-support-medical-device-applications-and-submissions-frequently-asked</a>, for more information.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
		i.	<p>For each clinical investigation conducted in the US, the submission includes a statement that the investigation was conducted in compliance with 21 CFR parts 50, 56, and 812 (or, with respect to part 56, that it was not subject to the regulations under 21 CFR 56.104 or 56.105).</p> <p><b>OR</b></p> <p>The submission includes a brief statement of the reason for noncompliance with 21 CFR parts 50,56 and/or 812.</p> <p><i>Select “N/A” if the clinical investigations were conducted solely OUS.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		ii.	<p>For each clinical investigation conducted OUS, the submission includes a statement that the clinical investigations were conducted in accordance with good clinical practice (GCP) as described in 21 CFR 812.28(a)(1).</p> <p><b>OR</b></p> <p>The submission includes a waiver request in accordance with 21 CFR 812.28(c).</p> <p><b>OR</b></p> <p>The submission includes a brief statement of the reason for not conducting the investigation in accordance with GCP and a description of steps taken to ensure that the data and results are credible and accurate and that the rights, safety, and well-being of subjects have been adequately protected.</p> <p><i>Select “N/A” if the clinical investigations were conducted solely inside the US.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
			Comments:				

*Contains Nonbinding Recommendations*

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>				Yes	No	N/A	*Page #		
	8.	<p>The submission identifies prior submissions for the same device included in the current submission (e.g., submission numbers for a prior not substantially equivalent [NSE] determination, prior deleted or withdrawn 510(k), Q-Submission, IDE, PMA, etc.).</p> <p><b>OR</b></p> <p>States that there were no prior submissions for the subject device.</p> <p><i>Prior submissions (or no prior submissions) for this device should be included in Section F (prior related submissions) of the CDRH Premarket Review Submission Cover Sheet form (<a href="#">Form 3514</a>, available at <a href="https://www.fda.gov/media/72421/download">https://www.fda.gov/media/72421/download</a>). This information may also be included in the Cover Letter (i.e., as a statement that there were no prior submissions for the device or a listing of the number(s) of the prior submissions).</i></p>	<input type="checkbox"/>	<input type="checkbox"/>					
	a.	<p>If there were prior submissions, the submitter has identified where in the current submission any issues related to a determination of substantial equivalence from prior submissions for this device are addressed.</p> <p><i>To address this criterion, it is recommended that the submission include a separate section with the prior submission number(s), a copy of the FDA feedback (e.g., letter, meeting minutes), and a statement of how or where in the submission this prior feedback was addressed. Note that adequacy of how the feedback was addressed will be assessed during the substantive review.</i></p> <p><i>Select “N/A” if the submitter states there were no prior submissions.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
		Comments:							
	<p><b>Combination Product Provisions – Per 503(g) of the FD&amp;C Act.</b>  Select “N/A” if the product is not a combination product. 21 CFR 3.2(e). The remaining criteria in this section will be omitted from the checklist if “N/A” is selected. If you are unsure if the product is a combination product, consult with the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.</p>				<input type="checkbox"/>				

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.			Yes	No	N/A	*Page #
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>						
	9.	Submission identifies the product as a combination product.	<input type="checkbox"/>	<input type="checkbox"/>		
	10.	The combination product contains as a constituent part an approved drug as defined in section 503(g)(5)(B) of the FD&C Act. Select “N/A” if the combination product does not contain as a constituent part an approved drug. Please also select “N/A” if a right of reference or use for the drug constituent part(s) is included with the submission. If “N/A” is selected, part a below is omitted from the checklist.	<input type="checkbox"/>		<input type="checkbox"/>	
	a.	The submission includes appropriate patent statement or certification and a statement that the submitter will give notice, as applicable. See section 503(g)(5)(A)&(C) of the FD&C Act.	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:				
<b>B.</b>	<b>Device Description</b>					
	11.	The device has a device-specific guidance document, special controls, and/or requirements in a device-specific classification regulation regarding the device description that is applicable to the subject device. <i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>	<input type="checkbox"/>		<input type="checkbox"/>	
	a.	<p>The submission addresses device description recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.			Yes	No	N/A	*Page #
<p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>						
	b.	<p>The submission includes device description information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:				
	12.	Descriptive information is present and consistent within the submission (e.g., the device description section is consistent with the device description in the labeling).	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:				
	13.	The submission includes descriptive information for the device, including the following:				
	a.	A description of the principle of operation or mechanism of action for achieving the intended effect.	<input type="checkbox"/>	<input type="checkbox"/>		
	b.	A description of proposed conditions of use, such as surgical technique for implants; anatomical location of use; user interface; how the device interacts with other devices; and/or how the device interacts with the patient.	<input type="checkbox"/>	<input type="checkbox"/>		
	c.	<p>A list and description of each device for which clearance is requested.</p> <p><i>Select “N/A” if there is only one device or model. “Device” may refer to models, part numbers, various sizes, etc.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		d.	<p>Submission contains representative engineering drawing(s), schematics, illustrations, photos and/or figures of the device.</p> <p><b>OR</b></p> <p>Submission includes a statement that engineering drawings, schematics, etc. are not applicable to the device (e.g., device is a reagent and figures are not pertinent to describe the device).</p> <p><i>In lieu of engineering drawings, schematics, etc. of each device to be marketed, “representative” drawings, etc. may be provided, where “representative” is intended to mean that the drawings, etc. provided capture the differences in design, size, and other important characteristics of the various models, sizes, or versions of the device(s) to be marketed.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:					
	14.		<p>Device is intended to be marketed with accessories and/or as part of a system.</p> <p><i>Select “N/A” if the device is not intended to be marketed with accessories, and/or as part of a system. If “N/A” is selected, parts a-c below are omitted from the checklist.</i></p>	<input type="checkbox"/>		<input type="checkbox"/>		
		a.	Submission includes a list of all accessories to be marketed with the subject device.	<input type="checkbox"/>	<input type="checkbox"/>			
		b.	<p>Submission includes a description (as detailed in item 13a., 13b., and 13d. above) of each accessory.</p> <p><i>Select “N/A” if the accessory(ies) has been previously cleared, or is exempt, and the proposed indications for use are consistent with the cleared indications.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	<p>A 510(k) number is provided for each accessory that received a prior 510(k) clearance</p> <p><b>AND</b></p> <p>A statement is provided that identifies accessories that have not received prior 510(k) clearance.</p>	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:					

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
<b>C. Substantial Equivalence Discussion</b>							
	15.		Submitter has identified a predicate device(s), including the following information:				
		a.	<p>Predicate device identifier provided (e.g., 510(k) number, De Novo number, reclassified PMA number, classification regulation reference, if exempt (e.g., 21 CFR 872.3710), or statement that the predicate is a preamendment device).</p> <p>For predicates that are preamendments devices, information is provided to document preamendments status.</p> <p><i>Information regarding <a href="#">documenting preamendment status</a> is available online (<a href="https://www.fda.gov/medical-devices/quality-and-compliance-medical-devices/preamendment-status">https://www.fda.gov/medical-devices/quality-and-compliance-medical-devices/preamendment-status</a>).</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	<p>The identified predicate(s) is consistent throughout the submission (e.g., the predicate(s) identified in the Substantial Equivalence section is the same as that listed in the 510(k) Summary (if applicable) and that used in comparative performance testing.</p>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:				
	16.		<p>Submission includes a comparison of the following for the predicate(s) and subject device and a discussion why any differences between the subject and predicate(s) do not impact safety and effectiveness [see section 513(i)(1)(A) of the FD&amp;C Act and 21 CFR 807.87(f)].</p> <p><i>See “<a href="#">The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)]</a>” guidance document, available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k</a> for more information on comparing intended use and technological characteristics.</i></p>				
		a.	<p>Indications for use</p> <p><i>If there are no differences between the subject device and the predicate(s) with respect to indications and intended use, this should be explicitly stated.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>							
		b.	Technology, including technical specifications, features, materials, and principles of operation  <i>Examples of technological characteristics include, but are not limited to design, features, materials, energy source, and principle of operation.</i>  <i>FDA recommends a tabular format for comparing technological characteristics. Any characteristic that is the same as the predicate(s) should be explicitly stated. Differences in technological characteristics should be identified and a rationale provided why they do not raise different questions of safety and effectiveness.</i>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:				
<b>D.</b>	<b>Proposed Labeling (see also 21 CFR parts 801 and 809 as applicable)</b>						
	17.		Submission includes proposed package labels and labeling (e.g., instructions for use, package insert, operator’s manual).	<input type="checkbox"/>	<input type="checkbox"/>		
		a.	Indications for use are stated in labeling and are identical to Indications for Use form and 510(k) Summary (if 510(k) Summary provided).	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	Labeling includes: <ul style="list-style-type: none"><li>- Statements of conditions, purposes or uses for which the device is intended (e.g., hazards, warnings, precautions, contraindications) (21 CFR 801.5) <b><u>AND</u></b></li><li>- Includes adequate directions for use (see 21 CFR 801.5) <b><u>OR</u></b></li><li>- Submission states that device qualifies for exemption per 21 CFR 801 Subpart D</li></ul>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:				
	18.		Labeling includes name and place of business of the manufacturer, packer, or distributor (21 CFR 801.1).	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
		<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>					
		Comments:					
	19.	<p>Labeling includes the prescription statement (see 21 CFR 801.109(b)(1)) or Rx Only symbol (see also Section 502(a) of the FD&amp;C Act and FDA’s final rule, “<a href="#">Use of Symbols in Labeling</a>” (81 FR 38911), available at <a href="https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling">https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling</a>).</p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		<p><i>Select “N/A” if not indicated for prescription use.</i></p>					
		Comments:					
	20.	<p>The device has a device-specific guidance document, special controls, and/or requirements in a device-specific classification regulation regarding labeling that is applicable to the subject device.</p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
	a.	<p>The submission addresses labeling recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

<p>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</p> <p>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</p>			Yes	No	N/A	*Page #
	b.	<p>The submission includes labeling information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:				
	21.	<p>If the device is an in vitro diagnostic device, provided labeling includes all applicable information required per 21 CFR 809.10.</p> <p><i>Select “N/A” if not an in vitro diagnostic device.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		Comment:				
E.	<p><b>Sterilization</b></p> <p><i>If an in vitro diagnostic (IVD) device and sterilization is not applicable, select “N/A.” The criteria in this section will be omitted from the checklist if “N/A” is selected.</i></p>			<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.	Yes	No	N/A	*Page #
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>	Yes	No	N/A	*Page #
<p>Submission states that the device and/or accessories, if applicable, are: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Provided sterile, intended to be single-use</p> <p><input type="checkbox"/> Requires processing during its use-life</p> <p><input type="checkbox"/> Non-sterile when used (and no processing required)</p> <p><input type="checkbox"/> Information regarding the sterility status of the device is not provided (if this box is checked, please also check one of the two boxes below)</p> <p style="margin-left: 20px;"><input type="checkbox"/> Sterility status not needed for this device (e.g., software-only device)</p> <p style="margin-left: 20px;"><input type="checkbox"/> Sterility status needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “non-sterile when used” or “not provided and not needed” is selected, the sterility-related criteria below are omitted from the checklist.</i></p> <p><i>If information on sterility status is not provided, and it is needed or the need for this information is unclear, select “No.”</i></p> <p><i>The “Requires processing during its use-life” option refers to devices falling into one of the four categories below:</i></p> <ul style="list-style-type: none"> <li>• <i>Supplied sterile and requires reprocessing prior to subsequent patient use</i></li> <li>• <i>Supplied non-sterile and requires user to process the device for initial use, as well as to reprocess the device after each use</i></li> <li>• <i>Reusable medical device (single-user) reprocessed between each use</i></li> <li>• <i>Single-use medical devices initially supplied as non-sterile to the user, and requiring the user to process the device prior to its use</i></li> </ul> <p><i>Please refer to the guidance document titled “<a href="#">Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/reprocessing-medical-devices-health-care-settings-validation-methods-and-labeling">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/reprocessing-medical-devices-health-care-settings-validation-methods-and-labeling</a>, for additional information.</i></p>				
Comments:				
22.	Assessment of the need for cleaning and subsequent disinfection or sterilization information.			

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.							
		a.	Identification of device and/or accessories, if applicable, that are provided sterile. <i>Select “N/A” if no part of the device or accessories are provided sterile.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		b.	Identification of device and/or accessories, if applicable, that are end user sterilized or disinfected. <i>Select “N/A” if no part of the device or accessories are end user sterilized or disinfected.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		c.	Identification of device and/or accessories, if applicable, that are reusable. <i>Select “N/A” if no part of the device or accessories are reusable.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
			Comments:				
	23.		If the device and/or accessories, if applicable, are provided sterile: <i>Select “N/A” if no part of the device or accessories are provided sterile, otherwise complete a-f below.</i>			<input type="checkbox"/>	
		a.	Sterilization method is stated for each device (including dose for radiation sterilization)	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	A description of method to validate the sterilization parameters is provided for each proposed sterilization method (e.g., half-cycle method and full citation of FDA-recognized standard, including date). <i>Note: the sterilization validation report is not required.</i>	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	For devices sterilized using chemical sterilants such as ethylene oxide (EO) and hydrogen peroxide, submission states maximum levels of sterilant residuals remaining on the device and sterilant residual limits. <i>Select “N/A” if not sterilized using chemical sterilants.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		d.	Sterility Assurance Level (SAL) stated	<input type="checkbox"/>	<input type="checkbox"/>		
		e.	Submission includes description of packaging	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		f.	<p>For products labeled “non-pyrogenic,” a description of the method used to make the determination stated (e.g., limulus amebocyte lysate [LAL]).</p> <p><i>Select “N/A” if not labeled “non-pyrogenic.”</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:					
	24.		<p>If the device and/or accessory, if applicable, is reusable or end user sterilized or disinfected:</p> <p><i>Select “N/A” if no part of the device or accessories are reusable or end user sterilized or disinfected, otherwise complete a-d below.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		a.	<p>Cleaning method is provided in labeling for each device and/or accessory, if applicable.</p> <p><i>Select “N/A” if not reusable and does not need cleaning prior to disinfection or sterilization.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	<p>Disinfection method is provided in labeling for each device and/or accessory, if applicable.</p> <p><i>Select “N/A” if not disinfected (i.e., undergoes terminal sterilization) prior to use.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	<p>Sterilization method is provided in labeling for each device and/or accessory, if applicable.</p> <p><i>Select “N/A” if not sterilized (i.e., undergoes disinfection) prior to use.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>		
		d.	<p>Device types in this submission are listed in the Federal Register Notice entitled “<a href="#">Validated Instructions for Use and Validation Data Requirements for Certain Reusable Medical Devices in Premarket Notifications</a>” (Reprocessing FR Notice, available at <a href="https://www.federalregister.gov/documents/2017/06/09/2017-12007/medical-devices-validated-instructions-for-use-and-validation-data-requirements-for-certain-reusable">https://www.federalregister.gov/documents/2017/06/09/2017-12007/medical-devices-validated-instructions-for-use-and-validation-data-requirements-for-certain-reusable</a>).</p> <p><i>Device types identified in the Reprocessing FR Notice devices posing a greater likelihood of microbial transmission and represent a high risk of infection. Select “N/A” if the device type in the submission is not included in the Reprocessing FR Notice.</i></p>	<input type="checkbox"/>		<input type="checkbox"/>			
		i.	<p>If device types in this submission are included in the Reprocessing FR Notice, the submission includes protocols and test reports for validating the reprocessing instructions.</p> <p><i>Select “N/A” if the device type in the submission is not included in the Reprocessing FR Notice.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:						
	25.		<p>The device has a device-specific guidance document, special controls, and/or requirement in a device-specific classification regulation regarding sterility and/or reprocessing that is applicable to the subject device</p> <p><i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i></p>	<input type="checkbox"/>		<input type="checkbox"/>			

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		a.	<p>The submission addresses sterility and/or reprocessing recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	<p>The submission includes sterility and/or reprocessing information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
			Comments:					
<b>F.</b>	<b>Shelf-Life</b>							
	26.	Proposed shelf-life/expiration date stated			<input type="checkbox"/>	<input type="checkbox"/>		
		<p><b>OR</b></p> <p>Statement that shelf-life is not applicable because of low likelihood of time-dependent product degradation</p>						

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>		<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
	Comments:				
	27. For a sterile device, submission includes summary of methods used to establish that device packaging will maintain a sterile barrier for the entirety of the proposed shelf-life. <i>Select “N/A” if the device is not provided sterile.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:				
	28. Submission includes summary of methods used to establish that device performance is maintained for the entirety of the proposed shelf-life (e.g., mechanical properties, coating integrity, pH, osmolality, etc.). <b>OR</b> Statement why performance data is not needed to establish maintenance of device performance characteristics over the shelf-life period.	<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:				
<b>G.</b>	<b>Biocompatibility</b> <i>If an in vitro diagnostic (IVD) device, select “N/A.” The criteria in this section will be omitted from the checklist if “N/A” is selected.</i>			<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #
	<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				
	<p>Submission states that there: (<i>one of the below must be checked</i>)</p> <ul style="list-style-type: none"> <li><input type="checkbox"/> Are direct or indirect tissue-contacting components</li> <li><input type="checkbox"/> Are no direct or indirect tissue-contacting components</li> <li><input type="checkbox"/> Information regarding tissue contact status of the device is not provided (if this box checked, please also check one of the two boxes below) <ul style="list-style-type: none"> <li><input type="checkbox"/> Tissue contact information not needed for this device (e.g., software-only device)</li> <li><input type="checkbox"/> Tissue contact information is needed or need unclear</li> </ul> </li> </ul> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “are no” or “not provided and not needed” is selected, the biocompatibility-related criteria below are omitted from the checklist. If information on the tissue-contact status is not provided, and contact information is needed or its contact status is unclear, select “No.”</i></p> <p><i>An example of a direct tissue-contacting device would be an implant that has direct contact with tissues during use. An example of an indirect tissue-contacting device would be fluid entering the body following passing through device/device components not in direct contact with the tissue.</i></p>				
	<p>Comments:</p>				
	<p>29.</p>	<p>Submission includes a list identifying each tissue-contacting device component (e.g., implant, delivery catheter) and associated materials of construction for each component, including identification of color additives, if present.</p>			
		<p>Comments:</p>			
	<p>30.</p>	<p>Submission identifies contact classification (e.g., surface-contacting, less than 24 hour duration) for each tissue-contacting device component (e.g., implant, delivery catheter).</p>			
		<p>Comments:</p>			

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #	
	<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p> <p>31. For a biocompatibility assessment of tissue-contacting components, submission includes:</p> <ul style="list-style-type: none"> <li>• Each relevant endpoint for the device (as identified in device-specific guidance, or Attachment A of the FDA guidance document entitled “<a href="#">Use of International Standard ISO 10993-1, ‘Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process,’</a>” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-international-standard-iso-10993-1-biological-evaluation-medical-devices-part-1-evaluation-and">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-international-standard-iso-10993-1-biological-evaluation-medical-devices-part-1-evaluation-and</a>), has been addressed.</li> <li>• For any testing performed, test protocol (including identification and description of test article including whether the test article is the device in its final finished form using the recommended approach in Attachment F of “<a href="#">Use of International Standard ISO 10993-1, ‘Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process,’</a>” methods, and pass/fail criteria), and analysis of results (including tables with data points and statistical analyses, where appropriate), as described in Attachment E of the guidance document entitled “<a href="#">Use of International Standard ISO 10993-1, ‘Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process’</a>” provided for each completed test.</li> </ul> <p><b>OR</b></p> <p>A statement that biocompatibility testing is not needed with a rationale that considers all relevant endpoints (e.g., materials and manufacturing/processing are identical to the predicate).</p>	<input type="checkbox"/>	<input type="checkbox"/>			
	Comments:					
H.	Software					

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #
	<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				
	<p>Submission states that the device: <i>(one of the below must be checked)</i></p>				
	<p><input type="checkbox"/> Does contain software/firmware</p>				
	<p><input type="checkbox"/> Does not contain software/firmware</p>				
	<p><input type="checkbox"/> Information on whether device contains software/firmware is not provided (if this box checked, please also check one of the two boxes below)</p>				
	<p><input type="checkbox"/> Software/firmware information not needed for this device (e.g., surgical suture, condom)</p>				
	<p><input type="checkbox"/> Software/firmware information is needed or need unclear</p>				
	<p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p>				
	<p><i>If “does not contain” or “not provided and not needed” is selected, the software-related criteria below are omitted from the checklist. If information on software is not provided, and this information is needed or the need is unclear, select “No.”</i></p>				
	<p>Comments:</p>				
	<p>32.</p>	<p>Submission includes a statement of software level of concern and rationale for the software level of concern.</p>	<input type="checkbox"/>	<input type="checkbox"/>	
	<p>Comments:</p>				
	<p>33.</p>	<p>All applicable software documentation provided based on level of concern identified by the submitter, as described in “<a href="#">Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-content-premarket-submissions-software-contained-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-content-premarket-submissions-software-contained-medical-devices</a>, or the submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through an alternative approach (i.e., the submitter has identified an alternate approach with a rationale).</p>	<input type="checkbox"/>	<input type="checkbox"/>	
	<p><i>Note: This element is also applicable to non-internally generated or off-the-shelf (OTS) software used in the device.</i></p>				

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		Comments:						
<b>I.</b>	<b>Cybersecurity</b>							
	<p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does contain any external wired and/or wireless communication interfaces (Wired: USB, ethernet, SD, CD, RGA, etc. or Wireless: Wi-Fi, Bluetooth, RF, inductive, Cloud, etc.)</p> <p><input type="checkbox"/> Does not contain external interfaces as described above</p> <p><input type="checkbox"/> Information on whether device has external interfaces is not provided (if this box is checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> Cybersecurity information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> Cybersecurity information is needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not contain” or “not provided and not needed” is selected, the cybersecurity criteria below are omitted from the checklist. If information on cybersecurity is not provided, and this information is needed or the need is unclear, select “No.”</i></p>				<input type="checkbox"/>			
	34.	<p>All applicable documentation identified by the submitter, as described in Guidance for the “<a href="#">Content of Premarket Submissions for Management of Cybersecurity in Medical Devices</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/content-premarket-submissions-management-cybersecurity-medical-devices-0">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/content-premarket-submissions-management-cybersecurity-medical-devices-0</a>.</p> <p><b>OR</b></p> <p>Submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through an alternative approach (i.e., the submitter has identified an alternate approach with a rationale).</p>			<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:						
<b>J.</b>	<b>Electrical Safety and EMC</b>							

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #		
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>							
<p>Electrical Safety:</p> <p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does require electrical safety evaluation</p> <p><input type="checkbox"/> Does not require electrical safety evaluation</p> <p><input type="checkbox"/> Information on whether device requires electrical safety evaluation not provided (if this box checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> Electrical safety information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> Electrical safety information needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not require” or “not provided and not needed” is selected, the electrical safety criteria below are omitted from the checklist. If information on electrical safety is not provided, and it is needed or the need for this information is unclear, select “No.”</i></p>							
<p>Comments:</p>							
<p>35. Submission includes evaluation of electrical safety (e.g., per IEC 60601-1, or equivalent FDA-recognized standard, and if applicable, a device-specific standard).</p> <p><b>OR</b></p> <p>Submission includes electrical safety evaluation using methods or standards that are not FDA-recognized and submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through this alternative approach (i.e., the submitter has identified alternate methods or standards with a rationale).</p>		<input type="checkbox"/>	<input type="checkbox"/>				
			<p>Comments:</p>				

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>					
	<p>EMC:</p> <p>Submission states that the device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Does require EMC evaluation</p> <p><input type="checkbox"/> Does not require EMC evaluation</p> <p><input type="checkbox"/> Information on whether device requires EMC evaluation not provided (if this box checked, please also check one of the two boxes below)</p> <p><input type="checkbox"/> EMC information not needed for this device (e.g., surgical suture, condom)</p> <p><input type="checkbox"/> EMC information needed or need unclear</p> <p>This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.</p> <p><i>If “does not require” or “not provided and not needed” is selected, the EMC criteria below are omitted from the checklist. If information on EMC is not provided, and it is needed or the need for this information is unclear, select “No.”</i></p>	<input type="checkbox"/>			
	Comments:				
36.	<p>Submission includes evaluation of electromagnetic compatibility (e.g., per IEC 60601-1-2 or equivalent FDA-recognized standard and if applicable, a device-specific standard).</p> <p><b>OR</b></p> <p>Submission includes electromagnetic compatibility evaluation using methods or standards that are not FDA-recognized and submission includes information to establish that the submitter has otherwise met the applicable statutory or regulatory criteria through this alternative approach (i.e., the submitter has identified alternate methods or standards with a rationale).</p>	<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:				

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>					
<b>K. Performance Data General</b>				<input type="checkbox"/>	
	<p><i>If an in vitro diagnostic (IVD) device, select “N/A.” The criteria in this section will be omitted from the checklist if “N/A” is selected.</i></p> <p><i>Performance data criteria relating to IVD devices is addressed in Section L.</i></p>				
Comments:					
37.	<p>Summaries of the non-clinical laboratory studies and full test reports* are provided.</p> <p>*Summary and full test report content recommendations can be found in FDA’s guidance “<a href="#">Recommended Content and Format of Non-Clinical Bench Performance Testing Information in Premarket Submissions</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/recommended-content-and-format-non-clinical-bench-performance-testing-information-premarket">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/recommended-content-and-format-non-clinical-bench-performance-testing-information-premarket</a>.</p> <p>If a submitter chooses to declare conformity to a voluntary consensus standard that FDA has recognized, submission of a full test report may not be necessary. Refer to Abbreviated Criteria #3. See FDA’s guidance “<a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices</a>.</p> <p><i>Select “N/A” if the submission appropriately does not include performance data or there are no completed tests without a Declaration of Conformity.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	a.	<p>Submission includes an explanation of how the data generated from each test supports a finding of substantial equivalence (e.g., comparison to predicate device testing, dimensional analysis, etc.).</p> <p><i>Select “N/A” if the submission does not include performance data.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
Comments:							
	38.	<p>The device has a device-specific guidance document, special controls, and/or requirement in a device-specific classification regulation regarding performance data that is applicable to the subject device.</p> <p><i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i></p>		<input type="checkbox"/>		<input type="checkbox"/>	
		<p>a.</p> <p>The submission addresses performance data recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		<p>b.</p> <p>The submission includes performance data that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

***Contains Nonbinding Recommendations***

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>
			Comments:				
	39.	<p>If literature is referenced in the submission, submission includes:</p> <p><i>Select “N/A” if the submission does not reference literature. If “N/A” is selected, parts a and b below are omitted from the checklist.</i></p> <p><i>Note that the applicability of the referenced article to support a substantial equivalence finding should be assessed during the substantive review; only the presence of a discussion is required to support acceptance.</i></p>					<input type="checkbox"/>
		a.	Legible reprints or a summary of each article.			<input type="checkbox"/>	<input type="checkbox"/>
		b.	Discussion of how each article is applicable to support the substantial equivalence of the subject device to the predicate.			<input type="checkbox"/>	<input type="checkbox"/>
			Comments:				
	40.	<p>For each completed animal study, the submission provides the following:</p> <p><i>Select “N/A” if no animal study was conducted. If “N/A” is selected, parts a-c below are omitted from the checklist. Note that this section does not address biocompatibility evaluations, which are assessed in Section G of the checklist.</i></p>					<input type="checkbox"/>
		a.	Submission includes a study protocol which includes all elements as outlined in 21 CFR 58.120			<input type="checkbox"/>	<input type="checkbox"/>
		b.	Submission includes final study report which includes all elements outlined in 21 CFR 58.185			<input type="checkbox"/>	<input type="checkbox"/>
		c.	Submission contains a statement that the study was conducted in compliance with applicable requirements in the GLP regulation (21 CFR Part 58), OR, if the study was not conducted in compliance with the GLP regulation, the submission explains why the noncompliance would not impact the validity of the study data provided to support a substantial equivalence determination.			<input type="checkbox"/>	<input type="checkbox"/>
			Comments:				

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>		
<b>L.</b>	<b>Performance Characteristics – In Vitro Diagnostic Devices Only (see also 21 CFR 809.10(b)(12))</b>								
	<p>Submission indicates that device: (<i>one of the below must be checked</i>)</p> <p><input type="checkbox"/> Is an in vitro diagnostic device</p> <p><input type="checkbox"/> Is not an in vitro diagnostic device</p> <p><i>If “is not” is selected, the performance data-related criteria below are omitted from the checklist.</i></p>								
	41.	Submission includes the following studies, as appropriate for the device type, including associated protocol descriptions, study results and line data:							
		a.	Precision/reproducibility				<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.	Accuracy (includes as appropriate linearity; calibrator or assay traceability; calibrator and/or assay stability protocol and acceptance criteria; assay cut-off; method comparison or comparison to clinical outcome; matrix comparison; and clinical reference range or cutoff).				<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	Sensitivity (detection limits, LoB, LoD, LoQ where relevant for the device type).				<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		d.	Analytical specificity				<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			Comments:						
	42.	The device has a device-specific guidance document, special controls, and/or requirement in a device-specific classification regulation regarding performance data that is applicable to the subject device.				<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		<i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>							

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information is located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>		
		a.	<p>The submission addresses performance data recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>			
		b.	<p>The submission includes performance data that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:						

<b>Digital Signature Concurrence Table</b>	
Reviewer Sign-Off	

*Contains Nonbinding Recommendations*

Management Sign-Off (digital signature optional)*	
---	--

\*Management review of checklist and concurrence with decision required.

## Appendix C. Acceptance Checklist for Special 510(k)s

**(Should be completed within 15 days of DCC receipt)**

The following information is not intended to serve as a comprehensive review. FDA recommends that the submitter include this completed checklist as part of the application.

**510(k)#:**

**Date Received by DCC:**

**510(k) Lead Reviewer:**

**Center:**

**Office:**

**Division:**

**Decision:** Accept \_\_\_\_\_ Refuse to Accept \_\_\_\_\_

**If Accept, notify the submitter.**

**If Refuse to Accept, notify submitter electronically and include a copy of this checklist.**

**Is an Addendum attached?: Yes No**

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete; it means the reviewer did not assess the element during the RTA review and that the element will be assessed during substantive review.

### Special 510(k) Factors

(See “[The Special 510\(k\) Program](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/special-510k-program),” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/special-510k-program>)

Please complete the below questions to determine if the 510(k) is appropriate for review as a Special 510(k). Complete the Refuse to Accept Checklist for a Traditional 510(k) if submission is converted.

		<b>Yes</b>	<b>No</b>	
1.	<b>510(k) is submitted to modify a legally marketed device (predicate) AND the Special 510(k) submission is submitted by the manufacturer legally authorized to market the predicate device.</b>	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:			
2.	<b>Performance data are needed to evaluate the change. If a manufacturer determines under their design control procedures that no additional verification or validation testing is necessary to evaluate a change, manufacturers may submit these changes as a Special 510(k) with a clear rationale supporting their conclusion that no performance data are necessary. When FDA does not agree with the manufacturer's assessment, FDA intends to</b>	<input type="checkbox"/>	<input type="checkbox"/>	

***Contains Nonbinding Recommendations***

	<i>continue with the additional Special 510(k) factors.</i>			
	Comments:			
3.	<b>There is a well-established method to evaluate the change.</b> <i>Well-established methods include those used in the previously cleared 510(k), an FDA-recognized consensus standard or FDA guidance document, qualified medical device development tools (MDDTs), are widely available and accepted, or found acceptable through a different premarket submission by the same manufacturer of the predicate.</i>	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:			
4.	<b>The data be reviewed in a summary or risk analysis format.</b> <i>The results from verification and validation associated with design or labeling changes should be able to be placed in a summary or risk analysis format without losing information necessary to support SE. Complete test reports should not be submitted in a Special 510(k). If complete test reports are submitted, FDA intends to assess whether the information can be reviewed in a summary format before converting to a Traditional 510(k).</i>	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments:			

**Is the submission appropriate for review as a Special 510(k)? Answer Yes if the change was submitted by the manufacturer of the predicate, well-established methods are available for any performance data necessary, and performance data can be reviewed in a summary or risk analysis format.**

- Yes, submission is appropriate for a Special 510(k). Continue checklist below.
- No, submission is not appropriate for a Special 510(k). Discontinue this RTA checklist, convert to a Traditional and apply the Traditional checklist.

<b>Organizational Elements</b>				
Failure to include these items should not result in an RTA designation.				
*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.		Yes	No	*Page #
1.	Submission contains a Table of Contents.	<input type="checkbox"/>	<input type="checkbox"/>	
2.	Each section is labeled (e.g., headings or tabs designating Device Description section, Labeling section, etc.).	<input type="checkbox"/>	<input type="checkbox"/>	
3.	All pages of the submission are numbered. <i>All pages should be numbered in such a manner that information can be referenced by page number. This may be done either by consecutively numbering the entire submission, or numbering the pages within a section</i>	<input type="checkbox"/>	<input type="checkbox"/>	

***Contains Nonbinding Recommendations***

	(e.g., 12-1, 12-2...).			
4.	Type of 510(k) is identified (i.e., Traditional, Abbreviated, or Special) <i>If type of 510(k) is not designated, review as a Traditional 510(k).</i>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:				

**Elements of a Complete Submission (RTA Items)**  
**(21 CFR 807.87 unless otherwise indicated)**

Submission should be designated RTA if not addressed

- Any “No” answer will result in a “Refuse to Accept” decision; however, FDA staff has discretion to determine whether missing items are needed to ensure that the submission is administratively complete to allow the submission to be accepted or to request missing checklist items interactively from submitters during the RTA review.
- Each element on the checklist should be addressed within the submission. The submitter may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criterion is considered present (Yes). An assessment of the rationale will be considered during the review of the submission.

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
A.	<b>Administrative</b>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	1.	All content used to support the submission is written in English (including translations of test reports, literature articles, etc.).	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:			
	2.	Submission identifies the following (FDA recommends use of the CDRH Premarket Review Submission Cover Sheet form ( <a href="#">Form 3514</a> , available at <a href="https://www.fda.gov/media/72421/download">https://www.fda.gov/media/72421/download</a> )):			
	a.	Device trade/proprietary name	<input type="checkbox"/>	<input type="checkbox"/>	
	b.	Device class and panel OR Classification regulation OR Statement that device has not been classified with rationale for that conclusion	<input type="checkbox"/>	<input type="checkbox"/>	
		Comments:			
	3.	Submission contains an Indications for Use Statement with Rx	<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.		Yes	No	N/A	*Page #
	<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				
	<p>and/or OTC designated (see also 21 CFR 801.109, and FDA’s final rule, “<a href="#">Use of Symbols in Labeling</a>” (81 FR 38911), available at <a href="https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling">https://www.federalregister.gov/documents/2016/06/15/2016-13989/use-of-symbols-in-labeling</a>).</p> <p><i>See recommended <a href="#">format</a> (<a href="https://www.fda.gov/media/86323/download">https://www.fda.gov/media/86323/download</a>).</i></p>				
	Comments:				
4.	<p>Submission contains a 510(k) Summary or 510(k) Statement.</p> <p><i>Refer to 21 CFR 807.92 and 21 CFR 807.93 for contents of 510(k) Summary and Statement, respectively. Adequacy of the content will be assessed during substantive review.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:				
5.	<p>Submission contains a Truthful and Accuracy Statement per 21 CFR 807.87(l).</p> <p><i>See recommended <a href="#">format</a> (<a href="https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-truthful-and-accurate-statement">https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-truthful-and-accurate-statement</a>).</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		
	Comments:				
6.	<p>Submission is a Class III 510(k) Device.</p> <p><i>Select “N/A” only if submission is not a Class III 510(k).</i></p>	<input type="checkbox"/>		<input type="checkbox"/>	
	<p>a.</p> <p>Contains Class III Summary and Certification per 21 CFR 807.87(k).</p> <p><i>See recommended <a href="#">content</a> (<a href="https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-class-iii-certification-and-summary">https://www.fda.gov/medical-devices/premarket-notification-510k/premarket-notification-class-iii-certification-and-summary</a>). Select “N/A” only if submission is not a Class III 510(k).</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	Comments				
7.	<p>The submission identifies prior submissions for the same device included in the current submission (e.g., submission numbers for a prior not substantially equivalent [NSE] determination, prior deleted or withdrawn 510(k), Q-Submission, IDE, PMA, etc.).</p> <p><b>OR</b></p> <p>States that there were no prior submissions for the subject device.</p>	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
		*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.					
		<i>Prior submissions (or no prior submissions) for this device should be included in Section F (prior related submissions) of the CDRH Premarket Review Submission Cover Sheet form (<a href="#">Form 3514</a>, available at <a href="https://www.fda.gov/media/72421/download">https://www.fda.gov/media/72421/download</a>). This information may also be included in the Cover Letter (i.e., as a statement that there were no prior submissions for the device or a listing of the number(s) of the prior submissions).</i>					
		<p>a. If there were prior submissions, the submitter has identified where in the current submission any issues related to a determination of substantial equivalence from prior submissions for this device are addressed.</p> <p><i>To address this criterion, it is recommended that the submission include a separate section with the prior submission number(s), a copy of the FDA feedback (e.g., letter, meeting minutes), and a statement of how or where in the submission this prior feedback was addressed. Note that adequacy of how the feedback was addressed will be assessed during the substantive review.</i></p> <p><i>Select “N/A” if the submitter states there were no prior submissions.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	8.	The submission utilizes voluntary consensus standard(s) (See section 514(c) of the FD&C Act). <i>This includes both FDA-recognized and non-recognized consensus standards. Select “N/A” if the submission does not utilize voluntary consensus standards.</i>	<input type="checkbox"/>		<input type="checkbox"/>		
	a.	The submission cites FDA-recognized voluntary consensus standard(s).	<input type="checkbox"/>		<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #	
*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.								
		i.	<p>The submission includes a Declaration of Conformity (DOC) as outlined in FDA’s guidance “<a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices</a>.</p> <p><b>OR</b></p> <p>If citing general use of a standard as noted in FDA’s guidance “<a href="#">Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices</a>,” the basis of such use is included along with the underlying information or data that supports how the standard was used.</p>	<input type="checkbox"/>	<input type="checkbox"/>			
		b.	The submission cites non-FDA-recognized voluntary consensus standard(s).	<input type="checkbox"/>		<input type="checkbox"/>		
		i.	The basis of use is included along with the underlying information or data that supports how the standard was used.	<input type="checkbox"/>	<input type="checkbox"/>			
			Comments:					
	<p><b>Combination Product Provisions – Per 503(g) of the FD&amp;C Act.</b>  Select N/A if the product is not a combination product. 21 CFR 3.2(e).  The remaining criteria in this section will be omitted from the checklist if “N/A” is selected. If you are unsure if the product is a combination product, consult with the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer.</p>					<input type="checkbox"/>		
	9.	Submission identifies the product as a combination product.		<input type="checkbox"/>	<input type="checkbox"/>			
	10.	The combination product contains as a constituent part an approved drug as defined in section 503(g)(5)(B) of the FD&C Act. Select “N/A” if the combination product does not contain as a constituent part an approved drug. Please also select “N/A” if a right of reference or use for the drug constituent part(s) is included with the submission. If “N/A” is selected, part a below is omitted from the checklist.		<input type="checkbox"/>		<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

<p><b>Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.</b></p> <p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>				<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>*Page #</b>	
		a.	The submission includes appropriate patent statement or certification and a statement that the submitter will give notice, as applicable. See section 503(g)(5)(A)&(C) of the FD&C Act.	<input type="checkbox"/>	<input type="checkbox"/>			
		Comments:						
<b>B.</b>	<b>Device Description</b>							
	11.	The device has a device-specific guidance document, special controls, and/or requirements in a device-specific classification regulation regarding the device description that is applicable to the subject device. <i>If “N/A” is selected, parts a and b below are omitted from the checklist.</i>			<input type="checkbox"/>	<input type="checkbox"/>		
		a.	<p>The submission addresses device description recommendations outlined in the device-specific guidance.</p> <p><b>OR</b></p> <p>The submission provides an alternative approach intended to address the applicable statutory and/or regulatory criteria.</p> <p><i>Select “N/A” if there is no applicable device-specific guidance. Select “No” if the submission does not include a rationale for any omitted information or any alternative approach as outlined above. Note that the adequacy of how recommendations in a device-specific guidance, etc., have been addressed should be assessed during the substantive review.</i></p>			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.	<p>The submission includes device description information that addresses relevant mitigation measures set forth in the special controls or device-specific classification regulation applicable to the device.</p> <p><b>OR</b></p> <p>The submission uses alternative mitigation measures and provides rationale why the alternative measures provide an equivalent assurance of safety and effectiveness.</p> <p><i>Select “N/A” if there are no applicable special controls or device-specific classification regulation. Select “No” if the submission does not include a rationale for any omitted</i></p>			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.							
			<i>information or any alternative approach as outlined above. Note that the adequacy of how such mitigation measures have been addressed should be assessed during the substantive review.</i>				
		Comments:					
	12.		Descriptive information is present and consistent within the submission (e.g., the device description section is consistent with the device description in the labeling).	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	13.		The submission includes descriptive information for the device, including the following:				
		a.	A description of the principle of operation or mechanism of action for achieving the intended effect.	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	A description of proposed conditions of use, such as surgical technique for implants; anatomical location of use; user interface; how the device interacts with other devices; and/or how the device interacts with the patient.	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	A list and description of each device for which clearance is requested. <i>Select “N/A” if there is only one device or model. “Device” may refer to models, part numbers, various sizes, etc.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
		d.	Submission contains representative engineering drawing(s), schematics, illustrations, photos and/or figures of the device. <b>OR</b> Submission includes a statement that engineering drawings, schematics, etc. are not applicable to the device (e.g., device is a reagent and figures are not pertinent to describe the device). <i>In lieu of engineering drawings, schematics, etc. of each device to be marketed, “representative” drawings, etc. may be provided, where “representative” is intended to mean that the drawings, etc. provided capture the differences in design, size, and other important characteristics of the various models, sizes, or versions of the device(s) to be marketed.</i>	<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.							
		Comments:					
	14.	A detailed description of all device modification(s) including rationale for each modification.  <i>When labeling or specific technological characteristics (e.g., materials, dimensions) are unchanged in comparison to the predicate, the submission should clearly state that no changes were made.</i>		<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	15.	Device is intended to be marketed with accessories and/or as part of a system.  <i>Select “N/A” if the device is not intended to be marketed with accessories and/or as part of a system. If “N/A” is selected, parts a-c below are omitted from the checklist.</i>		<input type="checkbox"/>		<input type="checkbox"/>	
	a.	Submission includes a list of all accessories to be marketed with the subject device.		<input type="checkbox"/>	<input type="checkbox"/>		
	b.	Submission includes a description (as detailed in item 13a., 13b., and 13d. above) of each accessory.  <i>Select “N/A” if the accessory(ies) has been previously cleared, or is exempt, and the proposed indications for use are consistent with the cleared indications.</i>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
	c.	A 510(k) number is provided for each accessory that received a prior 510(k) clearance  <b><u>AND</u></b> A statement is provided that identifies accessories that have not received prior 510(k) clearance.		<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
<b>C.</b>	<b>Substantial Equivalence Discussion</b>						
	16.	Submitter has identified a predicate device(s), including the following information:					
	a.	Predicate device identifier provided (e.g., 510(k) number, De Novo number, reclassified PMA number, classification regulation reference, if exempt (e.g., 21 CFR 872.3710), or		<input type="checkbox"/>	<input type="checkbox"/>		

***Contains Nonbinding Recommendations***

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
			<p>statement that the predicate is a preamendment device). For predicates that are preamendments devices, information is provided to document preamendments status.</p> <p><i>Information regarding <a href="#">documenting preamendment status</a> is available online (<a href="https://www.fda.gov/medical-devices/quality-and-compliance-medical-devices/preamendment-status">https://www.fda.gov/medical-devices/quality-and-compliance-medical-devices/preamendment-status</a>).</i></p>				
		b.	The identified predicate(s) is consistent throughout the submission (e.g., the predicate(s) identified in the Substantial Equivalence section is the same as that listed in the 510(k) Summary (if applicable) and that used in comparative performance testing.	<input type="checkbox"/>	<input type="checkbox"/>		
		Comments:					
	17.		<p>Submission includes a comparison of the following for the predicate(s) and subject device and a discussion why any differences between the subject and predicate(s) do not impact safety and effectiveness [see section 513(i)(1)(A) of the FD&amp;C Act and 21 CFR 807.87(f)]</p> <p><i>See the FDA guidance document “<a href="#">The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)]</a>,” available at <a href="https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k">https://www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-program-evaluating-substantial-equivalence-premarket-notifications-510k</a> for more information on comparing intended use and technological characteristics.</i></p>				
		a.	<p>Indications for use</p> <p><i>If there are no differences between the subject device and the predicate(s) with respect to indications and intended use, this should be explicitly stated.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	<p>Technology, including technical specifications, features, materials, and principles of operation</p> <p><i>Examples of technological characteristics include, but are not limited to design, features, materials, energy source, and principle of operation.</i></p>	<input type="checkbox"/>	<input type="checkbox"/>		

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.				Yes	No	N/A	*Page #
<p><b>*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b></p>							
			<p><i>FDA recommends a tabular format for comparing technological characteristics. Any characteristic that is the same as the predicate(s) should be explicitly stated. Differences in technological characteristics should be identified and a rationale provided why they do not raise different questions of safety and effectiveness.</i></p>				
<b>D.</b>	<b>Design Control Activities</b>						
18.	Design Control Activities Summary includes all of the following:						
		a.	Identification of risk analysis method(s) used to assess the impact of the modification on the device <b>AND</b> the results of the analysis.	<input type="checkbox"/>	<input type="checkbox"/>		
		b.	Identification of the device change(s).	<input type="checkbox"/>	<input type="checkbox"/>		
		c.	Identification of all risks associated with each device change, including identification of risks that are considered new because of the change; and	<input type="checkbox"/>	<input type="checkbox"/>		
		d.	Risk control measures to mitigate identified risks (e.g., labeling, verification).	<input type="checkbox"/>	<input type="checkbox"/>		
		e.	Based on the Risk Analysis, an identification of the verification and/or validation activities required to comply with 21 CFR 820.30. This identification includes a summary of test methods (including any protocol deviations), acceptance criteria, results in a summary or risk analysis format (e.g., basic descriptive statistics, where appropriate), and why each is adequate to establish substantial equivalence. If unchanged from a previous premarket submission, the manufacturer references the location of protocols and acceptance criteria by providing a submission and section numbers.	<input type="checkbox"/>	<input type="checkbox"/>		
		i.	For non-standardized test methods only: <ul style="list-style-type: none"> <li>• A reference to the protocol used for the existing device with an identification of any differences (e.g., protocol, test conditions, pre-defined acceptance criteria, sample size) from the previous 510(k). If protocol changes were             </li> </ul>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

*Contains Nonbinding Recommendations*

Check “Yes” if item is present, “N/A” if it is not needed and “No” if it is not included but needed.					Yes	No	N/A	*Page #	
<b>*Submitters including the checklist with their submission should identify the page numbers where requested information located. Use the comments section for an element if additional space is needed to identify the location of supporting information.</b>									
				made, the results summary describes why the test methods, acceptance criteria, and results support SE.					
		f.	A signed statement by the manufacturer’s designated individual(s) responsible for design control activities. Both items below must be present to answer “Yes.”	<ul style="list-style-type: none"> <li>i. Statement that, as required by the risk analysis, all verification and validation activities were performed by designated individual(s) and the results demonstrated that the predetermined acceptance criteria were met.</li> <li>ii. Statement that the submitter has complied and is not currently in violation of the design control procedure requirements as specified in 21 CFR 820.30 and the records are available for review, upon request.</li> </ul>	<input type="checkbox"/>	<input type="checkbox"/>			
Comments:									
<b>E.</b>	<b>Proposed Labeling (see also 21 CFR parts 801 and 809 as applicable)</b>								
	19.	Submission includes proposed package labels and labeling (e.g., instructions for use, package insert, operator’s manual).		<input type="checkbox"/>	<input type="checkbox"/>				
	a.	All changes in proposed labeling resulting from device modification(s) are highlighted or prominently identified. <i>FDA recommends clean and redlined copies be provided.</i>		<input type="checkbox"/>	<input type="checkbox"/>				
		Comments:							

Digital Signature Concurrence Table	
Reviewer Sign-Off	
Management Sign-Off (digital signature optional)*	

\*Management review of checklist and concurrence with decision required.