

The Least Burdensome Provisions: Concept and Principles

Draft Guidance for Industry and Food and Drug Administration Staff

DRAFT GUIDANCE

This draft guidance document is being distributed for comment purposes only.

Document issued on December 15, 2017.

You should submit comments and suggestions regarding this draft document within 60 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions about this document as applied to devices regulated by CDRH, contact the Office of the Center Director at (301) 796-6900. For questions regarding this document as applied to devices regulated by CBER, contact the Office of Communication, Outreach and Development in CBER at 1-800-835-4709 or 240-402-8010 or by email at ocod@fda.hhs.gov.

When final, this guidance will supersede “The Least Burdensome Provisions of the FDA Modernization Act of 1997: Concept and Principles,” issued on October 4, 2002.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Center for Biologics Evaluation and Research

Preface

40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56

Additional Copies

CDRH

Additional copies are available from the Internet. You may also send an e-mail request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please use the document number 1332 to identify the guidance you are requesting.

CBER

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., WO71, Room 3128, Silver Spring, MD 20903, or by calling 1-800-835-4709 or 240-402-8010, by email, ocod@fda.hhs.gov, or from the Internet at <https://www.fda.gov/BiologicsBloodVaccines/GuidanceComplianceRegulatoryInformation/Guidances/default.htm>.

Table of Contents

| | | |
|----|---|----|
| 57 | | |
| 58 | | |
| 59 | I. Introduction..... | 4 |
| 60 | II. Background..... | 5 |
| 61 | III. Scope..... | 7 |
| 62 | IV. Guiding Principles | 8 |
| 63 | V. Applications of Least Burdensome Principles | 9 |
| 64 | A. The minimum information necessary..... | 9 |
| 65 | (1) Less burdensome sources of clinical data | 10 |
| 66 | (2) Use of nonclinical data..... | 11 |
| 67 | (3) Acceptance of alternative approaches | 13 |
| 68 | B. The most efficient means | 14 |
| 69 | (1) Reducing the burden of traditional clinical studies | 14 |
| 70 | (2) Use of benefit-risk assessments..... | 16 |
| 71 | (3) Streamlining processes and reducing administrative burden | 16 |
| 72 | (4) Smart regulation..... | 18 |
| 73 | (5) Global harmonization..... | 18 |
| 74 | C. The right time..... | 19 |
| 75 | (1) Balancing premarket and postmarket information needs | 19 |
| 76 | VI. Compliance Policies that Support the Goals of the Least Burdensome Concept..... | 20 |
| 77 | VII. Conclusion | 21 |
| 78 | | |
| 79 | | |

The Least Burdensome Provisions: Concept and Principles

Draft Guidance for Industry and Food and Drug Administration Staff

This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.

I. Introduction

The Food and Drug Administration (FDA) is committed to helping patients gain more timely access to new medical devices and to maintaining continued access to existing medical devices that are high quality, safe, and effective, by expediting their development, assessment, review, and surveillance, consistent with the Agency’s statutory mission to protect and promote the public health.¹ By streamlining regulatory processes and removing or reducing unnecessary burdens associated with FDA regulatory activities, patients can have earlier and continued access to beneficial products.

Since the FDA Modernization Act of 1997 (FDAMA), Congress has directed FDA to take a least burdensome approach to medical device premarket evaluation in a manner that eliminates unnecessary burdens that may delay the marketing of beneficial new products, while maintaining the statutory requirements for clearance and approval. This draft guidance is intended to accurately reflect Congress’ intent by describing the guiding principles and recommended approach for FDA staff and industry to facilitate consistent application of least burdensome principles.

We define “least burdensome” to be **the minimum amount of information necessary to adequately address a regulatory question or issue through the most efficient manner at the**

¹ Section 1003 of the Federal Food, Drug, and Cosmetic Act (FD&C Act).

Contains Nonbinding Recommendations

Draft – Not for Implementation

114 **right time.** This concept applies to all products that meet the statutory definition of a device and
115 throughout the total product lifecycle (premarket and postmarket).²

116
117 FDA's guidance documents, including this draft guidance, do not establish legally enforceable
118 responsibilities. Instead, guidances describe the Agency's current thinking on a topic and should
119 be viewed only as recommendations, unless specific regulatory or statutory requirements are
120 cited. The use of the word *should* in Agency guidance means that something is suggested or
121 recommended, but not required. Throughout this guidance document, the terms *we*, *us* and
122 *our* refer to FDA staff from the Center for Devices and Radiological Health (CDRH)
123 or the Center for Biologics Evaluation and Research (CBER) involved in device regulation.
124

125 **II. Background**

126
127 Congress first added least burdensome provisions to the Federal Food, Drug, and Cosmetic Act
128 (FD&C Act) under FDAMA (Public Law 105-115). Congress enacted additional least
129 burdensome provisions to the FD&C Act through the FDA Safety and Innovation Act (Public
130 Law 112-144) (FDASIA) and the 21st Century Cures Act (Public Law 114-255) (Cures Act). The
131 least burdensome statutory provisions currently state:

- 132
- 133 • “Whenever the Secretary requests information to demonstrate that devices with differing
134 technological characteristics are substantially equivalent, the Secretary shall only request
135 information that is necessary to making substantial equivalence determinations. In
136 making such request, the Secretary shall consider the least burdensome means of
137 demonstrating substantial equivalence and request information accordingly.”³
138
 - 139 • “Any clinical data, including one or more well-controlled investigations, specified in
140 writing by the Secretary for demonstrating a reasonable assurance of device effectiveness
141 shall be specified as a result of a determination by the Secretary that such data are
142 necessary to establish device effectiveness. The Secretary shall consider, in consultation
143 with the applicant, the least burdensome appropriate means of evaluating device
144 effectiveness that would have a reasonable likelihood of resulting in approval.”⁴
145
 - 146 • In requesting additional information with respect to a premarket approval application
147 (PMA), “the Secretary shall consider the least burdensome appropriate means necessary
148 to demonstrate a reasonable assurance of device safety and effectiveness.”⁵
149
 - 150 • “[T]he Secretary shall consider the role of postmarket information in determining the
151 least burdensome means of demonstrating a reasonable assurance of device safety and

² Section 201(h) of the FD&C Act.

³ Section 513(i)(1)(D)(i) of the FD&C Act.

⁴ Section 513(a)(3)(D)(ii) of the FD&C Act.

⁵ Section 515(c)(5)(A) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

effectiveness.”⁶

- The term “necessary” in the least burdensome provisions means the “minimum required information” that would support a determination of substantial equivalence or a reasonable assurance of device safety and effectiveness.⁷
- The least burdensome provisions do not change the standards for premarket approval or substantial equivalence.⁸

FDA issued least burdensome guidance documents after the enactment of FDAMA. “Suggested Format for Developing and Responding to Deficiencies in Accordance with the Least Burdensome Provisions of FDAMA” (“Deficiencies Guidance”) was issued on November 2, 2000. In that guidance document, FDA recommended that its staff use a specific format for requests for additional information needed to make a decision on a medical device marketing submission (often called “deficiencies” or “deficiency letters”) to be in accordance with least burdensome principles. This format was intended to directly connect FDA requests to applicable statutory and regulatory criteria for a decision and optimize the time and effort of both industry and FDA. That guidance document also included a recommended format for industry responses to FDA deficiencies.

With the enactment of the Medical Device User Fee Amendments of 2017 (Public Law 115-52, §§ 201-210) (MDUFA IV), FDA committed to updating the Deficiencies Guidance. “Developing and Responding to Deficiencies in Accordance with the Least Burdensome Provisions” (<https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm073680.pdf>) was issued on September 29, 2017. The Deficiencies Guidance was updated to recommend that all deficiency letters include a statement regarding the basis for each deficiency and provides details regarding supervisory review, major/minor deficiencies, additional considerations, and prioritization of deficiencies.

“The Least Burdensome Provisions of the FDA Modernization Act of 1997: Concept and Principles” was issued on October 4, 2002 (“2002 Least Burdensome Guidance”). The guidance stated that, while the least burdensome provisions from FDAMA applied to PMA and 510(k) submissions, FDA believed that least burdensome principles should be implemented for all premarket regulatory activities. The document also defined the term “least burdensome” and included suggested approaches for industry and FDA staff to use least burdensome principles in PMA and 510(k) review, including focusing on the statutory and regulatory criteria for marketing authorization. The guidance also described general applications of least burdensome approaches to activities such as postmarket controls, and recommendations for how the Agency should communicate requests for additional information to industry.

⁶ Section 515(c)(5)(C) of the FD&C Act.

⁷ Sections 513(a)(3)(D)(iii), 513(i)(1)(D)(ii), and 515(c)(5)(B) of the FD&C Act.

⁸ Sections 513(a)(3)(D)(iv), 513(i)(1)(D)(iii), and 515(c)(5)(D) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

192 This guidance document, when finalized, will replace the 2002 Least Burdensome Guidance.
193 The statutory updates in FDASIA and the Cures Act clarified the original least burdensome
194 provisions and further recognized the role of postmarket activities as they relate to premarket
195 decisions. FDA believes, as a matter of policy, that least burdensome principles should be
196 consistently and widely applied to all activities in the premarket and postmarket settings to
197 remove or reduce unnecessary burdens so that patients can have earlier and continued access to
198 high quality, safe, and effective devices, and the Agency is applying tools in our implementation
199 of these principles, such as regular training of all Center employees on least burdensome
200 principles. This draft guidance, therefore, reflects FDA’s belief that least burdensome principles
201 should be applied throughout the medical device total product lifecycle. The least burdensome
202 concept remains the same in that the principles are based on sound science, the intent of the law,
203 the use of alternative approaches, and the efficient use of resources to effectively address
204 regulatory issues. We have provided contemporary examples for both premarket and postmarket
205 settings to demonstrate approaches that FDA and industry can take to ensure that least
206 burdensome principles are implemented for all device-related applications and interactions with
207 FDA.
208

III. Scope

209
210 The least burdensome concept and this guidance apply to all products that meet the statutory
211 definition of a device.⁹ The policy in this guidance applies to all activities (including premarket
212 and postmarket actions) pertaining to the regulation of medical devices. The policy in this
213 guidance applies, but is not limited, to:
214

- 215
- 216 • Premarket submissions, including PMAs, premarket notifications (510(k)s), De Novo
- 217 requests, humanitarian device exemption applications, and investigational device
- 218 exemption applications
- 219
- 220 • Additional Information and Major Deficiency Letters
- 221
- 222 • Q-Submissions
- 223
- 224 • Informal or interactive inquiries regarding device development
- 225
- 226 • Panel review and recommendations
- 227
- 228 • Postmarket surveillance and post-approval studies
- 229
- 230 • Reclassifications and exemptions
- 231

⁹ Section 201(h) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 232 • Guidance documents and their application
- 233
- 234 • Compliance-related interactions
- 235
- 236 • Regulation development
- 237

238 **IV. Guiding Principles**

239 FDA defines least burdensome to be the minimum amount of information necessary to
240 adequately address a regulatory question or issue through the most efficient manner at the right
241 time (e.g., need to know versus nice to know). Our least burdensome principles do not change
242 the applicable regulatory standards, such as the device approval or clearance standards, nor the
243 applicable requirements, including premarket submission content requirements and the
244 requirement for valid scientific evidence.¹⁰

245
246
247 FDA intends to, and industry should, apply the following guiding principles when taking a least
248 burdensome approach to a particular question or issue at any point in the total product lifecycle:

- 249
250 1. FDA intends to request the minimum information necessary to adequately address the
251 regulatory question or issue at hand.
- 252
253 2. Industry should submit material, including premarket submissions, to FDA that are least
254 burdensome for FDA to review.
 - 255 • Industry should submit well-organized, clear, and concise information.
 - 256
- 257 3. FDA intends to use the most efficient means to resolve regulatory questions and issues.
 - 258 • FDA intends to use all reasonable measures to streamline processes and policies,
259 as well as render regulatory decisions within appropriate timeframes, such as
260 MDUFA performance goals.
 - 261 • FDA intends to routinely use interactive approaches, whenever possible, to
262 resolve questions and issues.
 - 263 • FDA intends to, and industry should, use tailored approaches that have been
264 adapted to individual circumstances and needs to address regulatory questions and
265 issues.
 - 266 • FDA intends to take appropriate consideration of the time and resource
267 implications of its requests.
 - 268
- 269 4. The right information should be provided at the right time (e.g., just-in-time data
270 collection) to address the right questions.

¹⁰ Sections 513(i) and 515 of the FD&C Act and 21 CFR Part 807 Subpart E, Parts 812 and 814, and 860.7(c).

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 271 • FDA intends to, and industry should, consider the use of postmarket data
272 collection to reduce premarket data collection whenever appropriate and feasible.
273
- 274 5. Regulatory approaches should be designed to fit the technology, taking into account its
275 unique innovation cycles, evidence generation needs, and timely patient access.
276
- 277 6. FDA intends to leverage data from other countries and decisions by, or on behalf of,
278 other national medical device regulatory authorities to the extent appropriate and feasible.
279
- 280 7. FDA intends to apply least burdensome principles in international medical device
281 convergence and harmonization efforts.
282

283 Providing excellent customer service is critical to successfully applying least burdensome
284 principles. FDA strives for clear and concise communication of its requests, expectations,
285 processes, policies, and decisions, as well as the rationale behind them.
286

287 Industry can help us apply least burdensome principles by providing FDA with clear and concise
288 requests, premarket submissions, and responses, along with their rationales. Excellent customer
289 service and open lines of communication between FDA and its customers will help to provide
290 regulatory outcomes that best serve patients.
291

V. Applications of Least Burdensome Principles

292
293

294 This section provides numerous examples intended to represent the least burdensome concept
295 and implementation of least burdensome guiding principles as applied to medical device
296 regulation. This includes examples of using less burdensome sources of clinical data, using
297 nonclinical data, accepting alternative approaches, reducing the burden of traditional clinical
298 studies, using benefit-risk assessments, streamlining processes and reducing administrative
299 burden, engaging in smart regulation, participating in global harmonization, and balancing
300 premarket and postmarket information needs.

301

302 While some examples may only be applicable to FDA as a regulatory authority, implementation
303 of the least burdensome principles is a shared responsibility between FDA and industry.
304

305 These examples are provided for illustration purposes and are not intended to be an exhaustive
306 list. The examples are grouped by the elements of the least burdensome definition, although
307 some examples could reasonably be included under multiple categories.
308

A. The minimum information necessary

309

Contains Nonbinding Recommendations

Draft – Not for Implementation

(1) Less burdensome sources of clinical data

Reasonable assurance of safety and effectiveness of a device is determined on the basis of valid scientific evidence.¹¹ The evidence required may vary according to the characteristics of the device, its conditions of use, and the extent of experience with its use, among other factors.¹² Alternative sources of clinical data should be considered when appropriate, and, in many cases, may be the least burdensome means for assessing device safety and effectiveness and for other regulatory decision-making. Alternative sources of data may include peer-reviewed literature, outside the U.S. (OUS) data, real-world evidence (RWE), and well-documented case histories. These alternative sources of data should be considered by FDA staff and industry when determining the least burdensome approach to a regulatory requirement or decision. Isolated case reports, random experience, reports lacking sufficient details to permit scientific evaluation, and unsubstantiated opinions are not regarded as valid scientific evidence.¹³ Adverse event reports, although not generally considered valid scientific evidence, may be considered as part of a benefit-risk assessment.

Leveraging existing data

The use of existing data to inform regulatory decisions can be a scientifically valid application of the least burdensome principles. When available, appropriate, and relevant to the specific device or regulatory issue at hand, peer-reviewed literature, registry data, and OUS data may be used in lieu of, or to supplement, other data. For example, FDA approved a humanitarian device exemption (HDE) application to treat pediatric esophageal atresia based on a combination of published literature and well-documented compassionate use cases.¹⁴

Under appropriate circumstances, FDA and applicants may also leverage information contained in a previously filed PMA, including information from clinical or preclinical tests or studies that demonstrate a reasonable assurance of the safety and effectiveness of a device, but excluding descriptions of methods of manufacture, product composition, and other trade secrets. According to the “six-year rule,”¹⁵ while excluding trade secrets, FDA may use safety and effectiveness data from clinical or preclinical tests or studies, six years after PMA approval, in order to approve another applicant’s device, establish a performance standard or special control, or classify or reclassify another device under section 513 of the FD&C Act. The Agency decided to apply the six-year rule only to data in PMAs approved after November 28, 1990, the date of enactment of the Safe Medical Devices Act (Public Law 101-629) (SMDA). For example, FDA used the six-year rule upon our own initiative to support the reclassification of stair-climbing wheelchairs from Class III to Class II with the establishment of special controls.¹⁶ For more information on the six-year rule, see the FDA guidance document “Guidance on Section 216 of the Food and Drug Administration Modernization Act of 1997”

¹¹ 21 CFR 860.7(c)(1).

¹² 21 CFR 860.7(c)(2).

¹³ 21 CFR 860.7(c)(2).

¹⁴ Summary of Safety and Probable Benefit available at: https://www.accessdata.fda.gov/cdrh_docs/pdf15/H150003B.pdf.

¹⁵ Section 520(h)(4) of the FD&C Act.

¹⁶ The final reclassification order was issued on April 14, 2014 (79 FR 20779).

Contains Nonbinding Recommendations

Draft – Not for Implementation

347 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm073709.pdf)
348 [ments/ucm073709.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm073709.pdf)).

349
350 The extrapolation of existing clinical data from a studied patient population into a new pediatric
351 patient population should be considered when endpoints present in the existing data source are
352 relevant, there are no differences between adult and pediatric use that could impact safety and
353 effectiveness, and the quality of data is sufficient. For more information on when and how to
354 extrapolate adult data for pediatric populations, see the FDA guidance document “Leveraging
355 Existing Clinical Data for Extrapolation to Pediatric Uses of Medical Devices”
356 ([https://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-](https://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-gen/documents/document/ucm444591.pdf)
357 [gen/documents/document/ucm444591.pdf](https://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-gen/documents/document/ucm444591.pdf)).

358 **Real-world evidence (RWE)**

359 Real-World Data (RWD) are data relating to patient health status and/or the delivery of health
360 care routinely collected from a variety of sources. RWE is the clinical evidence regarding the
361 usage and potential benefits or risks of a medical product derived from analysis of RWD. RWD
362 may originate from electronic health records (EHRs), registries, and medical administrative
363 claims data. For more information about RWE, see the FDA guidance document “Use of Real-
364 World Evidence to Support Regulatory Decision-Making for Medical Devices”
365 ([https://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-](https://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-gen/documents/document/ucm513027.pdf)
366 [gen/documents/document/ucm513027.pdf](https://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-gen/documents/document/ucm513027.pdf)).

367
368
369 FDA and industry are using RWD sources to address both premarket and postmarket issues.
370 FDA approved a PMA for a permanent pacemaker electrode using clinical data captured through
371 a remote monitoring system in a prospective registry.¹⁷ Several registries initially designed for
372 postmarket surveillance have been used to support expanded indications for devices such as
373 ventricular support devices, vaginal mesh, and transcatheter valves. Finally, FDA and industry
374 leveraged RWE generated from a registry to support expanded indications for use for a
375 cryosurgical tool intended for benign and malignant lesion ablation.¹⁸

376

377 **(2) Use of nonclinical data**

378 Nonclinical data should also be considered as an alternative data source. In some cases, this may
379 include the use of benchtop models, nonclinical literature, use of tissue phantoms, or computer
380 modeling and simulations. Especially in situations where testing modalities are representative or
381 predictive of clinical performance, FDA frequently relies upon nonclinical testing in lieu of or to
382 supplement clinical data. For example, FDA has approved magnetic resonance (MR) conditional
383 labeling for pacemakers, cardiac resynchronization therapy devices, and implantable cardioverter
384 defibrillators based, in part, on validated computer modeling. FDA has also accepted cadaver
385 images in lieu of imaging from live subjects for certain imaging indications such as extremity
386 imaging.

¹⁷ Summary of Safety and Effectiveness Data available at:
https://www.accessdata.fda.gov/cdrh_docs/pdf12/P120017b.pdf.

¹⁸ 510(k) summary available at: https://www.accessdata.fda.gov/cdrh_docs/pdf17/K171626.pdf.

Contains Nonbinding Recommendations

Draft – Not for Implementation

387
388 While clinical data may sometimes be necessary to meet a regulatory requirement, nonclinical
389 data should be considered as a replacement for clinical data, when appropriate. The use of
390 descriptive information, *in vitro* studies, computer modeling and simulations, and/or animal
391 performance data that could be responsive to an outstanding regulatory question should be
392 considered before requesting clinical data.

393 394 **Bottom-up approach to data requests**

395 Data requests from FDA should follow a stepwise analytical process to ensure that each request
396 reflects the least burdensome approach. This logic has been used for 510(k) submissions to
397 ensure application of least burdensome principles. FDA should first consider whether descriptive
398 information is sufficient. While few 510(k) submissions rely solely on descriptive information,
399 FDA and industry should consider this approach. For example, dimensional analysis and
400 materials comparisons have been used to establish substantial equivalence for orthopedic bone
401 plate and screw 510(k) submissions.

402
403 When descriptive information is not sufficient, FDA and industry should then consider whether
404 nonclinical performance testing or analytical studies using clinical samples could address the
405 issue. Nonclinical animal and/or biocompatibility studies are typically requested when other
406 forms of nonclinical bench performance testing are not sufficient to demonstrate substantial
407 equivalence. When analytical or nonclinical bench testing, or nonclinical animal testing and/or
408 biocompatibility studies are insufficient, FDA may request clinical performance data. For more
409 information about how FDA has used this process in 510(k) submissions, see the FDA guidance
410 document “The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications
411 [510(k)]”
412 ([https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UC](https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM284443)
413 [M284443](https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM284443)).

414 415 **Use of nonclinical bench performance testing**

416 Bench performance testing should be considered to address preclinical or clinical endpoints,
417 when appropriate. This may include bench models for anatomy, such as evaluating tortuous
418 paths for catheters used across many clinical applications. The use of tissue phantoms has also
419 increased for evaluating the magnetic resonance imaging compatibility of implants and tissue
420 effects from high intensity therapeutic ultrasound. Bench performance testing may not be an
421 appropriate surrogate when the methods do not correspond with clinically-relevant scenarios.

422 423 **Computer modeling and simulations (CM&S)**

424 CM&S should be used to support medical device safety and effectiveness as alternatives to
425 traditional benchtop or animal performance testing in appropriate circumstances. For example,
426 CM&S has been used to predict mechanical properties for cardiovascular and orthopedic devices
427 under simulated loading conditions. Additionally, CM&S has been used to estimate the
428 radiofrequency energy absorbed by patients undergoing magnetic resonance imaging (MRI) to
429 assess medical device safety. FDA’s recommendations regarding the use of CM&S in
430 submissions are included in the FDA guidance document “Reporting of Computational

Contains Nonbinding Recommendations

Draft – Not for Implementation

431 Modeling Studies in Medical Device Submissions”
432 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM381813.pdf)
433 [ments/UCM381813.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM381813.pdf)).
434

(3) Acceptance of alternative approaches

435
436 Alternative approaches should be considered, when appropriate, to optimize the time and
437 resources of FDA and industry. Both FDA and industry should understand that there are often
438 multiple ways to satisfactorily address a particular regulatory issue. The resolution of the
439 regulatory issue should be based on a discussion about which method is least burdensome, while
440 still satisfactorily addressing the regulatory issue.

Resolution of scientific issues

441
442 The acceptance of alternative approaches for the evaluation of scientific issues identified during
443 the premarket review of a novel medical device is a common application of least burdensome
444 principles. FDA and industry should be flexible and open-minded in determining the most
445 efficient mechanism and minimum information necessary to address a specific issue. In some
446 cases, a justification, in lieu of additional data, may be acceptable to address an issue.

447
448 A common example of FDA accepting alternative approaches includes the biocompatibility of
449 medical devices. When appropriate, FDA and industry have leveraged OUS clinical data or large
450 animal safety studies to address certain biocompatibility endpoints. Additionally, the use of
451 rationales based on materials properties, chemistry, and processing have been leveraged as
452 alternatives to repeat testing. For example, manufacturers often modify the tips of their vascular
453 catheters and, instead of repeat testing, leverage prior testing conducted on another vascular
454 device for which the same materials and processing methods were used.

455
456 Another common example concerns the review of nonclinical testing in regulatory submissions.
457 Through requests for additional information, the Agency may identify one particular method for
458 addressing a scientific issue. When appropriate, FDA should identify when alternative
459 approaches or justifications would resolve the issue under discussion. Common examples include
460 leveraging existing device information such as mechanical testing, software validation, and
461 sterilization validation. For example, FDA often requests additional bench testing results to
462 address differences in technological characteristics for devices with multiple sizes or models to
463 support substantial equivalence determinations. In consultation with the applicant, FDA has
464 accepted alternative testing and scientific justifications in lieu of previously requested testing for
465 certain device types, for which worst-case testing scenarios can be reasonably justified based on
466 size. In other cases, FDA has requested mechanical testing for certain bone plates or screws, but
467 accepted alternative approaches that included detailed engineering analyses to establish
468 substantial equivalence.

Considering alternative labeling

469
470
471 Applicants propose labeling, including indications for use (IFU), in their regulatory applications.
472 If a labeling statement or proposed IFU is not supported by the submitted evidence and would
473 otherwise result in an adverse decision, such as a not substantially equivalent determination for a
474

Contains Nonbinding Recommendations

Draft – Not for Implementation

475 510(k), FDA staff and industry should discuss both (1) a labeling statement or an IFU, if any,
476 that can be supported by the information submitted to FDA, and (2) the minimum information
477 that would support the sought-after labeling statement or IFU. The applicant can then choose
478 which avenue they wish to pursue within statutory and MDUFA deadlines.

479
480 In other cases, the addition of specific warnings or precautions to the device labeling may
481 provide sufficient risk mitigation to support a favorable decision. For example, FDA has
482 accepted certain risk mitigations through labeling instead of fail-safe and failure alert
483 mechanisms in its review of CLIA Waiver applications for *in vitro* diagnostic devices (IVDs).

484

485 **B. The most efficient means**

486 **(1) Reducing the burden of traditional clinical studies**

487 When clinical data are necessary, FDA and industry should consider the most efficient means of
488 obtaining the evidence necessary to meet the regulatory need or standard. For example, the PMA
489 requirements for a Class III device include providing a reasonable assurance of safety and
490 effectiveness.¹⁹ In many cases, alternatives to randomized controlled studies may be sufficient.

491 In appropriate circumstances, FDA has accepted historical controls, the use of objective
492 performance criteria (OPC), performance goals (PGs), and alternative sources of data, including
493 evidence from registries, claims data, and published literature. For more information about
494 design considerations for clinical studies, see the FDA guidance document “Design
495 Considerations for Pivotal Clinical Investigations for Medical Devices”

496 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu
497 ments/UCM373766.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM373766.pdf)).

498

499 **Historical control groups**

500 The use of historical control groups involves quantitatively comparing the results of use of the
501 device with prior experience derived from the adequately documented natural history of a
502 disease or condition in comparable patients or populations who received no treatment or who
503 followed an established effective regimen (therapeutic, diagnostic, prophylactic).²⁰ The use of
504 historical control groups may reduce the number of patients enrolled in clinical studies while
505 retaining their strength as well-controlled clinical investigations. FDA and industry have used
506 historical control groups in the evaluation of generic types of devices including, but not limited
507 to, transcatheter aortic valves, hip resurfacing devices, total knee and ankle replacements,
508 neurostimulators, and diagnostic devices using single-arm clinical study designs to assess safety
509 and effectiveness.

510

511 **Non-comparative clinical outcome studies**

512 Non-comparative clinical outcome studies can include those using OPCs, PGs, observational
513 studies, registries, meta-analysis, and literature summaries. OPC refers to a target value derived

¹⁹ Section 515 of the FD&C Act.

²⁰ 21 CFR 860.7(f)(1)(iv)(d).

Contains Nonbinding Recommendations

Draft – Not for Implementation

514 from historical data within clinical studies or registries and is used in a pass/fail manner to assess
515 safety and effectiveness endpoints. PG is a numerical value used as a comparison for safety
516 and/or effectiveness endpoints that may be accepted or developed by a professional society,
517 standards development organization, or FDA. The use of single-group studies compared to an
518 OPC or PG can reduce the sample size necessary to support marketing submissions.

519
520 In device types where existing data can be leveraged to set OPCs, such as heart valves, clinical
521 studies are then routinely performed using those OPCs. In another case where less data was
522 available, FDA, in consultation with industry and the Obstetrics and Gynecology Devices
523 Advisory Panel, leveraged five PMA approvals with similar control data to establish an OPC for
524 endometrial ablation devices to give applicants the option to conduct a single-arm study.²¹ A
525 single arm study can reduce the study size necessary to demonstrate reasonable assurance of
526 safety and effectiveness.

Subject as own control

527
528 When possible, FDA and industry should consider when subjects in clinical studies can serve as
529 their own controls to minimize the number of enrolled patients. Cross-over study designs, where
530 each subject receives the treatment and control interventions sequentially in a randomized order,
531 have been used for clinical studies involving many different devices including obesity devices,
532 dermal fillers, and neurostimulators. Paired designs, where a patient serves as his/her own
533 concurrent control, have been used in split-face study designs to assess plastic surgery devices
534 and split-knees designs to assess orthopedic devices. IVD studies have also used patients as their
535 own controls, for example, to assess the long-term performance of colorectal cancer screening
536 tests.

Adaptive study designs

537
538 The use of adaptive study designs may reduce resource requirements, decrease time to study
539 completion, and/or increase the chance of study success. For more information, see the FDA
540 guidance document “Adaptive Designs for Medical Device Clinical Studies”
541 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM446729.pdf)
542 [ments/UCM446729.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM446729.pdf)). Adaptive study designs have been used to minimize the number of
543 study subjects for premarket and postmarket studies for neurological and cardiovascular device
544 types.

Use of alternatives to prospective sample collection

545
546 Certain circumstances can make prospective patient samples for IVDs impractical, such as the
547 low prevalence of a condition or the rarity of measuring certain concentration levels in a clinical
548 setting. In such cases, alternative approaches to sample collection should be considered, such as
549 the use of banked and retrospective samples, contrived samples, and surrogate samples or
550 biomarkers.

551
552
553
554

²¹ FDA Letter to Global Endometrial Ablation Manufacturers is available at:
<https://www.fda.gov/downloads/MedicalDevices/ResourcesforYou/Industry/UCM470246.pdf>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

555 **(2) Use of benefit-risk assessments**

556 Least burdensome principles are consistent with FDA’s approach to weighing benefits and risks
557 in regulatory decision-making. All regulatory processes involve some uncertainty about the
558 benefits and risks of a medical device. Greater uncertainty is appropriate in some circumstances,
559 such as when the probable benefits are high (e.g., a breakthrough device) or the probable risks of
560 the device are low.

561

562 **Marketing submissions**

563 In determining the safety and effectiveness of a device, FDA considers, among other factors, the
564 probable benefit to health from the use of the device weighed against any probable injury or
565 illness from such use.²² The extent (probability, magnitude/severity, and duration) of both
566 benefit and risk are considered along with uncertainty, patient-centric metrics and perspective,
567 and a characterization of the disease. A positive decision may be rendered when FDA determines
568 that the probable benefits to health outweigh any probable risks and that the device will provide
569 clinically significant results.²³ For example, despite the occurrence of serious adverse events and
570 death in clinical studies and OUS registries for a mitral valve repair device, FDA determined that
571 there was a narrow patient population with low life expectancy and quality of life for which the
572 probable benefits outweigh the probable risks.²⁴ This device provided an unmet clinical need for
573 treatment in patients who were not candidates for mitral valve surgery. Taking into account the
574 benefit-risk assessment, FDA determined that the device has a reasonable assurance of safety and
575 effectiveness for this narrow patient population.

576

577 For more information about using benefit-risk in PMA and De Novo request decisions, see the
578 FDA guidance document “Factors to Consider When Making Benefit-Risk Determinations in
579 Medical Device Premarket Approval and De Novo Classification”

580 (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm517504.pdf>).

581

582 **(3) Streamlining processes and reducing administrative burden**

583 Least burdensome principles also apply to streamlining regulatory processes to improve
584 efficiency. FDA has implemented several policies and practices to reduce administrative burden,
585 eliminate potential redundancies, and conserve both FDA and industry resources.

586

587 **Reducing redundancies**

588 The inclusion of multiple devices or indications within a bundled marketing submission or the
589 use of dual submissions can limit redundant submission and review of regulatory information by
590 FDA and industry. Bundling is appropriate for generic types of devices with scientific and
591 regulatory issues that can be most efficiently addressed during one review. For more information
592 about bundling, see the FDA guidance document “Bundling Multiple Devices or Multiple
593

²² 21 CFR 860.7(b)(3).

²³ 21 CFR 860.7(d)-(e).

²⁴ Summary of Safety and Effectiveness Data available at:
https://www.accessdata.fda.gov/cdrh_docs/pdf10/P100009B.pdf.

Contains Nonbinding Recommendations

Draft – Not for Implementation

594 Indications in a Single Submission”
595 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089732.pdf)
596 [ments/ucm089732.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089732.pdf)).
597

598 The dual 510(k)/CLIA Waiver permits the concurrent review of a 510(k) submission and CLIA
599 Waiver application. FDA and industry work collaboratively to develop comparison and
600 reproducibility study designs to generate one data set that should reduce study-related costs and
601 review time. For more information about the dual 510(k)/CLIA Waiver, see the FDA guidance
602 document “Administrative Procedures for CLIA Categorization”
603 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070889.pdf)
604 [ments/ucm070889.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070889.pdf)).
605

Marketing submission efficiencies

606 Special and Abbreviated 510(k) submissions rely on conformance with design controls and
607 conformity to FDA-recognized voluntary consensus standards. PMA annual reports can be used
608 to summarize design, labeling, and manufacturing changes that do not affect safety and
609 effectiveness.
610

Medical Device Development Tools (MDDTs)

611
612 An MDDT is a method, material, or measurement used to assess the effectiveness, safety, or
613 performance of a medical device. MDDTs are tools that can be qualified and used to streamline
614 device development and regulatory evaluation. The efficient use of MDDTs can reduce
615 development costs and review times after tool qualification. For more information about
616 MDDTs, see the FDA guidance document “Qualification of Medical Device Development
617 Tools”
618 ([https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocume](https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm374432.pdf)
619 [nts/ucm374432.pdf](https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm374432.pdf)).
620

Medical Device Reporting (MDR)

621
622 Reducing the burden of MDRs has been executed through enhancements to existing processes.
623 The Electronic MDR (eMDR) system has been implemented to fast-track the generation,
624 submission, and review of MDRs. The use of eMDR expedites report processing and reduces the
625 data entry burden on industry, FDA, healthcare facilities, and importers. For more information
626 about eMDR, see the FDA guidance document “Questions and Answers about eMDR –
627 Electronic Medical Device Reporting”
628 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm179471.pdf)
629 [ments/ucm179471.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm179471.pdf)).
630

631
632 Alternative summary MDR reporting can be requested by persons and entities that are not
633 exempt from mandatory reporting. FDA may grant an alternative, or full or partial exemption
634 from the MDR regulations.²⁵ For example, manufacturers may request that reports be submitted
635 quarterly, semiannually, or annually instead of 30 calendar days after becoming aware of the
636 reportable event. Additionally, manufacturers can request that reports only contain a subset of

²⁵ 21 CFR 803.19.

Contains Nonbinding Recommendations

Draft – Not for Implementation

637 the data required by the MDR regulations. Registry data used for postmarket surveillance has
638 allowed manufactures to apply for alternative summary reporting where only certain adverse
639 events must be reported to the FDA. In some cases, FDA has allowed manufactures to provide a
640 summary MDR report generated from a specific registry each quarter. For more information
641 about alternative summary and summary MDR reporting, see the FDA guidance documents
642 “Medical Device Reporting – Alternative Summary Reporting (ASR) Program”
643 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm072102.pdf)
644 [ments/ucm072102.pdf](https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm359566.pdf)) and “Medical Device Reporting for Manufacturers”
645 ([https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocume](https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm359566.pdf)
646 [nts/ucm359566.pdf](https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm359566.pdf)).
647

648 **(4) Smart regulation**

649 The application of least burdensome principles should include a regular reexamination of the
650 regulatory paradigm for medical devices to ensure that existing regulatory processes are still the
651 most efficient and request the minimum information necessary. As specific medical technologies
652 become better understood from a scientific and clinical perspective, FDA should periodically
653 assess the appropriateness of data requests in premarket submissions, evaluate premarket and
654 postmarket balance, and determine whether devices are candidates for reclassification, as such
655 evaluation may help FDA focus on issues of higher public health concern.
656

657 **Exemption from 510(k)**

658 Central to reexamination of regulatory processes is the consideration of whether premarket
659 submissions are necessary to reasonably assure a device’s safety and effectiveness. In
660 accordance with the FD&C Act, as amended by the Cures Act, FDA published notices
661 exempting numerous Class II and Class I device types from 510(k) requirements.²⁶ FDA is also
662 required to periodically publish a list of device types for which a 510(k) submission is no longer
663 necessary to provide a reasonable assurance of safety and effectiveness.²⁷
664

665 **(5) Global harmonization**

666 Efforts to advance international harmonization and regulatory convergence should be viewed as
667 applying the least burdensome concept by using the most efficient means to achieve regulatory
668 goals. While U.S. statutes and regulations may not be identical to those of other countries, FDA
669 should align itself with international regulatory authorities whenever practicable and possible.
670

671 **Reliance on voluntary consensus standards**

672 The development of voluntary consensus standards allows FDA, industry, and other stakeholders
673 to agree upon methods and acceptance criteria that can be used to support the safe and effective
674 use of medical devices. The recognition of standards by FDA can streamline interactions
675 between FDA and industry. When recognized and used by multiple regulatory authorities,

²⁶ Sections 510(l)(2) and (m)(1) of the FD&C Act. The final exemption notices for Class I and Class II devices were published in the Federal Registers of April 13, 2017 (82 FR 17841) and July 11, 2017 (82 FR 31976), respectively.

²⁷ Sections 510(l)(2) and (m)(1) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

676 standards can also support global harmonization by creating consistent approaches to medical
677 device evaluation. For more information, see section 514(c) of the FD&C Act and the FDA
678 guidance document “Recognition and Use of Consensus Standards”
679 ([https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocu](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm077295.pdf)
680 [ments/ucm077295.pdf](https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm077295.pdf)).

681
682 In the absence of a recognized consensus standard, evaluation of performance data involves the
683 submission and review of complete test protocols and data reports. By providing a declaration of
684 conformity to FDA-recognized standards with explicit testing methods, applicants and FDA do
685 not need to discuss whether test methods are scientifically valid and can focus their resources on
686 reviewing the test data. When consensus standards have both explicit test methods and either
687 performance limits and/or acceptance criteria, a declaration of conformity can potentially replace
688 the submission and review of both the test methods and complete data. FDA accepts declarations
689 of conformity for many generic device types to support regulatory submissions.

690

International Medical Device Regulators Forum (IMDRF)

691 FDA’s participation on the IMDRF to develop internationally recognized guidance documents,
692 standards, and auditing practices supports regulatory convergence. IMDRF is a voluntary group
693 of international regulatory authorities intended to build on previous work from the Global
694 Harmonization Task Force on Medical Devices (GHTF). This includes the piloting of both the
695 Summary Technical Documentation (STED) format and Regulated Product Submission (RPS)
696 for regulatory submissions. Harmonization of FDA’s content requirements for marketing
697 submissions with those of international regulatory authorities can streamline applicant efforts to
698 address a regulatory issue. FDA’s participation on IMDRF’s work items related to Software as a
699 Medical Device (SaMD) also seeks to harmonize FDA’s clinical evaluation of SaMD with those
700 of the international community.

701

Medical Device Single Audit Program (MDSAP)

702
703 MDSAP is a program that applies the least burdensome principles by allowing for one audit to
704 satisfy the requirements of multiple regulatory jurisdictions. The goal is to reduce regulatory
705 burden on industry by minimizing audits with potential redundant requests or disruption of
706 business by different international regulatory authorities.

707

708

C. The right time

709

(1) Balancing premarket and postmarket information needs

710 Striking the right balance between premarket and postmarket information needs is a key
711 principle of the least burdensome concept. This balance is intended to address obtaining the
712 minimum necessary information at the right time in the total product lifecycle. As discussed in
713 the Background (section II), the FD&C Act requires FDA to consider the role of postmarket
714 information when making a determination of the least burdensome means of demonstrating a
715

Contains Nonbinding Recommendations

Draft – Not for Implementation

716 reasonable assurance of safety and effectiveness for PMAs.²⁸ FDA and industry should consider
717 the appropriate balance between premarket and postmarket information needs for all medical
718 device regulatory issues, when applicable.
719

Reviewing only some changes

720 FDA and industry’s reliance on the Quality System (QS) Regulation (21 CFR Part 820) is
721 another example of the application of least burdensome principles that supports efficiency.
722 Manufacturers have the ability to make certain design changes to cleared devices and labeling
723 without reporting under section 510(k) of the FD&C Act. For more information, see the FDA
724 guidance documents “Deciding When to Submit a 510(k) for a Change to an Existing Device”
725 (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm514771.pdf>) and “Deciding When to Submit a 510(k) for a Software Change to an
726 Existing Device”
727 (<https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM514737.pdf>).
728
729
730
731

Total product lifecycle approach

732 FDA should only request information that is necessary to make a given regulatory decision.
733 When requesting information, FDA should assess the right time for obtaining necessary
734 information and determine whether a shift from premarket to postmarket evaluation is
735 appropriate while still reasonably assuring device safety and effectiveness. Reliance on
736 postmarket controls, such as the QS Regulation, post-approval studies (PAS), postmarket
737 surveillance, and MDR, should be considered when determining the suitability for devices for
738 the market. In some cases, FDA has determined that premarket review is not required to
739 reasonably assure a device’s safety and effectiveness. For example, some medical devices that
740 are exempt from premarket review rely on the QS Regulation and other postmarket controls to
741 reasonably assure their safety and effectiveness.
742

743 In other cases, certain safety and effectiveness questions may be appropriately and efficiently
744 answered in a postmarket setting. For example, long-term safety and effectiveness questions for
745 a leadless pacemaker were addressed through PAS. Analytical studies for long-term outcomes
746 regarding companion diagnostics have also been addressed in a postmarket setting, when
747 appropriate. For more information about FDA’s recommended approach to
748 premarket/postmarket balance, see the FDA guidance document “Balancing Premarket and
749 Postmarket Data Collection for Devices Subject to Premarket Approval”
750 (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm393994.pdf>).
751
752
753

VI. Compliance Policies that Support the Goals of the Least Burdensome Concept

754
755
756

²⁸ Section 515(c)(5)(C) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

757 The compliance policies below help support the goals of the least burdensome concept by
758 allowing for more efficient and effective use of resources by both FDA and industry.

759

Enforcement discretion policy

761 In some cases, FDA has published guidance documents communicating that the Agency does not
762 intend to examine whether certain products comply with premarket review and postmarket
763 regulatory requirements for devices under the FD&C Act and its implementing regulations,
764 including, but not limited to: registration and listing and 510(k) requirements; labeling
765 requirements; current good manufacturing practice requirements as set forth in the QS
766 Regulation; and MDR requirements.²⁹ Although these guidances do not change or otherwise
767 affect any requirements of the FD&C Act or any applicable regulations, FDA has used this
768 approach for products such as mobile medical applications and general wellness products so that
769 FDA can focus its oversight on those medical devices whose functionality could pose a higher
770 risk to patients. For more information, see the FDA guidance documents “Mobile Medical
771 Applications”

772 (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM263366>) and “General Wellness: Policy for Low Risk Devices”

773 (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm429674.pdf>).

774

Medical necessity for marketed devices

778 FDA recognizes that devices may have benefit even when the devices fail to meet all regulatory
779 requirements. When contemplating exercising enforcement discretion for a violative device,
780 FDA considers the needs of patients and clinicians. In cases when there are no alternative
781 devices, or the risk associated with changing to an alternative is greater than the risk associated
782 with the violative devices, FDA can determine the violative devices to be medically necessary
783 for some situations. For example, FDA may exercise discretion by not taking enforcement action
784 against a violative device, in order to address patient and clinician need. FDA bases this
785 determination on benefit-risk principles and revisits the analysis as new information becomes
786 available.

787 For more information about using benefit-risk in compliance/enforcement decisions, see the
788 FDA guidance document “Factors to Consider Regarding Benefit-Risk in Medical Device
789 Product Availability, Compliance, and Enforcement Decisions”

790 (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm506679.pdf>).

791

VII. Conclusion

792

793
794 When finalized, this draft guidance will reflect the principle that medical device regulation
795 should be least burdensome across the total product lifecycle. FDA intends to request the

²⁹ 21 CFR Part 807, Part 801 and 809.10, Part 820, and Part 803, respectively.

Contains Nonbinding Recommendations

Draft – Not for Implementation

796 minimum amount of information necessary to adequately address a regulatory question or issue
797 through the most efficient manner at the right time. Industry should provide information to FDA
798 that is least burdensome for FDA to review. Open lines of communication between FDA and
799 industry will provide regulatory outcomes that best serve patients. Successful application of least
800 burdensome principles will ensure that patients have access to high-quality, safe, and effective
801 medical devices.

DRAFT